


Data Path : Z:\svoasrv\HPCHEM1\BNA_F\Data\BF090920\
 Data File : BF121541.D
 Acq On : 9 Sep 2020 11:04
 Operator : JU/CG
 Sample : DFTPP
 Misc :
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 BNA_F
 ClientSampleId :
 DFTPP

Integration File: rteint.p

Method : Z:\SVOASRV\HPCHEM1\BNA_F\METHODS\8270-BF082720.M
 Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 Last Update : Wed Sep 02 17:52:18 2020



AutoFind: Scans 1619, 1620, 1621; Background Corrected with Scan 1614

Target Mass	Rel. to Mass	Lower Limit%	Upper Limit%	Rel. Abn%	Raw Abn	Result Pass/Fail
51	198	10	80	35.7	73874	PASS
68	69	0.00	2	1.6	1150	PASS
69	198	0.00	100	34.7	71864	PASS
70	69	0.00	2	0.5	394	PASS
127	198	10	80	47.0	97317	PASS
197	198	0.00	2	0.6	1238	PASS
198	198	100	100	100.0	206997	PASS
199	198	5	9	6.5	13395	PASS
275	198	10	60	27.2	56365	PASS
365	198	1	100	3.4	7072	PASS
441	198	0.01	100	14.9	30857	PASS
442	198	50	100	96.3	199312	PASS
443	442	15	24	19.5	38893	PASS