

Data Path : Z:\svoasrv\HPCHEM1\BNA_F\Data\BF090924\
 Data File : BF139280.D
 Acq On : 09 Sep 2024 12:03
 Operator : RC/JU
 Sample : SSTDCCC040
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_F
 LabSampleId :
 SSTDCCC040

Quant Time: Sep 10 02:21:40 2024
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_F\Methods\8270-BF090424.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Wed Sep 04 16:11:43 2024
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	Amount	Calc.	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	20.000	20.000	0.0	101	0.00
2	1,4-Dioxane	40.000	34.934	12.7	90	0.00
3	Pyridine	40.000	35.804	10.5	93	0.00
4	n-Nitrosodimethylamine	40.000	36.490	8.8	94	0.00
5 S	2-Fluorophenol	80.000	75.292	5.9	99	0.00
6	Aniline	40.000	37.339	6.7	97	0.00
7 S	Phenol-d6	80.000	74.704	6.6	97	0.00
8	2-Chlorophenol	40.000	38.845	2.9	103	0.00
9	Benzaldehyde	40.000	33.965	15.1	97	0.00
10 C	Phenol	40.000	37.556	6.1	98	0.00
11	bis(2-Chloroethyl)ether	40.000	38.550	3.6	102	0.00
12	1,3-Dichlorobenzene	40.000	37.744	5.6	99	0.00
13 C	1,4-Dichlorobenzene	40.000	38.184	4.5	100	0.00
14	1,2-Dichlorobenzene	40.000	38.242	4.4	100	0.00
15	Benzyl Alcohol	40.000	39.010	2.5	100	0.00
16	2,2'-oxybis(1-Chloropropane	40.000	37.265	6.8	97	0.00
17	2-Methylphenol	40.000	37.894	5.3	99	0.00
18	Hexachloroethane	40.000	39.039	2.4	102	0.00
19 P	n-Nitroso-di-n-propylamine	40.000	38.146	4.6	101	0.00
20	3+4-Methylphenols	40.000	38.131	4.7	99	0.00
21 I	Naphthalene-d8	20.000	20.000	0.0	102	0.00
22	Acetophenone	40.000	37.936	5.2	100	0.00
23 S	Nitrobenzene-d5	80.000	77.756	2.8	101	0.00
24	Nitrobenzene	40.000	38.065	4.8	100	0.00
25	Isophorone	40.000	38.654	3.4	101	0.00
26 C	2-Nitrophenol	40.000	41.184	-3.0	103	0.00
27	2,4-Dimethylphenol	40.000	39.113	2.2	101	0.00
28	bis(2-Chloroethoxy)methane	40.000	38.746	3.1	103	0.00
29 C	2,4-Dichlorophenol	40.000	39.052	2.4	101	0.00
30	1,2,4-Trichlorobenzene	40.000	38.918	2.7	102	0.00
31	Naphthalene	40.000	38.036	4.9	100	0.00
32	Benzoic acid	40.000	41.759	-4.4	109	0.00
33	4-Chloroaniline	40.000	37.854	5.4	101	0.00
34 C	Hexachlorobutadiene	40.000	39.452	1.4	105	0.00
35	Caprolactam	40.000	39.878	0.3	101	0.00
36 C	4-Chloro-3-methylphenol	40.000	39.234	1.9	101	0.00
37	2-Methylnaphthalene	40.000	38.149	4.6	101	0.00
38	1-Methylnaphthalene	40.000	37.770	5.6	100	0.00
39 I	Acenaphthene-d10	20.000	20.000	0.0	103	0.00
40	1,2,4,5-Tetrachlorobenzene	40.000	38.352	4.1	103	0.00
41 P	Hexachlorocyclopentadiene	40.000	40.400	-1.0	98	0.00
42 S	2,4,6-Tribromophenol	80.000	80.795	-1.0	106	0.00
43 C	2,4,6-Trichlorophenol	40.000	39.366	1.6	100	0.00
44	2,4,5-Trichlorophenol	40.000	38.640	3.4	102	0.00
45 S	2-Fluorobiphenyl	80.000	75.134	6.1	103	0.00
46	1,1'-Biphenyl	40.000	37.660	5.9	101	0.00
47	2-Chloronaphthalene	40.000	37.894	5.3	102	0.00

Data Path : Z:\svoasrv\HPCHEM1\BNA_F\Data\BF090924\
 Data File : BF139280.D
 Acq On : 09 Sep 2024 12:03
 Operator : RC/JU
 Sample : SSTDCCC040
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_F
 LabSampleId :
 SSTDCCC040

Quant Time: Sep 10 02:21:40 2024
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_F\Methods\8270-BF090424.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Wed Sep 04 16:11:43 2024
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	Amount	Calc.	%Dev	Area%	Dev(min)
48	2-Nitroaniline	40.000	40.100	-0.3	103	0.00
49	Acenaphthylene	40.000	38.376	4.1	102	0.00
50	Dimethylphthalate	40.000	39.180	2.1	105	0.00
51	2,6-Dinitrotoluene	40.000	40.126	-0.3	102	0.00
52 C	Acenaphthene	40.000	38.678	3.3	102	0.00
53	3-Nitroaniline	40.000	38.911	2.7	100	0.00
54 P	2,4-Dinitrophenol	40.000	40.202	-0.5	110	0.00
55	Dibenzofuran	40.000	38.529	3.7	103	0.00
56 P	4-Nitrophenol	40.000	39.788	0.5	98	0.00
57	2,4-Dinitrotoluene	40.000	41.001	-2.5	102	0.00
58	Fluorene	40.000	38.272	4.3	103	0.00
59	2,3,4,6-Tetrachlorophenol	40.000	39.826	0.4	104	0.00
60	Diethylphthalate	40.000	40.338	-0.8	106	0.00
61	4-Chlorophenyl-phenylether	40.000	38.176	4.6	101	0.00
62	4-Nitroaniline	40.000	41.064	-2.7	104	0.00
63	Azobenzene	40.000	39.217	2.0	105	0.00
64 I	Phenanthrene-d10	20.000	20.000	0.0	104	0.00
65	4,6-Dinitro-2-methylphenol	40.000	41.356	-3.4	104	0.00
66 c	n-Nitrosodiphenylamine	40.000	38.079	4.8	103	0.00
67	4-Bromophenyl-phenylether	40.000	39.373	1.6	105	0.00
68	Hexachlorobenzene	40.000	38.402	4.0	104	0.00
69	Atrazine	40.000	36.152	9.6	101	0.00
70 C	Pentachlorophenol	40.000	42.778	-6.9	105	0.00
71	Phenanthrene	40.000	37.843	5.4	103	0.00
72	Anthracene	40.000	38.410	4.0	103	0.00
73	Carbazole	40.000	38.891	2.8	104	0.00
74	Di-n-butylphthalate	40.000	42.507	-6.3	109	0.00
75 C	Fluoranthene	40.000	39.526	1.2	107	0.00
76 I	Chrysene-d12	20.000	20.000	0.0	106	0.00
77	Benzidine	40.000	49.652	-24.1	121	0.00
78	Pyrene	40.000	39.889	0.3	105	0.00
79 S	Terphenyl-d14	80.000	79.696	0.4	106	0.00
80	Butylbenzylphthalate	40.000	48.530	-21.3	120	0.00
81	Benzo(a)anthracene	40.000	40.124	-0.3	111	0.00
82	3,3'-Dichlorobenzidine	40.000	42.132	-5.3	109	0.00
83	Chrysene	40.000	37.536	6.2	99	0.00
84	Bis(2-ethylhexyl)phthalate	40.000	44.958	-12.4	110	0.00
85 c	Di-n-octyl phthalate	40.000	41.738	-4.3	105	0.00
86 I	Perylene-d12	20.000	20.000	0.0	104	0.00
87	Indeno(1,2,3-cd)pyrene	40.000	38.847	2.9	100	0.00
88	Benzo(b)fluoranthene	40.000	38.959	2.6	102	0.00
89	Benzo(k)fluoranthene	40.000	39.275	1.8	103	0.00
90 C	Benzo(a)pyrene	40.000	39.796	0.5	103	0.00
91	Dibenzo(a,h)anthracene	40.000	38.739	3.2	99	0.00
92	Benzo(g,h,i)perylene	40.000	38.791	3.0	101	0.00

Data Path : Z:\svoasrv\HPCHEM1\BNA_F\Data\BF090924\
Data File : BF139280.D
Acq On : 09 Sep 2024 12:03
Operator : RC/JU
Sample : SSTDCCC040
Misc :
ALS Vial : 2 Sample Multiplier: 1

Instrument :
BNA_F
LabSampleId :
SSTDCCC040

Quant Time: Sep 10 02:21:40 2024
Quant Method : Z:\svoasrv\HPCHEM1\BNA_F\Methods\8270-BF090424.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Wed Sep 04 16:11:43 2024
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

Compound	Amount	Calc.	%Dev	Area%	Dev(min)
----------	--------	-------	------	-------	----------

(#) = Out of Range

SPCC's out = 0 CCC's out = 0