


Data Path : Z:\svoasrv\HPCHEM1\BNA_F\Data\BF090924\
Data File : BF139279.D
Acq On : 09 Sep 2024 11:33
Operator : RC/JU
Sample : DFTPP
Misc :
ALS Vial : 1 Sample Multiplier: 1

Instrument :
BNA_F
ClientSampleId :
DFTPP

Integration File: rteint.p

Method : Z:\svoasrv\HPCHEM1\BNA_F\Methods\8270-BF090424.M
Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
Last Update : Wed Sep 04 16:11:43 2024






AutoFind: Scans 1626, 1627, 1628; Background Corrected with Scan 1620

Target Mass	Rel. to Mass	Lower Limit%	Upper Limit%	Rel. Abn%	Raw Abn	Result Pass/Fail
51	198	10	80	50.0	64208	PASS
68	69	0.00	2	1.8	976	PASS
69	198	0.00	100	41.7	53477	PASS
70	69	0.00	2	0.5	273	PASS
127	198	10	80	49.5	63496	PASS
197	198	0.00	2	0.0	0	PASS
198	198	100	100	100.0	128376	PASS
199	198	5	9	6.8	8692	PASS
275	198	10	60	26.6	34163	PASS
365	198	1	100	3.3	4186	PASS
441	198	0.01	100	12.8	16485	PASS
442	442	50	100	100.0	101784	PASS
443	442	15	24	19.5	19824	PASS

Data Path : Z:\svoasrv\HPCHEM1\BNA_F\Data\BF090924\
Data File : BF139279.D
Acq On : 09 Sep 2024 11:33
Operator : RC/JU
Sample : DFTPP
Misc :
ALS Vial : 1 Sample Multiplier: 1

Instrument :
BNA_F
ClientSampleId :
DFTPP

Quant Time: Sep 09 12:37:29 2024
Quant Method : Z:\svoasrv\HPCHEM1\BNA_F\Methods\8270-BF090424.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Wed Sep 04 16:11:43 2024
Response via : Initial Calibration



TIC: BF139279.D\data.ms

(70) Pentachlorophenol (C)

11.210min (-0.006) 34138.67 ng




response 179607

Ion	Exp%	Act%
265.70	100.00	100.00
268.00	62.10	62.18
264.00	61.50	62.91
0.00	0.00	0.00

Data Path : Z:\svoasrv\HPCHEM1\BNA_F\Data\BF090924\
 Data File : BF139279.D
 Acq On : 09 Sep 2024 11:33
 Operator : RC/JU
 Sample : DFTPP
 Misc :
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 BNA_F
 ClientSampleId :
 DFTPP

Quant Time: Sep 09 12:37:29 2024
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_F\Methods\8270-BF090424.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Wed Sep 04 16:11:43 2024
 Response via : Initial Calibration



TIC: BF139279.D\data.ms

(77) Benzidine

12.751min (-0.000) 0.00 ng

response 774783

Ion	Exp%	Act%
184.00	100.00	100.00
185.00	14.90	14.15
183.00	11.00	10.90
0.00	0.00	0.00

DDT Breakdown

Date	Instrument Name	DFTPP Data File
9/9/2024	BNA_F	<u>BF139279.D</u>
Compound Name	Response	Retention Time
DDT	369567	13.574
DDD	7131	13.268
DDE	354	12.933
SUM(DDD+DDE)	SUM(DDT+DDD+DDE)	% Breakdown Of DDT
7485	377052	1.99

Instrument :
BNA_F
ClientSampleId :
DFTPP