

Data Path : Z:\SVOASRV\HPCHEM1\BNA F\DATA\BF091018\
 Data File : BF108893.D
 Acq On : 10 Sep 2018 15:23
 Operator : JU/SJ
 Sample : J4821-06 2X
 Misc :
 ALS Vial : 8 Sample Multiplier: 1

Instrument :
 BNA_F
 ClientSampleId :
 ENY-BC-02

Quant Time: Sep 10 16:36:23 2018
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA F\METHODS\8270-BF090518.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Wed Sep 05 18:03:31 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Dichlorobenzene-d4	6.74	152	206736	20.00	ng	0.00
21) Naphthalene-d8	8.02	136	747188	20.00	ng	0.00
38) Acenaphthene-d10	9.77	164	316869	20.00	ng	0.00
63) Phenanthrene-d10	11.25	188	524293	20.00	ng	0.00
75) Chrysene-d12	13.87	240	517445	20.00	ng	0.00
86) Perylene-d12	15.28	264	432015	20.00	ng	0.02
System Monitoring Compounds						
5) 2-Fluorophenol	5.34	112	591710	47.42	ng	0.00
7) Phenol-d6	6.37	99	757701	47.19	ng	0.00
23) Nitrobenzene-d5	7.30	82	472847	39.60	ng	0.00
41) 2,4,6-Tribromophenol	10.55	330	126443	41.80	ng	0.00
44) 2-Fluorobiphenyl	9.10	172	768293	41.36	ng	0.00
78) Terphenyl-d14	12.82	244	588194	26.51	ng	0.00
Target Compounds						
49) Dimethylphthalate	9.48	163	209515	9.453	ng	Qvalue 99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\SVOASRV\HPCHEM1\BNA F\DATA\BF091018\
Data File : BF108893.D
Acq On : 10 Sep 2018 15:23
Operator : JU/SJ
Sample : J4821-06 2X
Misc :
ALS Vial : 8 Sample Multiplier: 1

Instrument :
BNA_F
ClientSampleId :
ENY-BC-02

Quant Time: Sep 10 16:36:23 2018
Quant Method : Z:\SVOASRV\HPCHEM1\BNA F\METHODS\8270-BF090518.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Wed Sep 05 18:03:31 2018
Response via : Initial Calibration

