



2-Nitrophenol

Response Ratio



Resp Ratio = $2.00 \times 10^{-1} \times \text{Amt} - 2.56 \times 10^{-2}$
Coef of Det (r^2) = 0.999 Curve Fit: Linear


Method Name: Z:\SVOASRV\HPCHEM1\BNA_F\METHODS\8270-BF091020.M
Calibration Table Last Updated: Thu Sep 10 16:47:54 2020



Method Name: Z:\SVOASRV\HPCHEM1\BNA_F\METHODS\8270-BF091020.M
Calibration Table Last Updated: Thu Sep 10 16:47:54 2020

2,4-Dinitrophenol


Response Ratio



Resp Ratio = $2.01 \times 10^{-1} \times \text{Amt} - 9.08 \times 10^{-2}$
Coef of Det (r^2) = 0.991 Curve Fit: Linear

Method Name: Z:\SVOASRV\HPCHEM1\BNA_F\METHODS\8270-BF091020.M
Calibration Table Last Updated: Thu Sep 10 16:47:54 2020

4,6-Dinitro-2-methylphenol



Resp Ratio = $1.43 \times 10^{-1} \times \text{Amt} - 4.90 \times 10^{-2}$
Coef of Det (r^2) = 0.997 Curve Fit: Linear

Method Name: Z:\SVOASRV\HPCHEM1\BNA_F\METHODS\8270-BF091020.M
Calibration Table Last Updated: Thu Sep 10 16:47:54 2020