


Data Path : Z:\svoasrv\HPCHEM1\BNA\_F\Data\BF091020\  
 Data File : BF121546.D  
 Acq On : 10 Sep 2020 12:05  
 Operator : JU/CG  
 Sample : DFTPP  
 Misc :  
 ALS Vial : 1 Sample Multiplier: 1

Instrument :  
 BNA\_F  
 ClientSampleId :  
 DFTPP

Integration File: rteint.p

Method : Z:\SVOASRV\HPCHEM1\BNA\_F\METHODS\8270-BF091020.M  
 Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION  
 Last Update : Thu Sep 10 16:47:54 2020



Spectrum Information: Average of 11.604 to 11.616 min.

| Target Mass | Rel. to Mass | Lower Limit% | Upper Limit% | Rel. Abn% | Raw Abn | Result Pass/Fail |
|-------------|--------------|--------------|--------------|-----------|---------|------------------|
| 51          | 198          | 10           | 80           | 39.9      | 50662   | PASS             |
| 68          | 69           | 0.00         | 2            | 1.4       | 627     | PASS             |
| 69          | 198          | 0.00         | 100          | 35.0      | 44420   | PASS             |
| 70          | 69           | 0.00         | 2            | 0.3       | 113     | PASS             |
| 127         | 198          | 10           | 80           | 47.2      | 59994   | PASS             |
| 197         | 198          | 0.00         | 2            | 0.6       | 783     | PASS             |
| 198         | 198          | 100          | 100          | 100.0     | 127021  | PASS             |
| 199         | 198          | 5            | 9            | 6.8       | 8580    | PASS             |
| 275         | 198          | 10           | 60           | 28.0      | 35538   | PASS             |
| 365         | 198          | 1            | 100          | 3.9       | 4969    | PASS             |
| 441         | 198          | 0.01         | 100          | 15.3      | 19484   | PASS             |
| 442         | 198          | 50           | 100          | 97.2      | 123525  | PASS             |
| 443         | 442          | 15           | 24           | 18.6      | 23009   | PASS             |