


Data Path : \\74.0.250.170\svoasrv\HPCHEM1\BNA F\Data\BF091117\
 Data File : BF098411.D
 Acq On : 11 Sep 2017 19:29
 Operator : SJ/JU
 Sample : DFTPP
 Misc :
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 BNA_F
 ClientSampleId :
 DFTPP

Integration File: rteint.p

Method : Z:\HPCHEM1\BNA_F\METHODS\8270-BF091117.M
 Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 Last Update : Tue Sep 12 02:14:50 2017



AutoFind: Scans 1589, 1590, 1591; Background Corrected with Scan 1584

Target Mass	Rel. to Mass	Lower Limit%	Upper Limit%	Rel. Abn%	Raw Abn	Result Pass/Fail
51	198	10	80	55.4	66018	PASS
68	69	0.00	2	1.9	979	PASS
69	198	0.00	100	43.4	51642	PASS
70	69	0.00	2	0.6	323	PASS
127	198	10	80	54.5	64920	PASS
197	198	0.00	2	0.6	704	PASS
198	198	100	100	100.0	119114	PASS
199	198	5	9	6.8	8091	PASS
275	198	10	60	20.1	23943	PASS
365	198	1	100	1.9	2287	PASS
441	442	0.01	24	17.5	11325	PASS
442	198	50	100	54.4	64845	PASS
443	442	15	24	19.9	12912	PASS