




Data Path : Z:\svoasrv\HPCHEM1\BNA_F\Data\BF091124\
Data File : BF139314.D
Acq On : 11 Sep 2024 09:52
Operator : RC/JU
Sample : DFTPP
Misc :
ALS Vial : 1 Sample Multiplier: 1

Quant Time: Sep 11 11:06:02 2024
Quant Method : Z:\svoasrv\HPCHEM1\BNA_F\Methods\8270-BF090424.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Wed Sep 04 16:11:43 2024
Response via : Initial Calibration



TIC: BF139314.D\data.ms

(70) Pentachlorophenol (C)




11.210min (-0.006) 43719.77 ng

response 275456

Ion	Exp%	Act%
265.70	100.00	100.00
268.00	62.10	63.08
264.00	61.50	62.27
0.00	0.00	0.00

Data Path : Z:\svoasrv\HPCHEM1\BNA_F\Data\BF091124\
Data File : BF139314.D
Acq On : 11 Sep 2024 09:52
Operator : RC/JU
Sample : DFTPP
Misc :
ALS Vial : 1 Sample Multiplier: 1

Quant Time: Sep 11 11:06:02 2024
Quant Method : Z:\svoasrv\HPCHEM1\BNA_F\Methods\8270-BF090424.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Wed Sep 04 16:11:43 2024
Response via : Initial Calibration



TIC: BF139314.D\data.ms

(77) Benzidine

12.745min (-0.006) 0.00 ng

response 906153

Ion	Exp%	Act%
184.00	100.00	100.00
185.00	14.90	14.02
183.00	11.00	11.42
0.00	0.00	0.00