

Data Path : \\74.0.250.170\SV0ASRV\HPCHEM1\BNA F\DATA\BF091217\
 Data File : BF098461.D
 Acq On : 12 Sep 2017 20:19
 Operator : SJ/JU
 Sample : I5140-01
 Misc :
 ALS Vial : 16 Sample Multiplier: 1

Instrument :
 BNA_F
 ClientSampleId :
 RL-5-0-13.75

Manual Integrations
 APPROVED

Sohil
 9/13/2017 5:00:08 PM

Quant Time: Sep 13 02:01:42 2017
 Quant Method : Z:\HPCHEM1\BNA F\METHODS\8270-BF091117.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Tue Sep 12 02:14:50 2017
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Dichlorobenzene-d4	6.66	152	197286	20.00	ng	0.00
21) Naphthalene-d8	7.95	136	770085	20.00	ng	0.00
38) Acenaphthene-d10	9.69	164	296347	20.00	ng	-0.01
63) Phenanthrene-d10	11.17	188	464458	20.00	ng	-0.01
75) Chrysene-d12	13.81	240	358085	20.00	ng	0.00
86) Perylene-d12	15.21	264	266597	20.00	ng	0.00
System Monitoring Compounds						
5) 2-Fluorophenol	5.27	112	690013	51.71	ng	0.00
7) Phenol-d6	6.31	99	861972	50.88	ng	-0.01
23) Nitrobenzene-d5	7.23	82	508295	36.80	ng	-0.01
41) 2,4,6-Tribromophenol	10.49	330	94159	47.61	ng	0.00
44) 2-Fluorobiphenyl	9.02	172	727107	41.59	ng	-0.01
78) Terphenyl-d14	12.76	244	436392	26.30	ng	-0.01
Target Compounds						
31) Naphthalene	7.97	128	179125	4.705	ng	99
49) Dimethylphthalate	9.42	163	102618	4.473	ng	100
70) Phenanthrene	11.20	178	67943	2.759	ng	98
74) Fluoranthene	12.39	202	94584	3.837	ng	98
77) Pyrene	12.62	202	83706	2.963	ng	98
87) Benzo(b)fluoranthene	14.82	252	39556m	2.360	ng	

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : \\74.0.250.170\SVOASRV\HPCHEM1\BNA F\DATA\BF091217\
Data File : BF098461.D
Acq On : 12 Sep 2017 20:19
Operator : SJ/JU
Sample : I5140-01
Misc :
ALS Vial : 16 Sample Multiplier: 1

Instrument :
BNA_F
ClientSampleId :
RL-5-0-13.75

Manual Integrations
APPROVED

Sohil
9/13/2017 5:00:08 PM

Quant Time: Sep 13 02:01:42 2017
Quant Method : Z:\HPCHEM1\BNA F\METHODS\8270-BF091117.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Tue Sep 12 02:14:50 2017
Response via : Initial Calibration

