

Data Path : \\74.0.250.170\SV0ASRV\HPCHEM1\BNA F\DATA\BF091217\
 Data File : BF098457.D
 Acq On : 12 Sep 2017 18:27
 Operator : SJ/JU
 Sample : PB102209BL
 Misc :
 ALS Vial : 12 Sample Multiplier: 1

Instrument :
 BNA_F
 ClientSampleId :
 PB102209BL

Integration Parameters: rteint.p
 Integrator: RTE
 Smoothing : ON
 Sampling : 1
 Start Thrs: 0.2
 Stop Thrs : 0
 Filtering: 5
 Min Area: 3 % of largest Peak
 Max Peaks: 100
 Peak Location: TOP

If leading or trailing edge < 100 prefer < Baseline drop else tangent >
 Peak separation: 5

Method : Z:\HPCHEM1\BNA_F\METHODS\8270-BF091117.M
 Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION

Signal : TIC

peak #	R.T. min	first scan	max scan	last scan	PK TY	peak height	corr. area	corr. % max.	% of total
1	4.869	469	473	488	rBV	379640	658252	9.51%	1.237%
2	5.281	538	543	558	rBV	5980063	6919819	100.00%	13.008%
3	6.316	713	719	722	rBV	5536586	6646190	96.05%	12.494%
4	6.439	735	740	743	rBV	6014828	6445890	93.15%	12.117%
5	6.657	774	777	781	rBV	1552041	1423626	20.57%	2.676%
6	6.816	800	804	808	rBV	5038242	5022835	72.59%	9.442%
7	7.233	869	875	878	rBV	4012522	4215625	60.92%	7.925%
8	7.945	991	996	999	rBV	2111770	1873826	27.08%	3.523%
9	9.022	1173	1179	1182	rBV	6132196	6743969	97.46%	12.678%
10	9.692	1288	1293	1297	rBV	1944980	1875281	27.10%	3.525%
11	10.486	1423	1428	1431	rBV	3198144	2949608	42.63%	5.545%
12	11.174	1541	1545	1549	rBV	1893740	1582596	22.87%	2.975%
13	12.762	1810	1815	1818	rBV	4597642	4429918	64.02%	8.328%
14	13.804	1988	1992	2001	rVB	1140006	1165906	16.85%	2.192%
15	15.204	2226	2230	2241	rVB	811991	1086374	15.70%	2.042%
16	15.621	2297	2301	2305	rBV2	50355	70451	1.02%	0.132%
17	16.074	2374	2378	2384	rVB2	50183	85007	1.23%	0.160%


Sum of corrected areas: 53195173

Data Path : \\74.0.250.170\SV0ASRV\HPCHEM1\BNA F\DATA\BF091217\
Data File : BF098457.D
Acq On : 12 Sep 2017 18:27
Operator : SJ/JU
Sample : PB102209BL
Misc :
ALS Vial : 12 Sample Multiplier: 1

Instrument :
BNA_F
ClientSampleId :
PB102209BL

Quant Method : Z:\HPCHEM1\BNA_F\METHODS\8270-BF091117.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION

TIC Library : C:\DATABASE\NIST11.L
TIC Integration Parameters: LSCINT.P



Data Path : \\74.0.250.170\SV0ASRV\HPCHEM1\BNA F\DATA\BF091217\
Data File : BF098457.D
Acq On : 12 Sep 2017 18:27
Operator : SJ/JU
Sample : PB102209BL
Misc :
ALS Vial : 12 Sample Multiplier: 1

Instrument :
BNA_F
ClientSampleId :
PB102209BL










Quant Method : Z:\HPCHEM1\BNA F\METHODS\8270-BF091117.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION

TIC Library : C:\DATABASE\NIST11.L
TIC Integration Parameters: LSCINT.P

Peak Number 1 2-Pentanone, 4-hydroxy-4-me... Concentration Rank 3

R.T.	EstConc	Area	Relative to ISTD	R.T.
4.87	9.25 ng	658252	1,4-Dichlorobenzene-d4	6.66

Hit#	of	5	Tentative ID	MW	MolForm	CAS#	Qual
1			2-Pentanone, 4-hydroxy-4-methyl-	116	C6H12O2	000123-42-2	59
2			Acetic acid, 1,1-dimethylethyl e...	116	C6H12O2	000540-88-5	38
3			Acetic acid, cyano-, 1,1-dimethy...	141	C7H11NO2	001116-98-9	37
4			2-Hexanol, 2-methyl-	116	C7H16O	000625-23-0	28
5			2,3-Butanedione, monooxime	101	C4H7NO2	000057-71-6	16



Data Path : \\74.0.250.170\SVOASRV\HPCHEM1\BNA F\DATA\BF091217\
Data File : BF098457.D
Acq On : 12 Sep 2017 18:27
Operator : SJ/JU
Sample : PB102209BL
Misc :
ALS Vial : 12 Sample Multiplier: 1

Instrument :
BNA_F
ClientSampleId :
PB102209BL








Quant Method : Z:\HPCHEM1\BNA F\METHODS\8270-BF091117.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION

TIC Library : C:\DATABASE\NIST11.L
TIC Integration Parameters: LSCINT.P

Peak Number 2 unknown6.44 Concentration Rank 1

R.T.	EstConc	Area	Relative to ISTD	R.T.
6.44	90.56 ng	6445890	1,4-Dichlorobenzene-d4	6.66

Hit#	of	Tentative ID	MW	MolForm	CAS#	Qual
1	5	3-Chloro-6-fluoro-pyrazine	132	C4H2ClFN2	1000146-10-7	27
2		5-Fluoro-2-chloropyrimidine	132	C4H2ClFN2	062802-42-0	16
3		Tranvlcypropine-propionyl	189	C12H15NO	1000123-86-3	12
4		1,3,5-Trifluorobenzene	132	C6H3F3	000372-38-3	9
5		(E)-3-Chloro-2-methyl-2-pentenal	132	C6H9ClO	031357-76-3	9



Data Path : \\74.0.250.170\SV0ASRV\HPCHEM1\BNA_F\DATA\BF091217\
Data File : BF098457.D
Acq On : 12 Sep 2017 18:27
Operator : SJ/JU
Sample : PB102209BL
Misc :
ALS Vial : 12 Sample Multiplier: 1

Instrument :
BNA_F
ClientSampleId :
PB102209BL

Quant Method : Z:\HPCHEM1\BNA_F\METHODS\8270-BF091117.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION

TIC Library : C:\DATABASE\NIST11.L
TIC Integration Parameters: LSCINT.P

TIC Top Hit name	RT	EstConc	Units	Response	--Internal Standard--			
					#	RT	Resp	Conc
2-Pentanone, 4-hy...	4.87	9.3	ng	658252	1	6.66	1423630	20.0
unknown6.44	6.44	90.6	ng	6445890	1	6.66	1423630	20.0