

Data Path : Z:\HPCHEM1\BNA_F\DATA\BF091415\
 Data File : BF081602.D
 Acq On : 14 Sep 2015 11:25
 Operator : UM/IZ
 Sample : SSTDCCC040
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_F
 ClientSampleId :
 SSTDCCC040

Quant Time: Sep 14 13:39:48 2015
 Quant Method : Z:\HPCHEM1\BNA_F\METHODS\8270-BF090115.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Thu Sep 10 16:31:29 2015
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Dichlorobenzene-d4	8.23	152	63074	20.00	ng	-0.03
21) Naphthalene-d8	11.13	136	256903	20.00	ng	-0.03
38) Acenaphthene-d10	15.27	164	125496	20.00	ng	-0.03
63) Phenanthrene-d10	18.77	188	254809	20.00	ng	-0.05
75) Chrysene-d12	23.40	240	208529	20.00	ng	-0.02
86) Perylene-d12	24.82	264	166412	20.00	ng	-0.02

System Monitoring Compounds						
5) 2-Fluorophenol	5.37	112	316085	77.75	ng	-0.05
7) Phenol-d6	7.66	99	431515	80.19	ng	-0.03
23) Nitrobenzene-d5	9.54	82	430705	80.89	ng	-0.02
41) 2,4,6-Tribromophenol	17.18	330	90070	80.61	ng	-0.03
44) 2-Fluorobiphenyl	13.78	172	773973	79.03	ng	-0.03
78) Terphenyl-d14	22.12	244	717744	80.12	ng	-0.03

Target Compounds						Qvalue
2) 1,4-Dioxane	1.57	88	83795	45.90	ng	85
3) Pyridine	2.08	79	198074	39.34	ng	# 75
4) n-Nitrosodimethylamine	2.06	42	121286	43.36	ng	# 61
6) Aniline	7.53	93	295197	38.85	ng	# 85
8) 2-Chlorophenol	7.78	128	177292	39.58	ng	# 81
9) Benzaldehyde	7.26	77	117513	34.85	ng	# 75
10) Phenol	7.68	94	223483	35.81	ng	# 53
11) bis(2-Chloroethyl)ether	7.77	93	180844	40.17	ng	# 64
12) 1,3-Dichlorobenzene	8.08	146	188112	39.30	ng	# 88
13) 1,4-Dichlorobenzene	8.27	146	188166	38.61	ng	# 91
14) 1,2-Dichlorobenzene	8.59	146	180484	41.13	ng	# 91
15) Benzyl Alcohol	8.67	79	182682	41.38	ng	# 72
16) 2,2'-oxybis(1-Chloropropan	8.99	45	384598	44.56	ng	80
17) 2-Methylphenol	9.02	107	144492	40.42	ng	# 81
18) Hexachloroethane	9.33	117	78617	41.46	ng	88
19) n-Nitroso-di-n-propylamine	9.31	70	160603	43.85	ng	# 88
20) 3+4-Methylphenols	9.39	107	192020	39.84	ng	86
22) Acetophenone	9.23	105	271165	38.64	ng	# 76
24) Nitrobenzene	9.58	77	229308	41.47	ng	# 63
25) Isophorone	10.18	82	410555	41.45	ng	# 84
26) 2-Nitrophenol	10.31	139	92958	38.57	ng	# 29
27) 2,4-Dimethylphenol	10.58	122	162047	38.93	ng	# 74
28) bis(2-Chloroethoxy)methane	10.78	93	236520	41.34	ng	# 91
29) 2,4-Dichlorophenol	10.91	162	142570	39.50	ng	90
30) 1,2,4-Trichlorobenzene	11.04	180	159881	40.77	ng	96
31) Naphthalene	11.19	128	537832	39.26	ng	99
32) Benzoic acid	11.10	122	97909	33.74	ng	# 62
33) 4-Chloroaniline	11.43	127	226134	40.47	ng	# 86
34) Hexachlorobutadiene	11.54	225	101745	42.63	ng	98
35) Caprolactam	12.39	113	34762	31.40	ng	# 23
36) 4-Chloro-3-methylphenol	12.73	107	175007	43.31	ng	# 74
37) 2-Methylnaphthalene	12.84	142	359066	40.27	ng	# 88
39) 1,2,4,5-Tetrachlorobenzene	13.25	216	154805	39.00	ng	99
40) Hexachlorocyclopentadiene	13.22	237	77913	40.00	ng	97


Data Path : Z:\HPCHEM1\BNA_F\DATA\BF091415\
 Data File : BF081602.D
 Acq On : 14 Sep 2015 11:25
 Operator : UM/IZ
 Sample : SSTDCCC040
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_F
 ClientSampleId :
 SSTDCCC040

Quant Time: Sep 14 13:39:48 2015
 Quant Method : Z:\HPCHEM1\BNA_F\METHODS\8270-BF090115.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Thu Sep 10 16:31:29 2015
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
42) 2,4,6-Trichlorophenol	13.59	196	110490	40.34	ng	98
43) 2,4,5-Trichlorophenol	13.69	196	112235	40.16	ng #	90
45) 1,1'-Biphenyl	13.99	154	438211	37.95	ng	96
46) 2-Chloronaphthalene	13.97	162	324390	38.91	ng #	87
47) 2-Nitroaniline	14.32	65	142558	41.95	ng #	64
48) Acenaphthylene	14.93	152	563728	38.11	ng	97
49) Dimethylphthalate	14.87	163	399030	40.92	ng #	96
50) 2,6-Dinitrotoluene	14.96	165	82455	37.26	ng #	39
51) Acenaphthene	15.35	154	316869	37.66	ng	88
52) 3-Nitroaniline	15.33	138	99284	39.41	ng #	59
53) 2,4-Dinitrophenol	15.60	184	41146	40.09	ng #	26
54) Dibenzofuran	15.77	168	470216	38.27	ng #	85
55) 4-Nitrophenol	15.93	139	62246	39.33	ng #	14
56) 2,4-Dinitrotoluene	15.91	165	113656	40.33	ng #	63
57) Fluorene	16.58	166	387120	41.52	ng	97
58) 2,3,4,6-Tetrachlorophenol	16.04	232	84257	39.49	ng	96
59) Diethylphthalate	16.56	149	431721	42.09	ng	95
60) 4-Chlorophenyl-phenylether	16.68	204	189150	43.52	ng #	84
61) 4-Nitroaniline	16.80	138	96178	37.79	ng #	32
62) Azobenzene	17.04	77	481776	42.25	ng	82
64) 4,6-Dinitro-2-methylphenol	16.87	198	60896	42.73	ng #	78
65) n-Nitrosodiphenylamine	17.00	169	346942	37.42	ng	90
66) 4-Bromophenyl-phenylether	17.82	248	105016	40.76	ng #	91
67) Hexachlorobenzene	17.86	284	106912	39.69	ng #	84
68) Atrazine	18.40	200	110804	41.30	ng	79
69) Pentachlorophenol	18.41	266	52233	44.22	ng	99
70) Phenanthrene	18.84	178	552566	39.01	ng	97
71) Anthracene	18.95	178	560341	38.50	ng	97
72) Carbazole	19.43	167	517136	37.87	ng	96
73) Di-n-butylphthalate	20.48	149	711409	44.11	ng #	98
74) Fluoranthene	21.44	202	603803	39.37	ng	88
76) Benzidine	21.75	184	295989	33.06	ng	98
77) Pyrene	21.78	202	598063	38.72	ng	98
79) Butylbenzylphthalate	22.81	149	309153	46.02	ng	89
80) Benzo(a)anthracene	23.37	228	515113	38.79	ng	96
81) 3,3'-Dichlorobenzidine	23.38	252	175938	39.28	ng #	96
82) Chrysene	23.42	228	446663	37.52	ng	95
83) Bis(2-ethylhexyl)phthalate	23.50	149	405219	45.71	ng #	92
84) Di-n-octyl phthalate	24.15	149	669804	45.88	ng	97
85) Indeno(1,2,3-cd)pyrene	25.88	276	378179	43.30	ng #	85
87) Benzo(b)fluoranthene	24.48	252	847581	71.34	ng #	86
88) Benzo(k)fluoranthene	24.48	252	847581	85.98	ng #	87
89) Benzo(a)pyrene	24.78	252	414250	41.15	ng #	81
90) Dibenzo(a,h)anthracene	25.89	278	320052	41.14	ng #	77
91) Benzo(g,h,i)perylene	26.18	276	293332	39.51	ng #	69

(#) = qualifier out of range (m) = manual integration (+) = signals summed



#1

1,4-Dichlorobenzene-d4

Concen: 20.00 ng

RT: 8.23 min Scan# 603

Delta R.T. -0.03 min

Lab File: BF081602.D

Acq: 14 Sep 2015 11:25

Instrument :

BNA_F

ClientSampleId :

SSTDCCC040





Tgt Ion:152 Resp: 63074

Ion Ratio Lower Upper

152 100

150 156.3 124.7 187.1

115 64.7 39.0 58.4#



#2

1,4-Dioxane

Concen: 45.90 ng

RT: 1.57 min Scan# 20

Delta R.T. -0.03 min

Lab File: BF081602.D

Acq: 14 Sep 2015 11:25




Tgt Ion: 88 Resp: 83795


Ion Ratio Lower Upper

88 100

58 79.0 77.7 116.5

43 36.2 26.6 40.0





#3

Pyridine

Concen: 39.34 ng

RT: 2.08 min Scan# 65

Delta R.T. -0.09 min

Lab File: BF081602.D

Acq: 14 Sep 2015 11:25

Instrument :

BNA_F

ClientSampleId :

SSTDCCC040


Tgt Ion: 79 Resp: 198074

Ion Ratio Lower Upper

79 100

52 119.2 76.8 115.2#




51 57.1 32.2 48.4#



Abundance Ion 79.00 (78.70 to 79.70): BF0

Ion 52.00 (51.70 to 52.70): BF0

Ion 51.00 (50.70 to 51.70): BF0



#4

n-Nitrosodimethylamine

Concen: 43.36 ng

RT: 2.06 min Scan# 63

Delta R.T. -0.03 min

Lab File: BF081602.D

Acq: 14 Sep 2015 11:25


Tgt Ion: 42 Resp: 121286

Ion Ratio Lower Upper

42 100

74 83.5 105.0 157.6#



44 7.1 6.6 10.0




Abundance Ion 42.00 (41.70 to 42.70): BF0

Ion 74.00 (73.70 to 74.70): BF0

Ion 44.00 (43.70 to 44.70): BF0





#5

2-Fluorophenol

Concen: 77.75 ng

RT: 5.37 min Scan# 353

Delta R.T. -0.05 min

Lab File: BF081602.D

Acq: 14 Sep 2015 11:25

Instrument :


BNA_F

ClientSampleId :

SSTDCCC040

Tgt Ion: 112 Resp: 316085




Ion	Ratio	Lower	Upper
112	100		
64	72.3	39.7	59.5#
63	39.5	17.4	26.0#



Abundance Ion 112.00 (111.70 to 112.70): E

Ion 64.00 (63.70 to 64.70): BF0

Ion 63.00 (62.70 to 63.70): BF0



#6

Aniline

Concen: 38.85 ng

RT: 7.53 min Scan# 542


Delta R.T. -0.05 min

Lab File: BF081602.D

Acq: 14 Sep 2015 11:25

Tgt Ion: 93 Resp: 295197



Ion	Ratio	Lower	Upper
93	100		
66	44.2	27.2	40.8#
65	23.2	14.7	22.1#




Abundance Ion 93.00 (92.70 to 93.70): BF0

Ion 66.00 (65.70 to 66.70): BF0

Ion 65.00 (64.70 to 65.70): BF0





#7

Phenol-d6

Concen: 80.19 ng

RT: 7.66 min Scan# 553

Delta R.T. -0.03 min

Lab File: BF081602.D

Acq: 14 Sep 2015 11:25

Instrument :

BNA_F

ClientSampleId :

SSTDCCC040


Tgt Ion: 99 Resp: 431515

Ion Ratio Lower Upper

99 100

42 25.6 13.7 20.5#




71 39.1 14.2 21.2#



Abundance Ion 99.00 (98.70 to 99.70): BF0

Ion 42.00 (41.70 to 42.70): BF0

Ion 71.00 (70.70 to 71.70): BF0



#8

2-Chlorophenol

Concen: 39.58 ng

RT: 7.78 min Scan# 564

Delta R.T. -0.03 min

Lab File: BF081602.D

Acq: 14 Sep 2015 11:25


Tgt Ion: 128 Resp: 177292

Ion Ratio Lower Upper

128 100

130 32.0 12.2 52.2



64 61.2 20.5 60.5#




Abundance Ion 128.00 (127.70 to 128.70): E

Ion 130.00 (129.70 to 130.70): E

Ion 64.00 (63.70 to 64.70): BF0





#9

Benzaldehyde

Concen: 34.85 ng

RT: 7.26 min Scan# 518

Delta R.T. -0.03 min

Lab File: BF081602.D

Acq: 14 Sep 2015 11:25

Instrument :

BNA_F

ClientSampleId :

SSTDCCC040





Tgt Ion: 77 Resp: 117513

Ion Ratio Lower Upper

77 100

105 92.2 91.6 131.6

106 87.0 102.6 142.6#



#10

Phenol

Concen: 35.81 ng

RT: 7.68 min Scan# 555

Delta R.T. -0.03 min

Lab File: BF081602.D

Acq: 14 Sep 2015 11:25




Tgt Ion: 94 Resp: 223483


Ion Ratio Lower Upper

94 100

65 35.8 0.7 40.7

66 49.1 0.8 40.8#









#11
bis(2-Chloroethyl)ether
Concen: 40.17 ng
RT: 7.77 min Scan# 563
Delta R.T. -0.03 min
Lab File: BF081602.D
Acq: 14 Sep 2015 11:25




Instrument :
BNA_F
ClientSampleId :
SSTDCCC040


Tgt Ion: 93 Resp: 180844
Ion Ratio Lower Upper
93 100
63 130.3 65.6 105.6#
95 31.3 13.6 53.6



#12
1,3-Dichlorobenzene
Concen: 39.30 ng
RT: 8.08 min Scan# 590
Delta R.T. -0.03 min
Lab File: BF081602.D
Acq: 14 Sep 2015 11:25

Tgt Ion: 146 Resp: 188112
Ion Ratio Lower Upper
146 100
148 62.1 51.0 76.4
75 40.1 15.0 22.6#









#13
 1,4-Dichlorobenzene
 Concen: 38.61 ng
 RT: 8.27 min Scan# 607
 Delta R.T. -0.02 min
 Lab File: BF081602.D
 Acq: 14 Sep 2015 11:25




Instrument :
 BNA_F
 ClientSampleId :
 SSTDCCC040


Tgt Ion:146 Resp: 188166
 Ion Ratio Lower Upper
 146 100
 148 65.2 51.8 77.6
 111 45.5 24.9 37.3#



#14
 1,2-Dichlorobenzene
 Concen: 41.13 ng
 RT: 8.59 min Scan# 635
 Delta R.T. -0.02 min
 Lab File: BF081602.D
 Acq: 14 Sep 2015 11:25

Tgt Ion:146 Resp: 180484
 Ion Ratio Lower Upper
 146 100
 148 64.3 52.7 79.1
 111 48.0 28.8 43.2#









#15
Benzyl Alcohol
Concen: 41.38 ng
RT: 8.67 min Scan# 642
Delta R.T. -0.05 min
Lab File: BF081602.D
Acq: 14 Sep 2015 11:25




Instrument :
BNA_F
ClientSampleId :
SSTDCCC040


Tgt Ion: 79 Resp: 182682
Ion Ratio Lower Upper
79 100
108 66.9 88.6 132.8#
77 61.5 47.0 70.4



#16
2,2'-oxybis(1-Chloropropane)
Concen: 44.56 ng
RT: 8.99 min Scan# 670
Delta R.T. -0.03 min
Lab File: BF081602.D
Acq: 14 Sep 2015 11:25

Tgt Ion: 45 Resp: 384598
Ion Ratio Lower Upper
45 100
77 15.3 0.0 28.1
79 12.8 0.0 26.0









#17
 2-Methylphenol
 Concen: 40.42 ng
 RT: 9.02 min Scan# 672
 Delta R.T. -0.03 min
 Lab File: BF081602.D
 Acq: 14 Sep 2015 11:25




Instrument :
 BNA_F
 ClientSampleId :
 SSTDCCC040


Tgt Ion:	107	Resp:	144492
Ion	Ratio	Lower	Upper
107	100		
108	117.2	96.6	145.0
77	58.3	23.3	34.9#
79	55.2	22.0	33.0#



#18
 Hexachloroethane
 Concen: 41.46 ng
 RT: 9.33 min Scan# 699
 Delta R.T. -0.03 min
 Lab File: BF081602.D
 Acq: 14 Sep 2015 11:25

Tgt Ion:	117	Resp:	78617
Ion	Ratio	Lower	Upper
117	100		
119	93.8	83.8	125.6
201	80.5	55.1	82.7









#19
n-Nitroso-di-n-propylamine
Concen: 43.85 ng
RT: 9.31 min Scan# 698
Delta R.T. -0.03 min
Lab File: BF081602.D
Acq: 14 Sep 2015 11:25




Instrument :
BNA_F
ClientSampleId :
SSTDCCC040


Tgt Ion: 70 Resp: 160603
Ion Ratio Lower Upper
70 100
42 74.3 63.8 95.6
101 8.7 9.2 13.8#
130 17.5 27.3 40.9#



#20
3+4-Methylphenols
Concen: 39.84 ng
RT: 9.39 min Scan# 705
Delta R.T. -0.03 min
Lab File: BF081602.D
Acq: 14 Sep 2015 11:25

Tgt Ion: 107 Resp: 192020
Ion Ratio Lower Upper
107 100
108 90.4 74.6 114.6
77 39.6 0.7 40.7
79 33.6 0.0 38.9









#21
Naphthalene-d8
Concen: 20.00 ng
RT: 11.13 min Scan# 857
Delta R.T. -0.03 min
Lab File: BF081602.D
Acq: 14 Sep 2015 11:25




Instrument :
BNA_F
ClientSampleId :
SSTDCCC040


Tgt Ion:	136	Resp:	256903
Ion	Ratio	Lower	Upper
136	100		
137	10.4	9.0	13.6
54	12.6	8.2	12.2#
68	5.0	5.8	8.6#



#22
Acetophenone
Concen: 38.64 ng
RT: 9.23 min Scan# 691
Delta R.T. -0.03 min
Lab File: BF081602.D
Acq: 14 Sep 2015 11:25

Tgt Ion:	105	Resp:	271165
Ion	Ratio	Lower	Upper
105	100		
71	0.0	4.7	7.1#
51	36.3	22.0	33.0#
120	19.4	30.1	45.1#









#23
 Nitrobenzene-d5
 Concen: 80.89 ng
 RT: 9.54 min Scan# 718
 Delta R.T. -0.02 min
 Lab File: BF081602.D
 Acq: 14 Sep 2015 11:25




Instrument :
 BNA_F
 ClientSampleId :
 SSTDCCC040


Tgt Ion: 82 Resp: 430705
 Ion Ratio Lower Upper
 82 100
 128 37.3 51.0 76.6#
 54 67.0 47.5 71.3



#24
 Nitrobenzene
 Concen: 41.47 ng
 RT: 9.58 min Scan# 721
 Delta R.T. -0.03 min
 Lab File: BF081602.D
 Acq: 14 Sep 2015 11:25

Tgt Ion: 77 Resp: 229308
 Ion Ratio Lower Upper
 77 100
 123 39.1 59.8 89.8#
 65 16.7 10.2 15.4#





#25

Isophorone

Concen: 41.45 ng

RT: 10.18 min Scan# 774

Delta R.T. -0.05 min

Lab File: BF081602.D

Acq: 14 Sep 2015 11:25

Instrument :

BNA_F

ClientSampleId :

SSTDCCC040


Tgt Ion: 82 Resp: 410555

Ion Ratio Lower Upper

82 100

95 7.9 4.0 6.0#




138 14.2 18.3 27.5#



Abundance Ion 82.00 (81.70 to 82.70): BF0

Ion 95.00 (94.70 to 95.70): BF0

Ion 138.00 (137.70 to 138.70): BF0



#26

2-Nitrophenol

Concen: 38.57 ng

RT: 10.31 min Scan# 785

Delta R.T. -0.03 min

Lab File: BF081602.D

Acq: 14 Sep 2015 11:25


Tgt Ion: 139 Resp: 92958

Ion Ratio Lower Upper

139 100

109 36.1 11.0 16.4#



65 73.3 24.7 37.1#




Abundance Ion 139.00 (138.70 to 139.70): E

Ion 109.00 (108.70 to 109.70): E

Ion 65.00 (64.70 to 65.70): BF0









#27
 2,4-Dimethylphenol
 Concen: 38.93 ng
 RT: 10.58 min Scan# 809
 Delta R.T. -0.03 min
 Lab File: BF081602.D
 Acq: 14 Sep 2015 11:25




Instrument :
 BNA_F
 ClientSampleId :
 SSTDCCC040


Tgt Ion:122 Resp: 162047
 Ion Ratio Lower Upper
 122 100
 107 125.0 71.8 107.6#
 121 57.7 42.3 63.5



#28
 bis(2-Chloroethoxy)methane
 Concen: 41.34 ng
 RT: 10.78 min Scan# 826
 Delta R.T. -0.02 min
 Lab File: BF081602.D
 Acq: 14 Sep 2015 11:25

Tgt Ion: 93 Resp: 236520
 Ion Ratio Lower Upper
 93 100
 95 31.7 27.5 41.3
 123 9.4 13.7 20.5#









#29
2,4-Dichlorophenol
Concen: 39.50 ng
RT: 10.91 min Scan# 838
Delta R.T. -0.03 min
Lab File: BF081602.D
Acq: 14 Sep 2015 11:25




Instrument :
BNA_F
ClientSampleId :
SSTDCCC040


Tgt Ion:162 Resp: 142570
Ion Ratio Lower Upper
162 100
164 62.7 44.9 84.9
98 47.2 13.0 53.0



#30
1,2,4-Trichlorobenzene
Concen: 40.77 ng
RT: 11.04 min Scan# 849
Delta R.T. -0.03 min
Lab File: BF081602.D
Acq: 14 Sep 2015 11:25

Tgt Ion:180 Resp: 159881
Ion Ratio Lower Upper
180 100
182 93.8 77.1 115.7
145 33.0 23.3 34.9





#31

Naphthalene

Concen: 39.26 ng

RT: 11.19 min Scan# 862

Delta R.T. -0.02 min

Lab File: BF081602.D

Acq: 14 Sep 2015 11:25

Instrument :

BNA_F

ClientSampleId :

SSTDCCC040


Tgt Ion:128 Resp: 537832

Ion Ratio Lower Upper

128 100

129 10.5 8.4 12.6




127 13.2 9.9 14.9



Abundance Ion 128.00 (127.70 to 128.70): E

Ion 129.00 (128.70 to 129.70): E

Ion 127.00 (126.70 to 127.70): E



#32

Benzoic acid

Concen: 33.74 ng

RT: 11.10 min Scan# 854

Delta R.T. 0.01 min

Lab File: BF081602.D

Acq: 14 Sep 2015 11:25


Tgt Ion:122 Resp: 97909

Ion Ratio Lower Upper

122 100

105 122.0 76.1 116.1#



77 103.3 40.5 80.5#




Abundance Ion 122.00 (121.70 to 122.70): E

Ion 105.00 (104.70 to 105.70): E

Ion 77.00 (76.70 to 77.70): BF0









#33
4-Chloroaniline
Concen: 40.47 ng
RT: 11.43 min Scan# 883
Delta R.T. -0.03 min
Lab File: BF081602.D
Acq: 14 Sep 2015 11:25

Instrument :
BNA_F
ClientSampleId :
SSTDCCC040

Tgt Ion:	127	Resp:	226134
Ion	Ratio	Lower	Upper
127	100		
129	32.8	25.8	38.6
65	36.5	17.9	26.9#
92	20.9	10.5	15.7#




Abundance Ion 127.00 (126.70 to 127.70): E
Ion 129.00 (128.70 to 129.70): E
Ion 65.00 (64.70 to 65.70): BF0
Ion 92.00 (91.70 to 92.70): BF0






#34
Hexachlorobutadiene
Concen: 42.63 ng
RT: 11.54 min Scan# 893
Delta R.T. -0.02 min
Lab File: BF081602.D
Acq: 14 Sep 2015 11:25

Tgt Ion:	225	Resp:	101745
Ion	Ratio	Lower	Upper
225	100		
223	62.6	50.2	75.2
227	62.8	52.2	78.2



Abundance Ion 225.00 (224.70 to 225.70): E
Ion 223.00 (222.70 to 223.70): E
Ion 227.00 (226.70 to 227.70): E





#35

Caprolactam

Concen: 31.40 ng

RT: 12.39 min Scan# 967

Delta R.T. -0.07 min

Lab File: BF081602.D

Acq: 14 Sep 2015 11:25

Instrument :

BNA_F

ClientSampleId :

SSTDCCC040


Tgt Ion:113 Resp: 34762

Ion Ratio Lower Upper

113 100

55 279.7 152.4 192.4#




56 211.2 104.6 144.6#



Abundance Ion 113.00 (112.70 to 113.70): E

Ion 55.00 (54.70 to 55.70): BF0

Ion 56.00 (55.70 to 56.70): BF0



#36

4-Chloro-3-methylphenol

Concen: 43.31 ng

RT: 12.73 min Scan# 997

Delta R.T. -0.03 min

Lab File: BF081602.D

Acq: 14 Sep 2015 11:25


Tgt Ion:107 Resp: 175007

Ion Ratio Lower Upper

107 100

144 21.6 25.4 38.0#



142 68.2 77.3 115.9#




Abundance Ion 107.00 (106.70 to 107.70): E

Ion 144.00 (143.70 to 144.70): E

Ion 142.00 (141.70 to 142.70): E





#37

2-Methylnaphthalene

Concen: 40.27 ng

RT: 12.84 min Scan# 1006

Delta R.T. -0.03 min

Lab File: BF081602.D

Acq: 14 Sep 2015 11:25

Instrument :

BNA_F

ClientSampleId :

SSTDCCC040


Tgt Ion:142 Resp: 359066

Ion Ratio Lower Upper

142 100

141 88.8 67.5 101.3




115 41.0 18.2 27.2#



Abundance Ion 142.00 (141.70 to 142.70): E

Ion 141.00 (140.70 to 141.70): E

Ion 115.00 (114.70 to 115.70): E



#38

Acenaphthene-d10

Concen: 20.00 ng

RT: 15.27 min Scan# 1219

Delta R.T. -0.03 min

Lab File: BF081602.D

Acq: 14 Sep 2015 11:25


Tgt Ion:164 Resp: 125496

Ion Ratio Lower Upper

164 100

162 100.1 75.2 112.8



160 44.6 31.5 47.3




Abundance Ion 164.00 (163.70 to 164.70): E

Ion 162.00 (161.70 to 162.70): E

Ion 160.00 (159.70 to 160.70): E









#39
 1,2,4,5-Tetrachlorobenzene
 Concen: 39.00 ng
 RT: 13.25 min Scan# 1042
 Delta R.T. -0.02 min
 Lab File: BF081602.D
 Acq: 14 Sep 2015 11:25

Instrument :
 BNA_F
 ClientSampleId :
 SSTDCCC040

Tgt Ion:	216	Resp:	154805
Ion	Ratio	Lower	Upper
216	100		
214	78.8	63.5	95.3
179	22.3	16.6	24.8
108	21.7	16.3	24.5




Abundance Ion 216.00 (215.70 to 216.70): E
 Ion 214.00 (213.70 to 214.70): E
 Ion 179.00 (178.70 to 179.70): E
 Ion 108.00 (107.70 to 108.70): E






#40
 Hexachlorocyclopentadiene
 Concen: 40.00 ng
 RT: 13.22 min Scan# 1040
 Delta R.T. -0.03 min
 Lab File: BF081602.D
 Acq: 14 Sep 2015 11:25

Tgt Ion:	237	Resp:	77913
Ion	Ratio	Lower	Upper
237	100		
235	63.3	42.0	82.0
272	11.9	0.0	36.1



Abundance Ion 236.80 (236.50 to 237.50): E
 Ion 234.90 (234.60 to 235.60): E
 Ion 271.75 (271.45 to 272.45): E









#41
 2,4,6-Tribromophenol
 Concen: 80.61 ng
 RT: 17.18 min Scan# 1386
 Delta R.T. -0.03 min
 Lab File: BF081602.D
 Acq: 14 Sep 2015 11:25




Instrument :
 BNA_F
 ClientSampleId :
 SSTDCCC040


Tgt Ion:330 Resp: 90070
 Ion Ratio Lower Upper
 330 100
 332 100.9 76.6 114.8
 141 51.1 29.7 44.5#



#42
 2,4,6-Trichlorophenol
 Concen: 40.34 ng
 RT: 13.59 min Scan# 1072
 Delta R.T. -0.03 min
 Lab File: BF081602.D
 Acq: 14 Sep 2015 11:25

Tgt Ion:196 Resp: 110490
 Ion Ratio Lower Upper
 196 100
 198 96.9 75.7 113.5
 200 30.2 23.9 35.9









#43
 2,4,5-Trichlorophenol
 Concen: 40.16 ng
 RT: 13.69 min Scan# 1081
 Delta R.T. -0.03 min
 Lab File: BF081602.D
 Acq: 14 Sep 2015 11:25

Instrument :
 BNA_F
 ClientSampleId :
 SSTDCCC040

Tgt Ion	Ratio	Lower	Upper
196	100		
198	95.7	75.4	113.0
97	63.6	33.3	49.9
132	30.5	24.6	37.0




Abundance Ion 196.00 (195.70 to 196.70): E
 Ion 198.00 (197.70 to 198.70): E
 Ion 97.00 (96.70 to 97.70): BF0
 Ion 132.00 (131.70 to 132.70): E






#44
 2-Fluorobiphenyl
 Concen: 79.03 ng
 RT: 13.78 min Scan# 1089
 Delta R.T. -0.03 min
 Lab File: BF081602.D
 Acq: 14 Sep 2015 11:25

Tgt Ion	Ratio	Lower	Upper
172	100		
171	36.0	26.1	39.1
170	23.8	17.4	26.2



Abundance Ion 172.00 (171.70 to 172.70): E
 Ion 171.00 (170.70 to 171.70): E
 Ion 170.00 (169.70 to 170.70): E









#45
 1,1'-Biphenyl
 Concen: 37.95 ng
 RT: 13.99 min Scan# 1107
 Delta R.T. -0.02 min
 Lab File: BF081602.D
 Acq: 14 Sep 2015 11:25




Instrument :
 BNA_F
 ClientSampleId :
 SSTDCCC040


Tgt Ion:154 Resp: 438211
 Ion Ratio Lower Upper
 154 100
 153 40.0 17.1 57.1
 76 10.4 0.0 31.6



#46
 2-Chloronaphthalene
 Concen: 38.91 ng
 RT: 13.97 min Scan# 1105
 Delta R.T. -0.03 min
 Lab File: BF081602.D
 Acq: 14 Sep 2015 11:25

Tgt Ion:162 Resp: 324390
 Ion Ratio Lower Upper
 162 100
 127 48.6 27.6 41.4#
 164 32.6 25.4 38.2





#47

2-Nitroaniline

Concen: 41.95 ng

RT: 14.32 min Scan# 1136

Delta R.T. -0.03 min

Lab File: BF081602.D

Acq: 14 Sep 2015 11:25

Instrument :

BNA_F

ClientSampleId :

SSTDCCC040





Tgt Ion: 65 Resp: 142558

Ion Ratio Lower Upper

65 100

92 60.0 55.4 83.0

138 85.8 115.5 173.3#



#48

Acenaphthylene

Concen: 38.11 ng

RT: 14.93 min Scan# 1189

Delta R.T. -0.03 min

Lab File: BF081602.D

Acq: 14 Sep 2015 11:25




Tgt Ion: 152 Resp: 563728


Ion Ratio Lower Upper

152 100

151 20.4 14.6 22.0

153 13.0 10.3 15.5





#49

Dimethylphthalate

Concen: 40.92 ng

RT: 14.87 min Scan# 1184

Delta R.T. -0.03 min

Lab File: BF081602.D

Acq: 14 Sep 2015 11:25

Instrument :

BNA_F

ClientSampleId :

SSTDCCC040


Tgt Ion:163 Resp: 399030

Ion Ratio Lower Upper

163 100

194 3.1 4.4 6.6#




164 9.5 8.3 12.5



Abundance Ion 163.00 (162.70 to 163.70): E

Ion 194.00 (193.70 to 194.70): E

Ion 164.00 (163.70 to 164.70): E



#50

2,6-Dinitrotoluene

Concen: 37.26 ng

RT: 14.96 min Scan# 1192

Delta R.T. -0.02 min

Lab File: BF081602.D

Acq: 14 Sep 2015 11:25


Tgt Ion:165 Resp: 82455

Ion Ratio Lower Upper

165 100

63 78.7 32.3 48.5#



89 71.6 28.6 43.0#




Abundance Ion 165.00 (164.70 to 165.70): E

Ion 63.00 (62.70 to 63.70): BF0

Ion 89.00 (88.70 to 89.70): BF0





#51

Acenaphthene

Concen: 37.66 ng

RT: 15.35 min Scan# 1226

Delta R.T. -0.03 min

Lab File: BF081602.D

Acq: 14 Sep 2015 11:25

Instrument :

BNA_F

ClientSampleId :

SSTDCCC040


Tgt Ion:154 Resp: 316869

Ion Ratio Lower Upper

154 100

153 114.3 82.1 123.1




152 55.1 37.9 56.9



Abundance Ion 154.00 (153.70 to 154.70): E

Ion 153.00 (152.70 to 153.70): E

Ion 152.00 (151.70 to 152.70): E



#52

3-Nitroaniline

Concen: 39.41 ng

RT: 15.33 min Scan# 1224

Delta R.T. -0.03 min

Lab File: BF081602.D

Acq: 14 Sep 2015 11:25


Tgt Ion:138 Resp: 99284

Ion Ratio Lower Upper

138 100

108 10.6 5.7 8.5#



92 124.1 67.7 101.5#




Abundance Ion 138.00 (137.70 to 138.70): E

Ion 108.00 (107.70 to 108.70): E

Ion 92.00 (91.70 to 92.70): BF0









#53
 2,4-Dinitrophenol
 Concen: 40.09 ng
 RT: 15.60 min Scan# 1248
 Delta R.T. -0.03 min
 Lab File: BF081602.D
 Acq: 14 Sep 2015 11:25




Instrument :
 BNA_F
 ClientSampleId :
 SSTDCCC040


Tgt Ion:184 Resp: 41146
 Ion Ratio Lower Upper
 184 100
 63 96.7 35.4 53.2#
 154 81.8 32.2 48.4#



#54
 Dibenzofuran
 Concen: 38.27 ng
 RT: 15.77 min Scan# 1263
 Delta R.T. -0.03 min
 Lab File: BF081602.D
 Acq: 14 Sep 2015 11:25

Tgt Ion:168 Resp: 470216
 Ion Ratio Lower Upper
 168 100
 139 42.0 24.2 36.2#
 169 12.7 10.6 15.8









#55
4-Nitrophenol
Concen: 39.33 ng
RT: 15.93 min Scan# 1277
Delta R.T. -0.05 min
Lab File: BF081602.D
Acq: 14 Sep 2015 11:25

Instrument :
BNA_F
ClientSampleId :
SSTDCCC040

Tgt Ion:139 Resp: 62246
Ion Ratio Lower Upper
139 100
109 87.9 12.7 52.7#
65 128.9 45.9 85.9#




Abundance Ion 139.00 (138.70 to 139.70): E
Ion 109.00 (108.70 to 109.70): E
Ion 65.00 (64.70 to 65.70): BF0






#56
2,4-Dinitrotoluene
Concen: 40.33 ng
RT: 15.91 min Scan# 1275
Delta R.T. -0.02 min
Lab File: BF081602.D
Acq: 14 Sep 2015 11:25

Tgt Ion:165 Resp: 113656
Ion Ratio Lower Upper
165 100
63 57.7 29.8 44.6#
89 84.8 44.5 66.7#
182 2.0 3.0 4.4#



Abundance Ion 165.00 (164.70 to 165.70): E
Ion 63.00 (62.70 to 63.70): BF0
Ion 89.00 (88.70 to 89.70): BF0
Ion 182.00 (181.70 to 182.70): E









#57
 Fluorene
 Concen: 41.52 ng
 RT: 16.58 min Scan# 1334
 Delta R.T. -0.03 min
 Lab File: BF081602.D
 Acq: 14 Sep 2015 11:25




Instrument :
 BNA_F
 ClientSampleId :
 SSTDCCC040


Tgt Ion:166 Resp: 387120
 Ion Ratio Lower Upper
 166 100
 165 94.4 72.6 109.0
 167 13.0 10.6 16.0



#58
 2,3,4,6-Tetrachlorophenol
 Concen: 39.49 ng
 RT: 16.04 min Scan# 1286
 Delta R.T. -0.14 min
 Lab File: BF081602.D
 Acq: 14 Sep 2015 11:25

Tgt Ion:232 Resp: 84257
 Ion Ratio Lower Upper
 232 100
 131 50.9 38.5 57.7
 130 2.6 1.8 2.6
 166 28.4 25.0 37.6









#59
Diethylphthalate
Concen: 42.09 ng
RT: 16.56 min Scan# 1332
Delta R.T. -0.03 min
Lab File: BF081602.D
Acq: 14 Sep 2015 11:25




Instrument :
BNA_F
ClientSampleId :
SSTDCCC040


Tgt Ion:149 Resp: 431721
Ion Ratio Lower Upper
149 100
177 19.0 17.5 26.3
150 13.0 9.4 14.2



#60
4-Chlorophenyl-phenylether
Concen: 43.52 ng
RT: 16.68 min Scan# 1342
Delta R.T. -0.03 min
Lab File: BF081602.D
Acq: 14 Sep 2015 11:25

Tgt Ion:204 Resp: 189150
Ion Ratio Lower Upper
204 100
206 31.4 24.8 37.2
141 70.0 42.2 63.4#





#61

4-Nitroaniline

Concen: 37.79 ng

RT: 16.80 min Scan# 1353

Delta R.T. -0.02 min

Lab File: BF081602.D

Acq: 14 Sep 2015 11:25

Instrument :

BNA_F

ClientSampleId :

SSTDCCC040


Tgt Ion:138 Resp: 96178

Ion Ratio Lower Upper

138 100

92 53.2 12.6 52.6#




108 88.3 12.4 52.4#



Abundance Ion 138.00 (137.70 to 138.70): E

Ion 92.00 (91.70 to 92.70): BF0

Ion 108.00 (107.70 to 108.70): E



#62

Azobenzene

Concen: 42.25 ng

RT: 17.04 min Scan# 1374

Delta R.T. -0.05 min

Lab File: BF081602.D

Acq: 14 Sep 2015 11:25

Tgt Ion: 77 Resp: 481776


Ion Ratio Lower Upper

77 100

182 17.9 15.1 55.1

105 19.6 3.4 43.4

51 39.2 12.0 52.0






Abundance Ion 77.00 (76.70 to 77.70): BF0

Ion 182.00 (181.70 to 182.70): E

Ion 105.00 (104.70 to 105.70): E

Ion 51.00 (50.70 to 51.70): BF0









#63
 Phenanthrene-d10
 Concen: 20.00 ng
 RT: 18.77 min Scan# 1525
 Delta R.T. -0.05 min
 Lab File: BF081602.D
 Acq: 14 Sep 2015 11:25




Instrument :
 BNA_F
 ClientSampleId :
 SSTDCCC040


Tgt Ion	Ratio	Lower	Upper
188	100		
94	8.6	11.1	16.7#
80	9.7	11.0	16.4#



#64
 4,6-Dinitro-2-methylphenol
 Concen: 42.73 ng
 RT: 16.87 min Scan# 1359
 Delta R.T. -0.02 min
 Lab File: BF081602.D
 Acq: 14 Sep 2015 11:25

Tgt Ion	Ratio	Lower	Upper
198	100		
51	82.6	39.6	79.6#
105	56.6	28.5	68.5





#65

n-Nitrosodiphenylamine

Concen: 37.42 ng

RT: 17.00 min Scan# 1370

Delta R.T. -0.03 min

Lab File: BF081602.D

Acq: 14 Sep 2015 11:25

Instrument :

BNA_F

ClientSampleId :

SSTDCCC040


Tgt Ion:169 Resp: 346942

Ion Ratio Lower Upper

169 100

168 68.7 48.9 73.3




167 37.7 26.2 39.4



Abundance Ion 169.00 (168.70 to 169.70): E

Ion 168.00 (167.70 to 168.70): E

Ion 167.00 (166.70 to 167.70): E



#66

4-Bromophenyl-phenylether

Concen: 40.76 ng

RT: 17.82 min Scan# 1442

Delta R.T. -0.03 min

Lab File: BF081602.D

Acq: 14 Sep 2015 11:25


Tgt Ion:248 Resp: 105016

Ion Ratio Lower Upper

248 100

250 99.1 78.5 117.7



141 90.2 59.0 88.6#




Abundance Ion 248.00 (247.70 to 248.70): E

Ion 250.00 (249.70 to 250.70): E

Ion 141.00 (140.70 to 141.70): E





#67

Hexachlorobenzene

Concen: 39.69 ng

RT: 17.86 min Scan# 1446

Delta R.T. -0.03 min

Lab File: BF081602.D

Acq: 14 Sep 2015 11:25

Instrument :

BNA_F

ClientSampleId :

SSTDCCC040


Tgt Ion:284 Resp: 106912

Ion Ratio Lower Upper

284 100

142 43.5 29.5 44.3




249 35.4 19.5 29.3#



Abundance Ion 283.80 (283.50 to 284.50): E

Ion 142.00 (141.70 to 142.70): E

Ion 249.00 (248.70 to 249.70): E



#68

Atrazine

Concen: 41.30 ng

RT: 18.40 min Scan# 1493

Delta R.T. -0.05 min

Lab File: BF081602.D

Acq: 14 Sep 2015 11:25


Tgt Ion:200 Resp: 110804

Ion Ratio Lower Upper

200 100

173 27.1 13.2 53.2



215 41.8 41.8 81.8




Abundance Ion 200.00 (199.70 to 200.70): E

Ion 173.00 (172.70 to 173.70): E

Ion 215.00 (214.70 to 215.70): E









#69
 Pentachlorophenol
 Concen: 44.22 ng
 RT: 18.41 min Scan# 1494
 Delta R.T. -0.03 min
 Lab File: BF081602.D
 Acq: 14 Sep 2015 11:25




Instrument :
 BNA_F
 ClientSampleId :
 SSTDCCC040


Tgt Ion: 266 Resp: 52233
 Ion Ratio Lower Upper
 266 100
 268 63.2 49.8 74.6
 264 62.2 50.2 75.2



#70
 Phenanthrene
 Concen: 39.01 ng
 RT: 18.84 min Scan# 1531
 Delta R.T. -0.03 min
 Lab File: BF081602.D
 Acq: 14 Sep 2015 11:25

Tgt Ion: 178 Resp: 552566
 Ion Ratio Lower Upper
 178 100
 176 19.2 13.8 20.8
 179 14.9 12.2 18.2









#71
 Anthracene
 Concen: 38.50 ng
 RT: 18.95 min Scan# 1541
 Delta R.T. -0.05 min
 Lab File: BF081602.D
 Acq: 14 Sep 2015 11:25

Instrument :
 BNA_F
 ClientSampleId :
 SSTDCCC040

Tgt Ion:178 Resp: 560341
 Ion Ratio Lower Upper
 178 100
 176 19.1 14.1 21.1
 179 16.0 12.2 18.2




Abundance Ion 178.00 (177.70 to 178.70): E
 Ion 176.00 (175.70 to 176.70): E
 Ion 179.00 (178.70 to 179.70): E






#72
 Carbazole
 Concen: 37.87 ng
 RT: 19.43 min Scan# 1583
 Delta R.T. -0.03 min
 Lab File: BF081602.D
 Acq: 14 Sep 2015 11:25

Tgt Ion:167 Resp: 517136
 Ion Ratio Lower Upper
 167 100
 166 21.2 15.7 23.5
 139 13.3 9.5 14.3



Abundance Ion 167.00 (166.70 to 167.70): E
 Ion 166.00 (165.70 to 166.70): E
 Ion 139.00 (138.70 to 139.70): E





#73

Di-n-butylphthalate

Concen: 44.11 ng

RT: 20.48 min Scan# 1675

Delta R.T. -0.03 min

Lab File: BF081602.D

Acq: 14 Sep 2015 11:25

Instrument :

BNA_F

ClientSampleId :

SSTDCCC040


Tgt Ion:149 Resp: 711409

Ion Ratio Lower Upper

149 100

150 8.8 7.4 11.2




104 4.9 2.4 3.6#



Abundance Ion 149.00 (148.70 to 149.70): E

Ion 150.00 (149.70 to 150.70): E

Ion 104.00 (103.70 to 104.70): E



#74

Fluoranthene

Concen: 39.37 ng

RT: 21.44 min Scan# 1759

Delta R.T. -0.03 min

Lab File: BF081602.D

Acq: 14 Sep 2015 11:25


Tgt Ion:202 Resp: 603803

Ion Ratio Lower Upper

202 100

101 8.3 0.0 38.3



203 16.8 0.0 37.3




Abundance Ion 202.00 (201.70 to 202.70): E

Ion 101.00 (100.70 to 101.70): E

Ion 203.00 (202.70 to 203.70): E









#75
Chrysene-d12
Concen: 20.00 ng
RT: 23.40 min Scan# 1930
Delta R.T. -0.02 min
Lab File: BF081602.D
Acq: 14 Sep 2015 11:25




Instrument :
BNA_F
ClientSampleId :
SSTDCCC040


Tgt Ion: 240 Resp: 208529
Ion Ratio Lower Upper
240 100
120 5.7 16.2 24.2#
236 25.6 18.4 27.6



#76
Benzidine
Concen: 33.06 ng
RT: 21.75 min Scan# 1786
Delta R.T. -0.03 min
Lab File: BF081602.D
Acq: 14 Sep 2015 11:25

Tgt Ion: 184 Resp: 295989
Ion Ratio Lower Upper
184 100
185 14.4 11.8 17.6
183 10.9 7.6 11.4









#77
Pyrene
 Concen: 38.72 ng
 RT: 21.78 min Scan# 1789
 Delta R.T. -0.03 min
 Lab File: BF081602.D
 Acq: 14 Sep 2015 11:25




Instrument :
 BNA_F
 ClientSampleId :
 SSTDCCC040


Tgt Ion	Ratio	Lower	Upper
202	100		
200	20.0	15.0	22.4
203	17.7	14.1	21.1



#78
Terphenyl-d14
 Concen: 80.12 ng
 RT: 22.12 min Scan# 1818
 Delta R.T. -0.03 min
 Lab File: BF081602.D
 Acq: 14 Sep 2015 11:25

Tgt Ion	Ratio	Lower	Upper
244	100		
212	8.0	4.3	6.5#
122	12.4	17.4	26.2#





#79

Butylbenzylphthalate

Concen: 46.02 ng

RT: 22.81 min Scan# 1879

Delta R.T. -0.02 min

Lab File: BF081602.D

Acq: 14 Sep 2015 11:25

Instrument :

BNA_F

ClientSampleId :

SSTDCCC040


Tgt Ion:149 Resp: 309153

Ion Ratio Lower Upper

149 100

91 71.6 48.6 72.8




206 17.8 14.9 22.3



Abundance Ion 149.00 (148.70 to 149.70): E

Ion 91.00 (90.70 to 91.70): BF0

Ion 206.00 (205.70 to 206.70): E



#80

Benzo(a)anthracene

Concen: 38.79 ng

RT: 23.37 min Scan# 1928

Delta R.T. -0.03 min

Lab File: BF081602.D

Acq: 14 Sep 2015 11:25


Tgt Ion:228 Resp: 515113

Ion Ratio Lower Upper

228 100

226 27.8 20.0 30.0



229 19.9 15.4 23.2




Abundance Ion 228.00 (227.70 to 228.70): E

Ion 226.00 (225.70 to 226.70): E

Ion 229.00 (228.70 to 229.70): E









#81
 3,3'-Dichlorobenzidine
 Concen: 39.28 ng
 RT: 23.38 min Scan# 1929
 Delta R.T. -0.02 min
 Lab File: BF081602.D
 Acq: 14 Sep 2015 11:25




Instrument :
 BNA_F
 ClientSampleId :
 SSTDCCC040


Tgt Ion	Ratio	Lower	Upper
252	100		
254	64.8	51.4	77.2
126	13.9	16.4	24.6#



#82
 Chrysene
 Concen: 37.52 ng
 RT: 23.42 min Scan# 1932
 Delta R.T. -0.02 min
 Lab File: BF081602.D
 Acq: 14 Sep 2015 11:25

Tgt Ion	Ratio	Lower	Upper
228	100		
226	30.4	21.0	31.4
229	20.4	15.6	23.4





#83

Bis(2-ethylhexyl)phthalate

Concen: 45.71 ng

RT: 23.50 min Scan# 1939

Delta R.T. -0.02 min

Lab File: BF081602.D

Acq: 14 Sep 2015 11:25

Instrument :

BNA_F

ClientSampleId :

SSTDCCC040


Tgt Ion:149 Resp: 405219

Ion Ratio Lower Upper

149 100

167 25.9 24.2 36.2




279 2.4 3.8 5.6#



Abundance Ion 149.00 (148.70 to 149.70): E

Ion 167.00 (166.70 to 167.70): E

Ion 279.00 (278.70 to 279.70): E



#84

Di-n-octyl phthalate

Concen: 45.88 ng

RT: 24.15 min Scan# 1996

Delta R.T. -0.03 min

Lab File: BF081602.D

Acq: 14 Sep 2015 11:25


Tgt Ion:149 Resp: 669804

Ion Ratio Lower Upper

149 100

167 1.4 1.0 1.6



43 13.9 10.2 15.2




Abundance Ion 149.00 (148.70 to 149.70): E

Ion 167.00 (166.70 to 167.70): E

Ion 43.00 (42.70 to 43.70): BF0









#85
 Indeno(1,2,3-cd)pyrene
 Concen: 43.30 ng
 RT: 25.88 min Scan# 2147
 Delta R.T. -0.03 min
 Lab File: BF081602.D
 Acq: 14 Sep 2015 11:25




Instrument :
 BNA_F
 ClientSampleId :
 SSTDCCC040


Tgt Ion	Ratio	Lower	Upper
276	100		
138	21.9	25.7	38.5#
277	24.5	15.6	23.4#



#86
 Perylene-d12
 Concen: 20.00 ng
 RT: 24.82 min Scan# 2055
 Delta R.T. -0.02 min
 Lab File: BF081602.D
 Acq: 14 Sep 2015 11:25

Tgt Ion	Ratio	Lower	Upper
264	100		
260	24.6	16.7	25.1
265	20.4	17.2	25.8









#87
Benzo(b)fluoranthene
Concen: 71.34 ng
RT: 24.48 min Scan# 2025
Delta R.T. -0.02 min
Lab File: BF081602.D
Acq: 14 Sep 2015 11:25




Instrument :
BNA_F
ClientSampleId :
SSTDCCC040


Tgt Ion: 252 Resp: 847581
Ion Ratio Lower Upper
252 100
253 22.5 17.5 26.3
125 8.8 17.1 25.7#



#88
Benzo(k)fluoranthene
Concen: 85.98 ng
RT: 24.48 min Scan# 2025
Delta R.T. -0.06 min
Lab File: BF081602.D
Acq: 14 Sep 2015 11:25

Tgt Ion: 252 Resp: 847581
Ion Ratio Lower Upper
252 100
253 22.5 17.0 25.4
125 8.8 16.0 24.0#





#89

Benzo(a)pyrene

Concen: 41.15 ng

RT: 24.78 min Scan# 2051

Delta R.T. -0.02 min

Lab File: BF081602.D

Acq: 14 Sep 2015 11:25

Instrument :

BNA_F

ClientSampleId :

SSTDCCC040


Tgt Ion:252 Resp: 414250

Ion Ratio Lower Upper

252 100

253 21.9 17.2 25.8




125 5.3 18.0 27.0#



Abundance Ion 252.00 (251.70 to 252.70): E

Ion 253.00 (252.70 to 253.70): E

Ion 125.00 (124.70 to 125.70): E



#90

Dibenzo(a,h)anthracene

Concen: 41.14 ng

RT: 25.89 min Scan# 2148

Delta R.T. -0.03 min

Lab File: BF081602.D

Acq: 14 Sep 2015 11:25


Tgt Ion:278 Resp: 320052

Ion Ratio Lower Upper

278 100

139 12.2 26.3 39.5#



279 24.0 18.2 27.4




Abundance Ion 278.00 (277.70 to 278.70): E

Ion 139.00 (138.70 to 139.70): E

Ion 279.00 (278.70 to 279.70): E





#91

Benzo(g,h,i)perylene

Concen: 39.51 ng

RT: 26.18 min Scan# 2174

Delta R.T. -0.03 min

Lab File: BF081602.D


Acq: 14 Sep 2015 11:25

Instrument :

BNA_F

ClientSampleId :

SSTDCCC040



Tgt Ion:	276	Resp:	293332
Ion Ratio	Lower	Upper	
276	100		
277	23.0	17.8	26.6
138	13.6	34.6	51.8#

