

Data Path : Z:\HPCHEM1\BNA_F\DATA\BF091415\
 Data File : BF081612.D
 Acq On : 14 Sep 2015 17:37
 Operator : UM/IZ
 Sample : G3597-06
 Misc :
 ALS Vial : 11 Sample Multiplier: 1

Instrument :
 BNA_F
 ClientSampleId :
 GP-6(0-5)

Quant Time: Sep 15 02:13:13 2015
 Quant Method : Z:\HPCHEM1\BNA_F\METHODS\8270-BF090115.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Mon Sep 14 13:51:56 2015
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Dichlorobenzene-d4	8.23	152	63051	20.00	ng	0.00
21) Naphthalene-d8	11.13	136	248104	20.00	ng	0.00
38) Acenaphthene-d10	15.27	164	111390	20.00	ng	0.00
63) Phenanthrene-d10	18.78	188	197424	20.00	ng	0.01
75) Chrysene-d12	23.40	240	122886	20.00	ng	0.00
86) Perylene-d12	24.84	264	97003	20.00	ng	0.01
System Monitoring Compounds						
5) 2-Fluorophenol	5.41	112	565391	139.13	ng	0.03
7) Phenol-d6	7.67	99	819669	152.38	ng	0.01
23) Nitrobenzene-d5	9.54	82	511940	99.55	ng	0.00
41) 2,4,6-Tribromophenol	17.18	330	146518	147.73	ng	0.00
44) 2-Fluorobiphenyl	13.78	172	869186	100.00	ng	0.00
78) Terphenyl-d14	22.13	244	512741	97.13	ng	0.01
Target Compounds						Qvalue
49) Dimethylphthalate	14.84	163	186066	21.50	ng	# 98
70) Phenanthrene	18.82	178	33549	3.06	ng	97
74) Fluoranthene	21.44	202	85776	7.22	ng	88
77) Pyrene	21.78	202	89481	9.83	ng	97
80) Benzo(a)anthracene	23.38	228	37047	4.73	ng	93
82) Chrysene	23.42	228	43352	6.18	ng	92
85) Indeno(1,2,3-cd)pyrene	25.89	276	17595	3.42	ng	# 82
87) Benzo(b)fluoranthene	24.49	252	54485m	7.87	ng	
88) Benzo(k)fluoranthene	24.52	252	15070m	2.62	ng	
89) Benzo(a)pyrene	24.78	252	34934m	5.95	ng	
91) Benzo(g,h,i)perylene	26.20	276	18146	4.19	ng	# 75

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\HPCHEM1\BNA_F\DATA\BF091415\
Data File : BF081612.D
Acq On : 14 Sep 2015 17:37
Operator : UM/IZ
Sample : G3597-06
Misc :
ALS Vial : 11 Sample Multiplier: 1

Instrument :
BNA_F
ClientSampleId :
GP-6(0-5)

Quant Time: Sep 15 02:13:13 2015
Quant Method : Z:\HPCHEM1\BNA_F\METHODS\8270-BF090115.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Mon Sep 14 13:51:56 2015
Response via : Initial Calibration

