


Data Path : Z:\HPCHEM1\BNA_F\Data\BF091415\
 Data File : BF081601.D
 Acq On : 14 Sep 2015 10:30
 Operator : UM/IZ
 Sample : DFTPP
 Misc :
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 BNA_F
 ClientSampleId :
 DFTPP

Integration File: rteint.p

Method : Z:\HPCHEM1\BNA_F\METHODS\8270-BF090115.M
 Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 Last Update : Mon Sep 14 13:51:56 2015



AutoFind: Scans 1614, 1615, 1616; Background Corrected with Scan 1609

Target Mass	Rel. to Mass	Lower Limit%	Upper Limit%	Rel. Abn%	Raw Abn	Result Pass/Fail
51	198	10	80	58.1	16830	PASS
68	69	0.00	2	0.0	0	PASS
69	198	0.00	100	45.7	13242	PASS
70	69	0.00	2	0.0	0	PASS
127	198	10	80	56.4	16328	PASS
197	198	0.00	2	0.4	116	PASS
198	198	100	100	100.0	28962	PASS
199	198	5	9	6.4	1867	PASS
275	198	10	60	22.6	6533	PASS
365	198	1	100	2.3	672	PASS
441	442	0.01	24	15.9	2753	PASS
442	198	50	100	59.8	17305	PASS
443	442	15	24	19.2	3331	PASS