

Data Path : Z:\SVOASRV\HPCHEM1\BNA F\DATA\BF091420\
 Data File : BF121566.D
 Acq On : 14 Sep 2020 18:05
 Operator : JU/CG
 Sample : L3869-05
 Misc :
 ALS Vial : 11 Sample Multiplier: 1

Instrument :
 BNA_F
 ClientSampleId :
 OR-03-091120

Quant Time: Sep 15 00:58:55 2020
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA F\METHODS\8270-BF091020.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Thu Sep 10 16:47:54 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Dichlorobenzene-d4	6.83	152	189386	20.00	ng	0.00
21) Naphthalene-d8	8.12	136	745095	20.00	ng	0.00
39) Acenaphthene-d10	9.87	164	387239	20.00	ng	0.00
64) Phenanthrene-d10	11.35	188	735397	20.00	ng	0.00
76) Chrysene-d12	13.99	240	590274	20.00	ng	0.00
86) Perylene-d12	15.44	264	456623	20.00	ng	0.00
System Monitoring Compounds						
5) 2-Fluorophenol	5.47	112	1120074	87.89	ng	0.03
7) Phenol-d6	6.48	99	1304667	78.04	ng	0.02
23) Nitrobenzene-d5	7.40	82	1075418	77.50	ng	0.00
42) 2,4,6-Tribromophenol	10.66	330	520335	108.11	ng	0.00
45) 2-Fluorobiphenyl	9.19	172	1824562	71.36	ng	0.00
79) Terphenyl-d14	12.95	244	2222214	76.27	ng	0.00
Target Compounds						
10) Phenol	6.49	94	40686	2.285	ng	Qvalue # 59
50) Dimethylphthalate	9.58	163	190973	6.804	ng	99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\SVOASRV\HPCHEM1\BNA F\DATA\BF091420\
Data File : BF121566.D
Acq On : 14 Sep 2020 18:05
Operator : JU/CG
Sample : L3869-05
Misc :
ALS Vial : 11 Sample Multiplier: 1

Instrument :
BNA_F
ClientSampleId :
OR-03-091120

Quant Time: Sep 15 00:58:55 2020
Quant Method : Z:\SVOASRV\HPCHEM1\BNA F\METHODS\8270-BF091020.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Thu Sep 10 16:47:54 2020
Response via : Initial Calibration

