

Data Path : Z:\svoasrv\HPCHEM1\BNA_F\Data\BF091421\
 Data File : BF125483.D
 Acq On : 14 Sep 2021 20:08
 Operator : JU/CG
 Sample : M3724-04
 Misc :
 ALS Vial : 9 Sample Multiplier: 1

Instrument :
 BNA_F
 ClientSampleId :
 31191

Quant Time: Sep 15 01:06:27 2021
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA_F\METHODS\8270-BF081821.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Tue Sep 14 17:17:24 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Dichlorobenzene-d4	6.875	152	148718	20.000	ng	0.00
21) Naphthalene-d8	8.157	136	577247	20.000	ng	# 0.00
39) Acenaphthene-d10	9.916	164	315494	20.000	ng	0.00
64) Phenanthrene-d10	11.404	188	559570	20.000	ng	0.00
76) Chrysene-d12	14.051	240	272647	20.000	ng	#-0.01
86) Perylene-d12	15.539	264	293135	20.000	ng	-0.01
System Monitoring Compounds						
5) 2-Fluorophenol	5.516	112	888926	90.472	ng	0.01
7) Phenol-d6	6.516	99	1011141	79.571	ng	0.00
23) Nitrobenzene-d5	7.440	82	850703	74.193	ng	-0.01
42) 2,4,6-Tribromophenol	10.710	330	377080	119.734	ng	0.00
45) 2-Fluorobiphenyl	9.234	172	1500833	66.301	ng	0.00
79) Terphenyl-d14	12.992	244	1528290	89.299	ng	0.00
Target Compounds						
32) Benzoic acid	7.940	122	100040	16.876	ng	Qvalue 86

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA_F\Data\BF091421\
Data File : BF125483.D
Acq On : 14 Sep 2021 20:08
Operator : JU/CG
Sample : M3724-04
Misc :
ALS Vial : 9 Sample Multiplier: 1

Instrument :
BNA_F
ClientSampleId :
31191

Quant Time: Sep 15 01:06:27 2021
Quant Method : Z:\SVOASRV\HPCHEM1\BNA_F\METHODS\8270-BF081821.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Tue Sep 14 17:17:24 2021
Response via : Initial Calibration

