

Data Path : Z:\HPCHEM1\BNA_F\Data\BF091516\
 Data File : BF090083.D
 Acq On : 15 Sep 2016 11:00
 Operator : UM/SJ
 Sample : PB93355BL
 Misc :
 ALS Vial : 3 Sample Multiplier: 1

Instrument :
 BNA_F
 ClientSampleId :
 PB93355BL

Quant Time: Sep 15 11:33:17 2016
 Quant Method : Z:\HPCHEM1\BNA_F\METHODS\8270-BF091516.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Thu Sep 15 11:18:43 2016
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Dichlorobenzene-d4	6.55	152	140022	20.00	ng	0.00
21) Naphthalene-d8	7.84	136	552681	20.00	ng	0.00
38) Acenaphthene-d10	9.58	164	300031	20.00	ng	-0.01
63) Phenanthrene-d10	11.05	188	494187	20.00	ng	0.00
75) Chrysene-d12	13.66	240	428134	20.00	ng	-0.01
86) Perylene-d12	14.98	264	299792	20.00	ng	-0.01
System Monitoring Compounds						
5) 2-Fluorophenol	5.12	112	1082535	124.34	ng	0.00
7) Phenol-d6	6.19	99	1427539	125.24	ng	-0.01
23) Nitrobenzene-d5	7.12	82	847537	86.22	ng	-0.01
41) 2,4,6-Tribromophenol	10.36	330	301566	100.06	ng	-0.01
44) 2-Fluorobiphenyl	8.91	172	1257774	65.84	ng	-0.01
78) Terphenyl-d14	12.63	244	1247988	69.22	ng	-0.01

Target Compounds	Qvalue
------------------	--------

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\HPCHEM1\BNA_F\Data\BF091516\
Data File : BF090083.D
Acq On : 15 Sep 2016 11:00
Operator : UM/SJ
Sample : PB93355BL
Misc :
ALS Vial : 3 Sample Multiplier: 1

Instrument :
BNA_F
ClientSampleId :
PB93355BL

Quant Time: Sep 15 11:33:17 2016
Quant Method : Z:\HPCHEM1\BNA_F\METHODS\8270-BF091516.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Thu Sep 15 11:18:43 2016
Response via : Initial Calibration

