

Data Path : Z:\HPCHEM1\BNA_F\DATA\BF091516\
 Data File : BF090083.D
 Acq On : 15 Sep 2016 11:00
 Operator : UM/SJ
 Sample : PB93355BL
 Misc :
 ALS Vial : 3 Sample Multiplier: 1

Instrument :
 BNA_F
 ClientSampleId :
 PB93355BL

Integration Parameters: rteint.p
 Integrator: RTE
 Smoothing : ON
 Sampling : 1
 Start Thrs: 0.2
 Stop Thrs : 0
 Filtering: 5
 Min Area: 3 % of largest Peak
 Max Peaks: 100
 Peak Location: TOP

If leading or trailing edge < 100 prefer < Baseline drop else tangent >
 Peak separation: 5

Method : Z:\HPCHEM1\BNA_F\METHODS\8270-BF091516.M
 Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION

Signal : TIC

peak #	R.T. min	first scan	max scan	last scan	PK TY	peak height	corr. area	corr. % max.	% of total
1	1.690	8	9	32	rVB	355844	939098	26.57%	3.025%
2	4.673	267	270	286	rBV	345986	645713	18.27%	2.080%
3	5.119	306	309	312	rBV	2169830	3087708	87.35%	9.946%
4	6.193	400	403	406	rBV	2155767	3156409	89.29%	10.168%
5	6.319	411	414	416	rBV	2421228	3272232	92.57%	10.541%
6	6.547	432	434	445	rBV	800726	793371	22.44%	2.556%
7	6.707	445	448	450	rBV	2328300	2685378	75.97%	8.650%
8	7.119	481	484	486	rBV	1897792	2064401	58.40%	6.650%
9	7.839	544	547	549	rBV	722612	985735	27.89%	3.175%
10	8.913	638	641	644	rBV	3203624	3471215	98.20%	11.182%
11	9.576	696	699	701	rBV	1338211	1208453	34.19%	3.893%
12	10.011	734	737	743	rVB	38842	62282	1.76%	0.201%
13	10.365	765	768	770	rBV2	1526369	2033625	57.53%	6.551%
14	11.051	825	828	830	rBV	817695	1144379	32.37%	3.686%
15	12.628	963	966	969	rBV	2643530	3534954	100.00%	11.387%
16	13.657	1054	1056	1066	rBV	1032128	1142707	32.33%	3.681%
17	14.983	1170	1172	1189	rBV	650054	816020	23.08%	2.629%


Sum of corrected areas: 31043680

Data Path : Z:\HPCHEM1\BNA_F\DATA\BF091516\
Data File : BF090083.D
Acq On : 15 Sep 2016 11:00
Operator : UM/SJ
Sample : PB93355BL
Misc :
ALS Vial : 3 Sample Multiplier: 1

Instrument :
BNA_F
ClientSampleId :
PB93355BL

Quant Method : Z:\HPCHEM1\BNA_F\METHODS\8270-BF091516.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION

TIC Library : C:\Database\NIST11.L
TIC Integration Parameters: LSCINT.P



Data Path : Z:\HPCHEM1\BNA_F\DATA\BF091516\
Data File : BF090083.D
Acq On : 15 Sep 2016 11:00
Operator : UM/SJ
Sample : PB93355BL
Misc :
ALS Vial : 3 Sample Multiplier: 1

Instrument :
BNA_F
ClientSampleId :
PB93355BL










Quant Method : Z:\HPCHEM1\BNA_F\METHODS\8270-BF091516.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION

TIC Library : C:\Database\NIST11.L
TIC Integration Parameters: LSCINT.P

Peak Number 2 2-Pentanone, 4-hydroxy-4-me... Concentration Rank 4

R.T.	EstConc	Area	Relative to ISTD	R.T.
4.67	16.28 ng	645713	1,4-Dichlorobenzene-d4	6.55

Hit#	of	Tentative ID	MW	MolForm	CAS#	Qual
1	5	2-Pentanone, 4-hydroxy-4-methyl-	116	C6H12O2	000123-42-2	50
2		Acetic acid, 1,1-dimethylethyl e...	116	C6H12O2	000540-88-5	38
3		Acetic acid, cyano-, 1,1-dimethy...	141	C7H11NO2	001116-98-9	17
4		Morpholine, 4-methyl-	101	C5H11NO	000109-02-4	9
5		Acetone	58	C3H6O	000067-64-1	9



Data Path : Z:\HPCHEM1\BNA_F\DATA\BF091516\
Data File : BF090083.D
Acq On : 15 Sep 2016 11:00
Operator : UM/SJ
Sample : PB93355BL
Misc :
ALS Vial : 3 Sample Multiplier: 1

Instrument :
BNA_F
ClientSampleId :
PB93355BL









Quant Method : Z:\HPCHEM1\BNA_F\METHODS\8270-BF091516.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION

TIC Library : C:\Database\NIST11.L
TIC Integration Parameters: LSCINT.P

Peak Number 3 unknown6.32 Concentration Rank 1

R.T.	EstConc	Area	Relative to ISTD	R.T.
6.32	82.49 ng	3272230	1,4-Dichlorobenzene-d4	6.55

Hit#	of	Tentative ID	MW	MolForm	CAS#	Qual
1	5	3-Chloro-6-fluoro-pyrazine	132	C4H2ClFN2	1000146-10-7	25
2		Tranylcypromine-propionyl	189	C12H15NO	1000123-86-3	16
3		(E)-3-Chloro-2-methyl-2-pentenal	132	C6H9ClO	031357-76-3	12
4		Amphetaminil	250	C17H18N2	017590-01-1	12
5		4-Chloro-6-fluoro-pyrimidine	132	C4H2ClFN2	051422-01-6	9



Data Path : Z:\HPCHEM1\BNA_F\DATA\BF091516\
Data File : BF090083.D
Acq On : 15 Sep 2016 11:00
Operator : UM/SJ
Sample : PB93355BL
Misc :
ALS Vial : 3 Sample Multiplier: 1

Instrument :
BNA_F
ClientSampleId :
PB93355BL

Quant Method : Z:\HPCHEM1\BNA_F\METHODS\8270-BF091516.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION

TIC Library : C:\Database\NIST11.L
TIC Integration Parameters: LSCINT.P

TIC Top Hit name	RT	EstConc	Units	Response	--Internal Standard--			
					#	RT	Resp	Conc
2-Pentanone, 4-hy...	4.67	16.3	ng	645713	1	6.55	793371	20.0
unknown6.32	6.32	82.5	ng	3272230	1	6.55	793371	20.0