

Data Path : Z:\HPCHEM1\BNA_F\DATA\BF091517\
 Data File : BF098570.D
 Acq On : 15 Sep 2017 15:33
 Operator : SJ/JU
 Sample : I5188-03
 Misc :
 ALS Vial : 14 Sample Multiplier: 1

Instrument :
 BNA_F
 ClientSampleId :
 RL-6-11-22

Manual Integrations
 APPROVED

Sohil
 9/18/2017 3:33:39 PM

Quant Time: Sep 15 16:30:38 2017
 Quant Method : Z:\HPCHEM1\BNA_F\METHODS\8270-BF091117.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Fri Sep 15 13:23:10 2017
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Dichlorobenzene-d4	6.65	152	139660	20.00	ng	0.00
21) Naphthalene-d8	7.93	136	559579	20.00	ng	0.00
38) Acenaphthene-d10	9.68	164	226199	20.00	ng	0.00
63) Phenanthrene-d10	11.16	188	347510	20.00	ng	0.00
75) Chrysene-d12	13.80	240	273848	20.00	ng	0.00
86) Perylene-d12	15.21	264	245757	20.00	ng	0.01
System Monitoring Compounds						
5) 2-Fluorophenol	5.26	112	1018696	107.85	ng	0.01
7) Phenol-d6	6.30	99	1300863	108.47	ng	0.00
23) Nitrobenzene-d5	7.22	82	759198	75.64	ng	0.00
41) 2,4,6-Tribromophenol	10.47	330	171808	113.80	ng	0.00
44) 2-Fluorobiphenyl	9.00	172	1207779	90.50	ng	0.00
78) Terphenyl-d14	12.75	244	717255	56.53	ng	0.00
Target Compounds						
10) Phenol	6.32	94	37764	2.711	ng	Qvalue # 47
20) 3+4-Methylphenols	7.09	107	39264	3.673	ng	# 57
31) Naphthalene	7.95	128	121138	4.379	ng	99
37) 2-Methylnaphthalene	8.64	142	42863	2.558	ng	99
49) Dimethylphthalate	9.40	163	164196	9.377	ng	99
70) Phenanthrene	11.19	178	81634	4.431	ng	99
74) Fluoranthene	12.37	202	108075	5.860	ng	98
77) Pyrene	12.60	202	99924	4.625	ng	97
80) Benzo(a)anthracene	13.79	228	46658	2.906	ng	# 85
82) Chrysene	13.83	228	46388	2.858	ng	# 92
87) Benzo(b)fluoranthene	14.82	252	64779m	4.192	ng	
89) Benzo(a)pyrene	15.15	252	40724	2.958	ng	# 73

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\HPCHEM1\BNA_F\DATA\BF091517\
Data File : BF098570.D
Acq On : 15 Sep 2017 15:33
Operator : SJ/JU
Sample : I5188-03
Misc :
ALS Vial : 14 Sample Multiplier: 1

Instrument :
BNA_F
ClientSampleId :
RL-6-11-22

Manual Integrations
APPROVED

Sohil
9/18/2017 3:33:39 PM

Quant Time: Sep 15 16:30:38 2017
Quant Method : Z:\HPCHEM1\BNA_F\METHODS\8270-BF091117.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Fri Sep 15 13:23:10 2017
Response via : Initial Calibration

