

Data Path : Z:\svoasrv\HPCHEM1\BNA_F\Data\BF091521\
 Data File : BF125507.D
 Acq On : 15 Sep 2021 16:40
 Operator : JU/CG
 Sample : SSTDCCC040
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_F
 LabSampleId :
 SSTDCCC040

Quant Time: Sep 15 18:27:59 2021
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA_F\METHODS\8270-BF091521.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Wed Sep 15 15:08:32 2021
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	Amount	Calc.	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	20.000	20.000	0.0	102	0.00
2	1,4-Dioxane	40.000	37.914	5.2	101	0.01
3	Pyridine	40.000	39.244	1.9	105	0.01
4	n-Nitrosodimethylamine	40.000	38.281	4.3	104	0.01
5 S	2-Fluorophenol	80.000	72.805	9.0	102	0.00
6	Aniline	40.000	37.763	5.6	104	0.00
7 S	Phenol-d6	80.000	73.528	8.1	103	0.00
8	2-Chlorophenol	40.000	37.072	7.3	104	0.00
9	Benzaldehyde	40.000	42.831	-7.1	103	0.00
10 C	Phenol	40.000	34.831	12.9	102	0.00
11	bis(2-Chloroethyl)ether	40.000	37.104	7.2	102	0.00
12	1,3-Dichlorobenzene	40.000	37.176	7.1	103	0.00
13 C	1,4-Dichlorobenzene	40.000	36.842	7.9	102	0.00
14	1,2-Dichlorobenzene	40.000	35.495	11.3	104	0.00
15	Benzyl Alcohol	40.000	37.900	5.3	103	0.00
16	2,2'-oxybis(1-Chloropropane	40.000	36.917	7.7	101	0.00
17	2-Methylphenol	40.000	36.880	7.8	103	0.00
18	Hexachloroethane	40.000	37.757	5.6	103	0.00
19 P	n-Nitroso-di-n-propylamine	40.000	36.372	9.1	103	0.00
20	3+4-Methylphenols	40.000	41.865	-4.7	102	0.00
21 I	Naphthalene-d8	20.000	20.000	0.0	103	0.00
22	Acetophenone	40.000	36.035	9.9	103	0.00
23 S	Nitrobenzene-d5	80.000	74.092	7.4	101	0.00
24	Nitrobenzene	40.000	37.185	7.0	102	0.00
25	Isophorone	40.000	37.018	7.5	103	0.00
26 C	2-Nitrophenol	40.000	38.698	3.3	103	0.00
27	2,4-Dimethylphenol	40.000	37.612	6.0	105	0.00
28	bis(2-Chloroethoxy)methane	40.000	37.086	7.3	102	0.00
29 C	2,4-Dichlorophenol	40.000	37.036	7.4	102	0.00
30	1,2,4-Trichlorobenzene	40.000	36.718	8.2	102	0.00
31	Naphthalene	40.000	36.449	8.9	102	0.00
32	Benzoic acid	40.000	40.108	-0.3	108	0.00
33	4-Chloroaniline	40.000	37.660	5.9	104	0.00
34 C	Hexachlorobutadiene	40.000	37.325	6.7	105	0.00
35	Caprolactam	40.000	37.389	6.5	104	0.00
36 C	4-Chloro-3-methylphenol	40.000	37.346	6.6	103	0.00
37	2-Methylnaphthalene	40.000	36.546	8.6	103	0.00
38	1-Methylnaphthalene	40.000	36.518	8.7	102	0.00
39 I	Acenaphthene-d10	20.000	20.000	0.0	104	0.00
40	1,2,4,5-Tetrachlorobenzene	40.000	36.832	7.9	104	0.00
41 P	Hexachlorocyclopentadiene	40.000	39.891	0.3	103	0.00
42 S	2,4,6-Tribromophenol	80.000	75.526	5.6	104	0.00
43 C	2,4,6-Trichlorophenol	40.000	37.962	5.1	104	0.00
44	2,4,5-Trichlorophenol	40.000	37.356	6.6	103	0.00
45 S	2-Fluorobiphenyl	80.000	70.411	12.0	102	0.00
46	1,1'-Biphenyl	40.000	36.287	9.3	104	0.00
47	2-Chloronaphthalene	40.000	36.419	9.0	102	0.00

Data Path : Z:\svoasrv\HPCHEM1\BNA_F\Data\BF091521\
 Data File : BF125507.D
 Acq On : 15 Sep 2021 16:40
 Operator : JU/CG
 Sample : SSTDCCC040
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_F
 LabSampleId :
 SSTDCCC040

Quant Time: Sep 15 18:27:59 2021
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA_F\METHODS\8270-BF091521.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Wed Sep 15 15:08:32 2021
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	Amount	Calc.	%Dev	Area%	Dev(min)
48	2-Nitroaniline	40.000	37.739	5.7	102	0.00
49	Acenaphthylene	40.000	36.199	9.5	103	0.00
50	Dimethylphthalate	40.000	36.804	8.0	104	0.00
51	2,6-Dinitrotoluene	40.000	37.673	5.8	103	0.00
52 C	Acenaphthene	40.000	36.189	9.5	103	0.00
53	3-Nitroaniline	40.000	37.882	5.3	104	0.00
54 P	2,4-Dinitrophenol	40.000	39.653	0.9	107	0.00
55	Dibenzofuran	40.000	35.794	10.5	103	0.00
56 P	4-Nitrophenol	40.000	39.968	0.1	105	0.00
57	2,4-Dinitrotoluene	40.000	38.257	4.4	104	0.00
58	Fluorene	40.000	36.013	10.0	104	0.00
59	2,3,4,6-Tetrachlorophenol	40.000	38.314	4.2	104	0.00
60	Diethylphthalate	40.000	36.496	8.8	104	0.00
61	4-Chlorophenyl-phenylether	40.000	36.452	8.9	105	0.00
62	4-Nitroaniline	40.000	37.400	6.5	104	0.00
63	Azobenzene	40.000	36.471	8.8	102	0.00
64 I	Phenanthrene-d10	20.000	20.000	0.0	103	0.00
65	4,6-Dinitro-2-methylphenol	40.000	41.419	-3.5	104	0.00
66 c	n-Nitrosodiphenylamine	40.000	36.821	7.9	104	0.00
67	4-Bromophenyl-phenylether	40.000	37.569	6.1	105	0.00
68	Hexachlorobenzene	40.000	37.362	6.6	103	0.00
69	Atrazine	40.000	38.711	3.2	104	0.00
70 C	Pentachlorophenol	40.000	43.587	-9.0	103	0.00
71	Phenanthrene	40.000	36.140	9.6	102	0.00
72	Anthracene	40.000	36.119	9.7	103	0.00
73	Carbazole	40.000	36.517	8.7	103	0.00
74	Di-n-butylphthalate	40.000	36.864	7.8	105	0.00
75 C	Fluoranthene	40.000	36.099	9.8	102	0.00
76 I	Chrysene-d12	20.000	20.000	0.0	101	0.00
77	Benzydine	40.000	39.491	1.3	103	0.00
78	Pyrene	40.000	37.532	6.2	103	0.00
79 S	Terphenyl-d14	80.000	73.552	8.1	103	0.00
80	Butylbenzylphthalate	40.000	38.110	4.7	103	0.00
81	Benzo(a)anthracene	40.000	37.667	5.8	103	0.00
82	3,3'-Dichlorobenzidine	40.000	36.205	9.5	97	0.00
83	Chrysene	40.000	37.079	7.3	102	0.00
84	Bis(2-ethylhexyl)phthalate	40.000	37.834	5.4	102	0.00
85 c	Di-n-octyl phthalate	40.000	37.535	6.2	101	0.00
86 I	Perylene-d12	20.000	20.000	0.0	99	0.00
87	Indeno(1,2,3-cd)pyrene	40.000	37.417	6.5	102	0.00
88	Benzo(b)fluoranthene	40.000	37.481	6.3	100	0.00
89	Benzo(k)fluoranthene	40.000	38.415	4.0	97	0.00
90 C	Benzo(a)pyrene	40.000	38.044	4.9	100	0.00
91	Dibenzo(a,h)anthracene	40.000	37.458	6.4	102	0.00
92	Benzo(g,h,i)perylene	40.000	37.840	5.4	102	0.00

Data Path : Z:\svoasrv\HPCHEM1\BNA_F\Data\BF091521\
Data File : BF125507.D
Acq On : 15 Sep 2021 16:40
Operator : JU/CG
Sample : SSTDCCC040
Misc :
ALS Vial : 2 Sample Multiplier: 1

Instrument :
BNA_F
LabSampleId :
SSTDCCC040

Quant Time: Sep 15 18:27:59 2021
Quant Method : Z:\SVOASRV\HPCHEM1\BNA_F\METHODS\8270-BF091521.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Wed Sep 15 15:08:32 2021
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

Compound	Amount	Calc.	%Dev	Area%	Dev(min)
----------	--------	-------	------	-------	----------

(#) = Out of Range SPCC's out = 0 CCC's out = 0