


Data Path : Z:\svoasrv\HPCHEM1\BNA_F\Data\BF091620\
 Data File : BF121588.D
 Acq On : 16 Sep 2020 10:06
 Operator : JU/CG
 Sample : DFTPP
 Misc :
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 BNA_F
 ClientSampleId :
 DFTPP

Integration File: rteint.p

Method : Z:\SVOASRV\HPCHEM1\BNA_F\METHODS\8270-BF091020.M
 Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 Last Update : Thu Sep 10 16:47:54 2020



AutoFind: Scans 1616, 1617, 1618; Background Corrected with Scan 1611

Target Mass	Rel. to Mass	Lower Limit%	Upper Limit%	Rel. Abn%	Raw Abn	Result Pass/Fail
51	198	10	80	42.0	46872	PASS
68	69	0.00	2	2.0	813	PASS
69	198	0.00	100	36.2	40338	PASS
70	69	0.00	2	0.0	0	PASS
127	198	10	80	47.4	52941	PASS
197	198	0.00	2	0.8	877	PASS
198	198	100	100	100.0	111576	PASS
199	198	5	9	6.5	7274	PASS
275	198	10	60	27.6	30768	PASS
365	198	1	100	3.5	3909	PASS
441	198	0.01	100	14.3	15969	PASS
442	198	50	100	92.5	103216	PASS
443	442	15	24	18.8	19417	PASS