

Data Path : Z:\svoasrv\HPCHEM1\BNA_F\Data\BF091624\
 Data File : BF139386.D
 Acq On : 16 Sep 2024 18:51
 Operator : RC/JU
 Sample : P3930-01
 Misc :
 ALS Vial : 12 Sample Multiplier: 1

Instrument :
 BNA_F
 ClientSampleId :
 DN-B-33

Quant Time: Sep 17 01:38:21 2024
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_F\Methods\8270-BF091324.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Sat Sep 14 02:58:23 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Dichlorobenzene-d4	6.875	152	58262	20.000	ng	0.00
21) Naphthalene-d8	8.157	136	211974	20.000	ng	0.00
39) Acenaphthene-d10	9.910	164	101610	20.000	ng	0.00
64) Phenanthrene-d10	11.392	188	154165	20.000	ng	0.00
76) Chrysene-d12	14.027	240	95887	20.000	ng	0.00
86) Perylene-d12	15.498	264	74008	20.000	ng	0.00
System Monitoring Compounds						
5) 2-Fluorophenol	5.493	112	280790	78.622	ng	0.00
7) Phenol-d6	6.504	99	364470	77.750	ng	0.00
23) Nitrobenzene-d5	7.440	82	248117	55.027	ng	0.00
42) 2,4,6-Tribromophenol	10.698	330	73644	68.078	ng	0.00
45) 2-Fluorobiphenyl	9.234	172	439447	60.657	ng	0.00
79) Terphenyl-d14	12.974	244	342766	58.675	ng	0.00

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA_F\Data\BF091624\
Data File : BF139386.D
Acq On : 16 Sep 2024 18:51
Operator : RC/JU
Sample : P3930-01
Misc :
ALS Vial : 12 Sample Multiplier: 1

Instrument :
BNA_F
ClientSampleId :
DN-B-33

Quant Time: Sep 17 01:38:21 2024
Quant Method : Z:\svoasrv\HPCHEM1\BNA_F\Methods\8270-BF091324.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Sat Sep 14 02:58:23 2024
Response via : Initial Calibration

