

Data Path : Z:\HPCHEM1\BNA_F\DATA\BF091715\
 Data File : BF081689.D
 Acq On : 17 Sep 2015 21:20
 Operator : UM/IZ
 Sample : G3664-01
 Misc :
 ALS Vial : 16 Sample Multiplier: 1

Instrument :
 BNA_F
 ClientSampleId :
 GP-1(0-5)

Quant Time: Sep 18 07:05:46 2015
 Quant Method : Z:\HPCHEM1\BNA_F\METHODS\8270-BF090115.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Tue Sep 15 14:03:59 2015
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Dichlorobenzene-d4	8.17	152	46318	20.00	ng	-0.04
21) Naphthalene-d8	11.07	136	178001	20.00	ng	-0.06
38) Acenaphthene-d10	15.19	164	81779	20.00	ng	-0.07
63) Phenanthrene-d10	18.70	188	162226	20.00	ng	-0.06
75) Chrysene-d12	23.35	240	94635	20.00	ng	-0.03
86) Perylene-d12	24.79	264	80758	20.00	ng	-0.02
System Monitoring Compounds						
5) 2-Fluorophenol	5.34	112	319116	106.89	ng	0.00
7) Phenol-d6	7.60	99	448446	113.48	ng	-0.04
23) Nitrobenzene-d5	9.48	82	299384	81.15	ng	-0.06
41) 2,4,6-Tribromophenol	17.10	330	86107	118.25	ng	-0.07
44) 2-Fluorobiphenyl	13.72	172	469801	73.62	ng	-0.06
78) Terphenyl-d14	22.07	244	343303	84.45	ng	-0.04
Target Compounds						
49) Dimethylphthalate	14.77	163	105110	16.54	ng	Qvalue
70) Phenanthrene	18.76	178	37131	4.12	ng	# 96
74) Fluoranthene	21.39	202	48291	4.95	ng	97
77) Pyrene	21.74	202	41504	5.92	ng	90
80) Benzo(a)anthracene	23.34	228	14185	2.35	ng	95
82) Chrysene	23.37	228	15760m	2.92	ng	
83) Bis(2-ethylhexyl)phthalate	23.47	149	11651	2.90	ng	# 90
85) Indeno(1,2,3-cd)pyrene	25.82	276	8499	2.14	ng	# 76
87) Benzo(b)fluoranthene	24.45	252	15742m	2.73	ng	
89) Benzo(a)pyrene	24.73	252	10885m	2.23	ng	
91) Benzo(g,h,i)perylene	26.13	276	7843	2.18	ng	# 69

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\HPCHEM1\BNA_F\DATA\BF091715\
Data File : BF081689.D
Acq On : 17 Sep 2015 21:20
Operator : UM/IZ
Sample : G3664-01
Misc :
ALS Vial : 16 Sample Multiplier: 1

Instrument :
BNA_F
ClientSampleId :
GP-1(0-5)

Quant Time: Sep 18 07:05:46 2015
Quant Method : Z:\HPCHEM1\BNA_F\METHODS\8270-BF090115.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Tue Sep 15 14:03:59 2015
Response via : Initial Calibration

