

Data Path : Z:\svoasrv\HPCHEM1\BNA_F\Data\BF091724\
 Data File : BF139394.D
 Acq On : 17 Sep 2024 11:08
 Operator : RC/JU
 Sample : PB163414BL
 Misc :
 ALS Vial : 3 Sample Multiplier: 1

Instrument :
 BNA_F
 ClientSampleId :
 PB163414BL

Quant Time: Sep 17 11:34:34 2024
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_F\Methods\8270-BF091324.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Sat Sep 14 02:58:23 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Dichlorobenzene-d4	6.875	152	72264	20.000	ng	0.00
21) Naphthalene-d8	8.157	136	271274	20.000	ng	0.00
39) Acenaphthene-d10	9.910	164	159150	20.000	ng	0.00
64) Phenanthrene-d10	11.398	188	316746	20.000	ng	0.00
76) Chrysene-d12	14.033	240	218080	20.000	ng	0.00
86) Perylene-d12	15.504	264	170230	20.000	ng	0.00
System Monitoring Compounds						
5) 2-Fluorophenol	5.510	112	528411	119.288	ng	0.02
7) Phenol-d6	6.510	99	691741	118.972	ng	0.00
23) Nitrobenzene-d5	7.439	82	473886	82.124	ng	0.00
42) 2,4,6-Tribromophenol	10.704	330	215601	127.247	ng	0.00
45) 2-Fluorobiphenyl	9.233	172	922004	81.252	ng	0.00
79) Terphenyl-d14	12.986	244	1108500	83.432	ng	0.00

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA_F\Data\BF091724\
Data File : BF139394.D
Acq On : 17 Sep 2024 11:08
Operator : RC/JU
Sample : PB163414BL
Misc :
ALS Vial : 3 Sample Multiplier: 1

Instrument :
BNA_F
ClientSampleId :
PB163414BL

Quant Time: Sep 17 11:34:34 2024
Quant Method : Z:\svoasrv\HPCHEM1\BNA_F\Methods\8270-BF091324.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Sat Sep 14 02:58:23 2024
Response via : Initial Calibration

