




Data Path : Z:\svoasrv\HPCHEM1\BNA_F\Data\BF091724\
Data File : BF139392.D
Acq On : 17 Sep 2024 09:30
Operator : RC/JU
Sample : DFTPP
Misc :
ALS Vial : 1 Sample Multiplier: 1

Quant Time: Sep 17 11:44:19 2024
Quant Method : Z:\svoasrv\HPCHEM1\BNA_F\Methods\8270-BF091324.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Sat Sep 14 02:58:23 2024
Response via : Initial Calibration



TIC: BF139392.D\data.ms

(70) Pentachlorophenol (C)




11.198min (+ 0.000) 37430.86 ng

response 98218

Ion	Exp%	Act%
265.70	100.00	100.00
268.00	63.90	62.99
264.00	64.20	61.60
0.00	0.00	0.00

Data Path : Z:\svoasrv\HPCHEM1\BNA_F\Data\BF091724\
Data File : BF139392.D
Acq On : 17 Sep 2024 09:30
Operator : RC/JU
Sample : DFTPP
Misc :
ALS Vial : 1 Sample Multiplier: 1

Quant Time: Sep 17 11:44:19 2024
Quant Method : Z:\svoasrv\HPCHEM1\BNA_F\Methods\8270-BF091324.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Sat Sep 14 02:58:23 2024
Response via : Initial Calibration



TIC: BF139392.D\data.ms

(77) Benzidine

12.733min (+ 0.006) 0.00 ng

response 432802

Ion	Exp%	Act%
184.00	100.00	100.00
185.00	14.40	13.84
183.00	11.10	11.61
0.00	0.00	0.00