

Data Path : Z:\HPCHEM1\BNA_F\DATA\BF091815\
 Data File : BF081694.D
 Acq On : 18 Sep 2015 12:56
 Operator : UM/IZ
 Sample : SSTDCCC040
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_F
 ClientSampleId :
 SSTDCCC040

Quant Time: Sep 18 22:56:47 2015
 Quant Method : Z:\HPCHEM1\BNA_F\METHODS\8270-BF090115.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Tue Sep 15 14:03:59 2015
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Dichlorobenzene-d4	8.17	152	63511	20.00	ng	-0.04
21) Naphthalene-d8	11.08	136	259156	20.00	ng	-0.04
38) Acenaphthene-d10	15.21	164	125610	20.00	ng	-0.04
63) Phenanthrene-d10	18.71	188	259375	20.00	ng	-0.04
75) Chrysene-d12	23.36	240	195673	20.00	ng	-0.02
86) Perylene-d12	24.79	264	137550	20.00	ng	-0.02

System Monitoring Compounds

5) 2-Fluorophenol	5.32	112	322267	78.73	ng	-0.03
7) Phenol-d6	7.61	99	445186	82.16	ng	-0.03
23) Nitrobenzene-d5	9.50	82	447528	83.32	ng	-0.03
41) 2,4,6-Tribromophenol	17.12	330	94687	84.66	ng	-0.04
44) 2-Fluorobiphenyl	13.73	172	805615	82.19	ng	-0.04
78) Terphenyl-d14	22.08	244	767565	91.31	ng	-0.03

Target Compounds

						Qvalue
2) 1,4-Dioxane	1.53	88	172193	93.66	ng	# 49
3) Pyridine	2.05	79	187582	37.00	ng	# 81
4) n-Nitrosodimethylamine	2.02	42	133633	47.45	ng	# 57
6) Aniline	7.49	93	302346	39.51	ng	# 83
8) 2-Chlorophenol	7.73	128	184216	40.84	ng	# 78
9) Benzaldehyde	7.20	77	117174	34.51	ng	# 73
10) Phenol	7.64	94	236403	37.62	ng	# 49
11) bis(2-Chloroethyl)ether	7.72	93	185033	40.81	ng	# 64
12) 1,3-Dichlorobenzene	8.02	146	190430	39.51	ng	# 87
13) 1,4-Dichlorobenzene	8.21	146	196799	40.10	ng	# 88
14) 1,2-Dichlorobenzene	8.54	146	184320	41.72	ng	# 91
15) Benzyl Alcohol	8.63	79	187594	42.20	ng	# 70
16) 2,2'-oxybis(1-Chloropropan	8.94	45	388557	44.71	ng	90
17) 2-Methylphenol	8.97	107	149713	41.60	ng	# 79
18) Hexachloroethane	9.27	117	81481	42.68	ng	# 87
19) n-Nitroso-di-n-propylamine	9.27	70	164453	44.59	ng	# 84
20) 3+4-Methylphenols	9.35	107	198176	40.84	ng	# 86
22) Acetophenone	9.18	105	278522	39.35	ng	# 72
24) Nitrobenzene	9.53	77	236649	42.42	ng	# 63
25) Isophorone	10.13	82	417016	41.74	ng	# 83
26) 2-Nitrophenol	10.25	139	95904	39.45	ng	# 21
27) 2,4-Dimethylphenol	10.53	122	164678	39.22	ng	# 72
28) bis(2-Chloroethoxy)methane	10.72	93	235448	40.80	ng	# 92
29) 2,4-Dichlorophenol	10.87	162	147881	40.61	ng	93
30) 1,2,4-Trichlorobenzene	10.98	180	164030	41.47	ng	98
31) Naphthalene	11.13	128	552019	39.94	ng	98
32) Benzoic acid	11.09	122	89339	30.76	ng	# 51
33) 4-Chloroaniline	11.37	127	232461	41.24	ng	# 85
34) Hexachlorobutadiene	11.48	225	105604	43.87	ng	99
35) Caprolactam	12.37	113	38047	34.07	ng	# 19
36) 4-Chloro-3-methylphenol	12.69	107	182780	44.84	ng	# 73
37) 2-Methylnaphthalene	12.78	142	370077	41.15	ng	# 89
39) 1,2,4,5-Tetrachlorobenzene	13.18	216	159249	40.09	ng	98
40) Hexachlorocyclopentadiene	13.16	237	64571	33.12	ng	97


Data Path : Z:\HPCHEM1\BNA_F\DATA\BF091815\
 Data File : BF081694.D
 Acq On : 18 Sep 2015 12:56
 Operator : UM/IZ
 Sample : SSTDCCC040
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_F
 ClientSampleId :
 SSTDCCC040

Quant Time: Sep 18 22:56:47 2015
 Quant Method : Z:\HPCHEM1\BNA_F\METHODS\8270-BF090115.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Tue Sep 15 14:03:59 2015
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
42) 2,4,6-Trichlorophenol	13.53	196	111078	40.52	ng	98
43) 2,4,5-Trichlorophenol	13.64	196	111336	39.81	ng #	87
45) 1,1'-Biphenyl	13.92	154	448044	38.77	ng	95
46) 2-Chloronaphthalene	13.91	162	333234	39.93	ng #	85
47) 2-Nitroaniline	14.27	65	146279	43.01	ng #	58
48) Acenaphthylene	14.86	152	586622	39.62	ng	97
49) Dimethylphthalate	14.81	163	407174	41.72	ng #	97
50) 2,6-Dinitrotoluene	14.91	165	85979	38.82	ng #	25
51) Acenaphthene	15.29	154	324676	38.55	ng	91
52) 3-Nitroaniline	15.27	138	97436	38.64	ng #	47
53) 2,4-Dinitrophenol	15.56	184	36786	36.53	ng #	21
54) Dibenzofuran	15.72	168	492262	40.03	ng #	86
55) 4-Nitrophenol	15.89	139	57358	36.21	ng #	4
56) 2,4-Dinitrotoluene	15.87	165	116255	41.22	ng #	66
57) Fluorene	16.53	166	398506	42.70	ng	96
58) 2,3,4,6-Tetrachlorophenol	16.08	232	86644	40.58	ng #	88
59) Diethylphthalate	16.51	149	447576	43.60	ng	96
60) 4-Chlorophenyl-phenylether	16.62	204	191019	43.91	ng #	87
61) 4-Nitroaniline	16.75	138	92776	36.42	ng #	18
62) Azobenzene	16.99	77	485139	42.51	ng	82
64) 4,6-Dinitro-2-methylphenol	16.83	198	56965	39.27	ng #	64
65) n-Nitrosodiphenylamine	16.94	169	353153	37.42	ng	92
66) 4-Bromophenyl-phenylether	17.76	248	108560	41.39	ng	97
67) Hexachlorobenzene	17.81	284	109627	39.98	ng #	81
68) Atrazine	18.36	200	104128	38.12	ng	81
69) Pentachlorophenol	18.36	266	41965	36.74	ng	96
70) Phenanthrene	18.77	178	575705	39.92	ng	97
71) Anthracene	18.89	178	570528	38.51	ng	98
72) Carbazole	19.37	167	533177	38.35	ng	96
73) Di-n-butylphthalate	20.41	149	735361	44.80	ng #	98
74) Fluoranthene	21.40	202	612831	39.26	ng	88
76) Benzidine	21.72	184	285645	34.01	ng	97
77) Pyrene	21.74	202	609262	42.03	ng	98
79) Butylbenzylphthalate	22.77	149	292218	46.36	ng #	73
80) Benzo(a)anthracene	23.34	228	496805	39.87	ng	98
81) 3,3'-Dichlorobenzidine	23.35	252	159291	37.90	ng #	95
82) Chrysene	23.39	228	432248	38.70	ng	95
83) Bis(2-ethylhexyl)phthalate	23.47	149	406885	48.91	ng #	91
84) Di-n-octyl phthalate	24.12	149	586938	42.85	ng	96
85) Indeno(1,2,3-cd)pyrene	25.83	276	328379	40.07	ng #	83
87) Benzo(b)fluoranthene	24.48	252	712901	72.60	ng #	81
88) Benzo(k)fluoranthene	24.48	252	712901	87.49	ng #	82
89) Benzo(a)pyrene	24.75	252	333631	40.09	ng #	83
90) Dibenzo(a,h)anthracene	25.84	278	279509	43.47	ng #	75
91) Benzo(g,h,i)perylene	26.13	276	255497	41.63	ng #	75

(#) = qualifier out of range (m) = manual integration (+) = signals summed



#1

1,4-Dichlorobenzene-d4

Concen: 20.00 ng

RT: 8.17 min Scan# 620

Delta R.T. -0.04 min

Lab File: BF081694.D

Acq: 18 Sep 2015 12:56

Instrument :

BNA_F

ClientSampleId :

SSTDCCC040


Tgt Ion:152 Resp: 63511

Ion Ratio Lower Upper

152 100

150 157.9 124.7 187.1




115 67.4 39.0 58.4#



Abundance Ion 152.00 (151.70 to 152.70): E

Ion 150.00 (149.70 to 150.70): E

Ion 115.00 (114.70 to 115.70): E



#2

1,4-Dioxane

Concen: 93.66 ng

RT: 1.53 min Scan# 39

Delta R.T. -0.03 min

Lab File: BF081694.D

Acq: 18 Sep 2015 12:56


Tgt Ion: 88 Resp: 172193

Ion Ratio Lower Upper

88 100

58 38.1 77.7 116.5#



43 18.5 26.6 40.0#




Abundance Ion 88.00 (87.70 to 88.70): BF0

Ion 58.00 (57.70 to 58.70): BF0

Ion 43.00 (42.70 to 43.70): BF0





#3

Pyridine

Concen: 37.00 ng

RT: 2.05 min Scan# 84

Delta R.T. -0.03 min

Lab File: BF081694.D

Acq: 18 Sep 2015 12:56

Instrument :

BNA_F

ClientSampleId :

SSTDCCC040





Tgt Ion: 79 Resp: 187582

Ion Ratio Lower Upper

79 100

52 111.2 76.8 115.2

51 57.5 32.2 48.4#



#4

n-Nitrosodimethylamine

Concen: 47.45 ng

RT: 2.02 min Scan# 82

Delta R.T. -0.03 min

Lab File: BF081694.D

Acq: 18 Sep 2015 12:56




Tgt Ion: 42 Resp: 133633


Ion Ratio Lower Upper

42 100

74 78.6 105.0 157.6#

44 6.7 6.6 10.0





#5

2-Fluorophenol

Concen: 78.73 ng

RT: 5.32 min Scan# 370

Delta R.T. -0.03 min

Lab File: BF081694.D

Acq: 18 Sep 2015 12:56

Instrument :

BNA_F

ClientSampleId :

SSTDCCC040


Tgt Ion: 112 Resp: 322267

Ion Ratio Lower Upper

112 100

64 74.7 39.7 59.5#




63 43.0 17.4 26.0#



Abundance Ion 112.00 (111.70 to 112.70): E

Ion 64.00 (63.70 to 64.70): BF0

Ion 63.00 (62.70 to 63.70): BF0



#6

Aniline

Concen: 39.51 ng

RT: 7.49 min Scan# 560

Delta R.T. -0.03 min

Lab File: BF081694.D

Acq: 18 Sep 2015 12:56


Tgt Ion: 93 Resp: 302346

Ion Ratio Lower Upper

93 100

66 45.7 27.2 40.8#



65 23.7 14.7 22.1#




Abundance Ion 93.00 (92.70 to 93.70): BF0

Ion 66.00 (65.70 to 66.70): BF0

Ion 65.00 (64.70 to 65.70): BF0





#7

Phenol-d6

Concen: 82.16 ng

RT: 7.61 min Scan# 571

Delta R.T. -0.03 min

Lab File: BF081694.D

Acq: 18 Sep 2015 12:56

Instrument :

BNA_F

ClientSampleId :

SSTDCCC040


Tgt Ion: 99 Resp: 445186

Ion Ratio Lower Upper

99 100

42 26.6 13.7 20.5#




71 41.5 14.2 21.2#



Abundance Ion 99.00 (98.70 to 99.70): BF0

Ion 42.00 (41.70 to 42.70): BF0

Ion 71.00 (70.70 to 71.70): BF0



#8

2-Chlorophenol

Concen: 40.84 ng

RT: 7.73 min Scan# 581

Delta R.T. -0.04 min

Lab File: BF081694.D

Acq: 18 Sep 2015 12:56


Tgt Ion: 128 Resp: 184216

Ion Ratio Lower Upper

128 100

130 31.5 12.2 52.2



64 64.8 20.5 60.5#



Abundance Ion 128.00 (127.70 to 128.70): E

Ion 130.00 (129.70 to 130.70): E

Ion 64.00 (63.70 to 64.70): BF0





#9

Benzaldehyde

Concen: 34.51 ng

RT: 7.20 min Scan# 535

Delta R.T. -0.03 min

Lab File: BF081694.D

Acq: 18 Sep 2015 12:56

Instrument :

BNA_F

ClientSampleId :

SSTDCCC040





Tgt Ion: 77 Resp: 117174

Ion Ratio Lower Upper

77 100

105 92.0 91.6 131.6

106 82.7 102.6 142.6#



#10

Phenol

Concen: 37.62 ng

RT: 7.64 min Scan# 573

Delta R.T. -0.03 min

Lab File: BF081694.D

Acq: 18 Sep 2015 12:56




Tgt Ion: 94 Resp: 236403

Ion Ratio Lower Upper

94 100

65 37.9 0.7 40.7

66 51.1 0.8 40.8#









#11
bis(2-Chloroethyl)ether
Concen: 40.81 ng
RT: 7.72 min Scan# 580
Delta R.T. -0.04 min
Lab File: BF081694.D
Acq: 18 Sep 2015 12:56




Instrument :
BNA_F
ClientSampleId :
SSTDCCC040


Tgt Ion: 93 Resp: 185033
Ion Ratio Lower Upper
93 100
63 130.0 65.6 105.6#
95 31.9 13.6 53.6



#12
1,3-Dichlorobenzene
Concen: 39.51 ng
RT: 8.02 min Scan# 607
Delta R.T. -0.04 min
Lab File: BF081694.D
Acq: 18 Sep 2015 12:56

Tgt Ion: 146 Resp: 190430
Ion Ratio Lower Upper
146 100
148 65.4 51.0 76.4
75 41.3 15.0 22.6#





#13

1,4-Dichlorobenzene

Concen: 40.10 ng

RT: 8.21 min Scan# 623

Delta R.T. -0.04 min

Lab File: BF081694.D

Acq: 18 Sep 2015 12:56

Instrument :

BNA_F

ClientSampleId :

SSTDCCC040


Tgt Ion:146 Resp: 196799

Ion Ratio Lower Upper

146 100

148 63.7 51.8 77.6




111 49.1 24.9 37.3#



Abundance Ion 146.00 (145.70 to 146.70): E

Ion 148.00 (147.70 to 148.70): E

Ion 111.00 (110.70 to 111.70): E



#14

1,2-Dichlorobenzene

Concen: 41.72 ng

RT: 8.54 min Scan# 652

Delta R.T. -0.03 min

Lab File: BF081694.D

Acq: 18 Sep 2015 12:56


Tgt Ion:146 Resp: 184320

Ion Ratio Lower Upper

146 100

148 64.2 52.7 79.1



111 49.2 28.8 43.2#



Abundance Ion 146.00 (145.70 to 146.70): E

Ion 148.00 (147.70 to 148.70): E

Ion 111.00 (110.70 to 111.70): E









#15
Benzyl Alcohol
Concen: 42.20 ng
RT: 8.63 min Scan# 660
Delta R.T. -0.03 min
Lab File: BF081694.D
Acq: 18 Sep 2015 12:56




Instrument :
BNA_F
ClientSampleId :
SSTDCCC040


Tgt Ion: 79 Resp: 187594
Ion Ratio Lower Upper
79 100
108 65.9 88.6 132.8#
77 62.9 47.0 70.4



#16
2,2'-oxybis(1-Chloropropane)
Concen: 44.71 ng
RT: 8.94 min Scan# 687
Delta R.T. -0.04 min
Lab File: BF081694.D
Acq: 18 Sep 2015 12:56

Tgt Ion: 45 Resp: 388557
Ion Ratio Lower Upper
45 100
77 11.8 0.0 28.1
79 9.1 0.0 26.0









#17
2-Methylphenol
Concen: 41.60 ng
RT: 8.97 min Scan# 690
Delta R.T. -0.03 min
Lab File: BF081694.D
Acq: 18 Sep 2015 12:56




Instrument :
BNA_F
ClientSampleId :
SSTDCCC040


Tgt Ion:107 Resp: 149713
Ion Ratio Lower Upper
107 100
108 112.5 96.6 145.0
77 55.7 23.3 34.9#
79 52.1 22.0 33.0#



#18
Hexachloroethane
Concen: 42.68 ng
RT: 9.27 min Scan# 716
Delta R.T. -0.04 min
Lab File: BF081694.D
Acq: 18 Sep 2015 12:56

Tgt Ion:117 Resp: 81481
Ion Ratio Lower Upper
117 100
119 96.0 83.8 125.6
201 84.8 55.1 82.7#









#19
n-Nitroso-di-n-propylamine
Concen: 44.59 ng
RT: 9.27 min Scan# 716
Delta R.T. -0.02 min
Lab File: BF081694.D
Acq: 18 Sep 2015 12:56




Instrument :
BNA_F
ClientSampleId :
SSTDCCC040


Tgt Ion: 70 Resp: 164453
Ion Ratio Lower Upper
70 100
42 68.9 63.8 95.6
101 8.8 9.2 13.8#
130 18.9 27.3 40.9#



#20
3+4-Methylphenols
Concen: 40.84 ng
RT: 9.35 min Scan# 723
Delta R.T. -0.03 min
Lab File: BF081694.D
Acq: 18 Sep 2015 12:56

Tgt Ion: 107 Resp: 198176
Ion Ratio Lower Upper
107 100
108 91.6 74.6 114.6
77 41.0 0.7 40.7#
79 35.8 0.0 38.9





#21

Naphthalene-d8

Concen: 20.00 ng

RT: 11.08 min Scan# 874

Delta R.T. -0.04 min

Lab File: BF081694.D

Acq: 18 Sep 2015 12:56

Instrument :

BNA_F

ClientSampleId :

SSTDCCC040

Tgt Ion:136 Resp: 259156


Ion Ratio Lower Upper

136 100

137 11.2 9.0 13.6

54 12.7 8.2 12.2#

68 5.0 5.8 8.6#






Abundance Ion 136.00 (135.70 to 136.70): E

Ion 137.00 (136.70 to 137.70): E

Ion 54.00 (53.70 to 54.70): BF0

Ion 68.00 (67.70 to 68.70): BF0



#22

Acetophenone

Concen: 39.35 ng

RT: 9.18 min Scan# 708

Delta R.T. -0.03 min

Lab File: BF081694.D

Acq: 18 Sep 2015 12:56

Tgt Ion:105 Resp: 278522


Ion Ratio Lower Upper

105 100

71 0.0 4.7 7.1#

51 40.8 22.0 33.0#

120 18.6 30.1 45.1#





Abundance Ion 105.00 (104.70 to 105.70): E

Ion 71.00 (70.70 to 71.70): BF0

Ion 51.00 (50.70 to 51.70): BF0

Ion 120.00 (119.70 to 120.70): E









#23
 Nitrobenzene-d5
 Concen: 83.32 ng
 RT: 9.50 min Scan# 736
 Delta R.T. -0.03 min
 Lab File: BF081694.D
 Acq: 18 Sep 2015 12:56




Instrument :
 BNA_F
 ClientSampleId :
 SSTDCCC040

Tgt Ion: 82 Resp: 447528
 Ion Ratio Lower Upper
 82 100
 128 38.4 51.0 76.6#
 54 67.9 47.5 71.3



#24
 Nitrobenzene
 Concen: 42.42 ng
 RT: 9.53 min Scan# 739
 Delta R.T. -0.03 min
 Lab File: BF081694.D
 Acq: 18 Sep 2015 12:56

Tgt Ion: 77 Resp: 236649
 Ion Ratio Lower Upper
 77 100
 123 38.7 59.8 89.8#
 65 15.8 10.2 15.4#





#25

Isophorone

Concen: 41.74 ng

RT: 10.13 min Scan# 791

Delta R.T. -0.03 min

Lab File: BF081694.D

Acq: 18 Sep 2015 12:56

Instrument :

BNA_F

ClientSampleId :

SSTDCCC040





Tgt Ion: 82 Resp: 417016

Ion Ratio Lower Upper

82 100

95 7.7 4.0 6.0#

138 13.8 18.3 27.5#



#26

2-Nitrophenol

Concen: 39.45 ng

RT: 10.25 min Scan# 802

Delta R.T. -0.04 min

Lab File: BF081694.D

Acq: 18 Sep 2015 12:56




Tgt Ion: 139 Resp: 95904

Ion Ratio Lower Upper

139 100

109 39.4 11.0 16.4#

65 78.1 24.7 37.1#









#27
 2,4-Dimethylphenol
 Concen: 39.22 ng
 RT: 10.53 min Scan# 826
 Delta R.T. -0.04 min
 Lab File: BF081694.D
 Acq: 18 Sep 2015 12:56




Instrument :
 BNA_F
 ClientSampleId :
 SSTDCCC040


Tgt Ion: 122 Resp: 164678
 Ion Ratio Lower Upper
 122 100
 107 127.9 71.8 107.6#
 121 57.2 42.3 63.5



#28
 bis(2-Chloroethoxy)methane
 Concen: 40.80 ng
 RT: 10.72 min Scan# 843
 Delta R.T. -0.03 min
 Lab File: BF081694.D
 Acq: 18 Sep 2015 12:56

Tgt Ion: 93 Resp: 235448
 Ion Ratio Lower Upper
 93 100
 95 32.2 27.5 41.3
 123 9.3 13.7 20.5#









#29
 2,4-Dichlorophenol
 Concen: 40.61 ng
 RT: 10.87 min Scan# 856
 Delta R.T. -0.03 min
 Lab File: BF081694.D
 Acq: 18 Sep 2015 12:56




Instrument :
 BNA_F
 ClientSampleId :
 SSTDCCC040


Tgt Ion	Resp	Lower	Upper
162	147881		
162	100		
164	62.5	44.9	84.9
98	40.9	13.0	53.0



#30
 1,2,4-Trichlorobenzene
 Concen: 41.47 ng
 RT: 10.98 min Scan# 866
 Delta R.T. -0.04 min
 Lab File: BF081694.D
 Acq: 18 Sep 2015 12:56

Tgt Ion	Resp	Lower	Upper
180	164030		
180	100		
182	95.9	77.1	115.7
145	33.6	23.3	34.9









#31
Naphthalene
Concen: 39.94 ng
RT: 11.13 min Scan# 879
Delta R.T. -0.03 min
Lab File: BF081694.D
Acq: 18 Sep 2015 12:56

Instrument :
BNA_F
ClientSampleId :
SSTDCCC040

Tgt Ion:128 Resp: 552019
Ion Ratio Lower Upper
128 100
129 10.7 8.4 12.6
127 13.4 9.9 14.9




Abundance Ion 128.00 (127.70 to 128.70): E
Ion 129.00 (128.70 to 129.70): E
Ion 127.00 (126.70 to 127.70): E






#32
Benzoic acid
Concen: 30.76 ng
RT: 11.09 min Scan# 875
Delta R.T. 0.00 min
Lab File: BF081694.D
Acq: 18 Sep 2015 12:56

Tgt Ion:122 Resp: 89339
Ion Ratio Lower Upper
122 100
105 133.0 76.1 116.1#
77 111.9 40.5 80.5#



Abundance Ion 122.00 (121.70 to 122.70): E
Ion 105.00 (104.70 to 105.70): E
Ion 77.00 (76.70 to 77.70): BF0






#33
4-Chloroaniline
Concen: 41.24 ng
RT: 11.37 min Scan# 900
Delta R.T. -0.03 min
Lab File: BF081694.D
Acq: 18 Sep 2015 12:56

Instrument :
BNA_F
ClientSampleId :
SSTDCCC040

Tgt Ion	Ratio	Lower	Upper
127	100		
129	32.4	25.8	38.6
65	38.5	17.9	26.9
92	20.9	10.5	15.7






Abundance

Ion 127.00 (126.70 to 127.70): E

Ion 129.00 (128.70 to 129.70): E


Ion 65.00 (64.70 to 65.70): BF0

Ion 92.00 (91.70 to 92.70): BF0



#34
Hexachlorobutadiene
Concen: 43.87 ng
RT: 11.48 min Scan# 909
Delta R.T. -0.04 min
Lab File: BF081694.D
Acq: 18 Sep 2015 12:56

Tgt Ion	Ratio	Lower	Upper
225	100		
223	62.8	50.2	75.2
227	64.3	52.2	78.2





Abundance

Ion 225.00 (224.70 to 225.70): E

Ion 223.00 (222.70 to 223.70): E

Ion 227.00 (226.70 to 227.70): E









#35
 Caprolactam
 Concen: 34.07 ng
 RT: 12.37 min Scan# 987
 Delta R.T. -0.02 min
 Lab File: BF081694.D
 Acq: 18 Sep 2015 12:56




Instrument :
 BNA_F
 ClientSampleId :
 SSTDCCC040


Tgt Ion: 113 Resp: 38047
 Ion Ratio Lower Upper
 113 100
 55 286.9 152.4 192.4#
 56 213.7 104.6 144.6#



#36
 4-Chloro-3-methylphenol
 Concen: 44.84 ng
 RT: 12.69 min Scan# 1015
 Delta R.T. -0.03 min
 Lab File: BF081694.D
 Acq: 18 Sep 2015 12:56

Tgt Ion: 107 Resp: 182780
 Ion Ratio Lower Upper
 107 100
 144 21.7 25.4 38.0#
 142 67.8 77.3 115.9#









#37
 2-Methylnaphthalene
 Concen: 41.15 ng
 RT: 12.78 min Scan# 1023
 Delta R.T. -0.04 min
 Lab File: BF081694.D
 Acq: 18 Sep 2015 12:56




Instrument :
 BNA_F
 ClientSampleId :
 SSTDCCC040

Tgt Ion:142 Resp: 370077
 Ion Ratio Lower Upper
 142 100
 141 87.8 67.5 101.3
 115 40.6 18.2 27.2#



#38
 Acenaphthene-d10
 Concen: 20.00 ng
 RT: 15.21 min Scan# 1236
 Delta R.T. -0.04 min
 Lab File: BF081694.D
 Acq: 18 Sep 2015 12:56

Tgt Ion:164 Resp: 125610
 Ion Ratio Lower Upper
 164 100
 162 106.6 75.2 112.8
 160 48.8 31.5 47.3#









#39
 1,2,4,5-Tetrachlorobenzene
 Concen: 40.09 ng
 RT: 13.18 min Scan# 1058
 Delta R.T. -0.04 min
 Lab File: BF081694.D
 Acq: 18 Sep 2015 12:56

Instrument :
 BNA_F
 ClientSampleId :
 SSTDCCC040

Tgt Ion:	216	Resp:	159249
Ion	Ratio	Lower	Upper
216	100		
214	78.7	63.5	95.3
179	23.0	16.6	24.8
108	21.9	16.3	24.5




Abundance Ion 216.00 (215.70 to 216.70): E
 Ion 214.00 (213.70 to 214.70): E
 Ion 179.00 (178.70 to 179.70): E
 Ion 108.00 (107.70 to 108.70): E






#40
 Hexachlorocyclopentadiene
 Concen: 33.12 ng
 RT: 13.16 min Scan# 1056
 Delta R.T. -0.06 min
 Lab File: BF081694.D
 Acq: 18 Sep 2015 12:56

Tgt Ion:	237	Resp:	64571
Ion	Ratio	Lower	Upper
237	100		
235	61.5	42.0	82.0
272	11.6	0.0	36.1



Abundance Ion 236.80 (236.50 to 237.50): E
 Ion 234.90 (234.60 to 235.60): E
 Ion 271.75 (271.45 to 272.45): E









#41
 2,4,6-Tribromophenol
 Concen: 84.66 ng
 RT: 17.12 min Scan# 1403
 Delta R.T. -0.04 min
 Lab File: BF081694.D
 Acq: 18 Sep 2015 12:56

Instrument :
 BNA_F
 ClientSampleId :
 SSTDCCC040

Tgt Ion: 330 Resp: 94687
 Ion Ratio Lower Upper
 330 100
 332 96.7 76.6 114.8
 141 49.3 29.7 44.5#




Abundance Ion 329.80 (329.50 to 330.50): E
 Ion 331.80 (331.50 to 332.50): E
 Ion 141.00 (140.70 to 141.70): E





#42
 2,4,6-Trichlorophenol
 Concen: 40.52 ng
 RT: 13.53 min Scan# 1089
 Delta R.T. -0.04 min
 Lab File: BF081694.D
 Acq: 18 Sep 2015 12:56

Tgt Ion: 196 Resp: 111078
 Ion Ratio Lower Upper
 196 100
 198 92.0 75.7 113.5
 200 30.0 23.9 35.9



Abundance Ion 196.00 (195.70 to 196.70): E
 Ion 198.00 (197.70 to 198.70): E
 Ion 200.00 (199.70 to 200.70): E









#43
 2,4,5-Trichlorophenol
 Concen: 39.81 ng
 RT: 13.64 min Scan# 1098
 Delta R.T. -0.04 min
 Lab File: BF081694.D
 Acq: 18 Sep 2015 12:56

Instrument :
 BNA_F
 ClientSampleId :
 SSTDCCC040

Tgt Ion:196 Resp: 111336
 Ion Ratio Lower Upper
 196 100
 198 96.4 75.4 113.0
 97 68.9 33.3 49.9#
 132 33.8 24.6 37.0




Abundance Ion 196.00 (195.70 to 196.70): E
 Ion 198.00 (197.70 to 198.70): E
 Ion 97.00 (96.70 to 97.70): BF0
 Ion 132.00 (131.70 to 132.70): E






#44
 2-Fluorobiphenyl
 Concen: 82.19 ng
 RT: 13.73 min Scan# 1106
 Delta R.T. -0.04 min
 Lab File: BF081694.D
 Acq: 18 Sep 2015 12:56

Tgt Ion:172 Resp: 805615
 Ion Ratio Lower Upper
 172 100
 171 35.2 26.1 39.1
 170 23.2 17.4 26.2



Abundance Ion 172.00 (171.70 to 172.70): E
 Ion 171.00 (170.70 to 171.70): E
 Ion 170.00 (169.70 to 170.70): E









#45
 1,1'-Biphenyl
 Concen: 38.77 ng
 RT: 13.92 min Scan# 1123
 Delta R.T. -0.04 min
 Lab File: BF081694.D
 Acq: 18 Sep 2015 12:56

Instrument :
 BNA_F
 ClientSampleId :
 SSTDCCC040

Tgt Ion:154 Resp: 448044
 Ion Ratio Lower Upper
 154 100
 153 40.0 17.1 57.1
 76 12.8 0.0 31.6




Abundance Ion 154.00 (153.70 to 154.70): E
 Ion 153.00 (152.70 to 153.70): E
 Ion 76.00 (75.70 to 76.70): BF0





#46
 2-Chloronaphthalene
 Concen: 39.93 ng
 RT: 13.91 min Scan# 1122
 Delta R.T. -0.04 min
 Lab File: BF081694.D
 Acq: 18 Sep 2015 12:56

Tgt Ion:162 Resp: 333234
 Ion Ratio Lower Upper
 162 100
 127 48.9 27.6 41.4#
 164 33.7 25.4 38.2



Abundance Ion 162.00 (161.70 to 162.70): E
 Ion 127.00 (126.70 to 127.70): E
 Ion 164.00 (163.70 to 164.70): E









#47
 2-Nitroaniline
 Concen: 43.01 ng
 RT: 14.27 min Scan# 1153
 Delta R.T. -0.04 min
 Lab File: BF081694.D
 Acq: 18 Sep 2015 12:56




Instrument :
 BNA_F
 ClientSampleId :
 SSTCCCC040

Tgt Ion: 65 Resp: 146279
 Ion Ratio Lower Upper
 65 100
 92 56.3 55.4 83.0
 138 75.8 115.5 173.3#



#48
 Acenaphthylene
 Concen: 39.62 ng
 RT: 14.86 min Scan# 1205
 Delta R.T. -0.06 min
 Lab File: BF081694.D
 Acq: 18 Sep 2015 12:56

Tgt Ion: 152 Resp: 586622
 Ion Ratio Lower Upper
 152 100
 151 20.1 14.6 22.0
 153 13.2 10.3 15.5









#49
Dimethylphthalate
Concen: 41.72 ng
RT: 14.81 min Scan# 1201
Delta R.T. -0.04 min
Lab File: BF081694.D
Acq: 18 Sep 2015 12:56




Instrument :
BNA_F
ClientSampleId :
SSTDCCC040

Tgt Ion:163 Resp: 407174
Ion Ratio Lower Upper
163 100
194 2.9 4.4 6.6#
164 10.1 8.3 12.5



#50
2,6-Dinitrotoluene
Concen: 38.82 ng
RT: 14.91 min Scan# 1209
Delta R.T. -0.04 min
Lab File: BF081694.D
Acq: 18 Sep 2015 12:56

Tgt Ion:165 Resp: 85979
Ion Ratio Lower Upper
165 100
63 91.3 32.3 48.5#
89 75.8 28.6 43.0#





#51

Acenaphthene

Concen: 38.55 ng

RT: 15.29 min Scan# 1243

Delta R.T. -0.04 min

Lab File: BF081694.D

Acq: 18 Sep 2015 12:56

Instrument :

BNA_F

ClientSampleId :

SSTDCCC040


Tgt Ion:154 Resp: 324676

Ion Ratio Lower Upper

154 100

153 112.3 82.1 123.1




152 53.7 37.9 56.9



Abundance Ion 154.00 (153.70 to 154.70): E

Ion 153.00 (152.70 to 153.70): E

Ion 152.00 (151.70 to 152.70): E



#52

3-Nitroaniline

Concen: 38.64 ng

RT: 15.27 min Scan# 1241

Delta R.T. -0.04 min

Lab File: BF081694.D

Acq: 18 Sep 2015 12:56


Tgt Ion:138 Resp: 97436

Ion Ratio Lower Upper

138 100

108 11.4 5.7 8.5#



92 135.5 67.7 101.5#



Abundance Ion 138.00 (137.70 to 138.70): E

Ion 108.00 (107.70 to 108.70): E

Ion 92.00 (91.70 to 92.70): BF0









#53
2,4-Dinitrophenol
Concen: 36.53 ng
RT: 15.56 min Scan# 1266
Delta R.T. -0.03 min
Lab File: BF081694.D
Acq: 18 Sep 2015 12:56




Instrument :
BNA_F
ClientSampleId :
SSTDCCC040


Tgt Ion:184 Resp: 36786
Ion Ratio Lower Upper
184 100
63 103.1 35.4 53.2#
154 81.8 32.2 48.4#



#54
Dibenzofuran
Concen: 40.03 ng
RT: 15.72 min Scan# 1280
Delta R.T. -0.04 min
Lab File: BF081694.D
Acq: 18 Sep 2015 12:56

Tgt Ion:168 Resp: 492262
Ion Ratio Lower Upper
168 100
139 40.7 24.2 36.2#
169 12.9 10.6 15.8





#55

4-Nitrophenol

Concen: 36.21 ng

RT: 15.89 min Scan# 1295

Delta R.T. -0.03 min

Lab File: BF081694.D

Acq: 18 Sep 2015 12:56

Instrument :

BNA_F

ClientSampleId :

SSTDCCC040


Tgt Ion:139 Resp: 57358

Ion Ratio Lower Upper

139 100

109 94.0 12.7 52.7#




65 136.6 45.9 85.9#



Abundance Ion 139.00 (138.70 to 139.70): E

Ion 109.00 (108.70 to 109.70): E

Ion 65.00 (64.70 to 65.70): BF0



#56

2,4-Dinitrotoluene

Concen: 41.22 ng

RT: 15.87 min Scan# 1293

Delta R.T. -0.03 min

Lab File: BF081694.D

Acq: 18 Sep 2015 12:56

Tgt Ion:165 Resp: 116255


Ion Ratio Lower Upper

165 100

63 56.4 29.8 44.6#

89 82.2 44.5 66.7#

182 1.9 3.0 4.4#





Abundance Ion 165.00 (164.70 to 165.70): E

Ion 63.00 (62.70 to 63.70): BF0

Ion 89.00 (88.70 to 89.70): BF0

Ion 182.00 (181.70 to 182.70): E









#57
 Fluorene
 Concen: 42.70 ng
 RT: 16.53 min Scan# 1351
 Delta R.T. -0.04 min
 Lab File: BF081694.D
 Acq: 18 Sep 2015 12:56




Instrument :
 BNA_F
 ClientSampleId :
 SSTCCCC040


Tgt Ion:166 Resp: 398506
 Ion Ratio Lower Upper
 166 100
 165 95.3 72.6 109.0
 167 13.2 10.6 16.0



#58
 2,3,4,6-Tetrachlorophenol
 Concen: 40.58 ng
 RT: 16.08 min Scan# 1312
 Delta R.T. -0.04 min
 Lab File: BF081694.D
 Acq: 18 Sep 2015 12:56

Tgt Ion:232 Resp: 86644
 Ion Ratio Lower Upper
 232 100
 131 59.0 38.5 57.7#
 130 2.3 1.8 2.6
 166 34.2 25.0 37.6





#59

Diethylphthalate

Concen: 43.60 ng

RT: 16.51 min Scan# 1349

Delta R.T. -0.04 min

Lab File: BF081694.D

Acq: 18 Sep 2015 12:56

Instrument :

BNA_F

ClientSampleId :

SSTDCCC040


Tgt Ion:149 Resp: 447576

Ion Ratio Lower Upper

149 100

177 19.5 17.5 26.3




150 12.3 9.4 14.2



Abundance Ion 149.00 (148.70 to 149.70): E

Ion 177.00 (176.70 to 177.70): E

Ion 150.00 (149.70 to 150.70): E



#60

4-Chlorophenyl-phenylether

Concen: 43.91 ng

RT: 16.62 min Scan# 1359

Delta R.T. -0.04 min

Lab File: BF081694.D

Acq: 18 Sep 2015 12:56


Tgt Ion:204 Resp: 191019

Ion Ratio Lower Upper

204 100

206 32.8 24.8 37.2



141 65.9 42.2 63.4#




Abundance Ion 204.00 (203.70 to 204.70): E

Ion 206.00 (205.70 to 206.70): E

Ion 141.00 (140.70 to 141.70): E









#61
 4-Nitroaniline
 Concen: 36.42 ng
 RT: 16.75 min Scan# 1370
 Delta R.T. -0.03 min
 Lab File: BF081694.D
 Acq: 18 Sep 2015 12:56




Instrument :
 BNA_F
 ClientSampleId :
 SSTDCCC040


Tgt Ion: 138 Resp: 92776
 Ion Ratio Lower Upper
 138 100
 92 60.1 12.6 52.6#
 108 96.9 12.4 52.4#



#62
 Azobenzene
 Concen: 42.51 ng
 RT: 16.99 min Scan# 1391
 Delta R.T. -0.04 min
 Lab File: BF081694.D
 Acq: 18 Sep 2015 12:56

Tgt Ion: 77 Resp: 485139
 Ion Ratio Lower Upper
 77 100
 182 18.5 15.1 55.1
 105 20.5 3.4 43.4
 51 39.8 12.0 52.0









#63
Phenanthrene-d10
Concen: 20.00 ng
RT: 18.71 min Scan# 1542
Delta R.T. -0.04 min
Lab File: BF081694.D
Acq: 18 Sep 2015 12:56




Instrument :
BNA_F
ClientSampleId :
SSTDCCC040


Tgt Ion	Ratio	Lower	Upper
188	100		
94	7.9	11.1	16.7#
80	8.6	11.0	16.4#



#64
4,6-Dinitro-2-methylphenol
Concen: 39.27 ng
RT: 16.83 min Scan# 1377
Delta R.T. -0.04 min
Lab File: BF081694.D
Acq: 18 Sep 2015 12:56

Tgt Ion	Ratio	Lower	Upper
198	100		
51	96.0	39.6	79.6#
105	62.3	28.5	68.5





#65

n-Nitrosodiphenylamine

Concen: 37.42 ng

RT: 16.94 min Scan# 1387

Delta R.T. -0.04 min

Lab File: BF081694.D

Acq: 18 Sep 2015 12:56

Instrument :

BNA_F

ClientSampleId :

SSTDCCC040


Tgt Ion:169 Resp: 353153

Ion Ratio Lower Upper




169 100

168 67.3 48.9 73.3

167 37.6 26.2 39.4



Abundance Ion 169.00 (168.70 to 169.70): E
Ion 168.00 (167.70 to 168.70): E
Ion 167.00 (166.70 to 167.70): E



#66

4-Bromophenyl-phenylether

Concen: 41.39 ng

RT: 17.76 min Scan# 1459

Delta R.T. -0.04 min

Lab File: BF081694.D

Acq: 18 Sep 2015 12:56


Tgt Ion:248 Resp: 108560

Ion Ratio Lower Upper



248 100

250 97.1 78.5 117.7

141 78.9 59.0 88.6



Abundance Ion 248.00 (247.70 to 248.70): E
Ion 250.00 (249.70 to 250.70): E
Ion 141.00 (140.70 to 141.70): E









#67
Hexachlorobenzene
Concen: 39.98 ng
RT: 17.81 min Scan# 1463
Delta R.T. -0.06 min
Lab File: BF081694.D
Acq: 18 Sep 2015 12:56

Instrument :
BNA_F
ClientSampleId :
SSTDCCC040

Tgt Ion: 284 Resp: 109627
Ion Ratio Lower Upper
284 100
142 46.0 29.5 44.3#
249 36.5 19.5 29.3#




Abundance Ion 283.80 (283.50 to 284.50): E
Ion 142.00 (141.70 to 142.70): E
Ion 249.00 (248.70 to 249.70): E





#68
Atrazine
Concen: 38.12 ng
RT: 18.36 min Scan# 1511
Delta R.T. -0.04 min
Lab File: BF081694.D
Acq: 18 Sep 2015 12:56

Tgt Ion: 200 Resp: 104128
Ion Ratio Lower Upper
200 100
173 25.5 13.2 53.2
215 44.5 41.8 81.8



Abundance Ion 200.00 (199.70 to 200.70): E
Ion 173.00 (172.70 to 173.70): E
Ion 215.00 (214.70 to 215.70): E









#69
 Pentachlorophenol
 Concen: 36.74 ng
 RT: 18.36 min Scan# 1511
 Delta R.T. -0.04 min
 Lab File: BF081694.D
 Acq: 18 Sep 2015 12:56

Instrument :
 BNA_F
 ClientSampleId :
 SSTDCCC040

Tgt Ion:266 Resp: 41965
 Ion Ratio Lower Upper
 266 100
 268 65.6 49.8 74.6
 264 64.9 50.2 75.2




Abundance Ion 265.70 (265.40 to 266.40): E
 Ion 268.00 (267.70 to 268.70): E
 Ion 264.00 (263.70 to 264.70): E





#70
 Phenanthrene
 Concen: 39.92 ng
 RT: 18.77 min Scan# 1547
 Delta R.T. -0.06 min
 Lab File: BF081694.D
 Acq: 18 Sep 2015 12:56

Tgt Ion:178 Resp: 575705
 Ion Ratio Lower Upper
 178 100
 176 19.4 13.8 20.8
 179 15.3 12.2 18.2



Abundance Ion 178.00 (177.70 to 178.70): E
 Ion 176.00 (175.70 to 176.70): E
 Ion 179.00 (178.70 to 179.70): E









#71
 Anthracene
 Concen: 38.51 ng
 RT: 18.89 min Scan# 1558
 Delta R.T. -0.06 min
 Lab File: BF081694.D
 Acq: 18 Sep 2015 12:56




Instrument :
 BNA_F
 ClientSampleId :
 SSTDCCC040

Tgt Ion:178 Resp: 570528
 Ion Ratio Lower Upper
 178 100
 176 18.7 14.1 21.1
 179 15.0 12.2 18.2



#72
 Carbazole
 Concen: 38.35 ng
 RT: 19.37 min Scan# 1600
 Delta R.T. -0.04 min
 Lab File: BF081694.D
 Acq: 18 Sep 2015 12:56

Tgt Ion:167 Resp: 533177
 Ion Ratio Lower Upper
 167 100
 166 21.5 15.7 23.5
 139 13.6 9.5 14.3





#73

Di-n-butylphthalate

Concen: 44.80 ng

RT: 20.41 min Scan# 1691

Delta R.T. -0.06 min

Lab File: BF081694.D

Acq: 18 Sep 2015 12:56

Instrument :

BNA_F

ClientSampleId :

SSTDCCC040


Tgt Ion:149 Resp: 735361

Ion Ratio Lower Upper

149 100

150 9.1 7.4 11.2




104 5.4 2.4 3.6#



Abundance Ion 149.00 (148.70 to 149.70): E

Ion 150.00 (149.70 to 150.70): E

Ion 104.00 (103.70 to 104.70): E



#74

Fluoranthene

Concen: 39.26 ng

RT: 21.40 min Scan# 1777

Delta R.T. -0.03 min

Lab File: BF081694.D

Acq: 18 Sep 2015 12:56


Tgt Ion:202 Resp: 612831

Ion Ratio Lower Upper

202 100

101 8.4 0.0 38.3



203 17.9 0.0 37.3




Abundance Ion 202.00 (201.70 to 202.70): E

Ion 101.00 (100.70 to 101.70): E

Ion 203.00 (202.70 to 203.70): E





#75

Chrysene-d12

Concen: 20.00 ng

RT: 23.36 min Scan# 1949

Delta R.T. -0.02 min

Lab File: BF081694.D

Acq: 18 Sep 2015 12:56

Instrument :

BNA_F

ClientSampleId :

SSTDCCC040





Tgt Ion: 240 Resp: 195673

Ion Ratio Lower Upper

240 100

120 4.6 16.2 24.2#

236 25.8 18.4 27.6



#76

Benzidine

Concen: 34.01 ng

RT: 21.72 min Scan# 1805

Delta R.T. -0.02 min

Lab File: BF081694.D

Acq: 18 Sep 2015 12:56




Tgt Ion: 184 Resp: 285645


Ion Ratio Lower Upper

184 100

185 13.5 11.8 17.6

183 11.0 7.6 11.4





#77

Pyrene

Concen: 42.03 ng

RT: 21.74 min Scan# 1807

Delta R.T. -0.04 min

Lab File: BF081694.D

Acq: 18 Sep 2015 12:56

Instrument :

BNA_F

ClientSampleId :

SSTDCCC040


Tgt Ion: 202 Resp: 609262

Ion Ratio Lower Upper

202 100

200 20.2 15.0 22.4




203 17.4 14.1 21.1



Abundance Ion 202.00 (201.70 to 202.70): E

Ion 200.00 (199.70 to 200.70): E

Ion 203.00 (202.70 to 203.70): E



#78

Terphenyl-d14

Concen: 91.31 ng

RT: 22.08 min Scan# 1837

Delta R.T. -0.03 min

Lab File: BF081694.D

Acq: 18 Sep 2015 12:56


Tgt Ion: 244 Resp: 767565

Ion Ratio Lower Upper

244 100

212 7.7 4.3 6.5#



122 10.5 17.4 26.2#




Abundance Ion 244.00 (243.70 to 244.70): E

Ion 212.00 (211.70 to 212.70): E

Ion 122.00 (121.70 to 122.70): E





#79

Butylbenzylphthalate

Concen: 46.36 ng

RT: 22.77 min Scan# 1897

Delta R.T. -0.03 min

Lab File: BF081694.D

Acq: 18 Sep 2015 12:56

Instrument :

BNA_F

ClientSampleId :

SSTDCCC040


Tgt Ion:149 Resp: 292218

Ion Ratio Lower Upper

149 100

91 85.4 48.6 72.8#




206 15.2 14.9 22.3



Abundance Ion 149.00 (148.70 to 149.70): E

Ion 91.00 (90.70 to 91.70): BF0

Ion 206.00 (205.70 to 206.70): E



#80

Benzo(a)anthracene

Concen: 39.87 ng

RT: 23.34 min Scan# 1947

Delta R.T. -0.03 min

Lab File: BF081694.D

Acq: 18 Sep 2015 12:56


Tgt Ion:228 Resp: 496805

Ion Ratio Lower Upper

228 100

226 26.5 20.0 30.0



229 19.7 15.4 23.2




Abundance Ion 228.00 (227.70 to 228.70): E

Ion 226.00 (225.70 to 226.70): E

Ion 229.00 (228.70 to 229.70): E









#81
 3,3'-Dichlorobenzidine
 Concen: 37.90 ng
 RT: 23.35 min Scan# 1948
 Delta R.T. -0.03 min
 Lab File: BF081694.D
 Acq: 18 Sep 2015 12:56




Instrument :
 BNA_F
 ClientSampleId :
 SSTDCCC040


Tgt Ion	Ratio	Lower	Upper
252	100		
254	63.4	51.4	77.2
126	13.0	16.4	24.6#



#82
 Chrysene
 Concen: 38.70 ng
 RT: 23.39 min Scan# 1951
 Delta R.T. -0.03 min
 Lab File: BF081694.D
 Acq: 18 Sep 2015 12:56

Tgt Ion	Ratio	Lower	Upper
228	100		
226	30.1	21.0	31.4
229	19.7	15.6	23.4









#83
 Bis(2-ethylhexyl)phthalate
 Concen: 48.91 ng
 RT: 23.47 min Scan# 1958
 Delta R.T. -0.03 min
 Lab File: BF081694.D
 Acq: 18 Sep 2015 12:56




Instrument :
 BNA_F
 ClientSampleId :
 SSTDCCC040


Tgt Ion:149 Resp: 406885
 Ion Ratio Lower Upper
 149 100
 167 25.0 24.2 36.2
 279 2.5 3.8 5.6#



#84
 Di-n-octyl phthalate
 Concen: 42.85 ng
 RT: 24.12 min Scan# 2015
 Delta R.T. -0.03 min
 Lab File: BF081694.D
 Acq: 18 Sep 2015 12:56

Tgt Ion:149 Resp: 586938
 Ion Ratio Lower Upper
 149 100
 167 1.5 1.0 1.6
 43 14.2 10.2 15.2









#85
 Indeno(1,2,3-cd)pyrene
 Concen: 40.07 ng
 RT: 25.83 min Scan# 2165
 Delta R.T. -0.04 min
 Lab File: BF081694.D
 Acq: 18 Sep 2015 12:56




Instrument :
 BNA_F
 ClientSampleId :
 SSTDCCC040


Tgt Ion	Ratio	Lower	Upper
276	100		
138	20.6	25.7	38.5#
277	24.8	15.6	23.4#



#86
 Perylene-d12
 Concen: 20.00 ng
 RT: 24.79 min Scan# 2074
 Delta R.T. -0.02 min
 Lab File: BF081694.D
 Acq: 18 Sep 2015 12:56

Tgt Ion	Ratio	Lower	Upper
264	100		
260	24.0	16.7	25.1
265	21.3	17.2	25.8





#87

Benzo(b)fluoranthene

Concen: 72.60 ng

RT: 24.48 min Scan# 2047

Delta R.T. 0.00 min





Lab File: BF081694.D

Acq: 18 Sep 2015 12:56

Instrument :
BNA_F
ClientSampleId :
SSTDCCC040

Tgt Ion: 252 Resp: 712901

Ion	Ratio	Lower	Upper
252	100		
253	22.2	17.5	26.3
125	3.8	17.1	25.7#



#88

Benzo(k)fluoranthene

Concen: 87.49 ng

RT: 24.48 min Scan# 2047




Delta R.T. -0.02 min


Lab File: BF081694.D

Acq: 18 Sep 2015 12:56

Tgt Ion: 252 Resp: 712901

Ion	Ratio	Lower	Upper
252	100		
253	22.2	17.0	25.4
125	3.8	16.0	24.0#





#89

Benzo(a)pyrene

Concen: 40.09 ng

RT: 24.75 min Scan# 2070

Delta R.T. -0.02 min

Lab File: BF081694.D

Acq: 18 Sep 2015 12:56

Instrument :

BNA_F

ClientSampleId :

SSTDCCC040


Tgt Ion:252 Resp: 333631

Ion Ratio Lower Upper

252 100

253 22.0 17.2 25.8




125 7.3 18.0 27.0#



Abundance Ion 252.00 (251.70 to 252.70): E

Ion 253.00 (252.70 to 253.70): E

Ion 125.00 (124.70 to 125.70): E



#90

Dibenzo(a,h)anthracene

Concen: 43.47 ng

RT: 25.84 min Scan# 2166

Delta R.T. -0.04 min

Lab File: BF081694.D

Acq: 18 Sep 2015 12:56


Tgt Ion:278 Resp: 279509

Ion Ratio Lower Upper

278 100

139 10.0 26.3 39.5#



279 24.2 18.2 27.4




Abundance Ion 278.00 (277.70 to 278.70): E

Ion 139.00 (138.70 to 139.70): E

Ion 279.00 (278.70 to 279.70): E





#91

Benzo(g,h,i)perylene

Concen: 41.63 ng

RT: 26.13 min Scan# 2191

Delta R.T. -0.04 min

Lab File: BF081694.D


Acq: 18 Sep 2015 12:56

Instrument :

BNA_F

ClientSampleId :

SSTDCCC040



Tgt Ion:	276	Resp:	255497
Ion	Ratio	Lower	Upper
276	100		
277	22.2	17.8	26.6
138	18.8	34.6	51.8#

