

Data Path : Z:\HPCHEM1\BNA_F\DATA\BF091815\
 Data File : BF081698.D
 Acq On : 18 Sep 2015 15:34
 Operator : UM/IZ
 Sample : PB85485BS
 Misc :
 ALS Vial : 7 Sample Multiplier: 1

Instrument :
 BNA_F
 ClientSampleId :
 PB85485BS

Quant Time: Sep 18 22:58:33 2015
 Quant Method : Z:\HPCHEM1\BNA_F\METHODS\8270-BF090115.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Tue Sep 15 14:03:59 2015
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Dichlorobenzene-d4	8.16	152	42356	20.00	ng	-0.06
21) Naphthalene-d8	11.06	136	170157	20.00	ng	-0.06
38) Acenaphthene-d10	15.20	164	81765	20.00	ng	-0.06
63) Phenanthrene-d10	18.70	188	172281	20.00	ng	-0.06
75) Chrysene-d12	23.35	240	145811	20.00	ng	-0.03
86) Perylene-d12	24.79	264	98036	20.00	ng	-0.02

System Monitoring Compounds

5) 2-Fluorophenol	5.34	112	321999	117.95	ng	-0.01
7) Phenol-d6	7.60	99	448388	124.08	ng	-0.04
23) Nitrobenzene-d5	9.48	82	293053	83.09	ng	-0.06
41) 2,4,6-Tribromophenol	17.11	330	96295	132.27	ng	-0.06
44) 2-Fluorobiphenyl	13.72	172	527268	82.64	ng	-0.06
78) Terphenyl-d14	22.07	244	531749	84.89	ng	-0.04

Target Compounds

						Qvalue
2) 1,4-Dioxane	1.52	88	34521	28.16	ng	# 11
3) Pyridine	2.12	79	116673	34.51	ng	# 77
4) n-Nitrosodimethylamine	2.09	42	87559	46.62	ng	# 59
6) Aniline	7.48	93	136200	26.69	ng	# 83
8) 2-Chlorophenol	7.73	128	125776	41.81	ng	# 84
9) Benzaldehyde	7.19	77	46659	20.61	ng	# 72
10) Phenol	7.62	94	165272	39.44	ng	# 52
11) bis(2-Chloroethyl)ether	7.70	93	127793	42.27	ng	# 75
12) 1,3-Dichlorobenzene	8.02	146	127264	39.59	ng	# 89
13) 1,4-Dichlorobenzene	8.21	146	128399	39.23	ng	# 90
14) 1,2-Dichlorobenzene	8.53	146	125503	42.59	ng	# 91
15) Benzyl Alcohol	8.62	79	126592	42.70	ng	# 72
16) 2,2'-oxybis(1-Chloropropan	8.93	45	271298	46.81	ng	# 95
17) 2-Methylphenol	8.96	107	100867	42.02	ng	# 80
18) Hexachloroethane	9.26	117	53761	42.22	ng	# 88
19) n-Nitroso-di-n-propylamine	9.24	70	106995	43.50	ng	# 90
20) 3+4-Methylphenols	9.34	107	133531	41.26	ng	# 85
22) Acetophenone	9.16	105	192628	41.45	ng	# 73
24) Nitrobenzene	9.51	77	154920	42.30	ng	# 61
25) Isophorone	10.10	82	271225	41.35	ng	# 83
26) 2-Nitrophenol	10.24	139	63701	39.91	ng	# 21
27) 2,4-Dimethylphenol	10.52	122	113989	41.34	ng	# 73
28) bis(2-Chloroethoxy)methane	10.71	93	160843	42.45	ng	# 91
29) 2,4-Dichlorophenol	10.86	162	101788	42.57	ng	# 93
30) 1,2,4-Trichlorobenzene	10.97	180	106652	41.06	ng	# 95
31) Naphthalene	11.12	128	369748	40.74	ng	# 99
32) Benzoic acid	11.03	122	55065	29.03	ng	# 48
33) 4-Chloroaniline	11.36	127	86573	23.39	ng	# 83
34) Hexachlorobutadiene	11.48	225	70716	44.74	ng	# 100
35) Caprolactam	12.29	113	18472	25.19	ng	# 23
36) 4-Chloro-3-methylphenol	12.68	107	126518	47.28	ng	# 73
37) 2-Methylnaphthalene	12.77	142	253298	42.89	ng	# 87
39) 1,2,4,5-Tetrachlorobenzene	13.18	216	110329	42.67	ng	# 97
40) Hexachlorocyclopentadiene	13.16	237	108503	85.50	ng	# 94


Data Path : Z:\HPCHEM1\BNA_F\DATA\BF091815\
 Data File : BF081698.D
 Acq On : 18 Sep 2015 15:34
 Operator : UM/IZ
 Sample : PB85485BS
 Misc :
 ALS Vial : 7 Sample Multiplier: 1

Instrument :
 BNA_F
 ClientSampleId :
 PB85485BS

Quant Time: Sep 18 22:58:33 2015
 Quant Method : Z:\HPCHEM1\BNA_F\METHODS\8270-BF090115.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Tue Sep 15 14:03:59 2015
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
42) 2,4,6-Trichlorophenol	13.52	196	72279	40.50	ng	97
43) 2,4,5-Trichlorophenol	13.64	196	76969	42.28	ng #	88
45) 1,1'-Biphenyl	13.91	154	309241	41.11	ng	96
46) 2-Chloronaphthalene	13.90	162	230227	42.38	ng #	86
47) 2-Nitroaniline	14.25	65	96896	43.76	ng #	61
48) Acenaphthylene	14.86	152	396191	41.11	ng	97
49) Dimethylphthalate	14.80	163	282310	44.44	ng #	96
50) 2,6-Dinitrotoluene	14.89	165	57390	39.80	ng #	31
51) Acenaphthene	15.28	154	225635	41.15	ng	87
52) 3-Nitroaniline	15.25	138	42994	26.19	ng #	41
53) 2,4-Dinitrophenol	15.55	184	55904	76.29	ng #	17
54) Dibenzofuran	15.71	168	351164	43.87	ng #	85
55) 4-Nitrophenol	15.88	139	77689	75.34	ng #	4
56) 2,4-Dinitrotoluene	15.84	165	81989	44.66	ng #	47
57) Fluorene	16.52	166	283087	46.60	ng	96
58) 2,3,4,6-Tetrachlorophenol	16.07	232	66014	47.49	ng	93
59) Diethylphthalate	16.49	149	318439	47.65	ng	95
60) 4-Chlorophenyl-phenylether	16.61	204	135567	47.88	ng #	86
61) 4-Nitroaniline	16.72	138	60694	36.60	ng #	19
62) Azobenzene	16.97	77	356657	48.01	ng	81
64) 4,6-Dinitro-2-methylphenol	16.81	198	39579	41.07	ng	81
65) n-Nitrosodiphenylamine	16.93	169	245155	39.11	ng	94
66) 4-Bromophenyl-phenylether	17.75	248	77443	44.46	ng	94
67) Hexachlorobenzene	17.80	284	76996	42.27	ng #	80
68) Atrazine	18.35	200	89660	49.42	ng	78
69) Pentachlorophenol	18.35	266	62115	71.14	ng	99
70) Phenanthrene	18.76	178	405234	42.31	ng	98
71) Anthracene	18.88	178	413714	42.04	ng	98
72) Carbazole	19.36	167	380647	41.22	ng	96
73) Di-n-butylphthalate	20.41	149	519933	47.68	ng #	97
74) Fluoranthene	21.39	202	448597	43.27	ng	90
76) Benzidine	21.71	184	189394	30.26	ng	97
77) Pyrene	21.74	202	460653	42.65	ng	98
79) Butylbenzylphthalate	22.77	149	209726	44.65	ng #	77
80) Benzo(a)anthracene	23.34	228	371683	40.03	ng	96
81) 3,3'-Dichlorobenzidine	23.35	252	86296	27.55	ng #	93
82) Chrysene	23.39	228	359382	43.17	ng	95
83) Bis(2-ethylhexyl)phthalate	23.47	149	299959	48.39	ng #	91
84) Di-n-octyl phthalate	24.12	149	446374	43.73	ng	95
85) Indeno(1,2,3-cd)pyrene	25.82	276	228919	37.48	ng #	83
87) Benzo(b)fluoranthene	24.45	252	545857	77.99	ng #	86
88) Benzo(k)fluoranthene	24.45	252	545857	93.99	ng #	87
89) Benzo(a)pyrene	24.75	252	255295	43.04	ng #	81
90) Dibenzo(a,h)anthracene	25.83	278	193683	42.26	ng #	80
91) Benzo(g,h,i)perylene	26.13	276	182143	41.64	ng #	68

(#) = qualifier out of range (m) = manual integration (+) = signals summed



#1

1,4-Dichlorobenzene-d4

Concen: 20.00 ng

RT: 8.16 min Scan# 619

Delta R.T. -0.06 min

Lab File: BF081698.D

Acq: 18 Sep 2015 15:34

Instrument :

BNA_F

ClientSampled :

PB85485BS


Tgt Ion:152 Resp: 42356

Ion Ratio Lower Upper

152 100

150 160.3 124.7 187.1




115 71.0 39.0 58.4#



Abundance Ion 152.00 (151.70 to 152.70): E

Ion 150.00 (149.70 to 150.70): E

Ion 115.00 (114.70 to 115.70): E



#2

1,4-Dioxane

Concen: 28.16 ng

RT: 1.52 min Scan# 38

Delta R.T. -0.04 min

Lab File: BF081698.D

Acq: 18 Sep 2015 15:34


Tgt Ion: 88 Resp: 34521

Ion Ratio Lower Upper

88 100

58 0.0 77.7 116.5#



43 0.0 26.6 40.0#




Abundance Ion 88.00 (87.70 to 88.70): BF0

Ion 58.00 (57.70 to 58.70): BF0

Ion 43.00 (42.70 to 43.70): BF0





#3

Pyridine

Concen: 34.51 ng

RT: 2.12 min Scan# 90

Delta R.T. 0.04 min

Lab File: BF081698.D

Acq: 18 Sep 2015 15:34

Instrument :

BNA_F

ClientSampleId :

PB85485BS


Tgt Ion: 79 Resp: 116673

Ion Ratio Lower Upper




79 100

52 115.9 76.8 115.2#

51 57.5 32.2 48.4#



Abundance Ion 79.00 (78.70 to 79.70): BF0
Ion 52.00 (51.70 to 52.70): BF0
Ion 51.00 (50.70 to 51.70): BF0



#4

n-Nitrosodimethylamine

Concen: 46.62 ng

RT: 2.09 min Scan# 88

Delta R.T. 0.04 min

Lab File: BF081698.D

Acq: 18 Sep 2015 15:34


Tgt Ion: 42 Resp: 87559

Ion Ratio Lower Upper



42 100


74 80.1 105.0 157.6#

44 7.1 6.6 10.0



Abundance Ion 42.00 (41.70 to 42.70): BF0
Ion 74.00 (73.70 to 74.70): BF0
Ion 44.00 (43.70 to 44.70): BF0





#5

2-Fluorophenol

Concen: 117.95 ng

RT: 5.34 min Scan# 372

Delta R.T. -0.01 min

Lab File: BF081698.D

Acq: 18 Sep 2015 15:34

Instrument :

BNA_F

ClientSampleId :

PB85485BS


Tgt Ion: 112 Resp: 321999

Ion Ratio Lower Upper

112 100

64 78.2 39.7 59.5#




63 44.3 17.4 26.0#



Abundance Ion 112.00 (111.70 to 112.70): E

Ion 64.00 (63.70 to 64.70): BF0

Ion 63.00 (62.70 to 63.70): BF0



#6

Aniline

Concen: 26.69 ng

RT: 7.48 min Scan# 559

Delta R.T. -0.04 min

Lab File: BF081698.D

Acq: 18 Sep 2015 15:34


Tgt Ion: 93 Resp: 136200

Ion Ratio Lower Upper

93 100

66 45.1 27.2 40.8#



65 23.6 14.7 22.1#




Abundance Ion 93.00 (92.70 to 93.70): BF0

Ion 66.00 (65.70 to 66.70): BF0

Ion 65.00 (64.70 to 65.70): BF0





#7

Phenol-d6

Concen: 124.08 ng

RT: 7.60 min Scan# 570

Delta R.T. -0.04 min

Lab File: BF081698.D

Acq: 18 Sep 2015 15:34

Instrument :

BNA_F

ClientSampleId :

PB85485BS


Tgt Ion: 99 Resp: 448388

Ion Ratio Lower Upper

99 100

42 26.5 13.7 20.5#




71 40.7 14.2 21.2#



Abundance Ion 99.00 (98.70 to 99.70): BF0

Ion 42.00 (41.70 to 42.70): BF0

Ion 71.00 (70.70 to 71.70): BF0



#8

2-Chlorophenol

Concen: 41.81 ng

RT: 7.73 min Scan# 581

Delta R.T. -0.04 min

Lab File: BF081698.D

Acq: 18 Sep 2015 15:34


Tgt Ion: 128 Resp: 125776

Ion Ratio Lower Upper

128 100

130 32.7 12.2 52.2



64 58.3 20.5 60.5




Abundance Ion 128.00 (127.70 to 128.70): E

Ion 130.00 (129.70 to 130.70): E

Ion 64.00 (63.70 to 64.70): BF0





#9

Benzaldehyde

Concen: 20.61 ng

RT: 7.19 min Scan# 534

Delta R.T. -0.04 min

Lab File: BF081698.D

Acq: 18 Sep 2015 15:34

Instrument :

BNA_F

ClientSampleId :

PB85485BS





Tgt Ion: 77 Resp: 46659

Ion Ratio Lower Upper

77 100

105 91.1 91.6 131.6#

106 81.8 102.6 142.6#



#10

Phenol

Concen: 39.44 ng

RT: 7.62 min Scan# 572

Delta R.T. -0.04 min

Lab File: BF081698.D

Acq: 18 Sep 2015 15:34




Tgt Ion: 94 Resp: 165272


Ion Ratio Lower Upper

94 100

65 37.1 0.7 40.7

66 49.5 0.8 40.8#









#11
 bis(2-Chloroethyl)ether
 Concen: 42.27 ng
 RT: 7.70 min Scan# 579
 Delta R.T. -0.06 min
 Lab File: BF081698.D
 Acq: 18 Sep 2015 15:34




Instrument :
 BNA_F
 ClientSampleId :
 PB85485BS


Tgt Ion: 93 Resp: 127793
 Ion Ratio Lower Upper
 93 100
 63 116.9 65.6 105.6#
 95 33.1 13.6 53.6



#12
 1,3-Dichlorobenzene
 Concen: 39.59 ng
 RT: 8.02 min Scan# 607
 Delta R.T. -0.04 min
 Lab File: BF081698.D
 Acq: 18 Sep 2015 15:34

Tgt Ion: 146 Resp: 127264
 Ion Ratio Lower Upper
 146 100
 148 63.1 51.0 76.4
 75 38.5 15.0 22.6#









#13
 1,4-Dichlorobenzene
 Concen: 39.23 ng
 RT: 8.21 min Scan# 623
 Delta R.T. -0.04 min
 Lab File: BF081698.D
 Acq: 18 Sep 2015 15:34




Instrument :
 BNA_F
 ClientSampleId :
 PB85485BS


Tgt Ion:146 Resp: 128399
 Ion Ratio Lower Upper
 146 100
 148 62.6 51.8 77.6
 111 44.4 24.9 37.3#



#14
 1,2-Dichlorobenzene
 Concen: 42.59 ng
 RT: 8.53 min Scan# 651
 Delta R.T. -0.04 min
 Lab File: BF081698.D
 Acq: 18 Sep 2015 15:34

Tgt Ion:146 Resp: 125503
 Ion Ratio Lower Upper
 146 100
 148 64.2 52.7 79.1
 111 48.9 28.8 43.2#





#15

Benzyl Alcohol

Concen: 42.70 ng

RT: 8.62 min Scan# 659

Delta R.T. -0.04 min

Lab File: BF081698.D

Acq: 18 Sep 2015 15:34

Instrument :

BNA_F

ClientSampleId :

PB85485BS





Tgt Ion: 79 Resp: 126592

Ion Ratio Lower Upper

79 100

108 68.0 88.6 132.8#

77 62.6 47.0 70.4



#16

2,2'-oxybis(1-Chloropropane)

Concen: 46.81 ng

RT: 8.93 min Scan# 686

Delta R.T. -0.06 min

Lab File: BF081698.D

Acq: 18 Sep 2015 15:34




Tgt Ion: 45 Resp: 271298


Ion Ratio Lower Upper

45 100

77 10.0 0.0 28.1

79 7.8 0.0 26.0









#17
2-Methylphenol
Concen: 42.02 ng
RT: 8.96 min Scan# 689
Delta R.T. -0.04 min
Lab File: BF081698.D
Acq: 18 Sep 2015 15:34




Instrument :
BNA_F
ClientSampleId :
PB85485BS


Tgt Ion	Ratio	Lower	Upper
107	100		
108	115.7	96.6	145.0
77	57.4	23.3	34.9
79	53.5	22.0	33.0



#18
Hexachloroethane
Concen: 42.22 ng
RT: 9.26 min Scan# 715
Delta R.T. -0.06 min
Lab File: BF081698.D
Acq: 18 Sep 2015 15:34

Tgt Ion	Ratio	Lower	Upper
117	100		
119	93.8	83.8	125.6
201	79.4	55.1	82.7








#19
n-Nitroso-di-n-propylamine
Concen: 43.50 ng
RT: 9.24 min Scan# 713
Delta R.T. -0.06 min
Lab File: BF081698.D
Acq: 18 Sep 2015 15:34





Instrument :
BNA_F
ClientSampleId :
PB85485BS


Tgt Ion: 70 Resp: 106995
Ion Ratio Lower Upper
70 100
42 78.6 63.8 95.6
101 8.7 9.2 13.8#
130 16.3 27.3 40.9#



#20
3+4-Methylphenols
Concen: 41.26 ng
RT: 9.34 min Scan# 722
Delta R.T. -0.04 min
Lab File: BF081698.D
Acq: 18 Sep 2015 15:34

Tgt Ion:107 Resp: 133531
Ion Ratio Lower Upper
107 100
108 88.7 74.6 114.6
77 39.9 0.7 40.7
79 32.8 0.0 38.9





#21

Naphthalene-d8

Concen: 20.00 ng

RT: 11.06 min Scan# 873

Delta R.T. -0.06 min

Lab File: BF081698.D

Acq: 18 Sep 2015 15:34

Instrument :

BNA_F

ClientSampleId :

PB85485BS

Tgt Ion:136 Resp: 170157





Ion Ratio Lower Upper

136 100

137 11.0 9.0 13.6

54 12.9 8.2 12.2#

68 5.0 5.8 8.6#



#22

Acetophenone

Concen: 41.45 ng

RT: 9.16 min Scan# 706

Delta R.T. -0.06 min

Lab File: BF081698.D

Acq: 18 Sep 2015 15:34

Tgt Ion:105 Resp: 192628




Ion Ratio Lower Upper


105 100

71 0.0 4.7 7.1#

51 39.6 22.0 33.0#

120 18.8 30.1 45.1#









#23
 Nitrobenzene-d5
 Concen: 83.09 ng
 RT: 9.48 min Scan# 734
 Delta R.T. -0.06 min
 Lab File: BF081698.D
 Acq: 18 Sep 2015 15:34




Instrument :
 BNA_F
 ClientSampleId :
 PB85485BS


Tgt Ion: 82 Resp: 293053
 Ion Ratio Lower Upper
 82 100
 128 36.8 51.0 76.6#
 54 68.5 47.5 71.3



#24
 Nitrobenzene
 Concen: 42.30 ng
 RT: 9.51 min Scan# 737
 Delta R.T. -0.06 min
 Lab File: BF081698.D
 Acq: 18 Sep 2015 15:34

Tgt Ion: 77 Resp: 154920
 Ion Ratio Lower Upper
 77 100
 123 37.2 59.8 89.8#
 65 16.1 10.2 15.4#





#25

Isophorone

Concen: 41.35 ng

RT: 10.10 min Scan# 789

Delta R.T. -0.06 min

Lab File: BF081698.D

Acq: 18 Sep 2015 15:34

Instrument :

BNA_F

ClientSampleId :

PB85485BS





Tgt Ion: 82 Resp: 271225

Ion Ratio Lower Upper

82 100

95 8.0 4.0 6.0#

138 13.8 18.3 27.5#



#26

2-Nitrophenol

Concen: 39.91 ng

RT: 10.24 min Scan# 801

Delta R.T. -0.06 min

Lab File: BF081698.D

Acq: 18 Sep 2015 15:34




Tgt Ion: 139 Resp: 63701


Ion Ratio Lower Upper

139 100

109 39.9 11.0 16.4#

65 77.5 24.7 37.1#









#27
 2,4-Dimethylphenol
 Concen: 41.34 ng
 RT: 10.52 min Scan# 825
 Delta R.T. -0.06 min
 Lab File: BF081698.D
 Acq: 18 Sep 2015 15:34




Instrument :
 BNA_F
 ClientSampled :
 PB85485BS


Tgt Ion:122 Resp: 113989
 Ion Ratio Lower Upper
 122 100
 107 126.6 71.8 107.6#
 121 57.0 42.3 63.5



#28
 bis(2-Chloroethoxy)methane
 Concen: 42.45 ng
 RT: 10.71 min Scan# 842
 Delta R.T. -0.04 min
 Lab File: BF081698.D
 Acq: 18 Sep 2015 15:34

Tgt Ion: 93 Resp: 160843
 Ion Ratio Lower Upper
 93 100
 95 31.7 27.5 41.3
 123 9.2 13.7 20.5#









#29
 2,4-Dichlorophenol
 Concen: 42.57 ng
 RT: 10.86 min Scan# 855
 Delta R.T. -0.04 min
 Lab File: BF081698.D
 Acq: 18 Sep 2015 15:34




Instrument :
 BNA_F
 ClientSampleId :
 PB85485BS


Tgt Ion:162 Resp: 101788
 Ion Ratio Lower Upper
 162 100
 164 63.5 44.9 84.9
 98 43.5 13.0 53.0



#30
 1,2,4-Trichlorobenzene
 Concen: 41.06 ng
 RT: 10.97 min Scan# 865
 Delta R.T. -0.06 min
 Lab File: BF081698.D
 Acq: 18 Sep 2015 15:34

Tgt Ion:180 Resp: 106652
 Ion Ratio Lower Upper
 180 100
 182 92.9 77.1 115.7
 145 33.9 23.3 34.9





#31

Naphthalene

Concen: 40.74 ng

RT: 11.12 min Scan# 878

Delta R.T. -0.04 min

Lab File: BF081698.D

Acq: 18 Sep 2015 15:34

Instrument :

BNA_F

ClientSampleId :

PB85485BS


Tgt Ion:128 Resp: 369748

Ion Ratio Lower Upper

128 100

129 10.6 8.4 12.6




127 13.0 9.9 14.9



Abundance Ion 128.00 (127.70 to 128.70): E

Ion 129.00 (128.70 to 129.70): E

Ion 127.00 (126.70 to 127.70): E



#32

Benzoic acid

Concen: 29.03 ng

RT: 11.03 min Scan# 870

Delta R.T. -0.06 min

Lab File: BF081698.D

Acq: 18 Sep 2015 15:34


Tgt Ion:122 Resp: 55065

Ion Ratio Lower Upper

122 100

105 132.6 76.1 116.1#



77 117.0 40.5 80.5#




Abundance Ion 122.00 (121.70 to 122.70): E

Ion 105.00 (104.70 to 105.70): E

Ion 77.00 (76.70 to 77.70): BF0









#33
4-Chloroaniline
Concen: 23.39 ng
RT: 11.36 min Scan# 899
Delta R.T. -0.04 min
Lab File: BF081698.D
Acq: 18 Sep 2015 15:34

Instrument :
BNA_F
ClientSampleId :
PB85485BS

Tgt Ion:127 Resp: 86573
Ion Ratio Lower Upper
127 100
129 32.0 25.8 38.6
65 40.1 17.9 26.9#
92 23.2 10.5 15.7#




Abundance Ion 127.00 (126.70 to 127.70): E
Ion 129.00 (128.70 to 129.70): E
Ion 65.00 (64.70 to 65.70): BF0
Ion 92.00 (91.70 to 92.70): BF0






#34
Hexachlorobutadiene
Concen: 44.74 ng
RT: 11.48 min Scan# 909
Delta R.T. -0.04 min
Lab File: BF081698.D
Acq: 18 Sep 2015 15:34

Tgt Ion:225 Resp: 70716
Ion Ratio Lower Upper
225 100
223 62.8 50.2 75.2
227 64.7 52.2 78.2



Abundance Ion 225.00 (224.70 to 225.70): E
Ion 223.00 (222.70 to 223.70): E
Ion 227.00 (226.70 to 227.70): E









#35
 Caprolactam
 Concen: 25.19 ng
 RT: 12.29 min Scan# 980
 Delta R.T. -0.10 min
 Lab File: BF081698.D
 Acq: 18 Sep 2015 15:34

Instrument :
 BNA_F
 ClientSampleId :
 PB85485BS

Tgt Ion:113 Resp: 18472
 Ion Ratio Lower Upper
 113 100
 55 280.5 152.4 192.4#
 56 210.6 104.6 144.6#




Abundance Ion 113.00 (112.70 to 113.70): E
 Ion 55.00 (54.70 to 55.70): BF0
 Ion 56.00 (55.70 to 56.70): BF0






#36
 4-Chloro-3-methylphenol
 Concen: 47.28 ng
 RT: 12.68 min Scan# 1014
 Delta R.T. -0.04 min
 Lab File: BF081698.D
 Acq: 18 Sep 2015 15:34

Tgt Ion:107 Resp: 126518
 Ion Ratio Lower Upper
 107 100
 144 21.8 25.4 38.0#
 142 66.6 77.3 115.9#



Abundance Ion 107.00 (106.70 to 107.70): E
 Ion 144.00 (143.70 to 144.70): E
 Ion 142.00 (141.70 to 142.70): E





#37

2-Methylnaphthalene

Concen: 42.89 ng

RT: 12.77 min Scan# 1022

Delta R.T. -0.06 min

Lab File: BF081698.D

Acq: 18 Sep 2015 15:34

Instrument :

BNA_F

ClientSampleId :

PB85485BS


Tgt Ion:142 Resp: 253298

Ion Ratio Lower Upper

142 100

141 89.4 67.5 101.3




115 42.2 18.2 27.2#



Abundance Ion 142.00 (141.70 to 142.70): E

Ion 141.00 (140.70 to 141.70): E

Ion 115.00 (114.70 to 115.70): E



#38

Acenaphthene-d10

Concen: 20.00 ng

RT: 15.20 min Scan# 1235

Delta R.T. -0.06 min

Lab File: BF081698.D

Acq: 18 Sep 2015 15:34


Tgt Ion:164 Resp: 81765

Ion Ratio Lower Upper

164 100

162 106.6 75.2 112.8



160 47.0 31.5 47.3




Abundance Ion 164.00 (163.70 to 164.70): E

Ion 162.00 (161.70 to 162.70): E

Ion 160.00 (159.70 to 160.70): E





#39

1,2,4,5-Tetrachlorobenzene

Concen: 42.67 ng

RT: 13.18 min Scan# 1058

Delta R.T. -0.04 min

Lab File: BF081698.D

Acq: 18 Sep 2015 15:34

Instrument :

BNA_F

ClientSampleId :

PB85485BS

Tgt Ion:216 Resp: 110329


Ion Ratio Lower Upper

216 100

214 77.4 63.5 95.3

179 22.6 16.6 24.8

108 23.5 16.3 24.5






Abundance Ion 216.00 (215.70 to 216.70): E

Ion 214.00 (213.70 to 214.70): E

Ion 179.00 (178.70 to 179.70): E

Ion 108.00 (107.70 to 108.70): E



#40

Hexachlorocyclopentadiene

Concen: 85.50 ng

RT: 13.16 min Scan# 1056

Delta R.T. -0.06 min

Lab File: BF081698.D

Acq: 18 Sep 2015 15:34


Tgt Ion:237 Resp: 108503

Ion Ratio Lower Upper

237 100

235 65.7 42.0 82.0


272 11.5 0.0 36.1




Abundance Ion 236.80 (236.50 to 237.50): E


Ion 234.90 (234.60 to 235.60): E

Ion 271.75 (271.45 to 272.45): E



Time-->









#41
 2,4,6-Tribromophenol
 Concen: 132.27 ng
 RT: 17.11 min Scan# 1402
 Delta R.T. -0.06 min
 Lab File: BF081698.D
 Acq: 18 Sep 2015 15:34

Instrument :
 BNA_F
 ClientSampleId :
 PB85485BS

Tgt Ion: 330 Resp: 96295
 Ion Ratio Lower Upper
 330 100
 332 99.0 76.6 114.8
 141 49.9 29.7 44.5#




Abundance Ion 329.80 (329.50 to 330.50): E
 Ion 331.80 (331.50 to 332.50): E
 Ion 141.00 (140.70 to 141.70): E






#42
 2,4,6-Trichlorophenol
 Concen: 40.50 ng
 RT: 13.52 min Scan# 1088
 Delta R.T. -0.06 min
 Lab File: BF081698.D
 Acq: 18 Sep 2015 15:34

Tgt Ion: 196 Resp: 72279
 Ion Ratio Lower Upper
 196 100
 198 90.5 75.7 113.5
 200 30.0 23.9 35.9



Abundance Ion 196.00 (195.70 to 196.70): E
 Ion 198.00 (197.70 to 198.70): E
 Ion 200.00 (199.70 to 200.70): E









#43
 2,4,5-Trichlorophenol
 Concen: 42.28 ng
 RT: 13.64 min Scan# 1098
 Delta R.T. -0.04 min
 Lab File: BF081698.D
 Acq: 18 Sep 2015 15:34

Instrument :
 BNA_F
 ClientSampleId :
 PB85485BS

Tgt Ion:196 Resp: 76969
 Ion Ratio Lower Upper
 196 100
 198 92.6 75.4 113.0
 97 63.6 33.3 49.9#
 132 36.3 24.6 37.0




Abundance Ion 196.00 (195.70 to 196.70): E
 Ion 198.00 (197.70 to 198.70): E
 Ion 97.00 (96.70 to 97.70): BF0
 Ion 132.00 (131.70 to 132.70): E






#44
 2-Fluorobiphenyl
 Concen: 82.64 ng
 RT: 13.72 min Scan# 1105
 Delta R.T. -0.06 min
 Lab File: BF081698.D
 Acq: 18 Sep 2015 15:34

Tgt Ion:172 Resp: 527268
 Ion Ratio Lower Upper
 172 100
 171 35.6 26.1 39.1
 170 23.6 17.4 26.2



Abundance Ion 172.00 (171.70 to 172.70): E
 Ion 171.00 (170.70 to 171.70): E
 Ion 170.00 (169.70 to 170.70): E





#45

1,1'-Biphenyl

Concen: 41.11 ng

RT: 13.91 min Scan# 1122

Delta R.T. -0.06 min

Lab File: BF081698.D

Acq: 18 Sep 2015 15:34

Instrument :

BNA_F

ClientSampleId :

PB85485BS


Tgt Ion:154 Resp: 309241

Ion Ratio Lower Upper

154 100

153 39.5 17.1 57.1




76 12.8 0.0 31.6



Abundance Ion 154.00 (153.70 to 154.70): E

Ion 153.00 (152.70 to 153.70): E

Ion 76.00 (75.70 to 76.70): BF0



#46

2-Chloronaphthalene

Concen: 42.38 ng

RT: 13.90 min Scan# 1121

Delta R.T. -0.06 min

Lab File: BF081698.D

Acq: 18 Sep 2015 15:34


Tgt Ion:162 Resp: 230227

Ion Ratio Lower Upper

162 100

127 48.7 27.6 41.4#



164 33.2 25.4 38.2




Abundance Ion 162.00 (161.70 to 162.70): E

Ion 127.00 (126.70 to 127.70): E

Ion 164.00 (163.70 to 164.70): E









#47
 2-Nitroaniline
 Concen: 43.76 ng
 RT: 14.25 min Scan# 1152
 Delta R.T. -0.06 min
 Lab File: BF081698.D
 Acq: 18 Sep 2015 15:34




Instrument :
 BNA_F
 ClientSampleId :
 PB85485BS


Tgt Ion: 65 Resp: 96896
 Ion Ratio Lower Upper
 65 100
 92 59.3 55.4 83.0
 138 80.8 115.5 173.3#



#48
 Acenaphthylene
 Concen: 41.11 ng
 RT: 14.86 min Scan# 1205
 Delta R.T. -0.06 min
 Lab File: BF081698.D
 Acq: 18 Sep 2015 15:34

Tgt Ion: 152 Resp: 396191
 Ion Ratio Lower Upper
 152 100
 151 20.2 14.6 22.0
 153 12.7 10.3 15.5









#49
Dimethylphthalate
Concen: 44.44 ng
RT: 14.80 min Scan# 1200
Delta R.T. -0.06 min
Lab File: BF081698.D
Acq: 18 Sep 2015 15:34




Instrument :
BNA_F
ClientSampleId :
PB85485BS


Tgt Ion:163 Resp: 282310
Ion Ratio Lower Upper
163 100
194 3.3 4.4 6.6#
164 9.5 8.3 12.5



#50
2,6-Dinitrotoluene
Concen: 39.80 ng
RT: 14.89 min Scan# 1208
Delta R.T. -0.06 min
Lab File: BF081698.D
Acq: 18 Sep 2015 15:34

Tgt Ion:165 Resp: 57390
Ion Ratio Lower Upper
165 100
63 84.9 32.3 48.5#
89 74.5 28.6 43.0#





#51

Acenaphthene

Concen: 41.15 ng

RT: 15.28 min Scan# 1242

Delta R.T. -0.06 min

Lab File: BF081698.D

Acq: 18 Sep 2015 15:34

Instrument :

BNA_F

ClientSampleId :

PB85485BS


Tgt Ion:154 Resp: 225635

Ion Ratio Lower Upper

154 100

153 116.3 82.1 123.1




152 55.7 37.9 56.9



Abundance Ion 154.00 (153.70 to 154.70): E

Ion 153.00 (152.70 to 153.70): E

Ion 152.00 (151.70 to 152.70): E



#52

3-Nitroaniline

Concen: 26.19 ng

RT: 15.25 min Scan# 1239

Delta R.T. -0.07 min

Lab File: BF081698.D

Acq: 18 Sep 2015 15:34


Tgt Ion:138 Resp: 42994

Ion Ratio Lower Upper

138 100

108 11.0 5.7 8.5#



92 141.5 67.7 101.5#




Abundance Ion 138.00 (137.70 to 138.70): E

Ion 108.00 (107.70 to 108.70): E

Ion 92.00 (91.70 to 92.70): BF0









#53
2,4-Dinitrophenol
Concen: 76.29 ng
RT: 15.55 min Scan# 1265
Delta R.T. -0.04 min
Lab File: BF081698.D
Acq: 18 Sep 2015 15:34




Instrument :
BNA_F
ClientSampleId :
PB85485BS


Tgt Ion:184 Resp: 55904
Ion Ratio Lower Upper
184 100
63 109.1 35.4 53.2#
154 80.8 32.2 48.4#



#54
Dibenzofuran
Concen: 43.87 ng
RT: 15.71 min Scan# 1279
Delta R.T. -0.06 min
Lab File: BF081698.D
Acq: 18 Sep 2015 15:34

Tgt Ion:168 Resp: 351164
Ion Ratio Lower Upper
168 100
139 42.0 24.2 36.2#
169 13.1 10.6 15.8









#55
4-Nitrophenol
Concen: 75.34 ng
RT: 15.88 min Scan# 1294
Delta R.T. -0.04 min
Lab File: BF081698.D
Acq: 18 Sep 2015 15:34




Instrument :
BNA_F
ClientSampleId :
PB85485BS


Tgt Ion:139 Resp: 77689
Ion Ratio Lower Upper
139 100
109 91.2 12.7 52.7#
65 138.7 45.9 85.9#



#56
2,4-Dinitrotoluene
Concen: 44.66 ng
RT: 15.84 min Scan# 1291
Delta R.T. -0.06 min
Lab File: BF081698.D
Acq: 18 Sep 2015 15:34

Tgt Ion:165 Resp: 81989
Ion Ratio Lower Upper
165 100
63 70.2 29.8 44.6#
89 95.3 44.5 66.7#
182 1.9 3.0 4.4#









#57
 Fluorene
 Concen: 46.60 ng
 RT: 16.52 min Scan# 1350
 Delta R.T. -0.06 min
 Lab File: BF081698.D
 Acq: 18 Sep 2015 15:34




Instrument :
 BNA_F
 ClientSampleId :
 PB85485BS


Tgt Ion:166 Resp: 283087
 Ion Ratio Lower Upper
 166 100
 165 94.9 72.6 109.0
 167 13.5 10.6 16.0



#58
 2,3,4,6-Tetrachlorophenol
 Concen: 47.49 ng
 RT: 16.07 min Scan# 1311
 Delta R.T. -0.06 min
 Lab File: BF081698.D
 Acq: 18 Sep 2015 15:34

Tgt Ion:232 Resp: 66014
 Ion Ratio Lower Upper
 232 100
 131 55.3 38.5 57.7
 130 2.4 1.8 2.6
 166 32.1 25.0 37.6









#59
Diethylphthalate
Concen: 47.65 ng
RT: 16.49 min Scan# 1348
Delta R.T. -0.06 min
Lab File: BF081698.D
Acq: 18 Sep 2015 15:34




Instrument :
BNA_F
ClientSampleId :
PB85485BS


Tgt Ion:149 Resp: 318439
Ion Ratio Lower Upper
149 100
177 18.7 17.5 26.3
150 12.7 9.4 14.2



#60
4-Chlorophenyl-phenylether
Concen: 47.88 ng
RT: 16.61 min Scan# 1358
Delta R.T. -0.06 min
Lab File: BF081698.D
Acq: 18 Sep 2015 15:34

Tgt Ion:204 Resp: 135567
Ion Ratio Lower Upper
204 100
206 32.8 24.8 37.2
141 67.5 42.2 63.4#









#61
4-Nitroaniline
Concen: 36.60 ng
RT: 16.72 min Scan# 1368
Delta R.T. -0.06 min
Lab File: BF081698.D
Acq: 18 Sep 2015 15:34

Instrument :
BNA_F
ClientSampleId :
PB85485BS

Tgt Ion:138 Resp: 60694
Ion Ratio Lower Upper
138 100
92 60.3 12.6 52.6#
108 95.4 12.4 52.4#




Abundance Ion 138.00 (137.70 to 138.70): E
Ion 92.00 (91.70 to 92.70): BF0
Ion 108.00 (107.70 to 108.70): E






#62
Azobenzene
Concen: 48.01 ng
RT: 16.97 min Scan# 1390
Delta R.T. -0.06 min
Lab File: BF081698.D
Acq: 18 Sep 2015 15:34

Tgt Ion: 77 Resp: 356657
Ion Ratio Lower Upper
77 100
182 17.9 15.1 55.1
105 19.7 3.4 43.4
51 40.1 12.0 52.0



Abundance Ion 77.00 (76.70 to 77.70): BF0
Ion 182.00 (181.70 to 182.70): E
Ion 105.00 (104.70 to 105.70): E
Ion 51.00 (50.70 to 51.70): BF0









#63
 Phenanthrene-d10
 Concen: 20.00 ng
 RT: 18.70 min Scan# 1541
 Delta R.T. -0.06 min
 Lab File: BF081698.D
 Acq: 18 Sep 2015 15:34

Instrument :
 BNA_F
 ClientSampleId :
 PB85485BS

Tgt Ion:188 Resp: 172281
 Ion Ratio Lower Upper
 188 100
 94 8.5 11.1 16.7#
 80 10.1 11.0 16.4#




Abundance Ion 188.00 (187.70 to 188.70): E
 Ion 94.00 (93.70 to 94.70): BF0
 Ion 80.00 (79.70 to 80.70): BF0






#64
 4,6-Dinitro-2-methylphenol
 Concen: 41.07 ng
 RT: 16.81 min Scan# 1376
 Delta R.T. -0.06 min
 Lab File: BF081698.D
 Acq: 18 Sep 2015 15:34

Tgt Ion:198 Resp: 39579
 Ion Ratio Lower Upper
 198 100
 51 79.2 39.6 79.6
 105 55.9 28.5 68.5



Abundance Ion 198.00 (197.70 to 198.70): E
 Ion 51.00 (50.70 to 51.70): BF0
 Ion 105.00 (104.70 to 105.70): E





#65

n-Nitrosodiphenylamine

Concen: 39.11 ng

RT: 16.93 min Scan# 1386

Delta R.T. -0.06 min

Lab File: BF081698.D

Acq: 18 Sep 2015 15:34

Instrument :

BNA_F

ClientSampleId :

PB85485BS


Tgt Ion:169 Resp: 245155

Ion Ratio Lower Upper

169 100

168 65.4 48.9 73.3




167 36.5 26.2 39.4



Abundance Ion 169.00 (168.70 to 169.70): E

Ion 168.00 (167.70 to 168.70): E

Ion 167.00 (166.70 to 167.70): E



#66

4-Bromophenyl-phenylether

Concen: 44.46 ng

RT: 17.75 min Scan# 1458

Delta R.T. -0.06 min

Lab File: BF081698.D

Acq: 18 Sep 2015 15:34


Tgt Ion:248 Resp: 77443

Ion Ratio Lower Upper

248 100

250 96.9 78.5 117.7



141 85.0 59.0 88.6




Abundance Ion 248.00 (247.70 to 248.70): E

Ion 250.00 (249.70 to 250.70): E

Ion 141.00 (140.70 to 141.70): E





#67

Hexachlorobenzene

Concen: 42.27 ng

RT: 17.80 min Scan# 1462

Delta R.T. -0.07 min

Lab File: BF081698.D

Acq: 18 Sep 2015 15:34

Instrument :

BNA_F

ClientSampleId :

PB85485BS


Tgt Ion:284 Resp: 76996

Ion Ratio Lower Upper

284 100

142 47.5 29.5 44.3#




249 36.2 19.5 29.3#



Abundance Ion 283.80 (283.50 to 284.50): E

Ion 142.00 (141.70 to 142.70): E

Ion 249.00 (248.70 to 249.70): E



#68

Atrazine

Concen: 49.42 ng

RT: 18.35 min Scan# 1510

Delta R.T. -0.06 min

Lab File: BF081698.D

Acq: 18 Sep 2015 15:34


Tgt Ion:200 Resp: 89660

Ion Ratio Lower Upper

200 100

173 25.1 13.2 53.2



215 41.9 41.8 81.8




Abundance Ion 200.00 (199.70 to 200.70): E

Ion 173.00 (172.70 to 173.70): E

Ion 215.00 (214.70 to 215.70): E









#69
 Pentachlorophenol
 Concen: 71.14 ng
 RT: 18.35 min Scan# 1510
 Delta R.T. -0.06 min
 Lab File: BF081698.D
 Acq: 18 Sep 2015 15:34

Instrument :
 BNA_F
 ClientSampleId :
 PB85485BS

Tgt Ion: 266 Resp: 62115
 Ion Ratio Lower Upper
 266 100
 268 61.0 49.8 74.6
 264 62.7 50.2 75.2




Abundance Ion 265.70 (265.40 to 266.40): E
 Ion 268.00 (267.70 to 268.70): E
 Ion 264.00 (263.70 to 264.70): E






#70
 Phenanthrene
 Concen: 42.31 ng
 RT: 18.76 min Scan# 1546
 Delta R.T. -0.07 min
 Lab File: BF081698.D
 Acq: 18 Sep 2015 15:34

Tgt Ion: 178 Resp: 405234
 Ion Ratio Lower Upper
 178 100
 176 18.8 13.8 20.8
 179 15.5 12.2 18.2



Abundance Ion 178.00 (177.70 to 178.70): E
 Ion 176.00 (175.70 to 176.70): E
 Ion 179.00 (178.70 to 179.70): E









#71
 Anthracene
 Concen: 42.04 ng
 RT: 18.88 min Scan# 1557
 Delta R.T. -0.07 min
 Lab File: BF081698.D
 Acq: 18 Sep 2015 15:34

Instrument :
 BNA_F
 ClientSampleId :
 PB85485BS

Tgt Ion:178 Resp: 413714
 Ion Ratio Lower Upper
 178 100
 176 19.0 14.1 21.1
 179 15.3 12.2 18.2




Abundance Ion 178.00 (177.70 to 178.70): E
 Ion 176.00 (175.70 to 176.70): E
 Ion 179.00 (178.70 to 179.70): E






#72
 Carbazole
 Concen: 41.22 ng
 RT: 19.36 min Scan# 1599
 Delta R.T. -0.06 min
 Lab File: BF081698.D
 Acq: 18 Sep 2015 15:34

Tgt Ion:167 Resp: 380647
 Ion Ratio Lower Upper
 167 100
 166 21.1 15.7 23.5
 139 13.5 9.5 14.3



Abundance Ion 167.00 (166.70 to 167.70): E
 Ion 166.00 (165.70 to 166.70): E
 Ion 139.00 (138.70 to 139.70): E









#73
Di-n-butylphthalate
Concen: 47.68 ng
RT: 20.41 min Scan# 1691
Delta R.T. -0.06 min
Lab File: BF081698.D
Acq: 18 Sep 2015 15:34




Instrument :
BNA_F
ClientSampleId :
PB85485BS


Tgt Ion:149 Resp: 519933
Ion Ratio Lower Upper
149 100
150 8.8 7.4 11.2
104 5.1 2.4 3.6#



#74
Fluoranthene
Concen: 43.27 ng
RT: 21.39 min Scan# 1776
Delta R.T. -0.04 min
Lab File: BF081698.D
Acq: 18 Sep 2015 15:34

Tgt Ion:202 Resp: 448597
Ion Ratio Lower Upper
202 100
101 9.8 0.0 38.3
203 17.4 0.0 37.3





#75

Chrysene-d12

Concen: 20.00 ng

RT: 23.35 min Scan# 1948

Delta R.T. -0.03 min

Lab File: BF081698.D

Acq: 18 Sep 2015 15:34

Instrument :

BNA_F

ClientSampleId :

PB85485BS


Tgt Ion:240 Resp: 145811

Ion Ratio Lower Upper

240 100

120 10.0 16.2 24.2#




236 26.1 18.4 27.6



Abundance Ion 240.00 (239.70 to 240.70): E

Ion 120.00 (119.70 to 120.70): E

Ion 236.00 (235.70 to 236.70): E



#76

Benzidine

Concen: 30.26 ng

RT: 21.71 min Scan# 1804

Delta R.T. -0.03 min

Lab File: BF081698.D

Acq: 18 Sep 2015 15:34


Tgt Ion:184 Resp: 189394

Ion Ratio Lower Upper

184 100

185 13.7 11.8 17.6



183 10.5 7.6 11.4




Abundance Ion 184.00 (183.70 to 184.70): E

Ion 185.00 (184.70 to 185.70): E

Ion 183.00 (182.70 to 183.70): E









#77
Pyrene
 Concen: 42.65 ng
 RT: 21.74 min Scan# 1807
 Delta R.T. -0.04 min
 Lab File: BF081698.D
 Acq: 18 Sep 2015 15:34

Instrument :
 BNA_F
 ClientSampleId :
 PB85485BS

Tgt Ion	Ratio	Lower	Upper
202	100		
200	20.4	15.0	22.4
203	18.0	14.1	21.1




Abundance Ion 202.00 (201.70 to 202.70): E
 Ion 200.00 (199.70 to 200.70): E
 Ion 203.00 (202.70 to 203.70): E






#78
Terphenyl-d14
 Concen: 84.89 ng
 RT: 22.07 min Scan# 1836
 Delta R.T. -0.04 min
 Lab File: BF081698.D
 Acq: 18 Sep 2015 15:34

Tgt Ion	Ratio	Lower	Upper
244	100		
212	7.7	4.3	6.5#
122	13.1	17.4	26.2#



Abundance Ion 244.00 (243.70 to 244.70): E
 Ion 212.00 (211.70 to 212.70): E
 Ion 122.00 (121.70 to 122.70): E





#79

Butylbenzylphthalate

Concen: 44.65 ng

RT: 22.77 min Scan# 1897

Delta R.T. -0.03 min

Lab File: BF081698.D

Acq: 18 Sep 2015 15:34

Instrument :

BNA_F

ClientSampleId :

PB85485BS


Tgt Ion:149 Resp: 209726

Ion Ratio Lower Upper

149 100

91 81.5 48.6 72.8#




206 15.1 14.9 22.3



Abundance Ion 149.00 (148.70 to 149.70): E

Ion 91.00 (90.70 to 91.70): BF0

Ion 206.00 (205.70 to 206.70): E



#80

Benzo(a)anthracene

Concen: 40.03 ng

RT: 23.34 min Scan# 1947

Delta R.T. -0.03 min

Lab File: BF081698.D

Acq: 18 Sep 2015 15:34


Tgt Ion:228 Resp: 371683

Ion Ratio Lower Upper

228 100

226 27.7 20.0 30.0



229 20.1 15.4 23.2




Abundance Ion 228.00 (227.70 to 228.70): E

Ion 226.00 (225.70 to 226.70): E

Ion 229.00 (228.70 to 229.70): E









#81
 3,3'-Dichlorobenzidine
 Concen: 27.55 ng
 RT: 23.35 min Scan# 1948
 Delta R.T. -0.03 min
 Lab File: BF081698.D
 Acq: 18 Sep 2015 15:34

Instrument :
 BNA_F
 ClientSampleId :
 PB85485BS

Tgt Ion	Ratio	Lower	Upper
252	100		
254	64.9	51.4	77.2
126	8.2	16.4	24.6




Abundance Ion 252.00 (251.70 to 252.70): E
 Ion 254.00 (253.70 to 254.70): E
 Ion 126.00 (125.70 to 126.70): E






#82
 Chrysene
 Concen: 43.17 ng
 RT: 23.39 min Scan# 1951
 Delta R.T. -0.03 min
 Lab File: BF081698.D
 Acq: 18 Sep 2015 15:34

Tgt Ion	Ratio	Lower	Upper
228	100		
226	30.8	21.0	31.4
229	19.8	15.6	23.4



Abundance Ion 228.00 (227.70 to 228.70): E
 Ion 226.00 (225.70 to 226.70): E
 Ion 229.00 (228.70 to 229.70): E





#83

Bis(2-ethylhexyl)phthalate

Concen: 48.39 ng

RT: 23.47 min Scan# 1958

Delta R.T. -0.03 min

Lab File: BF081698.D

Acq: 18 Sep 2015 15:34

Instrument :

BNA_F

ClientSampleId :

PB85485BS


Tgt Ion:149 Resp: 299959

Ion Ratio Lower Upper

149 100

167 25.1 24.2 36.2




279 2.6 3.8 5.6#



Abundance Ion 149.00 (148.70 to 149.70): E

Ion 167.00 (166.70 to 167.70): E

Ion 279.00 (278.70 to 279.70): E



#84

Di-n-octyl phthalate

Concen: 43.73 ng

RT: 24.12 min Scan# 2015

Delta R.T. -0.03 min

Lab File: BF081698.D

Acq: 18 Sep 2015 15:34


Tgt Ion:149 Resp: 446374

Ion Ratio Lower Upper

149 100

167 1.4 1.0 1.6



43 14.7 10.2 15.2




Abundance Ion 149.00 (148.70 to 149.70): E

Ion 167.00 (166.70 to 167.70): E

Ion 43.00 (42.70 to 43.70): BF0









#85
 Indeno(1,2,3-cd)pyrene
 Concen: 37.48 ng
 RT: 25.82 min Scan# 2164
 Delta R.T. -0.06 min
 Lab File: BF081698.D
 Acq: 18 Sep 2015 15:34




Instrument :
 BNA_F
 ClientSampleId :
 PB85485BS


Tgt Ion	Ratio	Lower	Upper
276	100		
138	20.9	25.7	38.5#
277	24.9	15.6	23.4#



#86
 Perylene-d12
 Concen: 20.00 ng
 RT: 24.79 min Scan# 2074
 Delta R.T. -0.02 min
 Lab File: BF081698.D
 Acq: 18 Sep 2015 15:34

Tgt Ion	Ratio	Lower	Upper
264	100		
260	23.1	16.7	25.1
265	20.9	17.2	25.8





#87

Benzo(b)fluoranthene

Concen: 77.99 ng

RT: 24.45 min Scan# 2044

Delta R.T. -0.03 min

Lab File: BF081698.D

Acq: 18 Sep 2015 15:34





Instrument :

BNA_F

ClientSampleId :

PB85485BS

Tgt Ion:	252	Resp:	545857
Ion Ratio	Lower	Upper	
252	100		
253	21.8	17.5	26.3
125	8.3	17.1	25.7#



#88

Benzo(k)fluoranthene

Concen: 93.99 ng




RT: 24.45 min Scan# 2044


Delta R.T. -0.06 min

Lab File: BF081698.D

Acq: 18 Sep 2015 15:34

Tgt Ion:	252	Resp:	545857
Ion Ratio	Lower	Upper	
252	100		
253	21.8	17.0	25.4
125	8.3	16.0	24.0#





#89

Benzo(a)pyrene

Concen: 43.04 ng

RT: 24.75 min Scan# 2070

Delta R.T. -0.02 min

Lab File: BF081698.D

Acq: 18 Sep 2015 15:34

Instrument :

BNA_F

ClientSampleId :

PB85485BS


Tgt Ion:252 Resp: 255295

Ion Ratio Lower Upper

252 100

253 22.0 17.2 25.8




125 4.8 18.0 27.0#



Abundance Ion 252.00 (251.70 to 252.70): E

Ion 253.00 (252.70 to 253.70): E

Ion 125.00 (124.70 to 125.70): E



#90

Dibenzo(a,h)anthracene

Concen: 42.26 ng

RT: 25.83 min Scan# 2165

Delta R.T. -0.06 min

Lab File: BF081698.D

Acq: 18 Sep 2015 15:34


Tgt Ion:278 Resp: 193683

Ion Ratio Lower Upper

278 100

139 14.1 26.3 39.5#



279 23.2 18.2 27.4






Abundance Ion 278.00 (277.70 to 278.70): E

Ion 139.00 (138.70 to 139.70): E

Ion 279.00 (278.70 to 279.70): E





#91

Benzo(g,h,i)perylene

Concen: 41.64 ng

RT: 26.13 min Scan# 2191

Delta R.T. -0.04 min

Lab File: BF081698.D

Acq: 18 Sep 2015 15:34

Instrument :

BNA_F

ClientSampleId :

PB85485BS

Tgt Ion: 276 Resp: 182143

Ion Ratio Lower Upper

276 100

277 23.6 17.8 26.6

138 13.5 34.6 51.8#

