

Data Path : Z:\HPCHEM1\BNA_F\DATA\BF091815\
 Data File : BF081708.D
 Acq On : 18 Sep 2015 21:46
 Operator : UM/IZ
 Sample : G3704-13MSD
 Misc :
 ALS Vial : 17 Sample Multiplier: 1

Instrument :
 BNA_F
 ClientSampleId :
 MW-12-9-15-15MSD

Quant Time: Sep 18 23:03:35 2015
 Quant Method : Z:\HPCHEM1\BNA_F\METHODS\8270-BF090115.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Tue Sep 15 14:03:59 2015
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Dichlorobenzene-d4	8.16	152	41276	20.00	ng	-0.06
21) Naphthalene-d8	11.06	136	167974	20.00	ng	-0.06
38) Acenaphthene-d10	15.20	164	78127	20.00	ng	-0.06
63) Phenanthrene-d10	18.70	188	162732	20.00	ng	-0.06
75) Chrysene-d12	23.35	240	140763	20.00	ng	-0.03
86) Perylene-d12	24.78	264	91862	20.00	ng	-0.03

System Monitoring Compounds

5) 2-Fluorophenol	5.32	112	158984	59.76	ng	-0.03
7) Phenol-d6	7.59	99	130205	36.97	ng	-0.06
23) Nitrobenzene-d5	9.48	82	368891	105.96	ng	-0.06
41) 2,4,6-Tribromophenol	17.11	330	115580	166.15	ng	-0.06
44) 2-Fluorobiphenyl	13.72	172	651627	106.88	ng	-0.06
78) Terphenyl-d14	22.07	244	590606	97.67	ng	-0.04

Target Compounds

						Qvalue
2) 1,4-Dioxane	1.57	88	35107	29.38	ng	# 81
3) Pyridine	2.12	79	29575	8.98	ng	# 75
4) n-Nitrosodimethylamine	2.05	42	36523	19.95	ng	# 59
6) Aniline	7.48	93	79359	15.96	ng	# 84
8) 2-Chlorophenol	7.72	128	110675	37.75	ng	# 79
9) Benzaldehyde	7.19	77	56891	25.78	ng	# 73
10) Phenol	7.61	94	40556	9.93	ng	# 44
11) bis(2-Chloroethyl)ether	7.70	93	139939	47.50	ng	# 80
12) 1,3-Dichlorobenzene	8.01	146	129007	41.19	ng	# 87
13) 1,4-Dichlorobenzene	8.21	146	132552	41.56	ng	# 90
14) 1,2-Dichlorobenzene	8.53	146	129942	45.25	ng	# 91
15) Benzyl Alcohol	8.61	79	80020	27.70	ng	# 68
16) 2,2'-oxybis(1-Chloropropan	8.93	45	298359	52.83	ng	95
17) 2-Methylphenol	8.96	107	71521	30.58	ng	# 79
18) Hexachloroethane	9.26	117	54343	43.80	ng	88
19) n-Nitroso-di-n-propylamine	9.25	70	119815	49.99	ng	# 89
20) 3+4-Methylphenols	9.34	107	80691	25.59	ng	# 80
22) Acetophenone	9.16	105	213267	46.48	ng	# 73
24) Nitrobenzene	9.51	77	177671	49.14	ng	# 61
25) Isophorone	10.10	82	309912	47.86	ng	# 84
26) 2-Nitrophenol	10.24	139	68720	43.61	ng	# 21
27) 2,4-Dimethylphenol	10.52	122	107240	39.40	ng	# 74
28) bis(2-Chloroethoxy)methane	10.71	93	182402	48.76	ng	# 91
29) 2,4-Dichlorophenol	10.86	162	102834	43.57	ng	93
30) 1,2,4-Trichlorobenzene	10.97	180	116333	45.37	ng	# 97
31) Naphthalene	11.12	128	409136	45.67	ng	99
32) Benzoic acid	11.12	122	2010	3.52	ng	# 1
33) 4-Chloroaniline	11.36	127	62375	17.07	ng	# 82
34) Hexachlorobutadiene	11.46	225	73298	46.97	ng	99
35) Caprolactam	12.29	113	2767	3.82	ng	# 28
36) 4-Chloro-3-methylphenol	12.66	107	116662	44.16	ng	# 76
37) 2-Methylnaphthalene	12.77	142	285718	49.01	ng	# 91
39) 1,2,4,5-Tetrachlorobenzene	13.18	216	122871	49.73	ng	96
40) Hexachlorocyclopentadiene	13.16	237	116100	95.75	ng	97


Data Path : Z:\HPCHEM1\BNA_F\DATA\BF091815\
 Data File : BF081708.D
 Acq On : 18 Sep 2015 21:46
 Operator : UM/IZ
 Sample : G3704-13MSD
 Misc :
 ALS Vial : 17 Sample Multiplier: 1

Instrument :
 BNA_F
 ClientSampleId :
 MW-12-9-15-15MSD

Quant Time: Sep 18 23:03:35 2015
 Quant Method : Z:\HPCHEM1\BNA_F\METHODS\8270-BF090115.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Tue Sep 15 14:03:59 2015
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
42) 2,4,6-Trichlorophenol	13.52	196	78115	45.81	ng	98
43) 2,4,5-Trichlorophenol	13.64	196	81544	46.87	ng #	83
45) 1,1'-Biphenyl	13.91	154	348908	48.54	ng	96
46) 2-Chloronaphthalene	13.90	162	258155	49.73	ng #	87
47) 2-Nitroaniline	14.25	65	110026	52.01	ng #	59
48) Acenaphthylene	14.86	152	443007	48.11	ng	97
49) Dimethylphthalate	14.80	163	337564	55.61	ng #	97
50) 2,6-Dinitrotoluene	14.89	165	63690	46.23	ng #	21
51) Acenaphthene	15.28	154	253424	48.38	ng	90
52) 3-Nitroaniline	15.25	138	36098	23.02	ng #	47
53) 2,4-Dinitrophenol	15.55	184	61174	86.40	ng #	24
54) Dibenzofuran	15.71	168	398079	52.04	ng #	85
55) 4-Nitrophenol	15.90	139	24078	24.44	ng #	3
56) 2,4-Dinitrotoluene	15.85	165	91925	52.40	ng #	67
57) Fluorene	16.52	166	312183	53.78	ng	96
58) 2,3,4,6-Tetrachlorophenol	16.07	232	70329	52.95	ng	93
59) Diethylphthalate	16.49	149	347855	54.48	ng	95
60) 4-Chlorophenyl-phenylether	16.61	204	148743	54.98	ng #	86
61) 4-Nitroaniline	16.72	138	61504	38.81	ng #	13
62) Azobenzene	16.97	77	394049	55.51	ng	81
64) 4,6-Dinitro-2-methylphenol	16.81	198	43332	47.61	ng #	74
65) n-Nitrosodiphenylamine	16.93	169	266797	45.06	ng	92
66) 4-Bromophenyl-phenylether	17.75	248	83735	50.89	ng #	91
67) Hexachlorobenzene	17.80	284	84866	49.33	ng #	82
68) Atrazine	18.35	200	95471	55.71	ng #	78
69) Pentachlorophenol	18.35	266	65593	78.50	ng	98
70) Phenanthrene	18.77	178	446931	49.40	ng	97
71) Anthracene	18.88	178	452901	48.73	ng	98
72) Carbazole	19.36	167	418843	48.02	ng	96
73) Di-n-butylphthalate	20.41	149	560667	54.44	ng #	98
74) Fluoranthene	21.39	202	488465	49.88	ng	89
76) Benzidine	21.71	184	86104	14.25	ng #	97
77) Pyrene	21.74	202	505642	48.49	ng	99
79) Butylbenzylphthalate	22.77	149	229205	50.54	ng #	77
80) Benzo(a)anthracene	23.34	228	407841	45.50	ng	96
81) 3,3'-Dichlorobenzidine	23.35	252	82401	27.25	ng #	93
82) Chrysene	23.39	228	392185	48.81	ng	95
83) Bis(2-ethylhexyl)phthalate	23.47	149	333041	55.65	ng #	93
84) Di-n-octyl phthalate	24.12	149	510480	51.81	ng	96
85) Indeno(1,2,3-cd)pyrene	25.82	276	266621	45.22	ng #	84
87) Benzo(b)fluoranthene	24.45	252	610595	93.10	ng #	86
88) Benzo(k)fluoranthene	24.45	252	610595	112.20	ng #	87
89) Benzo(a)pyrene	24.73	252	289323	52.06	ng #	88
90) Dibenzo(a,h)anthracene	25.83	278	224932	52.38	ng #	79
91) Benzo(g,h,i)perylene	26.12	276	216193	52.75	ng #	76





(#) = qualifier out of range (m) = manual integration (+) = signals summed



#1
 1,4-Dichlorobenzene-d4
 Concen: 20.00 ng
 RT: 8.16 min Scan# 619
 Delta R.T. -0.06 min
 Lab File: BF081708.D
 Acq: 18 Sep 2015 21:46




Instrument :
 BNA_F
 ClientSampleId :
 MW-12-9-15-15MSD


Tgt Ion:152 Resp: 41276
 Ion Ratio Lower Upper
 152 100
 150 160.5 124.7 187.1
 115 69.3 39.0 58.4#



#2
 1,4-Dioxane
 Concen: 29.38 ng
 RT: 1.57 min Scan# 42
 Delta R.T. 0.00 min
 Lab File: BF081708.D
 Acq: 18 Sep 2015 21:46

Tgt Ion: 88 Resp: 35107
 Ion Ratio Lower Upper
 88 100
 58 92.3 77.7 116.5
 43 0.0 26.6 40.0#





#3

Pyridine

Concen: 8.98 ng

RT: 2.12 min Scan# 90

Delta R.T. 0.04 min

Lab File: BF081708.D

Acq: 18 Sep 2015 21:46

Instrument :

BNA_F

ClientSampleId :

MW-12-9-15-15MSD





Tgt Ion: 79 Resp: 29575

Ion Ratio Lower Upper

79 100

52 116.1 76.8 115.2#

51 62.7 32.2 48.4#



#4

n-Nitrosodimethylamine

Concen: 19.95 ng

RT: 2.05 min Scan# 84

Delta R.T. -0.01 min

Lab File: BF081708.D

Acq: 18 Sep 2015 21:46




Tgt Ion: 42 Resp: 36523


Ion Ratio Lower Upper

42 100

74 81.2 105.0 157.6#

44 6.4 6.6 10.0#





#5

2-Fluorophenol

Concen: 59.76 ng

RT: 5.32 min Scan# 370

Delta R.T. -0.03 min

Lab File: BF081708.D

Acq: 18 Sep 2015 21:46


Instrument :

BNA_F

ClientSampleId :

MW-12-9-15-15MSD

Tgt Ion:	112	Resp:	158984
Ion	Ratio	Lower	Upper
112	100		
64	79.4	39.7	59.5#
63	43.1	17.4	26.0#





Abundance Ion 112.00 (111.70 to 112.70): E

Ion 64.00 (63.70 to 64.70): BF0

Ion 63.00 (62.70 to 63.70): BF0

5.32

Time-->



#6

Aniline

Concen: 15.96 ng


RT: 7.48 min Scan# 559

Delta R.T. -0.04 min

Lab File: BF081708.D

Acq: 18 Sep 2015 21:46

Tgt Ion:	93	Resp:	79359
Ion	Ratio	Lower	Upper
93	100		
66	44.8	27.2	40.8#
65	23.2	14.7	22.1#




Abundance Ion 93.00 (92.70 to 93.70): BF0


Ion 66.00 (65.70 to 66.70): BF0

Ion 65.00 (64.70 to 65.70): BF0

7.48

Time-->





#7

Phenol-d6

Concen: 36.97 ng

RT: 7.59 min Scan# 569

Delta R.T. -0.06 min

Lab File: BF081708.D

Acq: 18 Sep 2015 21:46

Instrument :

BNA_F

ClientSampleId :

MW-12-9-15-15MSD


Tgt Ion: 99 Resp: 130205

Ion Ratio Lower Upper




99 100

42 26.5 13.7 20.5#

71 39.7 14.2 21.2#



Abundance Ion 99.00 (98.70 to 99.70): BF0
Ion 42.00 (41.70 to 42.70): BF0
Ion 71.00 (70.70 to 71.70): BF0



#8

2-Chlorophenol

Concen: 37.75 ng

RT: 7.72 min Scan# 580

Delta R.T. -0.06 min

Lab File: BF081708.D

Acq: 18 Sep 2015 21:46


Tgt Ion: 128 Resp: 110675

Ion Ratio Lower Upper



128 100


130 32.6 12.2 52.2

64 63.6 20.5 60.5#



Abundance Ion 128.00 (127.70 to 128.70): E
Ion 130.00 (129.70 to 130.70): E
Ion 64.00 (63.70 to 64.70): BF0





#9

Benzaldehyde

Concen: 25.78 ng

RT: 7.19 min Scan# 534

Delta R.T. -0.04 min

Lab File: BF081708.D

Acq: 18 Sep 2015 21:46

Instrument :

BNA_F

ClientSampleId :

MW-12-9-15-15MSD





Tgt Ion: 77 Resp: 56891

Ion Ratio Lower Upper

77 100

105 90.7 91.6 131.6#

106 85.5 102.6 142.6#



#10

Phenol

Concen: 9.93 ng

RT: 7.61 min Scan# 571

Delta R.T. -0.06 min

Lab File: BF081708.D

Acq: 18 Sep 2015 21:46




Tgt Ion: 94 Resp: 40556


Ion Ratio Lower Upper

94 100

65 37.8 0.7 40.7

66 56.1 0.8 40.8#









#11
bis(2-Chloroethyl)ether
Concen: 47.50 ng
RT: 7.70 min Scan# 579
Delta R.T. -0.06 min
Lab File: BF081708.D
Acq: 18 Sep 2015 21:46




Instrument :
BNA_F
ClientSampleId :
MW-12-9-15-15MSD


Tgt Ion: 93 Resp: 139939
Ion Ratio Lower Upper
93 100
63 110.7 65.6 105.6#
95 32.1 13.6 53.6



#12
1,3-Dichlorobenzene
Concen: 41.19 ng
RT: 8.01 min Scan# 606
Delta R.T. -0.06 min
Lab File: BF081708.D
Acq: 18 Sep 2015 21:46

Tgt Ion: 146 Resp: 129007
Ion Ratio Lower Upper
146 100
148 63.0 51.0 76.4
75 42.4 15.0 22.6#





#13

1,4-Dichlorobenzene

Concen: 41.56 ng

RT: 8.21 min Scan# 623

Delta R.T. -0.04 min

Lab File: BF081708.D

Acq: 18 Sep 2015 21:46

Instrument :

BNA_F

ClientSampleId :

MW-12-9-15-15MSD


Tgt Ion:146 Resp: 132552

Ion Ratio Lower Upper

146 100

148 63.7 51.8 77.6




111 45.9 24.9 37.3#



Abundance Ion 146.00 (145.70 to 146.70): E

Ion 148.00 (147.70 to 148.70): E

Ion 111.00 (110.70 to 111.70): E



#14

1,2-Dichlorobenzene

Concen: 45.25 ng

RT: 8.53 min Scan# 651

Delta R.T. -0.04 min

Lab File: BF081708.D

Acq: 18 Sep 2015 21:46


Tgt Ion:146 Resp: 129942

Ion Ratio Lower Upper

146 100

148 63.9 52.7 79.1



111 47.8 28.8 43.2#




Abundance Ion 146.00 (145.70 to 146.70): E

Ion 148.00 (147.70 to 148.70): E

Ion 111.00 (110.70 to 111.70): E





#15

Benzyl Alcohol

Concen: 27.70 ng

RT: 8.61 min Scan# 658

Delta R.T. -0.06 min

Lab File: BF081708.D

Acq: 18 Sep 2015 21:46

Instrument :

BNA_F

ClientSampleId :

MW-12-9-15-15MSD





Tgt Ion: 79 Resp: 80020

Ion Ratio Lower Upper

79 100

108 63.7 88.6 132.8#

77 64.3 47.0 70.4



#16

2,2'-oxybis(1-Chloropropane)

Concen: 52.83 ng

RT: 8.93 min Scan# 686

Delta R.T. -0.06 min

Lab File: BF081708.D

Acq: 18 Sep 2015 21:46




Tgt Ion: 45 Resp: 298359


Ion Ratio Lower Upper

45 100

77 10.1 0.0 28.1

79 7.8 0.0 26.0









#17
2-Methylphenol
Concen: 30.58 ng
RT: 8.96 min Scan# 689
Delta R.T. -0.04 min
Lab File: BF081708.D
Acq: 18 Sep 2015 21:46

Instrument :
BNA_F
ClientSampleId :
MW-12-9-15-15MSD

Tgt Ion:107 Resp: 71521
Ion Ratio Lower Upper
107 100
108 113.8 96.6 145.0
77 57.4 23.3 34.9#
79 55.4 22.0 33.0#




Abundance Ion 107.00 (106.70 to 107.70): E
150000
Ion 108.00 (107.70 to 108.70): E
Ion 77.00 (76.70 to 77.70): BF0
Ion 79.00 (78.70 to 79.70): BF0








#18
Hexachloroethane
Concen: 43.80 ng
RT: 9.26 min Scan# 715
Delta R.T. -0.06 min
Lab File: BF081708.D
Acq: 18 Sep 2015 21:46

Tgt Ion:117 Resp: 54343
Ion Ratio Lower Upper
117 100
119 91.1 83.8 125.6
201 76.9 55.1 82.7



Abundance Ion 117.00 (116.70 to 117.70): E
40000
Ion 119.00 (118.70 to 119.70): E
Ion 201.00 (200.70 to 201.70): E





#19

n-Nitroso-di-n-propylamine

Concen: 49.99 ng

RT: 9.25 min Scan# 714

Delta R.T. -0.04 min

Lab File: BF081708.D

Acq: 18 Sep 2015 21:46

Instrument :

BNA_F

ClientSampleId :

MW-12-9-15-15MSD

Tgt Ion: 70 Resp: 119815





Ion Ratio Lower Upper

70 100

42 74.8 63.8 95.6

101 9.6 9.2 13.8

130 18.5 27.3 40.9#



#20

3+4-Methylphenols

Concen: 25.59 ng

RT: 9.34 min Scan# 722

Delta R.T. -0.04 min

Lab File: BF081708.D

Acq: 18 Sep 2015 21:46

Tgt Ion: 107 Resp: 80691


Ion Ratio Lower Upper


107 100

108 83.4 74.6 114.6

77 41.6 0.7 40.7#

79 33.2 0.0 38.9









#21
Naphthalene-d8
Concen: 20.00 ng
RT: 11.06 min Scan# 873
Delta R.T. -0.06 min
Lab File: BF081708.D
Acq: 18 Sep 2015 21:46

Instrument :
BNA_F
ClientSampleId :
MW-12-9-15-15MSD

Tgt Ion:	136	Resp:	167974
Ion	Ratio	Lower	Upper
136	100		
137	10.6	9.0	13.6
54	12.2	8.2	12.2
68	5.0	5.8	8.6#




Abundance Ion 136.00 (135.70 to 136.70): E
Ion 137.00 (136.70 to 137.70): E
Ion 54.00 (53.70 to 54.70): BF0
Ion 68.00 (67.70 to 68.70): BF0






#22
Acetophenone
Concen: 46.48 ng
RT: 9.16 min Scan# 706
Delta R.T. -0.06 min
Lab File: BF081708.D
Acq: 18 Sep 2015 21:46

Tgt Ion:	105	Resp:	213267
Ion	Ratio	Lower	Upper
105	100		
71	0.0	4.7	7.1#
51	40.3	22.0	33.0#
120	18.9	30.1	45.1#



Abundance Ion 105.00 (104.70 to 105.70): E
Ion 71.00 (70.70 to 71.70): BF0
Ion 51.00 (50.70 to 51.70): BF0
Ion 120.00 (119.70 to 120.70): E









#23
 Nitrobenzene-d5
 Concen: 105.96 ng
 RT: 9.48 min Scan# 734
 Delta R.T. -0.06 min
 Lab File: BF081708.D
 Acq: 18 Sep 2015 21:46




Instrument :
 BNA_F
 ClientSampleId :
 MW-12-9-15-15MSD


Tgt Ion: 82 Resp: 368891
 Ion Ratio Lower Upper
 82 100
 128 36.9 51.0 76.6#
 54 69.2 47.5 71.3



#24
 Nitrobenzene
 Concen: 49.14 ng
 RT: 9.51 min Scan# 737
 Delta R.T. -0.06 min
 Lab File: BF081708.D
 Acq: 18 Sep 2015 21:46

Tgt Ion: 77 Resp: 177671
 Ion Ratio Lower Upper
 77 100
 123 37.4 59.8 89.8#
 65 16.1 10.2 15.4#





#25

Isophorone

Concen: 47.86 ng

RT: 10.10 min Scan# 789

Delta R.T. -0.06 min

Lab File: BF081708.D

Acq: 18 Sep 2015 21:46

Instrument :

BNA_F

ClientSampleId :

MW-12-9-15-15MSD


Tgt Ion: 82 Resp: 309912

Ion Ratio Lower Upper




82 100

95 7.7 4.0 6.0#

138 14.2 18.3 27.5#



Abundance Ion 82.00 (81.70 to 82.70): BF0
Ion 95.00 (94.70 to 95.70): BF0
Ion 138.00 (137.70 to 138.70): BF0



#26

2-Nitrophenol

Concen: 43.61 ng

RT: 10.24 min Scan# 801

Delta R.T. -0.06 min

Lab File: BF081708.D

Acq: 18 Sep 2015 21:46


Tgt Ion: 139 Resp: 68720

Ion Ratio Lower Upper



139 100


109 39.2 11.0 16.4#

65 78.4 24.7 37.1#



Abundance Ion 139.00 (138.70 to 139.70): E
Ion 109.00 (108.70 to 109.70): E
Ion 65.00 (64.70 to 65.70): BF0









#27
 2,4-Dimethylphenol
 Concen: 39.40 ng
 RT: 10.52 min Scan# 825
 Delta R.T. -0.06 min
 Lab File: BF081708.D
 Acq: 18 Sep 2015 21:46




Instrument :
 BNA_F
 ClientSampleId :
 MW-12-9-15-15MSD


Tgt Ion:122 Resp: 107240
 Ion Ratio Lower Upper
 122 100
 107 125.7 71.8 107.6#
 121 56.5 42.3 63.5



#28
 bis(2-Chloroethoxy)methane
 Concen: 48.76 ng
 RT: 10.71 min Scan# 842
 Delta R.T. -0.04 min
 Lab File: BF081708.D
 Acq: 18 Sep 2015 21:46

Tgt Ion: 93 Resp: 182402
 Ion Ratio Lower Upper
 93 100
 95 31.8 27.5 41.3
 123 9.0 13.7 20.5#









#29
 2,4-Dichlorophenol
 Concen: 43.57 ng
 RT: 10.86 min Scan# 855
 Delta R.T. -0.04 min
 Lab File: BF081708.D
 Acq: 18 Sep 2015 21:46

Instrument :
 BNA_F
 ClientSampleId :
 MW-12-9-15-15MSD

Tgt Ion	Ratio	Lower	Upper
162	100		
164	64.3	44.9	84.9
98	43.1	13.0	53.0




Abundance Ion 162.00 (161.70 to 162.70): E
 Ion 164.00 (163.70 to 164.70): E
 Ion 98.00 (97.70 to 98.70): BF0






#30
 1,2,4-Trichlorobenzene
 Concen: 45.37 ng
 RT: 10.97 min Scan# 865
 Delta R.T. -0.06 min
 Lab File: BF081708.D
 Acq: 18 Sep 2015 21:46

Tgt Ion	Ratio	Lower	Upper
180	100		
182	95.9	77.1	115.7
145	35.0	23.3	34.9#



Abundance Ion 180.00 (179.70 to 180.70): E
 Ion 182.00 (181.70 to 182.70): E
 Ion 145.00 (144.70 to 145.70): E





#31

Naphthalene

Concen: 45.67 ng

RT: 11.12 min Scan# 878

Delta R.T. -0.04 min

Lab File: BF081708.D

Acq: 18 Sep 2015 21:46

Instrument :

BNA_F

ClientSampleId :

MW-12-9-15-15MSD


Tgt Ion:128 Resp: 409136

Ion Ratio Lower Upper

128 100

129 10.7 8.4 12.6




127 12.7 9.9 14.9



Abundance Ion 128.00 (127.70 to 128.70): E

Ion 129.00 (128.70 to 129.70): E

Ion 127.00 (126.70 to 127.70): E



#32

Benzoic acid

Concen: 3.52 ng

RT: 11.12 min Scan# 878

Delta R.T. 0.04 min

Lab File: BF081708.D

Acq: 18 Sep 2015 21:46


Tgt Ion:122 Resp: 2010

Ion Ratio Lower Upper

122 100

105 46.0 76.1 116.1#



77 1253.8 40.5 80.5#




Abundance Ion 122.00 (121.70 to 122.70): E

Ion 105.00 (104.70 to 105.70): E

Ion 77.00 (76.70 to 77.70): BF0









#33
4-Chloroaniline
Concen: 17.07 ng
RT: 11.36 min Scan# 899
Delta R.T. -0.04 min
Lab File: BF081708.D
Acq: 18 Sep 2015 21:46




Instrument :
BNA_F
ClientSampleId :
MW-12-9-15-15MSD


Tgt Ion:	127	Resp:	62375
Ion	Ratio	Lower	Upper
127	100		
129	32.8	25.8	38.6
65	41.7	17.9	26.9#
92	21.7	10.5	15.7#



#34
Hexachlorobutadiene
Concen: 46.97 ng
RT: 11.46 min Scan# 908
Delta R.T. -0.06 min
Lab File: BF081708.D
Acq: 18 Sep 2015 21:46

Tgt Ion:	225	Resp:	73298
Ion	Ratio	Lower	Upper
225	100		
223	63.1	50.2	75.2
227	64.3	52.2	78.2





#35

Caprolactam

Concen: 3.82 ng

RT: 12.29 min Scan# 980

Delta R.T. -0.10 min

Lab File: BF081708.D

Acq: 18 Sep 2015 21:46

Instrument :

BNA_F

ClientSampleId :

MW-12-9-15-15MSD


Tgt Ion:113 Resp: 2767

Ion Ratio Lower Upper

113 100

55 265.0 152.4 192.4#




56 213.5 104.6 144.6#



Abundance Ion 113.00 (112.70 to 113.70): E

Ion 55.00 (54.70 to 55.70): BF0

Ion 56.00 (55.70 to 56.70): BF0



#36

4-Chloro-3-methylphenol

Concen: 44.16 ng

RT: 12.66 min Scan# 1013

Delta R.T. -0.06 min

Lab File: BF081708.D

Acq: 18 Sep 2015 21:46


Tgt Ion:107 Resp: 116662

Ion Ratio Lower Upper

107 100

144 22.5 25.4 38.0#



142 70.5 77.3 115.9#




Abundance Ion 107.00 (106.70 to 107.70): E

Ion 144.00 (143.70 to 144.70): E

Ion 142.00 (141.70 to 142.70): E









#37
2-Methylnaphthalene
Concen: 49.01 ng
RT: 12.77 min Scan# 1022
Delta R.T. -0.06 min
Lab File: BF081708.D
Acq: 18 Sep 2015 21:46




Instrument :
BNA_F
ClientSampleId :
MW-12-9-15-15MSD


Tgt Ion:142 Resp: 285718
Ion Ratio Lower Upper
142 100
141 86.8 67.5 101.3
115 39.6 18.2 27.2#



#38
Acenaphthene-d10
Concen: 20.00 ng
RT: 15.20 min Scan# 1235
Delta R.T. -0.06 min
Lab File: BF081708.D
Acq: 18 Sep 2015 21:46

Tgt Ion:164 Resp: 78127
Ion Ratio Lower Upper
164 100
162 105.2 75.2 112.8
160 46.9 31.5 47.3





#39

1,2,4,5-Tetrachlorobenzene

Concen: 49.73 ng

RT: 13.18 min Scan# 1058

Delta R.T. -0.04 min

Lab File: BF081708.D

Acq: 18 Sep 2015 21:46

Instrument :

BNA_F

ClientSampleId :

MW-12-9-15-15MSD

Tgt Ion:216 Resp: 122871


Ion Ratio Lower Upper

216 100

214 76.9 63.5 95.3

179 22.9 16.6 24.8

108 23.2 16.3 24.5






Abundance Ion 216.00 (215.70 to 216.70): E

Ion 214.00 (213.70 to 214.70): E

Ion 179.00 (178.70 to 179.70): E

Ion 108.00 (107.70 to 108.70): E



#40

Hexachlorocyclopentadiene

Concen: 95.75 ng

RT: 13.16 min Scan# 1056

Delta R.T. -0.06 min

Lab File: BF081708.D

Acq: 18 Sep 2015 21:46


Tgt Ion:237 Resp: 116100

Ion Ratio Lower Upper

237 100

235 61.8 42.0 82.0


272 10.6 0.0 36.1




Abundance Ion 236.80 (236.50 to 237.50): E


Ion 234.90 (234.60 to 235.60): E

Ion 271.75 (271.45 to 272.45): E



Time-->









#41
 2,4,6-Tribromophenol
 Concen: 166.15 ng
 RT: 17.11 min Scan# 1402
 Delta R.T. -0.06 min
 Lab File: BF081708.D
 Acq: 18 Sep 2015 21:46




Instrument :
 BNA_F
 ClientSampleId :
 MW-12-9-15-15MSD


Tgt Ion: 330 Resp: 115580
 Ion Ratio Lower Upper
 330 100
 332 95.9 76.6 114.8
 141 47.9 29.7 44.5#



#42
 2,4,6-Trichlorophenol
 Concen: 45.81 ng
 RT: 13.52 min Scan# 1088
 Delta R.T. -0.06 min
 Lab File: BF081708.D
 Acq: 18 Sep 2015 21:46

Tgt Ion: 196 Resp: 78115
 Ion Ratio Lower Upper
 196 100
 198 93.2 75.7 113.5
 200 28.8 23.9 35.9









#43
 2,4,5-Trichlorophenol
 Concen: 46.87 ng
 RT: 13.64 min Scan# 1098
 Delta R.T. -0.04 min
 Lab File: BF081708.D
 Acq: 18 Sep 2015 21:46

Instrument :
 BNA_F
 ClientSampleId :
 MW-12-9-15-15MSD

Tgt Ion:196 Resp: 81544
 Ion Ratio Lower Upper
 196 100
 198 91.5 75.4 113.0
 97 69.8 33.3 49.9#
 132 43.4 24.6 37.0#




Abundance Ion 196.00 (195.70 to 196.70): E
 Ion 198.00 (197.70 to 198.70): E
 Ion 97.00 (96.70 to 97.70): BF0
 Ion 132.00 (131.70 to 132.70): E






#44
 2-Fluorobiphenyl
 Concen: 106.88 ng
 RT: 13.72 min Scan# 1105
 Delta R.T. -0.06 min
 Lab File: BF081708.D
 Acq: 18 Sep 2015 21:46

Tgt Ion:172 Resp: 651627
 Ion Ratio Lower Upper
 172 100
 171 35.7 26.1 39.1
 170 23.8 17.4 26.2



Abundance Ion 172.00 (171.70 to 172.70): E
 Ion 171.00 (170.70 to 171.70): E
 Ion 170.00 (169.70 to 170.70): E









#45
 1,1'-Biphenyl
 Concen: 48.54 ng
 RT: 13.91 min Scan# 1122
 Delta R.T. -0.06 min
 Lab File: BF081708.D
 Acq: 18 Sep 2015 21:46

Instrument :
 BNA_F
 ClientSampleId :
 MW-12-9-15-15MSD

Tgt Ion	Ratio	Lower	Upper
154	100		
153	39.2	17.1	57.1
76	13.4	0.0	31.6




Abundance Ion 154.00 (153.70 to 154.70): E
 Ion 153.00 (152.70 to 153.70): E
 Ion 76.00 (75.70 to 76.70): BF0






#46
 2-Chloronaphthalene
 Concen: 49.73 ng
 RT: 13.90 min Scan# 1121
 Delta R.T. -0.06 min
 Lab File: BF081708.D
 Acq: 18 Sep 2015 21:46

Tgt Ion	Ratio	Lower	Upper
162	100		
127	48.8	27.6	41.4#
164	31.8	25.4	38.2



Abundance Ion 162.00 (161.70 to 162.70): E
 Ion 127.00 (126.70 to 127.70): E
 Ion 164.00 (163.70 to 164.70): E









#47
2-Nitroaniline
Concen: 52.01 ng
RT: 14.25 min Scan# 1152
Delta R.T. -0.06 min
Lab File: BF081708.D
Acq: 18 Sep 2015 21:46




Instrument :
BNA_F
ClientSampleId :
MW-12-9-15-15MSD


Tgt Ion: 65 Resp: 110026
Ion Ratio Lower Upper
65 100
92 59.1 55.4 83.0
138 76.0 115.5 173.3#



#48
Acenaphthylene
Concen: 48.11 ng
RT: 14.86 min Scan# 1205
Delta R.T. -0.06 min
Lab File: BF081708.D
Acq: 18 Sep 2015 21:46

Tgt Ion: 152 Resp: 443007
Ion Ratio Lower Upper
152 100
151 20.4 14.6 22.0
153 12.4 10.3 15.5









#49
Dimethylphthalate
Concen: 55.61 ng
RT: 14.80 min Scan# 1200
Delta R.T. -0.06 min
Lab File: BF081708.D
Acq: 18 Sep 2015 21:46




Instrument :
BNA_F
ClientSampleId :
MW-12-9-15-15MSD


Tgt Ion:163 Resp: 337564
Ion Ratio Lower Upper
163 100
194 3.3 4.4 6.6#
164 10.0 8.3 12.5



#50
2,6-Dinitrotoluene
Concen: 46.23 ng
RT: 14.89 min Scan# 1208
Delta R.T. -0.06 min
Lab File: BF081708.D
Acq: 18 Sep 2015 21:46

Tgt Ion:165 Resp: 63690
Ion Ratio Lower Upper
165 100
63 91.7 32.3 48.5#
89 79.1 28.6 43.0#





#51

Acenaphthene

Concen: 48.38 ng

RT: 15.28 min Scan# 1242

Delta R.T. -0.06 min

Lab File: BF081708.D

Acq: 18 Sep 2015 21:46

Instrument :

BNA_F

ClientSampleId :

MW-12-9-15-15MSD


Tgt Ion:154 Resp: 253424

Ion Ratio Lower Upper

154 100

153 113.2 82.1 123.1




152 54.3 37.9 56.9



Abundance Ion 154.00 (153.70 to 154.70): E

Ion 153.00 (152.70 to 153.70): E

Ion 152.00 (151.70 to 152.70): E



#52

3-Nitroaniline

Concen: 23.02 ng

RT: 15.25 min Scan# 1239

Delta R.T. -0.07 min

Lab File: BF081708.D

Acq: 18 Sep 2015 21:46


Tgt Ion:138 Resp: 36098

Ion Ratio Lower Upper

138 100

108 11.0 5.7 8.5#



92 136.4 67.7 101.5#




Abundance Ion 138.00 (137.70 to 138.70): E

Ion 108.00 (107.70 to 108.70): E

Ion 92.00 (91.70 to 92.70): BF0









#53
2,4-Dinitrophenol
Concen: 86.40 ng
RT: 15.55 min Scan# 1265
Delta R.T. -0.04 min
Lab File: BF081708.D
Acq: 18 Sep 2015 21:46




Instrument :
BNA_F
ClientSampleId :
MW-12-9-15-15MSD


Tgt Ion:184 Resp: 61174
Ion Ratio Lower Upper
184 100
63 106.5 35.4 53.2#
154 74.3 32.2 48.4#



#54
Dibenzofuran
Concen: 52.04 ng
RT: 15.71 min Scan# 1279
Delta R.T. -0.06 min
Lab File: BF081708.D
Acq: 18 Sep 2015 21:46

Tgt Ion:168 Resp: 398079
Ion Ratio Lower Upper
168 100
139 41.9 24.2 36.2#
169 12.8 10.6 15.8









#55
4-Nitrophenol
Concen: 24.44 ng
RT: 15.90 min Scan# 1296
Delta R.T. -0.02 min
Lab File: BF081708.D
Acq: 18 Sep 2015 21:46




Instrument :
BNA_F
ClientSampleId :
MW-12-9-15-15MSD


Tgt Ion:139 Resp: 24078
Ion Ratio Lower Upper
139 100
109 93.1 12.7 52.7#
65 138.6 45.9 85.9#



#56
2,4-Dinitrotoluene
Concen: 52.40 ng
RT: 15.85 min Scan# 1292
Delta R.T. -0.04 min
Lab File: BF081708.D
Acq: 18 Sep 2015 21:46

Tgt Ion:165 Resp: 91925
Ion Ratio Lower Upper
165 100
63 56.8 29.8 44.6#
89 80.9 44.5 66.7#
182 1.9 3.0 4.4#







#57
 Fluorene
 Concen: 53.78 ng
 RT: 16.52 min Scan# 1350
 Delta R.T. -0.06 min
 Lab File: BF081708.D
 Acq: 18 Sep 2015 21:46

Instrument :
 BNA_F
 ClientSampleId :
 MW-12-9-15-15MSD



Tgt Ion:166 Resp: 312183
 Ion Ratio Lower Upper
 166 100
 165 94.9 72.6 109.0
 167 13.3 10.6 16.0



Abundance Ion 166.00 (165.70 to 166.70): E
 Ion 165.00 (164.70 to 165.70): E
 Ion 167.00 (166.70 to 167.70): E




Time-->




#58
 2,3,4,6-Tetrachlorophenol
 Concen: 52.95 ng
 RT: 16.07 min Scan# 1311
 Delta R.T. -0.06 min
 Lab File: BF081708.D
 Acq: 18 Sep 2015 21:46


Tgt Ion:232 Resp: 70329
 Ion Ratio Lower Upper
 232 100
 131 55.6 38.5 57.7
 130 2.1 1.8 2.6
 166 32.4 25.0 37.6




Abundance Ion 232.00 (231.70 to 232.70): E
 Ion 131.00 (130.70 to 131.70): E
 Ion 130.00 (129.70 to 130.70): E
 Ion 166.00 (165.70 to 166.70): E



Time-->









#59
Diethylphthalate
Concen: 54.48 ng
RT: 16.49 min Scan# 1348
Delta R.T. -0.06 min
Lab File: BF081708.D
Acq: 18 Sep 2015 21:46

Instrument :
BNA_F
ClientSampleId :
MW-12-9-15-15MSD

Tgt Ion:149 Resp: 347855
Ion Ratio Lower Upper
149 100
177 18.4 17.5 26.3
150 12.1 9.4 14.2




Abundance Ion 149.00 (148.70 to 149.70): E
Ion 177.00 (176.70 to 177.70): E
Ion 150.00 (149.70 to 150.70): E






#60
4-Chlorophenyl-phenylether
Concen: 54.98 ng
RT: 16.61 min Scan# 1358
Delta R.T. -0.06 min
Lab File: BF081708.D
Acq: 18 Sep 2015 21:46

Tgt Ion:204 Resp: 148743
Ion Ratio Lower Upper
204 100
206 32.9 24.8 37.2
141 67.6 42.2 63.4#



Abundance Ion 204.00 (203.70 to 204.70): E
Ion 206.00 (205.70 to 206.70): E
Ion 141.00 (140.70 to 141.70): E









#61
4-Nitroaniline
Concen: 38.81 ng
RT: 16.72 min Scan# 1368
Delta R.T. -0.06 min
Lab File: BF081708.D
Acq: 18 Sep 2015 21:46




Instrument :
BNA_F
ClientSampleId :
MW-12-9-15-15MSD


Tgt Ion:138 Resp: 61504
Ion Ratio Lower Upper
138 100
92 60.3 12.6 52.6#
108 102.1 12.4 52.4#



#62
Azobenzene
Concen: 55.51 ng
RT: 16.97 min Scan# 1390
Delta R.T. -0.06 min
Lab File: BF081708.D
Acq: 18 Sep 2015 21:46

Tgt Ion: 77 Resp: 394049
Ion Ratio Lower Upper
77 100
182 17.3 15.1 55.1
105 19.2 3.4 43.4
51 40.0 12.0 52.0









#63
 Phenanthrene-d10
 Concen: 20.00 ng
 RT: 18.70 min Scan# 1541
 Delta R.T. -0.06 min
 Lab File: BF081708.D
 Acq: 18 Sep 2015 21:46

Instrument :
 BNA_F
 ClientSampleId :
 MW-12-9-15-15MSD

Tgt Ion	Ratio	Lower	Upper
188	100		
94	8.1	11.1	16.7#
80	9.5	11.0	16.4#




Abundance Ion 188.00 (187.70 to 188.70): E
 Ion 94.00 (93.70 to 94.70): BF0
 Ion 80.00 (79.70 to 80.70): BF0






#64
 4,6-Dinitro-2-methylphenol
 Concen: 47.61 ng
 RT: 16.81 min Scan# 1376
 Delta R.T. -0.06 min
 Lab File: BF081708.D
 Acq: 18 Sep 2015 21:46

Tgt Ion	Ratio	Lower	Upper
198	100		
51	85.9	39.6	79.6#
105	58.2	28.5	68.5



Abundance Ion 198.00 (197.70 to 198.70): E
 Ion 51.00 (50.70 to 51.70): BF0
 Ion 105.00 (104.70 to 105.70): E





#65

n-Nitrosodiphenylamine

Concen: 45.06 ng

RT: 16.93 min Scan# 1386

Delta R.T. -0.06 min

Lab File: BF081708.D

Acq: 18 Sep 2015 21:46

Instrument :

BNA_F

ClientSampleId :

MW-12-9-15-15MSD


Tgt Ion:169 Resp: 266797

Ion Ratio Lower Upper

169 100

168 67.8 48.9 73.3




167 37.1 26.2 39.4



Abundance Ion 169.00 (168.70 to 169.70): E

Ion 168.00 (167.70 to 168.70): E

Ion 167.00 (166.70 to 167.70): E



#66

4-Bromophenyl-phenylether

Concen: 50.89 ng

RT: 17.75 min Scan# 1458

Delta R.T. -0.06 min

Lab File: BF081708.D

Acq: 18 Sep 2015 21:46


Tgt Ion:248 Resp: 83735

Ion Ratio Lower Upper

248 100

250 96.1 78.5 117.7



141 89.5 59.0 88.6#




Abundance Ion 248.00 (247.70 to 248.70): E

Ion 250.00 (249.70 to 250.70): E

Ion 141.00 (140.70 to 141.70): E









#67
Hexachlorobenzene
Concen: 49.33 ng
RT: 17.80 min Scan# 1462
Delta R.T. -0.07 min
Lab File: BF081708.D
Acq: 18 Sep 2015 21:46

Instrument :
BNA_F
ClientSampleId :
MW-12-9-15-15MSD

Tgt Ion: 284 Resp: 84866
Ion Ratio Lower Upper
284 100
142 45.4 29.5 44.3#
249 36.6 19.5 29.3#




Abundance Ion 283.80 (283.50 to 284.50): E
Ion 142.00 (141.70 to 142.70): E
Ion 249.00 (248.70 to 249.70): E






#68
Atrazine
Concen: 55.71 ng
RT: 18.35 min Scan# 1510
Delta R.T. -0.06 min
Lab File: BF081708.D
Acq: 18 Sep 2015 21:46

Tgt Ion: 200 Resp: 95471
Ion Ratio Lower Upper
200 100
173 25.5 13.2 53.2
215 41.2 41.8 81.8#



Abundance Ion 200.00 (199.70 to 200.70): E
Ion 173.00 (172.70 to 173.70): E
Ion 215.00 (214.70 to 215.70): E









#69
 Pentachlorophenol
 Concen: 78.50 ng
 RT: 18.35 min Scan# 1510
 Delta R.T. -0.06 min
 Lab File: BF081708.D
 Acq: 18 Sep 2015 21:46

Instrument :
 BNA_F
 ClientSampleId :
 MW-12-9-15-15MSD

Tgt Ion: 266 Resp: 65593
 Ion Ratio Lower Upper
 266 100
 268 60.0 49.8 74.6
 264 62.6 50.2 75.2




Abundance Ion 265.70 (265.40 to 266.40): E
 Ion 268.00 (267.70 to 268.70): E
 Ion 264.00 (263.70 to 264.70): E






#70
 Phenanthrene
 Concen: 49.40 ng
 RT: 18.77 min Scan# 1547
 Delta R.T. -0.06 min
 Lab File: BF081708.D
 Acq: 18 Sep 2015 21:46

Tgt Ion: 178 Resp: 446931
 Ion Ratio Lower Upper
 178 100
 176 19.0 13.8 20.8
 179 14.7 12.2 18.2



Abundance Ion 178.00 (177.70 to 178.70): E
 Ion 176.00 (175.70 to 176.70): E
 Ion 179.00 (178.70 to 179.70): E









#71
 Anthracene
 Concen: 48.73 ng
 RT: 18.88 min Scan# 1557
 Delta R.T. -0.07 min
 Lab File: BF081708.D
 Acq: 18 Sep 2015 21:46




Instrument :
 BNA_F
 ClientSampleId :
 MW-12-9-15-15MSD


Tgt Ion:178 Resp: 452901
 Ion Ratio Lower Upper
 178 100
 176 19.2 14.1 21.1
 179 15.4 12.2 18.2



#72
 Carbazole
 Concen: 48.02 ng
 RT: 19.36 min Scan# 1599
 Delta R.T. -0.06 min
 Lab File: BF081708.D
 Acq: 18 Sep 2015 21:46

Tgt Ion:167 Resp: 418843
 Ion Ratio Lower Upper
 167 100
 166 20.7 15.7 23.5
 139 14.2 9.5 14.3





#73

Di-n-butylphthalate

Concen: 54.44 ng

RT: 20.41 min Scan# 1691

Delta R.T. -0.06 min

Lab File: BF081708.D

Acq: 18 Sep 2015 21:46

Instrument :

BNA_F

ClientSampleId :

MW-12-9-15-15MSD


Tgt Ion:149 Resp: 560667

Ion Ratio Lower Upper

149 100

150 9.0 7.4 11.2




104 5.2 2.4 3.6#



Abundance Ion 149.00 (148.70 to 149.70): E

Ion 150.00 (149.70 to 150.70): E

Ion 104.00 (103.70 to 104.70): E



#74

Fluoranthene

Concen: 49.88 ng

RT: 21.39 min Scan# 1776

Delta R.T. -0.04 min

Lab File: BF081708.D

Acq: 18 Sep 2015 21:46


Tgt Ion:202 Resp: 488465

Ion Ratio Lower Upper

202 100

101 9.3 0.0 38.3



203 17.2 0.0 37.3




Abundance Ion 202.00 (201.70 to 202.70): E

Ion 101.00 (100.70 to 101.70): E

Ion 203.00 (202.70 to 203.70): E





#75

Chrysene-d12

Concen: 20.00 ng

RT: 23.35 min Scan# 1948

Delta R.T. -0.03 min

Lab File: BF081708.D

Acq: 18 Sep 2015 21:46

Instrument :

BNA_F

ClientSampleId :

MW-12-9-15-15MSD


Tgt Ion:240 Resp: 140763

Ion Ratio Lower Upper

240 100

120 9.2 16.2 24.2#




236 25.3 18.4 27.6



Abundance Ion 240.00 (239.70 to 240.70): E

Ion 120.00 (119.70 to 120.70): E

Ion 236.00 (235.70 to 236.70): E



#76

Benzidine

Concen: 14.25 ng

RT: 21.71 min Scan# 1804

Delta R.T. -0.03 min

Lab File: BF081708.D

Acq: 18 Sep 2015 21:46


Tgt Ion:184 Resp: 86104

Ion Ratio Lower Upper

184 100

185 14.4 11.8 17.6



183 12.0 7.6 11.4#




Abundance Ion 184.00 (183.70 to 184.70): E

Ion 185.00 (184.70 to 185.70): E

Ion 183.00 (182.70 to 183.70): E





#77

Pyrene

Concen: 48.49 ng

RT: 21.74 min Scan# 1807

Delta R.T. -0.04 min

Lab File: BF081708.D

Acq: 18 Sep 2015 21:46

Instrument :

BNA_F

ClientSampleId :

MW-12-9-15-15MSD


Tgt Ion: 202 Resp: 505642

Ion Ratio Lower Upper

202 100

200 19.7 15.0 22.4




203 17.6 14.1 21.1



Abundance Ion 202.00 (201.70 to 202.70): E

Ion 200.00 (199.70 to 200.70): E

Ion 203.00 (202.70 to 203.70): E



#78

Terphenyl-d14

Concen: 97.67 ng

RT: 22.07 min Scan# 1836

Delta R.T. -0.04 min

Lab File: BF081708.D

Acq: 18 Sep 2015 21:46


Tgt Ion: 244 Resp: 590606

Ion Ratio Lower Upper

244 100

212 7.7 4.3 6.5#



122 12.6 17.4 26.2#




Abundance Ion 244.00 (243.70 to 244.70): E

Ion 212.00 (211.70 to 212.70): E

Ion 122.00 (121.70 to 122.70): E





#79

Butylbenzylphthalate

Concen: 50.54 ng

RT: 22.77 min Scan# 1897

Delta R.T. -0.03 min

Lab File: BF081708.D

Acq: 18 Sep 2015 21:46

Instrument :

BNA_F

ClientSampleId :

MW-12-9-15-15MSD


Tgt Ion:149 Resp: 229205

Ion Ratio Lower Upper

149 100

91 82.2 48.6 72.8#




206 15.7 14.9 22.3



Abundance Ion 149.00 (148.70 to 149.70): E

Ion 91.00 (90.70 to 91.70): BF0

Ion 206.00 (205.70 to 206.70): E



#80

Benzo(a)anthracene

Concen: 45.50 ng

RT: 23.34 min Scan# 1947

Delta R.T. -0.03 min

Lab File: BF081708.D

Acq: 18 Sep 2015 21:46


Tgt Ion:228 Resp: 407841

Ion Ratio Lower Upper

228 100

226 28.0 20.0 30.0



229 20.1 15.4 23.2




Abundance Ion 228.00 (227.70 to 228.70): E

Ion 226.00 (225.70 to 226.70): E

Ion 229.00 (228.70 to 229.70): E









#81
 3,3'-Dichlorobenzidine
 Concen: 27.25 ng
 RT: 23.35 min Scan# 1948
 Delta R.T. -0.03 min
 Lab File: BF081708.D
 Acq: 18 Sep 2015 21:46




Instrument :
 BNA_F
 ClientSampleId :
 MW-12-9-15-15MSD


Tgt Ion: 252 Resp: 82401
 Ion Ratio Lower Upper
 252 100
 254 64.9 51.4 77.2
 126 8.7 16.4 24.6#



#82
 Chrysene
 Concen: 48.81 ng
 RT: 23.39 min Scan# 1951
 Delta R.T. -0.03 min
 Lab File: BF081708.D
 Acq: 18 Sep 2015 21:46

Tgt Ion: 228 Resp: 392185
 Ion Ratio Lower Upper
 228 100
 226 30.0 21.0 31.4
 229 19.8 15.6 23.4









#83
 Bis(2-ethylhexyl)phthalate
 Concen: 55.65 ng
 RT: 23.47 min Scan# 1958
 Delta R.T. -0.03 min
 Lab File: BF081708.D
 Acq: 18 Sep 2015 21:46




Instrument :
 BNA_F
 ClientSampleId :
 MW-12-9-15-15MSD


Tgt Ion:149 Resp: 333041
 Ion Ratio Lower Upper
 149 100
 167 26.1 24.2 36.2
 279 3.3 3.8 5.6#



#84
 Di-n-octyl phthalate
 Concen: 51.81 ng
 RT: 24.12 min Scan# 2015
 Delta R.T. -0.03 min
 Lab File: BF081708.D
 Acq: 18 Sep 2015 21:46

Tgt Ion:149 Resp: 510480
 Ion Ratio Lower Upper
 149 100
 167 1.4 1.0 1.6
 43 14.6 10.2 15.2









#85
 Indeno(1,2,3-cd)pyrene
 Concen: 45.22 ng
 RT: 25.82 min Scan# 2164
 Delta R.T. -0.06 min
 Lab File: BF081708.D
 Acq: 18 Sep 2015 21:46




Instrument :
 BNA_F
 ClientSampleId :
 MW-12-9-15-15MSD


Tgt Ion: 276 Resp: 266621
 Ion Ratio Lower Upper
 276 100
 138 21.5 25.7 38.5#
 277 25.1 15.6 23.4#



#86
 Perylene-d12
 Concen: 20.00 ng
 RT: 24.78 min Scan# 2073
 Delta R.T. -0.03 min
 Lab File: BF081708.D
 Acq: 18 Sep 2015 21:46

Tgt Ion: 264 Resp: 91862
 Ion Ratio Lower Upper
 264 100
 260 24.0 16.7 25.1
 265 21.0 17.2 25.8









#87
Benzo(b)fluoranthene
Concen: 93.10 ng
RT: 24.45 min Scan# 2044
Delta R.T. -0.03 min
Lab File: BF081708.D
Acq: 18 Sep 2015 21:46




Instrument :
BNA_F
ClientSampleId :
MW-12-9-15-15MSD


Tgt Ion: 252 Resp: 610595
Ion Ratio Lower Upper
252 100
253 21.9 17.5 26.3
125 8.0 17.1 25.7#



#88
Benzo(k)fluoranthene
Concen: 112.20 ng
RT: 24.45 min Scan# 2044
Delta R.T. -0.06 min
Lab File: BF081708.D
Acq: 18 Sep 2015 21:46

Tgt Ion: 252 Resp: 610595
Ion Ratio Lower Upper
252 100
253 21.9 17.0 25.4
125 8.0 16.0 24.0#





#89

Benzo(a)pyrene

Concen: 52.06 ng

RT: 24.73 min Scan# 2069

Delta R.T. -0.03 min

Lab File: BF081708.D

Acq: 18 Sep 2015 21:46

Instrument :

BNA_F

ClientSampleId :

MW-12-9-15-15MSD


Tgt Ion:252 Resp: 289323

Ion Ratio Lower Upper

252 100

253 22.3 17.2 25.8




125 11.6 18.0 27.0#



Abundance Ion 252.00 (251.70 to 252.70): E

Ion 253.00 (252.70 to 253.70): E

Ion 125.00 (124.70 to 125.70): E



#90

Dibenzo(a,h)anthracene

Concen: 52.38 ng

RT: 25.83 min Scan# 2165

Delta R.T. -0.06 min

Lab File: BF081708.D

Acq: 18 Sep 2015 21:46


Tgt Ion:278 Resp: 224932

Ion Ratio Lower Upper

278 100

139 13.7 26.3 39.5#



279 24.0 18.2 27.4




Abundance Ion 278.00 (277.70 to 278.70): E

Ion 139.00 (138.70 to 139.70): E

Ion 279.00 (278.70 to 279.70): E





#91

Benzo(g,h,i)perylene

Concen: 52.75 ng

RT: 26.12 min Scan# 2190

Delta R.T. -0.06 min

Lab File: BF081708.D

Acq: 18 Sep 2015 21:46

Instrument :

BNA_F

ClientSampleId :

MW-12-9-15-15MSD


Tgt Ion: 276 Resp: 216193

Ion Ratio Lower Upper

276 100

277 22.4 17.8 26.6

138 20.5 34.6 51.8#



Abundance Ion 276.00 (275.70 to 276.70): E

Ion 277.00 (276.70 to 277.70): E

Ion 138.00 (137.70 to 138.70): E

