

Data Path : Z:\HPCHEM1\BNA F\DATA\BF091817\
 Data File : BF098605.D
 Acq On : 18 Sep 2017 16:40
 Operator : SJ/JU
 Sample : I5250-01 2X
 Misc :
 ALS Vial : 11 Sample Multiplier: 1

Instrument :
 BNA_F
 ClientSampleId :
 OK-01-091417

Manual Integrations
 APPROVED

Sohil
 9/19/2017 5:31:43 PM

Quant Time: Sep 19 06:27:03 2017
 Quant Method : Z:\HPCHEM1\BNA F\METHODS\8270-BF091117.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Mon Sep 18 13:37:24 2017
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Dichlorobenzene-d4	6.63	152	130713	20.00	ng	0.00
21) Naphthalene-d8	7.92	136	529098	20.00	ng	0.00
38) Acenaphthene-d10	9.67	164	229569	20.00	ng	0.00
63) Phenanthrene-d10	11.15	188	367597	20.00	ng	0.00
75) Chrysene-d12	13.79	240	228647	20.00	ng	0.00
86) Perylene-d12	15.19	264	200550	20.00	ng	0.01
System Monitoring Compounds						
5) 2-Fluorophenol	5.23	112	498440	56.38	ng	0.00
7) Phenol-d6	6.29	99	618956	55.14	ng	0.00
23) Nitrobenzene-d5	7.20	82	375246	39.54	ng	0.00
41) 2,4,6-Tribromophenol	10.46	330	102297	66.77	ng	0.00
44) 2-Fluorobiphenyl	8.99	172	645614	47.67	ng	0.00
78) Terphenyl-d14	12.73	244	419630	39.61	ng	0.00
Target Compounds						
48) Acenaphthylene	9.53	152	62541	2.866	ng	96
49) Dimethylphthalate	9.39	163	74788	4.209	ng	100
57) Fluorene	10.22	166	34333	2.270	ng	98
70) Phenanthrene	11.17	178	428501	21.989	ng	98
71) Anthracene	11.22	178	99236	4.960	ng	99
74) Fluoranthene	12.36	202	621368	31.851	ng	95
77) Pyrene	12.59	202	553620	30.693	ng	100
80) Benzo(a)anthracene	13.77	228	240280	17.924	ng	97
82) Chrysene	13.81	228	219474	16.197	ng	97
85) Indeno(1,2,3-cd)pyrene	16.52	276	86936	9.145	ng	# 86
87) Benzo(b)fluoranthene	14.79	252	262392m	20.808	ng	
88) Benzo(k)fluoranthene	14.82	252	82051m	6.970	ng	
89) Benzo(a)pyrene	15.13	252	174780	15.558	ng	96
90) Dibenzo(a,h)anthracene	16.52	278	21933	2.684	ng	# 73
91) Benzo(a,h,i)perylene	16.92	276	79774	9.705	ng	93

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\HPCHEM1\BNA F\DATA\BF091817\
Data File : BF098605.D
Acq On : 18 Sep 2017 16:40
Operator : SJ/JU
Sample : I5250-01 2X
Misc :
ALS Vial : 11 Sample Multiplier: 1

Instrument :
BNA_F
ClientSampleId :
OK-01-091417

Manual Integrations
APPROVED

Sohil
9/19/2017 5:31:43 PM

Quant Time: Sep 19 06:27:03 2017
Quant Method : Z:\HPCHEM1\BNA F\METHODS\8270-BF091117.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Mon Sep 18 13:37:24 2017
Response via : Initial Calibration

