

Data Path : Z:\svoasrv\HPCHEM1\BNA_F\Data\BF091824\
 Data File : BF139431.D
 Acq On : 18 Sep 2024 17:52
 Operator : RC/JU
 Sample : P4030-02
 Misc :
 ALS Vial : 5 Sample Multiplier: 1

Instrument :
 BNA_F
 ClientSampleId :
 WC-A-08-202409

Quant Time: Sep 19 01:27:57 2024
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_F\Methods\8270-BF091324.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Sat Sep 14 02:58:23 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Dichlorobenzene-d4	6.875	152	60436	20.000	ng	0.00
21) Naphthalene-d8	8.157	136	226595	20.000	ng	0.00
39) Acenaphthene-d10	9.910	164	127416	20.000	ng	0.00
64) Phenanthrene-d10	11.392	188	232750	20.000	ng	0.00
76) Chrysene-d12	14.027	240	105843	20.000	ng	0.00
86) Perylene-d12	15.498	264	127244	20.000	ng	0.00
System Monitoring Compounds						
5) 2-Fluorophenol	5.504	112	495088	133.639	ng	0.01
7) Phenol-d6	6.504	99	600478	123.488	ng	0.00
23) Nitrobenzene-d5	7.439	82	491606	101.993	ng	0.00
42) 2,4,6-Tribromophenol	10.704	330	203426	149.964	ng	0.00
45) 2-Fluorobiphenyl	9.233	172	884669	97.379	ng	0.00
79) Terphenyl-d14	12.980	244	864792	134.110	ng	0.00
Target Compounds						
70) Pentachlorophenol	11.198	266	29995	18.467	ng	Qvalue 96

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA_F\Data\BF091824\
Data File : BF139431.D
Acq On : 18 Sep 2024 17:52
Operator : RC/JU
Sample : P4030-02
Misc :
ALS Vial : 5 Sample Multiplier: 1

Instrument :
BNA_F
ClientSampleId :
WC-A-08-202409

Quant Time: Sep 19 01:27:57 2024
Quant Method : Z:\svoasrv\HPCHEM1\BNA_F\Methods\8270-BF091324.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Sat Sep 14 02:58:23 2024
Response via : Initial Calibration

