


Data Path : Z:\svoasrv\HPCHEM1\BNA_F\Data\BF091824\
Data File : BF139427.D
Acq On : 18 Sep 2024 15:52
Operator : RC/JU
Sample : DFTPP
Misc :
ALS Vial : 1 Sample Multiplier: 1

Instrument :
BNA_F
ClientSampleId :
DFTPP

Integration File: rteint.p

Method : Z:\svoasrv\HPCHEM1\BNA_F\Methods\8270-BF091324.M
Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
Last Update : Sat Sep 14 02:58:23 2024






AutoFind: Scans 1624, 1625, 1626; Background Corrected with Scan 1617

Target Mass	Rel. to Mass	Lower Limit%	Upper Limit%	Rel. Abn%	Raw Abn	Result Pass/Fail
51	198	10	80	47.1	54800	PASS
68	69	0.00	2	1.7	826	PASS
69	198	0.00	100	40.8	47445	PASS
70	69	0.00	2	0.6	285	PASS
127	198	10	80	48.7	56675	PASS
197	198	0.00	2	0.0	0	PASS
198	198	100	100	100.0	116333	PASS
199	198	5	9	6.5	7578	PASS
275	198	10	60	28.0	32517	PASS
365	198	1	100	3.6	4240	PASS
441	198	0.01	100	15.0	17494	PASS
442	442	50	100	100.0	110285	PASS
443	442	15	24	19.2	21197	PASS

Data Path : Z:\svoasrv\HPCHEM1\BNA_F\Data\BF091824\
Data File : BF139427.D
Acq On : 18 Sep 2024 15:52
Operator : RC/JU
Sample : DFTPP
Misc :
ALS Vial : 1 Sample Multiplier: 1

Instrument :
BNA_F
ClientSampleId :
DFTPP

Quant Time: Sep 18 16:54:49 2024
Quant Method : Z:\svoasrv\HPCHEM1\BNA_F\Methods\8270-BF091324.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Sat Sep 14 02:58:23 2024
Response via : Initial Calibration



TIC: BF139427.D\data.ms

(70) Pentachlorophenol (C)

11.198min (-0.000) 85281.36 ng




response 137480

Ion	Exp%	Act%
265.70	100.00	100.00
268.00	63.90	64.62
264.00	64.20	63.86
0.00	0.00	0.00

Data Path : Z:\svoasrv\HPCHEM1\BNA_F\Data\BF091824\
Data File : BF139427.D
Acq On : 18 Sep 2024 15:52
Operator : RC/JU
Sample : DFTPP
Misc :
ALS Vial : 1 Sample Multiplier: 1

Instrument :
BNA_F
ClientSampleId :
DFTPP

Quant Time: Sep 18 16:54:49 2024
Quant Method : Z:\svoasrv\HPCHEM1\BNA_F\Methods\8270-BF091324.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Sat Sep 14 02:58:23 2024
Response via : Initial Calibration



TIC: BF139427.D\data.ms

(77) Benzidine

12.733min (+ 0.006) 341478.03 ng

response 616440

Ion	Exp%	Act%
184.00	100.00	100.00
185.00	14.40	13.99
183.00	11.10	11.34
0.00	0.00	0.00

DDT Breakdown

Date	Instrument Name	DFTPP Data File
9/18/2024	BNA_F	<u>BF139427.D</u>
Compound Name	Response	Retention Time
DDT	322562	13.557
DDD	9676	13.251
DDE	660	12.916
SUM(DDD+DDE)	SUM(DDT+DDD+DDE)	% Breakdown Of DDT
10336	332898	3.10

Instrument :
BNA_F
ClientSampleId :
DFTPP