

Data Path : Z:\HPCHEM1\BNA_F\DATA\BF091916\
 Data File : BF090123.D
 Acq On : 19 Sep 2016 16:31
 Operator : UM/SJ
 Sample : H4716-02RE
 Misc :
 ALS Vial : 5 Sample Multiplier: 1

Instrument :
 BNA_F
 ClientSampleId :
 EFFRE

Manual Integrations
 APPROVED

umangi
 9/19/2016 7:19:59 PM

Quant Time: Sep 19 18:05:02 2016
 Quant Method : Z:\HPCHEM1\BNA_F\METHODS\8270-BF091516.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Thu Sep 15 11:18:43 2016
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Dichlorobenzene-d4	6.54	152	173460m	20.00	ng	-0.01
21) Naphthalene-d8	7.83	136	641386	20.00	ng	-0.01
38) Acenaphthene-d10	9.56	164	308812	20.00	ng	-0.02
63) Phenanthrene-d10	11.04	188	543538	20.00	ng	-0.01
75) Chrysene-d12	13.66	240	176016	20.00	ng	-0.01
86) Perylene-d12	14.98	264	345302	20.00	ng	-0.01
System Monitoring Compounds						
5) 2-Fluorophenol	5.11	112	392059	36.35	ng	-0.01
7) Phenol-d6	6.23	99	329314	23.32	ng	0.02
23) Nitrobenzene-d5	7.11	82	885294	77.61	ng	-0.02
41) 2,4,6-Tribromophenol	10.35	330	202574	65.30	ng	-0.02
44) 2-Fluorobiphenyl	8.91	172	1613608	82.06	ng	-0.01
78) Terphenyl-d14	12.63	244	781154	105.39	ng	-0.01
Target Compounds						
15) Benzyl Alcohol	6.74	79	236632	29.76	ng	# 87
22) Acetophenone	6.96	105	77051	4.96	ng	# 91
32) Benzoic acid	8.14	122	15222977m	1793.98	ng	
49) Dimethylphthalate	9.31	163	311091	13.56	ng	100
83) Bis(2-ethylhexyl)phthalate	13.68	149	17581	2.22	ng	# 93

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\HPCHEM1\BNA_F\DATA\BF091916\
Data File : BF090123.D
Acq On : 19 Sep 2016 16:31
Operator : UM/SJ
Sample : H4716-02RE
Misc :
ALS Vial : 5 Sample Multiplier: 1

Instrument :
BNA_F
ClientSampleId :
EFFRE

Manual Integrations
APPROVED

umangi
9/19/2016 7:19:59 PM

Quant Time: Sep 19 18:05:02 2016
Quant Method : Z:\HPCHEM1\BNA_F\METHODS\8270-BF091516.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Thu Sep 15 11:18:43 2016
Response via : Initial Calibration

