


Data Path : Z:\HPCHEM1\BNA_F\DATA\BF091916\
 Data File : BF090120.D
 Acq On : 19 Sep 2016 14:29
 Operator : UM/SJ
 Sample : DFTPP
 Misc :
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 BNA_F
 ClientSampleId :
 DFTPP

Integration File: rteint.p

Method : Z:\HPCHEM1\BNA_F\METHODS\8270-BF091516.M
 Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 Last Update : Thu Sep 15 11:18:43 2016



AutoFind: Scans 851, 852, 853; Background Corrected with Scan 848

Target Mass	Rel. to Mass	Lower Limit%	Upper Limit%	Rel. Abn%	Raw Abn	Result Pass/Fail
51	198	10	80	33.3	17076	PASS
68	69	0.00	2	1.2	235	PASS
69	198	0.00	100	36.7	18828	PASS
70	69	0.00	2	0.0	0	PASS
127	198	10	80	53.1	27273	PASS
197	198	0.00	2	0.5	235	PASS
198	198	100	100	100.0	51346	PASS
199	198	5	9	6.6	3371	PASS
275	198	10	60	23.7	12166	PASS
365	198	1	100	2.8	1438	PASS
441	442	0.01	24	15.5	6047	PASS
442	198	50	100	76.2	39116	PASS
443	442	15	24	19.8	7741	PASS