




Data Path : Z:\svoasrv\HPCHEM1\BNA_F\Data\BF091924\
Data File : BF139452.D
Acq On : 19 Sep 2024 09:11
Operator : RC/JU
Sample : DFTPP
Misc :
ALS Vial : 1 Sample Multiplier: 1

Quant Time: Sep 19 11:37:46 2024
Quant Method : Z:\svoasrv\HPCHEM1\BNA_F\Methods\8270-BF091324.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Sat Sep 14 02:58:23 2024
Response via : Initial Calibration



TIC: BF139452.D\data.ms

(70) Pentachlorophenol (C)




11.204min (+ 0.006) 28629.20 ng

response 149446

Ion	Exp%	Act%
265.70	100.00	100.00
268.00	63.90	63.30
264.00	64.20	62.93
0.00	0.00	0.00

Data Path : Z:\svoasrv\HPCHEM1\BNA_F\Data\BF091924\
Data File : BF139452.D
Acq On : 19 Sep 2024 09:11
Operator : RC/JU
Sample : DFTPP
Misc :
ALS Vial : 1 Sample Multiplier: 1

Quant Time: Sep 19 11:37:46 2024
Quant Method : Z:\svoasrv\HPCHEM1\BNA_F\Methods\8270-BF091324.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Sat Sep 14 02:58:23 2024
Response via : Initial Calibration



TIC: BF139452.D\data.ms

(77) Benzidine

12.739min (+ 0.012) 0.00 ng

response 660773

Ion	Exp%	Act%
184.00	100.00	100.00
185.00	14.40	13.80
183.00	11.10	11.44
0.00	0.00	0.00