

Data Path : Z:\svoasrv\HPCHEM1\BNA_F\Data\BF091924\
 Data File : BF139469.D
 Acq On : 19 Sep 2024 17:49
 Operator : RC/JU
 Sample : PB163446TB
 Misc :
 ALS Vial : 3 Sample Multiplier: 1

Instrument :
 BNA_F
 ClientSampleId :
 PB163446TB

Quant Time: Sep 20 01:36:17 2024
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_F\Methods\8270-BF091324.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Sat Sep 14 02:58:23 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Dichlorobenzene-d4	6.875	152	66084	20.000	ng	0.00
21) Naphthalene-d8	8.157	136	239522	20.000	ng	# 0.00
39) Acenaphthene-d10	9.910	164	127200	20.000	ng	0.00
64) Phenanthrene-d10	11.392	188	235896	20.000	ng	0.00
76) Chrysene-d12	14.033	240	168810	20.000	ng	0.00
86) Perylene-d12	15.498	264	140255	20.000	ng	0.00
System Monitoring Compounds						
5) 2-Fluorophenol	5.504	112	564056	139.243	ng	0.01
7) Phenol-d6	6.504	99	721400	135.676	ng	0.00
23) Nitrobenzene-d5	7.439	82	490749	96.320	ng	0.00
42) 2,4,6-Tribromophenol	10.698	330	170019	125.550	ng	0.00
45) 2-Fluorobiphenyl	9.233	172	860945	94.929	ng	0.00
79) Terphenyl-d14	12.980	244	931897	90.611	ng	0.00


Target Compounds	Qvalue


(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA_F\Data\BF091924\
Data File : BF139469.D
Acq On : 19 Sep 2024 17:49
Operator : RC/JU
Sample : PB163446TB
Misc :
ALS Vial : 3 Sample Multiplier: 1

Instrument :
BNA_F
ClientSampleId :
PB163446TB

Quant Time: Sep 20 01:36:17 2024
Quant Method : Z:\svoasrv\HPCHEM1\BNA_F\Methods\8270-BF091324.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Sat Sep 14 02:58:23 2024
Response via : Initial Calibration









#1
1,4-Dichlorobenzene-d4
Concen: 20.000 ng
RT: 6.875 min Scan# 814
Delta R.T. -0.006 min
Lab File: BF139469.D
Acq: 19 Sep 2024 17:49




Instrument :
BNA_F
ClientSampleId :
PB163446TB


Tgt Ion:152 Resp: 66084
Ion Ratio Lower Upper
152 100
150 155.0 124.1 186.1
115 61.1 48.1 72.1



#5
2-Fluorophenol
Concen: 139.243 ng
RT: 5.504 min Scan# 580
Delta R.T. 0.011 min
Lab File: BF139469.D
Acq: 19 Sep 2024 17:49


Tgt Ion:112 Resp: 564056
Ion Ratio Lower Upper
112 100
64 68.1 54.8 82.2
63 38.7 33.0 49.4








#7
Phenol-d6
Concen: 135.676 ng
RT: 6.504 min Scan# 71
Delta R.T. -0.000 min
Lab File: BF139469.D
Acq: 19 Sep 2024 17:49

Instrument :
BNA_F
ClientSampleId :
PB163446TB



Tgt Ion: 99 Resp: 721400




Ion	Ratio	Lower	Upper
99	100		
42	25.1	21.9	32.9
71	39.8	34.6	52.0




#21
Naphthalene-d8
Concen: 20.000 ng
RT: 8.157 min Scan# 1031
Delta R.T. -0.006 min
Lab File: BF139469.D
Acq: 19 Sep 2024 17:49

Tgt Ion: 136 Resp: 239522

Ion	Ratio	Lower	Upper
136	100		
137	10.9	8.6	12.8
54	9.8	7.8	11.8
68	5.6	3.6	5.4





#23

Nitrobenzene-d5

Concen: 96.320 ng

RT: 7.439 min Scan# 90

Delta R.T. -0.006 min

Lab File: BF139469.D

Acq: 19 Sep 2024 17:49

Instrument :

BNA_F

ClientSampleId :

PB163446TB





Tgt Ion: 82 Resp: 490749

Ion Ratio Lower Upper

82 100

128 40.2 31.8 47.6

54 60.6 50.6 75.8



#39

Acenaphthene-d10

Concen: 20.000 ng

RT: 9.910 min Scan# 1329

Delta R.T. -0.006 min

Lab File: BF139469.D

Acq: 19 Sep 2024 17:49




Tgt Ion:164 Resp: 127200


Ion Ratio Lower Upper

164 100

162 101.5 81.3 121.9

160 45.7 36.4 54.6





#42

2,4,6-Tribromophenol

Concen: 125.550 ng

RT: 10.698 min Scan# 1463

Delta R.T. -0.000 min

Lab File: BF139469.D


Acq: 19 Sep 2024 17:49

Instrument :

BNA_F

ClientSampleId :

PB163446TB






Tgt Ion:330 Resp: 170019

Ion Ratio Lower Upper

330 100

332 94.2 75.2 112.8

141 37.0 31.2 46.8



#45

2-Fluorobiphenyl

Concen: 94.929 ng

RT: 9.233 min Scan# 1214

Delta R.T. -0.000 min

Lab File: BF139469.D

Acq: 19 Sep 2024 17:49




Tgt Ion:172 Resp: 860945


Ion Ratio Lower Upper

172 100

171 35.1 28.3 42.5

170 23.4 19.0 28.6









#64
Phenanthrene-d10
Concen: 20.000 ng
RT: 11.392 min Scan# 11
Delta R.T. -0.006 min
Lab File: BF139469.D
Acq: 19 Sep 2024 17:49




Instrument :
BNA_F
ClientSampleId :
PB163446TB


Tgt Ion:188 Resp: 235896
Ion Ratio Lower Upper
188 100
94 8.1 5.5 8.3
80 9.3 6.6 9.8



#76
Chrysene-d12
Concen: 20.000 ng
RT: 14.033 min Scan# 2030
Delta R.T. -0.000 min
Lab File: BF139469.D
Acq: 19 Sep 2024 17:49


Tgt Ion:240 Resp: 168810
Ion Ratio Lower Upper
240 100
120 8.7 6.2 9.2
236 25.4 20.9 31.3








#79
 Terphenyl-d14
 Concen: 90.611 ng
 RT: 12.980 min Scan# 1851
 Delta R.T. -0.000 min
 Lab File: BF139469.D
 Acq: 19 Sep 2024 17:49

Instrument :
 BNA_F
 ClientSampleId :
 PB163446TB



Tgt Ion:244 Resp: 931897
 Ion Ratio Lower Upper
 244 100
 212 7.3 6.0 9.0
 122 10.3 7.6 11.4



#86
 Perylene-d12
 Concen: 20.000 ng
 RT: 15.498 min Scan# 2279
 Delta R.T. -0.000 min
 Lab File: BF139469.D
 Acq: 19 Sep 2024 17:49

Tgt Ion:264 Resp: 140255
 Ion Ratio Lower Upper
 264 100
 260 24.0 19.5 29.3
 265 21.4 16.6 25.0

