


Data Path : Z:\svoasrv\HPCHEM1\BNA_F\Data\BF091924\
Data File : BF139467.D
Acq On : 19 Sep 2024 16:50
Operator : RC/JU
Sample : DFTPP
Misc :
ALS Vial : 1 Sample Multiplier: 1

Instrument :
BNA_F
ClientSampleId :
DFTPP

Integration File: rteint.p

Method : Z:\svoasrv\HPCHEM1\BNA_F\Methods\8270-BF091324.M
Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
Last Update : Sat Sep 14 02:58:23 2024






AutoFind: Scans 1624, 1625, 1626; Background Corrected with Scan 1617

Target Mass	Rel. to Mass	Lower Limit%	Upper Limit%	Rel. Abn%	Raw Abn	Result Pass/Fail
51	198	10	80	46.8	45651	PASS
68	69	0.00	2	1.8	704	PASS
69	198	0.00	100	40.0	39059	PASS
70	69	0.00	2	0.4	167	PASS
127	198	10	80	47.9	46763	PASS
197	198	0.00	2	0.0	0	PASS
198	198	100	100	100.0	97624	PASS
199	198	5	9	6.7	6588	PASS
275	198	10	60	26.9	26288	PASS
365	198	1	100	3.9	3826	PASS
441	198	0.01	100	14.5	14166	PASS
442	442	50	100	100.0	90547	PASS
443	442	15	24	19.2	17399	PASS

Data Path : Z:\svoasrv\HPCHEM1\BNA_F\Data\BF091924\
Data File : BF139467.D
Acq On : 19 Sep 2024 16:50
Operator : RC/JU
Sample : DFTPP
Misc :
ALS Vial : 1 Sample Multiplier: 1

Instrument :
BNA_F
ClientSampleId :
DFTPP

Quant Time: Sep 19 18:45:12 2024
Quant Method : Z:\svoasrv\HPCHEM1\BNA_F\Methods\8270-BF091324.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Sat Sep 14 02:58:23 2024
Response via : Initial Calibration



TIC: BF139467.D\data.ms

(70) Pentachlorophenol (C)

11.198min (-0.000) 50487.09 ng




response 131068

Ion	Exp%	Act%
265.70	100.00	100.00
268.00	63.90	62.60
264.00	64.20	61.07
0.00	0.00	0.00

Data Path : Z:\svoasrv\HPCHEM1\BNA_F\Data\BF091924\
Data File : BF139467.D
Acq On : 19 Sep 2024 16:50
Operator : RC/JU
Sample : DFTPP
Misc :
ALS Vial : 1 Sample Multiplier: 1

Instrument :
BNA_F
ClientSampleId :
DFTPP

Quant Time: Sep 19 18:45:12 2024
Quant Method : Z:\svoasrv\HPCHEM1\BNA_F\Methods\8270-BF091324.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Sat Sep 14 02:58:23 2024
Response via : Initial Calibration



TIC: BF139467.D\data.ms

(77) Benzidine

12.733min (+ 0.006) 343272.25 ng m

response 678696

Ion	Exp%	Act%
184.00	100.00	100.00
185.00	14.40	14.06
183.00	11.10	11.19
0.00	0.00	0.00

DDT Breakdown

Date	Instrument Name	DFTPP Data File
9/19/2024	BNA_F	<u>BF139467.D</u>
Compound Name	Response	Retention Time
DDT	314419	13.551
DDD	5439	13.286
DDE	370	12.916
SUM(DDD+DDE)	SUM(DDT+DDD+DDE)	% Breakdown Of DDT
5809	320228	1.81

Instrument :
BNA_F
ClientSampleId :
DFTPP