

Data Path : Z:\HPCHEM1\BNA\_F\DATA\BF092016\  
 Data File : BF090141.D  
 Acq On : 20 Sep 2016 18:35  
 Operator : UM/SJ  
 Sample : H4854-01  
 Misc :  
 ALS Vial : 5 Sample Multiplier: 1

Instrument :  
 BNA\_F  
 ClientSampleId :  
 FK-BPM-01-A1

Quant Time: Sep 21 02:00:01 2016  
 Quant Method : Z:\HPCHEM1\BNA\_F\METHODS\8270-BF092016.M  
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION  
 QLast Update : Tue Sep 20 17:58:29 2016  
 Response via : Initial Calibration

| Internal Standards          | R.T.  | QIon | Response | Conc   | Units | Dev(Min)      |
|-----------------------------|-------|------|----------|--------|-------|---------------|
| 1) 1,4-Dichlorobenzene-d4   | 6.51  | 152  | 162946   | 20.00  | ng    | -0.01         |
| 21) Naphthalene-d8          | 7.80  | 136  | 658665   | 20.00  | ng    | -0.01         |
| 38) Acenaphthene-d10        | 9.55  | 164  | 362165   | 20.00  | ng    | 0.00          |
| 63) Phenanthrene-d10        | 11.02 | 188  | 575452   | 20.00  | ng    | -0.01         |
| 75) Chrysene-d12            | 13.62 | 240  | 470499   | 20.00  | ng    | -0.01         |
| 86) Perylene-d12            | 14.95 | 264  | 301788   | 20.00  | ng    | -0.01         |
| System Monitoring Compounds |       |      |          |        |       |               |
| 5) 2-Fluorophenol           | 5.08  | 112  | 1049793  | 105.01 | ng    | 0.01          |
| 7) Phenol-d6                | 6.17  | 99   | 1305674  | 105.76 | ng    | 0.00          |
| 23) Nitrobenzene-d5         | 7.08  | 82   | 795829   | 70.04  | ng    | -0.01         |
| 41) 2,4,6-Tribromophenol    | 10.33 | 330  | 266463   | 85.76  | ng    | -0.01         |
| 44) 2-Fluorobiphenyl        | 8.89  | 172  | 1448366  | 78.51  | ng    | 0.00          |
| 78) Terphenyl-d14           | 12.61 | 244  | 1176480  | 63.48  | ng    | 0.00          |
| Target Compounds            |       |      |          |        |       |               |
| 49) Dimethylphthalate       | 9.29  | 163  | 645698   | 27.99  | ng    | Qvalue<br>100 |

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\HPCHEM1\BNA\_F\DATA\BF092016\  
Data File : BF090141.D  
Acq On : 20 Sep 2016 18:35  
Operator : UM/SJ  
Sample : H4854-01  
Misc :  
ALS Vial : 5 Sample Multiplier: 1

Instrument :  
BNA\_F  
ClientSampleId :  
FK-BPM-01-A1

Quant Time: Sep 21 02:00:01 2016  
Quant Method : Z:\HPCHEM1\BNA\_F\METHODS\8270-BF092016.M  
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION  
QLast Update : Tue Sep 20 17:58:29 2016  
Response via : Initial Calibration

