

Data Path : Z:\HPCHEM1\BNA_F\DATA\BF092016\
 Data File : BF090144.D
 Acq On : 20 Sep 2016 20:01
 Operator : UM/SJ
 Sample : H4854-03
 Misc :
 ALS Vial : 8 Sample Multiplier: 1

Instrument :
 BNA_F
 ClientSampleId :
 FK-BPM-01-A2

Quant Time: Sep 21 02:10:28 2016
 Quant Method : Z:\HPCHEM1\BNA_F\METHODS\8270-BF092016.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Tue Sep 20 17:58:29 2016
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Dichlorobenzene-d4	6.51	152	162507	20.00	ng	-0.01
21) Naphthalene-d8	7.80	136	662214	20.00	ng	-0.01
38) Acenaphthene-d10	9.55	164	363612	20.00	ng	0.00
63) Phenanthrene-d10	11.02	188	531550	20.00	ng	-0.01
75) Chrysene-d12	13.62	240	417659	20.00	ng	-0.01
86) Perylene-d12	14.95	264	315507	20.00	ng	-0.01
System Monitoring Compounds						
5) 2-Fluorophenol	5.08	112	963819	96.67	ng	0.01
7) Phenol-d6	6.17	99	1303787	105.89	ng	0.00
23) Nitrobenzene-d5	7.08	82	744338	65.16	ng	-0.01
41) 2,4,6-Tribromophenol	10.33	330	249930	80.12	ng	-0.01
44) 2-Fluorobiphenyl	8.89	172	1369568	73.94	ng	0.00
78) Terphenyl-d14	12.60	244	1088842	66.19	ng	0.00
Target Compounds						
49) Dimethylphthalate	9.29	163	591991	25.56	ng	Qvalue 100

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\HPCHEM1\BNA_F\DATA\BF092016\
Data File : BF090144.D
Acq On : 20 Sep 2016 20:01
Operator : UM/SJ
Sample : H4854-03
Misc :
ALS Vial : 8 Sample Multiplier: 1

Instrument :
BNA_F
ClientSampleId :
FK-BPM-01-A2

Quant Time: Sep 21 02:10:28 2016
Quant Method : Z:\HPCHEM1\BNA_F\METHODS\8270-BF092016.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Tue Sep 20 17:58:29 2016
Response via : Initial Calibration

