

Data Path : Z:\HPCHEM1\BNA_F\DATA\BF092016\
 Data File : BF090138.D
 Acq On : 20 Sep 2016 17:10
 Operator : UM/SJ
 Sample : SSTDCCC040
 Misc :
 ALS Vial : 5 Sample Multiplier: 1

Instrument :
 BNA_F
 LabSampleId :
 SSTDCCC040

Quant Time: Sep 21 01:50:35 2016
 Quant Method : Z:\HPCHEM1\BNA_F\METHODS\8270-BF092016.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Tue Sep 20 17:58:29 2016
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	Amount	Calc.	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	20.000	20.000	0.0	103	0.00
2	1,4-Dioxane	40.000	35.679	10.8	100	0.00
3	Pyridine	40.000	38.141	4.6	102	0.00
4	n-Nitrosodimethylamine	40.000	37.386	6.5	101	0.00
5 S	2-Fluorophenol	80.000	74.650	6.7	99	0.00
6	Aniline	40.000	38.798	3.0	104	0.00
7 S	Phenol-d6	80.000	78.653	1.7	103	0.00
8	2-Chlorophenol	40.000	38.123	4.7	102	0.00
9	Benzaldehyde	40.000	38.540	3.7	104	0.00
10 C	Phenol	40.000	37.110	7.2	95	0.00
11	bis(2-Chloroethyl)ether	40.000	38.805	3.0	102	0.00
12	1,3-Dichlorobenzene	40.000	36.895	7.8	102	0.00
13 C	1,4-Dichlorobenzene	40.000	36.593	8.5	103	0.00
14	1,2-Dichlorobenzene	40.000	36.406	9.0	98	0.00
15	Benzyl Alcohol	40.000	37.670	5.8	98	0.00
16	2,2'-oxybis(1-Chloropropane)	40.000	38.860	2.9	105	0.00
17	2-Methylphenol	40.000	37.953	5.1	101	-0.01
18	Hexachloroethane	40.000	38.191	4.5	101	0.00
19 P	n-Nitroso-di-n-propylamine	40.000	38.845	2.9	103	0.00
20	3+4-Methylphenols	40.000	38.992	2.5	103	0.00
21 I	Naphthalene-d8	20.000	20.000	0.0	100	0.00
22	Acetophenone	40.000	40.732	-1.8	106	0.00
23 S	Nitrobenzene-d5	80.000	74.926	6.3	101	0.00
24	Nitrobenzene	40.000	40.590	-1.5	103	0.00
25	Isophorone	40.000	36.842	7.9	100	0.00
26 C	2-Nitrophenol	40.000	38.546	3.6	101	0.00
27	2,4-Dimethylphenol	40.000	37.755	5.6	101	0.00
28	bis(2-Chloroethoxy)methane	40.000	38.498	3.8	101	0.00
29 C	2,4-Dichlorophenol	40.000	40.330	-0.8	104	0.00
30	1,2,4-Trichlorobenzene	40.000	39.273	1.8	102	0.00
31	Naphthalene	40.000	38.277	4.3	100	0.00
32	Benzoic acid	40.000	43.526	-8.8	105	0.00
33	4-Chloroaniline	40.000	39.600	1.0	101	0.00
34 C	Hexachlorobutadiene	40.000	37.934	5.2	102	0.00
35	Caprolactam	40.000	37.125	7.2	102	0.00
36 C	4-Chloro-3-methylphenol	40.000	39.930	0.2	109	0.00
37	2-Methylnaphthalene	40.000	39.559	1.1	105	0.00
38 I	Acenaphthene-d10	20.000	20.000	0.0	103	0.00
39	1,2,4,5-Tetrachlorobenzene	40.000	37.164	7.1	100	0.00
40 P	Hexachlorocyclopentadiene	40.000	40.384	-1.0	98	0.00
41 S	2,4,6-Tribromophenol	80.000	77.193	3.5	100	0.00
42 C	2,4,6-Trichlorophenol	40.000	40.956	-2.4	102	0.00
43	2,4,5-Trichlorophenol	40.000	38.006	5.0	102	0.00
44 S	2-Fluorobiphenyl	80.000	71.958	10.1	99	0.00

Data Path : Z:\HPCHEM1\BNA_F\DATA\BF092016\
 Data File : BF090138.D
 Acq On : 20 Sep 2016 17:10
 Operator : UM/SJ
 Sample : SSTDCCC040
 Misc :
 ALS Vial : 5 Sample Multiplier: 1

Instrument :
 BNA_F
 LabSampleId :
 SSTDCCC040

Quant Time: Sep 21 01:50:35 2016
 Quant Method : Z:\HPCHEM1\BNA_F\METHODS\8270-BF092016.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Tue Sep 20 17:58:29 2016
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	Amount	Calc.	%Dev	Area%	Dev(min)
45	1,1'-Biphenyl	40.000	40.252	-0.6	102	0.00
46	2-Chloronaphthalene	40.000	36.582	8.5	104	0.00
47	2-Nitroaniline	40.000	37.779	5.6	104	0.00
48	Acenaphthylene	40.000	36.627	8.4	102	0.00
49	Dimethylphthalate	40.000	35.661	10.8	100	0.00
50	2,6-Dinitrotoluene	40.000	40.742	-1.9	103	0.00
51 C	Acenaphthene	40.000	38.326	4.2	102	0.00
52	3-Nitroaniline	40.000	37.408	6.5	101	0.00
53 P	2,4-Dinitrophenol	40.000	35.752	10.6	98	0.00
54	Dibenzofuran	40.000	37.212	7.0	105	0.00
55 P	4-Nitrophenol	40.000	39.814	0.5	99	0.00
56	2,4-Dinitrotoluene	40.000	40.650	-1.6	101	0.00
57	Fluorene	40.000	40.599	-1.5	103	0.00
58	2,3,4,6-Tetrachlorophenol	40.000	39.520	1.2	99	0.00
59	Diethylphthalate	40.000	39.210	2.0	103	0.00
60	4-Chlorophenyl-phenylether	40.000	36.307	9.2	101	0.00
61	4-Nitroaniline	40.000	40.570	-1.4	101	0.00
62	Azobenzene	40.000	38.229	4.4	98	0.00
63 I	Phenanthrene-d10	20.000	20.000	0.0	99	0.00
64	4,6-Dinitro-2-methylphenol	40.000	38.545	3.6	100	0.00
65 c	n-Nitrosodiphenylamine	40.000	36.971	7.6	98	0.00
66	4-Bromophenyl-phenylether	40.000	36.845	7.9	99	0.00
67	Hexachlorobenzene	40.000	35.260	11.9	100	0.00
68	Atrazine	40.000	35.033	12.4	100	0.00
69 C	Pentachlorophenol	40.000	40.528	-1.3	100	0.00
70	Phenanthrene	40.000	41.253	-3.1	101	0.00
71	Anthracene	40.000	37.193	7.0	100	0.00
72	Carbazole	40.000	37.279	6.8	102	0.00
73	Di-n-butylphthalate	40.000	39.749	0.6	104	0.00
74 C	Fluoranthene	40.000	35.953	10.1	96	0.00
75 I	Chrysene-d12	20.000	20.000	0.0	106	0.00
76	Benzidine	40.000	38.011	5.0	100	0.00
77	Pyrene	40.000	35.271	11.8	100	0.00
78 S	Terphenyl-d14	80.000	67.618	15.5	102	0.00
79	Butylbenzylphthalate	40.000	36.266	9.3	100	0.00
80	Benzo(a)anthracene	40.000	38.373	4.1	101	0.00
81	3,3'-Dichlorobenzidine	40.000	38.116	4.7	104	0.00
82	Chrysene	40.000	36.120	9.7	97	0.00
83	Bis(2-ethylhexyl)phthalate	40.000	37.723	5.7	98	0.00
84 c	Di-n-octyl phthalate	40.000	37.871	5.3	95	0.00
85	Indeno(1,2,3-cd)pyrene	40.000	38.562	3.6	98	0.00
86 I	Perylene-d12	20.000	20.000	0.0	98	0.00
87	Benzo(b)fluoranthene	40.000	41.207	-3.0	100	0.00

Data Path : Z:\HPCHEM1\BNA_F\DATA\BF092016\
Data File : BF090138.D
Acq On : 20 Sep 2016 17:10
Operator : UM/SJ
Sample : SSTDCCC040
Misc :
ALS Vial : 5 Sample Multiplier: 1

Instrument :
BNA_F
LabSampleId :
SSTDCCC040

Quant Time: Sep 21 01:50:35 2016
Quant Method : Z:\HPCHEM1\BNA_F\METHODS\8270-BF092016.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Tue Sep 20 17:58:29 2016
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	Amount	Calc.	%Dev	Area%	Dev(min)
88	Benzo(k)fluoranthene	40.000	33.928	15.2	91	0.00
89 C	Benzo(a)pyrene	40.000	37.512	6.2	96	0.00
90	Dibenzo(a,h)anthracene	40.000	39.111	2.2	99	0.00
91	Benzo(g,h,i)perylene	40.000	37.895	5.3	98	0.00

(#) = Out of Range

SPCC's out = 0 CCC's out = 0