

Data Path : Z:\svoasrv\HPCHEM1\BNA_F\Data\BF092024\
 Data File : BF139513.D
 Acq On : 20 Sep 2024 20:45
 Operator : RC/JU
 Sample : P4120-14
 Misc :
 ALS Vial : 22 Sample Multiplier: 1

Instrument :
 BNA_F
 ClientSampleId :
 RB24085

Quant Time: Sep 20 23:23:16 2024
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_F\Methods\8270-BF091324.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Sat Sep 14 02:58:23 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Dichlorobenzene-d4	6.881	152	85041	20.000	ng	0.00
21) Naphthalene-d8	8.157	136	304799	20.000	ng	# 0.00
39) Acenaphthene-d10	9.910	164	149285	20.000	ng	0.00
64) Phenanthrene-d10	11.398	188	264907	20.000	ng	0.00
76) Chrysene-d12	14.033	240	138468	20.000	ng	# 0.00
86) Perylene-d12	15.504	264	110184	20.000	ng	0.00
System Monitoring Compounds						
5) 2-Fluorophenol	5.504	112	548797	105.276	ng	0.01
7) Phenol-d6	6.510	99	678151	99.111	ng	0.00
23) Nitrobenzene-d5	7.439	82	516707	79.696	ng	0.00
42) 2,4,6-Tribromophenol	10.704	330	169821	106.851	ng	0.00
45) 2-Fluorobiphenyl	9.233	172	880212	82.695	ng	0.00
79) Terphenyl-d14	12.986	244	857868	101.691	ng	0.00


Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA_F\Data\BF092024\
Data File : BF139513.D
Acq On : 20 Sep 2024 20:45
Operator : RC/JU
Sample : P4120-14
Misc :
ALS Vial : 22 Sample Multiplier: 1

Instrument :
BNA_F
ClientSampleId :
RB24085


Quant Time: Sep 20 23:23:16 2024
Quant Method : Z:\svoasrv\HPCHEM1\BNA_F\Methods\8270-BF091324.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Sat Sep 14 02:58:23 2024
Response via : Initial Calibration








#1
1,4-Dichlorobenzene-d4
Concen: 20.000 ng
RT: 6.881 min Scan# 814
Delta R.T. -0.000 min
Lab File: BF139513.D
Acq: 20 Sep 2024 20:45

Instrument :
BNA_F
ClientSampleId :
RB24085







Tgt Ion:152 Resp: 85041
Ion Ratio Lower Upper
152 100
150 155.8 124.1 186.1
115 58.8 48.1 72.1



#5
2-Fluorophenol
Concen: 105.276 ng
RT: 5.504 min Scan# 580
Delta R.T. 0.011 min
Lab File: BF139513.D
Acq: 20 Sep 2024 20:45

Tgt Ion:112 Resp: 548797
Ion Ratio Lower Upper
112 100
64 70.5 54.8 82.2
63 39.3 33.0 49.4





#7

Phenol-d6

Concen: 99.111 ng

RT: 6.510 min Scan# 70

Delta R.T. 0.006 min

Lab File: BF139513.D

Acq: 20 Sep 2024 20:45

Instrument :

BNA_F

ClientSampleId :

RB24085





Tgt Ion: 99 Resp: 678151

Ion Ratio Lower Upper

99 100

42 24.3 21.9 32.9

71 38.5 34.6 52.0



#21

Naphthalene-d8

Concen: 20.000 ng

RT: 8.157 min Scan# 1031

Delta R.T. -0.006 min

Lab File: BF139513.D

Acq: 20 Sep 2024 20:45

Tgt Ion: 136 Resp: 304799




Ion Ratio Lower Upper

136 100

137 10.6 8.6 12.8

54 9.9 7.8 11.8

68 5.4 3.6 5.4#





#23

Nitrobenzene-d5

Concen: 79.696 ng

RT: 7.439 min Scan# 90

Delta R.T. -0.006 min

Lab File: BF139513.D

Acq: 20 Sep 2024 20:45

Instrument :

BNA_F

ClientSampleId :

RB24085





Tgt Ion: 82 Resp: 516707

Ion Ratio Lower Upper

82 100

128 40.7 31.8 47.6

54 61.7 50.6 75.8



#39

Acenaphthene-d10

Concen: 20.000 ng

RT: 9.910 min Scan# 1329

Delta R.T. -0.006 min

Lab File: BF139513.D

Acq: 20 Sep 2024 20:45




Tgt Ion:164 Resp: 149285


Ion Ratio Lower Upper

164 100

162 101.2 81.3 121.9


160 43.7 36.4 54.6








#42
2,4,6-Tribromophenol
Concen: 106.851 ng
RT: 10.704 min Scan# 1464
Delta R.T. 0.006 min
Lab File: BF139513.D
Acq: 20 Sep 2024 20:45

Instrument :
BNA_F
ClientSampleId :
RB24085






Tgt Ion:330 Resp: 169821
Ion Ratio Lower Upper
330 100
332 97.3 75.2 112.8
141 37.1 31.2 46.8



#45
2-Fluorobiphenyl
Concen: 82.695 ng
RT: 9.233 min Scan# 1214
Delta R.T. -0.000 min
Lab File: BF139513.D
Acq: 20 Sep 2024 20:45

Tgt Ion:172 Resp: 880212
Ion Ratio Lower Upper
172 100
171 35.3 28.3 42.5
170 23.4 19.0 28.6









#64
Phenanthrene-d10
Concen: 20.000 ng
RT: 11.398 min Scan# 11
Delta R.T. -0.000 min
Lab File: BF139513.D
Acq: 20 Sep 2024 20:45




Instrument :
BNA_F
ClientSampleId :
RB24085


Tgt Ion:188 Resp: 264907
Ion Ratio Lower Upper
188 100
94 8.1 5.5 8.3
80 8.6 6.6 9.8



#76
Chrysene-d12
Concen: 20.000 ng
RT: 14.033 min Scan# 2030
Delta R.T. -0.000 min
Lab File: BF139513.D
Acq: 20 Sep 2024 20:45


Tgt Ion:240 Resp: 138468
Ion Ratio Lower Upper
240 100
120 9.6 6.2 9.2#
236 24.7 20.9 31.3








#79
 Terphenyl-d14
 Concen: 101.691 ng
 RT: 12.986 min Scan# 11
 Delta R.T. 0.006 min
 Lab File: BF139513.D
 Acq: 20 Sep 2024 20:45

Instrument :
 BNA_F
 ClientSampleId :
 RB24085



Tgt Ion:244 Resp: 857868
 Ion Ratio Lower Upper
 244 100
 212 7.1 6.0 9.0
 122 9.8 7.6 11.4



#86
 Perylene-d12
 Concen: 20.000 ng
 RT: 15.504 min Scan# 2280
 Delta R.T. 0.006 min
 Lab File: BF139513.D
 Acq: 20 Sep 2024 20:45

Tgt Ion:264 Resp: 110184
 Ion Ratio Lower Upper
 264 100
 260 23.0 19.5 29.3
 265 21.1 16.6 25.0

