

Data Path : Z:\HPCHEM1\BNA_F\DATA\BF092115\
 Data File : BF081710.D
 Acq On : 21 Sep 2015 12:20
 Operator : UM/IZ
 Sample : SSTDCCC040
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_F
 ClientSampleId :
 SSTDCCC040

Quant Time: Sep 22 00:59:36 2015
 Quant Method : Z:\HPCHEM1\BNA_F\METHODS\8270-BF090115.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Tue Sep 15 14:03:59 2015
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Dichlorobenzene-d4	8.17	152	95741	20.00	ng	-0.04
21) Naphthalene-d8	11.08	136	377898	20.00	ng	-0.04
38) Acenaphthene-d10	15.21	164	182616	20.00	ng	-0.04
63) Phenanthrene-d10	18.71	188	375335	20.00	ng	-0.04
75) Chrysene-d12	23.36	240	308643	20.00	ng	-0.02
86) Perylene-d12	24.79	264	205541	20.00	ng	-0.02

System Monitoring Compounds

5) 2-Fluorophenol	5.32	112	453149	73.43	ng	-0.03
7) Phenol-d6	7.64	99	658622	80.63	ng	-0.01
23) Nitrobenzene-d5	9.51	82	662917	84.64	ng	-0.02
41) 2,4,6-Tribromophenol	17.13	330	140096	86.16	ng	-0.03
44) 2-Fluorobiphenyl	13.74	172	1186197	83.24	ng	-0.03
78) Terphenyl-d14	22.08	244	1064419	80.28	ng	-0.03

Target Compounds

						Qvalue
2) 1,4-Dioxane	1.52	88	178364	64.36	ng	# 57
3) Pyridine	2.02	79	242372	31.71	ng	# 83
4) n-Nitrosodimethylamine	2.02	42	200169	47.15	ng	# 61
6) Aniline	7.50	93	448724	38.90	ng	# 82
8) 2-Chlorophenol	7.74	128	272637	40.09	ng	# 84
9) Benzaldehyde	7.20	77	165367	32.31	ng	# 73
10) Phenol	7.66	94	355130	37.49	ng	# 49
11) bis(2-Chloroethyl)ether	7.73	93	271995	39.80	ng	# 66
12) 1,3-Dichlorobenzene	8.02	146	285814	39.34	ng	# 90
13) 1,4-Dichlorobenzene	8.21	146	290732	39.30	ng	# 89
14) 1,2-Dichlorobenzene	8.54	146	279974	42.03	ng	# 91
15) Benzyl Alcohol	8.64	79	280246	41.82	ng	# 69
16) 2,2'-oxybis(1-Chloropropan	8.94	45	536838	40.98	ng	# 93
17) 2-Methylphenol	8.98	107	222153	40.94	ng	# 77
18) Hexachloroethane	9.26	117	120261	41.79	ng	# 87
19) n-Nitroso-di-n-propylamine	9.28	70	236374	42.52	ng	# 91
20) 3+4-Methylphenols	9.36	107	290591	39.72	ng	# 85
22) Acetophenone	9.19	105	407223	39.45	ng	# 72
24) Nitrobenzene	9.54	77	341298	41.96	ng	# 63
25) Isophorone	10.14	82	606865	41.66	ng	# 84
26) 2-Nitrophenol	10.26	139	144060	40.64	ng	# 37
27) 2,4-Dimethylphenol	10.55	122	246118	40.19	ng	# 77
28) bis(2-Chloroethoxy)methane	10.72	93	339498	40.34	ng	# 93
29) 2,4-Dichlorophenol	10.87	162	212724	40.06	ng	# 90
30) 1,2,4-Trichlorobenzene	10.98	180	242178	41.98	ng	# 96
31) Naphthalene	11.13	128	808767	40.13	ng	# 98
32) Benzoic acid	11.09	122	69773	17.66	ng	# 44
33) 4-Chloroaniline	11.37	127	334276	40.66	ng	# 80
34) Hexachlorobutadiene	11.48	225	154686	44.06	ng	# 100
35) Caprolactam	12.44	113	54652	33.56	ng	# 14
36) 4-Chloro-3-methylphenol	12.71	107	266808	44.89	ng	# 76
37) 2-Methylnaphthalene	12.78	142	542931	41.40	ng	# 90
39) 1,2,4,5-Tetrachlorobenzene	13.19	216	240139	41.58	ng	# 98
40) Hexachlorocyclopentadiene	13.16	237	107950	38.09	ng	# 97


Data Path : Z:\HPCHEM1\BNA_F\DATA\BF092115\
 Data File : BF081710.D
 Acq On : 21 Sep 2015 12:20
 Operator : UM/IZ
 Sample : SSTDCCC040
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_F
 ClientSampleId :
 SSTDCCC040

Quant Time: Sep 22 00:59:36 2015
 Quant Method : Z:\HPCHEM1\BNA_F\METHODS\8270-BF090115.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Tue Sep 15 14:03:59 2015
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
42) 2,4,6-Trichlorophenol	13.55	196	160878	40.36	ng	99
43) 2,4,5-Trichlorophenol	13.66	196	162750	40.02	ng #	91
45) 1,1'-Biphenyl	13.93	154	654078	38.93	ng	96
46) 2-Chloronaphthalene	13.92	162	487689	40.20	ng #	84
47) 2-Nitroaniline	14.29	65	220971	44.69	ng #	62
48) Acenaphthylene	14.87	152	844014	39.21	ng	98
49) Dimethylphthalate	14.84	163	585003	41.23	ng #	97
50) 2,6-Dinitrotoluene	14.93	165	119121	36.99	ng #	18
51) Acenaphthene	15.31	154	475540	38.84	ng	90
52) 3-Nitroaniline	15.29	138	153022	41.74	ng #	45
53) 2,4-Dinitrophenol	15.58	184	50372	34.80	ng #	25
54) Dibenzofuran	15.72	168	710050	39.71	ng #	83
55) 4-Nitrophenol	15.91	139	75372	32.73	ng #	1
56) 2,4-Dinitrotoluene	15.89	165	173863	42.40	ng #	72
57) Fluorene	16.53	166	581997	42.89	ng	95
58) 2,3,4,6-Tetrachlorophenol	16.09	232	129398	41.68	ng	91
59) Diethylphthalate	16.52	149	647496	43.38	ng	96
60) 4-Chlorophenyl-phenylether	16.62	204	273686	43.28	ng #	84
61) 4-Nitroaniline	16.80	138	140427	37.91	ng #	22
62) Azobenzene	17.00	77	692436	41.73	ng	84
64) 4,6-Dinitro-2-methylphenol	16.86	198	85366	40.66	ng	84
65) n-Nitrosodiphenylamine	16.96	169	520294	38.10	ng	94
66) 4-Bromophenyl-phenylether	17.76	248	155850	41.07	ng #	88
67) Hexachlorobenzene	17.82	284	158881	40.04	ng #	91
68) Atrazine	18.37	200	140933	35.66	ng	80
69) Pentachlorophenol	18.37	266	53586	33.45	ng	97
70) Phenanthrene	18.78	178	823337	39.46	ng	97
71) Anthracene	18.91	178	821152	38.30	ng	98
72) Carbazole	19.39	167	768725	38.21	ng	95
73) Di-n-butylphthalate	20.43	149	1012588	42.63	ng #	98
74) Fluoranthene	21.40	202	919421	40.70	ng	91
76) Benzidine	21.72	184	430183	32.47	ng	97
77) Pyrene	21.75	202	867170	37.93	ng	98
79) Butylbenzylphthalate	22.78	149	435231	43.77	ng	92
80) Benzo(a)anthracene	23.35	228	764598	38.90	ng	97
81) 3,3'-Dichlorobenzidine	23.36	252	255603	38.55	ng #	92
82) Chrysene	23.40	228	685170	38.89	ng	94
83) Bis(2-ethylhexyl)phthalate	23.47	149	559176	42.61	ng #	91
84) Di-n-octyl phthalate	24.13	149	882803	40.86	ng	98
85) Indeno(1,2,3-cd)pyrene	25.84	276	524169	40.55	ng #	85
87) Benzo(b)fluoranthene	24.48	252	1061624	72.35	ng #	88
88) Benzo(k)fluoranthene	24.48	252	1061624	87.19	ng #	89
89) Benzo(a)pyrene	24.75	252	468411	37.67	ng #	88
90) Dibenzo(a,h)anthracene	25.85	278	441219	45.92	ng #	79
91) Benzo(g,h,i)perylene	26.14	276	414947	45.25	ng #	74





(#) = qualifier out of range (m) = manual integration (+) = signals summed



#1
 1,4-Dichlorobenzene-d4
 Concen: 20.00 ng
 RT: 8.17 min Scan# 620
 Delta R.T. -0.04 min
 Lab File: BF081710.D
 Acq: 21 Sep 2015 12:20




Instrument :
 BNA_F
 ClientSampleId :
 SSTDCCC040


Tgt Ion:152 Resp: 95741
 Ion Ratio Lower Upper
 152 100
 150 151.6 124.7 187.1
 115 63.4 39.0 58.4#



#2
 1,4-Dioxane
 Concen: 64.36 ng
 RT: 1.52 min Scan# 38
 Delta R.T. -0.04 min
 Lab File: BF081710.D
 Acq: 21 Sep 2015 12:20

Tgt Ion: 88 Resp: 178364
 Ion Ratio Lower Upper
 88 100
 58 46.6 77.7 116.5#
 43 22.9 26.6 40.0#





#3

Pyridine

Concen: 31.71 ng

RT: 2.02 min Scan# 82

Delta R.T. -0.06 min

Lab File: BF081710.D

Acq: 21 Sep 2015 12:20

Instrument :

BNA_F

ClientSampleId :

SSTDCCC040


Tgt Ion: 79 Resp: 242372

Ion Ratio Lower Upper

79 100

52 108.7 76.8 115.2




51 56.0 32.2 48.4#



Abundance Ion 79.00 (78.70 to 79.70): BF0

Ion 52.00 (51.70 to 52.70): BF0

Ion 51.00 (50.70 to 51.70): BF0



#4

n-Nitrosodimethylamine

Concen: 47.15 ng

RT: 2.02 min Scan# 82

Delta R.T. -0.03 min

Lab File: BF081710.D

Acq: 21 Sep 2015 12:20


Tgt Ion: 42 Resp: 200169

Ion Ratio Lower Upper

42 100

74 83.6 105.0 157.6#



44 6.4 6.6 10.0#



Abundance Ion 42.00 (41.70 to 42.70): BF0

Ion 74.00 (73.70 to 74.70): BF0

Ion 44.00 (43.70 to 44.70): BF0





#5

2-Fluorophenol

Concen: 73.43 ng

RT: 5.32 min Scan# 370

Delta R.T. -0.03 min

Lab File: BF081710.D

Acq: 21 Sep 2015 12:20

Instrument :

BNA_F

ClientSampleId :

SSTDCCC040


Tgt Ion: 112 Resp: 453149

Ion Ratio Lower Upper

112 100

64 75.2 39.7 59.5#




63 42.5 17.4 26.0#



Abundance Ion 112.00 (111.70 to 112.70): E

Ion 64.00 (63.70 to 64.70): BF0

Ion 63.00 (62.70 to 63.70): BF0



#6

Aniline

Concen: 38.90 ng

RT: 7.50 min Scan# 561

Delta R.T. -0.02 min

Lab File: BF081710.D

Acq: 21 Sep 2015 12:20


Tgt Ion: 93 Resp: 448724

Ion Ratio Lower Upper

93 100

66 45.9 27.2 40.8#



65 23.9 14.7 22.1#




Abundance Ion 93.00 (92.70 to 93.70): BF0

Ion 66.00 (65.70 to 66.70): BF0

Ion 65.00 (64.70 to 65.70): BF0





#7

Phenol-d6

Concen: 80.63 ng

RT: 7.64 min Scan# 573

Delta R.T. -0.01 min

Lab File: BF081710.D

Acq: 21 Sep 2015 12:20

Instrument :

BNA_F

ClientSampleId :

SSTDCCC040


Tgt Ion: 99 Resp: 658622

Ion Ratio Lower Upper

99 100

42 26.0 13.7 20.5#




71 40.1 14.2 21.2#



Abundance Ion 99.00 (98.70 to 99.70): BF0

Ion 42.00 (41.70 to 42.70): BF0

Ion 71.00 (70.70 to 71.70): BF0



#8

2-Chlorophenol

Concen: 40.09 ng

RT: 7.74 min Scan# 582

Delta R.T. -0.03 min

Lab File: BF081710.D

Acq: 21 Sep 2015 12:20


Tgt Ion: 128 Resp: 272637

Ion Ratio Lower Upper

128 100

130 32.1 12.2 52.2



64 58.6 20.5 60.5




Abundance Ion 128.00 (127.70 to 128.70): E

Ion 130.00 (129.70 to 130.70): E

Ion 64.00 (63.70 to 64.70): BF0





#9

Benzaldehyde

Concen: 32.31 ng

RT: 7.20 min Scan# 535

Delta R.T. -0.03 min

Lab File: BF081710.D

Acq: 21 Sep 2015 12:20

Instrument :

BNA_F

ClientSampleId :

SSTDCCC040





Tgt Ion: 77 Resp: 165367

Ion Ratio Lower Upper

77 100

105 88.1 91.6 131.6#

106 86.1 102.6 142.6#



#10

Phenol

Concen: 37.49 ng

RT: 7.66 min Scan# 575

Delta R.T. -0.01 min

Lab File: BF081710.D

Acq: 21 Sep 2015 12:20




Tgt Ion: 94 Resp: 355130

Ion Ratio Lower Upper

94 100

65 36.6 0.7 40.7

66 51.9 0.8 40.8#









#11
 bis(2-Chloroethyl)ether
 Concen: 39.80 ng
 RT: 7.73 min Scan# 581
 Delta R.T. -0.03 min
 Lab File: BF081710.D
 Acq: 21 Sep 2015 12:20




Instrument :
 BNA_F
 ClientSampleId :
 SSTDCCC040


Tgt Ion: 93 Resp: 271995
 Ion Ratio Lower Upper
 93 100
 63 128.3 65.6 105.6#
 95 31.7 13.6 53.6



#12
 1,3-Dichlorobenzene
 Concen: 39.34 ng
 RT: 8.02 min Scan# 607
 Delta R.T. -0.04 min
 Lab File: BF081710.D
 Acq: 21 Sep 2015 12:20

Tgt Ion: 146 Resp: 285814
 Ion Ratio Lower Upper
 146 100
 148 64.0 51.0 76.4
 75 38.4 15.0 22.6#





#13

1,4-Dichlorobenzene

Concen: 39.30 ng

RT: 8.21 min Scan# 623

Delta R.T. -0.04 min

Lab File: BF081710.D

Acq: 21 Sep 2015 12:20

Instrument :

BNA_F

ClientSampleId :

SSTDCCC040


Tgt Ion:146 Resp: 290732

Ion Ratio Lower Upper




146 100

148 63.8 51.8 77.6

111 47.8 24.9 37.3#



Abundance Ion 146.00 (145.70 to 146.70): E
Ion 148.00 (147.70 to 148.70): E
Ion 111.00 (110.70 to 111.70): E



#14

1,2-Dichlorobenzene

Concen: 42.03 ng

RT: 8.54 min Scan# 652

Delta R.T. -0.03 min

Lab File: BF081710.D

Acq: 21 Sep 2015 12:20


Tgt Ion:146 Resp: 279974

Ion Ratio Lower Upper



146 100

148 63.0 52.7 79.1

111 46.9 28.8 43.2#



Abundance Ion 146.00 (145.70 to 146.70): E
Ion 148.00 (147.70 to 148.70): E
Ion 111.00 (110.70 to 111.70): E









#15
Benzyl Alcohol
Concen: 41.82 ng
RT: 8.64 min Scan# 661
Delta R.T. -0.02 min
Lab File: BF081710.D
Acq: 21 Sep 2015 12:20




Instrument :
BNA_F
ClientSampleId :
SSTDCCC040


Tgt Ion: 79 Resp: 280246
Ion Ratio Lower Upper
79 100
108 65.0 88.6 132.8#
77 64.6 47.0 70.4



#16
2,2'-oxybis(1-Chloropropane)
Concen: 40.98 ng
RT: 8.94 min Scan# 687
Delta R.T. -0.04 min
Lab File: BF081710.D
Acq: 21 Sep 2015 12:20

Tgt Ion: 45 Resp: 536838
Ion Ratio Lower Upper
45 100
77 10.4 0.0 28.1
79 8.3 0.0 26.0









#17
 2-Methylphenol
 Concen: 40.94 ng
 RT: 8.98 min Scan# 691
 Delta R.T. -0.02 min
 Lab File: BF081710.D
 Acq: 21 Sep 2015 12:20




Instrument :
 BNA_F
 ClientSampleId :
 SSTDCCC040


Tgt Ion	Ratio	Lower	Upper
107	100		
108	112.9	96.6	145.0
77	60.2	23.3	34.9#
79	55.5	22.0	33.0#



#18
 Hexachloroethane
 Concen: 41.79 ng
 RT: 9.26 min Scan# 715
 Delta R.T. -0.06 min
 Lab File: BF081710.D
 Acq: 21 Sep 2015 12:20

Tgt Ion	Ratio	Lower	Upper
117	100		
119	94.2	83.8	125.6
201	83.9	55.1	82.7#





#19

n-Nitroso-di-n-propylamine

Concen: 42.52 ng

RT: 9.28 min Scan# 717

Delta R.T. -0.01 min

Lab File: BF081710.D

Acq: 21 Sep 2015 12:20

Instrument :

BNA_F

ClientSampleId :

SSTDCCC040

Tgt Ion: 70 Resp: 236374


Ion Ratio Lower Upper

70 100

42 79.3 63.8 95.6

101 8.4 9.2 13.8#

130 16.3 27.3 40.9#






Abundance Ion 70.00 (69.70 to 70.70): BF0

Ion 42.00 (41.70 to 42.70): BF0

Ion 101.00 (100.70 to 101.70): E

Ion 130.00 (129.70 to 130.70): E



#20

3+4-Methylphenols

Concen: 39.72 ng

RT: 9.36 min Scan# 724

Delta R.T. -0.02 min

Lab File: BF081710.D

Acq: 21 Sep 2015 12:20

Tgt Ion: 107 Resp: 290591


Ion Ratio Lower Upper

107 100

108 89.3 74.6 114.6

77 40.6 0.7 40.7

79 34.9 0.0 38.9






Abundance Ion 107.00 (106.70 to 107.70): E

Ion 108.00 (107.70 to 108.70): E

Ion 77.00 (76.70 to 77.70): BF0

Ion 79.00 (78.70 to 79.70): BF0





#21

Naphthalene-d8

Concen: 20.00 ng

RT: 11.08 min Scan# 874

Delta R.T. -0.04 min

Lab File: BF081710.D

Acq: 21 Sep 2015 12:20

Instrument :

BNA_F

ClientSampleId :

SSTDCCC040

Tgt Ion:136 Resp: 377898


Ion Ratio Lower Upper

136 100

137 11.2 9.0 13.6

54 11.9 8.2 12.2

68 5.0 5.8 8.6#






Abundance Ion 136.00 (135.70 to 136.70): E

Ion 137.00 (136.70 to 137.70): E

Ion 54.00 (53.70 to 54.70): BF0

Ion 68.00 (67.70 to 68.70): BF0



#22

Acetophenone

Concen: 39.45 ng

RT: 9.19 min Scan# 709

Delta R.T. -0.02 min

Lab File: BF081710.D

Acq: 21 Sep 2015 12:20

Tgt Ion:105 Resp: 407223


Ion Ratio Lower Upper

105 100

71 0.0 4.7 7.1#

51 40.1 22.0 33.0#

120 18.5 30.1 45.1#






Abundance Ion 105.00 (104.70 to 105.70): E

Ion 71.00 (70.70 to 71.70): BF0

Ion 51.00 (50.70 to 51.70): BF0

Ion 120.00 (119.70 to 120.70): E





#23

Nitrobenzene-d5

Concen: 84.64 ng

RT: 9.51 min Scan# 737

Delta R.T. -0.02 min

Lab File: BF081710.D

Acq: 21 Sep 2015 12:20





Instrument :

BNA_F

ClientSampleId :

SSTDCCC040

Tgt Ion:	82	Resp:	662917
Ion	Ratio	Lower	Upper
82	100		
128	38.3	51.0	76.6#
54	66.5	47.5	71.3



#24

Nitrobenzene

Concen: 41.96 ng




RT: 9.54 min Scan# 740

Delta R.T. -0.02 min

Lab File: BF081710.D

Acq: 21 Sep 2015 12:20

Tgt Ion:	77	Resp:	341298
Ion	Ratio	Lower	Upper
77	100		
123	39.2	59.8	89.8#
65	16.1	10.2	15.4#





#25

Isophorone

Concen: 41.66 ng

RT: 10.14 min Scan# 792

Delta R.T. -0.02 min

Lab File: BF081710.D

Acq: 21 Sep 2015 12:20

Instrument :

BNA_F

ClientSampleId :

SSTDCCC040


Tgt Ion: 82 Resp: 606865

Ion Ratio Lower Upper

82 100

95 8.2 4.0 6.0#




138 14.5 18.3 27.5#



Abundance Ion 82.00 (81.70 to 82.70): BF0

Ion 95.00 (94.70 to 95.70): BF0

Ion 138.00 (137.70 to 138.70): BF0



#26

2-Nitrophenol

Concen: 40.64 ng

RT: 10.26 min Scan# 803

Delta R.T. -0.03 min

Lab File: BF081710.D

Acq: 21 Sep 2015 12:20


Tgt Ion: 139 Resp: 144060

Ion Ratio Lower Upper

139 100

109 38.4 11.0 16.4#



65 66.2 24.7 37.1#



Abundance Ion 139.00 (138.70 to 139.70): E

Ion 109.00 (108.70 to 109.70): E

Ion 65.00 (64.70 to 65.70): BF0









#27
 2,4-Dimethylphenol
 Concen: 40.19 ng
 RT: 10.55 min Scan# 828
 Delta R.T. -0.02 min
 Lab File: BF081710.D
 Acq: 21 Sep 2015 12:20




Instrument :
 BNA_F
 ClientSampleId :
 SSTDCCC040


Tgt Ion:122 Resp: 246118
 Ion Ratio Lower Upper
 122 100
 107 122.2 71.8 107.6#
 121 56.2 42.3 63.5



#28
 bis(2-Chloroethoxy)methane
 Concen: 40.34 ng
 RT: 10.72 min Scan# 843
 Delta R.T. -0.03 min
 Lab File: BF081710.D
 Acq: 21 Sep 2015 12:20

Tgt Ion: 93 Resp: 339498
 Ion Ratio Lower Upper
 93 100
 95 33.3 27.5 41.3
 123 9.2 13.7 20.5#





#29

2,4-Dichlorophenol

Concen: 40.06 ng

RT: 10.87 min Scan# 856

Delta R.T. -0.03 min

Lab File: BF081710.D

Acq: 21 Sep 2015 12:20

Instrument :

BNA_F

ClientSampleId :

SSTDCCC040


Tgt Ion:162 Resp: 212724

Ion Ratio Lower Upper

162 100

164 62.3 44.9 84.9




98 45.3 13.0 53.0



Abundance Ion 162.00 (161.70 to 162.70): E

Ion 164.00 (163.70 to 164.70): E

Ion 98.00 (97.70 to 98.70): BF0



#30

1,2,4-Trichlorobenzene

Concen: 41.98 ng

RT: 10.98 min Scan# 866

Delta R.T. -0.04 min

Lab File: BF081710.D

Acq: 21 Sep 2015 12:20


Tgt Ion:180 Resp: 242178

Ion Ratio Lower Upper

180 100

182 94.0 77.1 115.7



145 33.3 23.3 34.9




Abundance Ion 180.00 (179.70 to 180.70): E

Ion 182.00 (181.70 to 182.70): E

Ion 145.00 (144.70 to 145.70): E





#31

Naphthalene

Concen: 40.13 ng

RT: 11.13 min Scan# 879

Delta R.T. -0.03 min

Lab File: BF081710.D

Acq: 21 Sep 2015 12:20

Instrument :

BNA_F

ClientSampleId :

SSTDCCC040


Tgt Ion:128 Resp: 808767

Ion Ratio Lower Upper

128 100

129 11.1 8.4 12.6




127 13.3 9.9 14.9



Abundance Ion 128.00 (127.70 to 128.70): E

Ion 129.00 (128.70 to 129.70): E

Ion 127.00 (126.70 to 127.70): E



#32

Benzoic acid

Concen: 17.66 ng

RT: 11.09 min Scan# 875

Delta R.T. 0.00 min

Lab File: BF081710.D

Acq: 21 Sep 2015 12:20


Tgt Ion:122 Resp: 69773

Ion Ratio Lower Upper

122 100

105 137.3 76.1 116.1#



77 120.4 40.5 80.5#




Abundance Ion 122.00 (121.70 to 122.70): E

Ion 105.00 (104.70 to 105.70): E

Ion 77.00 (76.70 to 77.70): BF0









#33
4-Chloroaniline
Concen: 40.66 ng
RT: 11.37 min Scan# 900
Delta R.T. -0.03 min
Lab File: BF081710.D
Acq: 21 Sep 2015 12:20

Instrument :
BNA_F
ClientSampleId :
SSTDCCC040

Tgt Ion:127 Resp: 334276
Ion Ratio Lower Upper
127 100
129 32.7 25.8 38.6
65 43.4 17.9 26.9#
92 22.8 10.5 15.7#




Abundance Ion 127.00 (126.70 to 127.70): E
Ion 129.00 (128.70 to 129.70): E
Ion 65.00 (64.70 to 65.70): BF0
Ion 92.00 (91.70 to 92.70): BF0





#34
Hexachlorobutadiene
Concen: 44.06 ng
RT: 11.48 min Scan# 909
Delta R.T. -0.04 min
Lab File: BF081710.D
Acq: 21 Sep 2015 12:20

Tgt Ion:225 Resp: 154686
Ion Ratio Lower Upper
225 100
223 62.7 50.2 75.2
227 64.9 52.2 78.2



Abundance Ion 225.00 (224.70 to 225.70): E
Ion 223.00 (222.70 to 223.70): E
Ion 227.00 (226.70 to 227.70): E









#35
 Caprolactam
 Concen: 33.56 ng
 RT: 12.44 min Scan# 993
 Delta R.T. 0.05 min
 Lab File: BF081710.D
 Acq: 21 Sep 2015 12:20

Instrument :
 BNA_F
 ClientSampleId :
 SSTDCCC040

Tgt Ion:113 Resp: 54652
 Ion Ratio Lower Upper
 113 100
 55 294.7 152.4 192.4#
 56 217.2 104.6 144.6#




Abundance Ion 113.00 (112.70 to 113.70): E
 Ion 55.00 (54.70 to 55.70): BF0
 Ion 56.00 (55.70 to 56.70): BF0






#36
 4-Chloro-3-methylphenol
 Concen: 44.89 ng
 RT: 12.71 min Scan# 1017
 Delta R.T. -0.01 min
 Lab File: BF081710.D
 Acq: 21 Sep 2015 12:20

Tgt Ion:107 Resp: 266808
 Ion Ratio Lower Upper
 107 100
 144 21.9 25.4 38.0#
 142 70.5 77.3 115.9#



Abundance Ion 107.00 (106.70 to 107.70): E
 Ion 144.00 (143.70 to 144.70): E
 Ion 142.00 (141.70 to 142.70): E









#37
 2-Methylnaphthalene
 Concen: 41.40 ng
 RT: 12.78 min Scan# 1023
 Delta R.T. -0.04 min
 Lab File: BF081710.D
 Acq: 21 Sep 2015 12:20




Instrument :
 BNA_F
 ClientSampleId :
 SSTDCCC040


Tgt Ion:142 Resp: 542931
 Ion Ratio Lower Upper
 142 100
 141 87.2 67.5 101.3
 115 39.6 18.2 27.2#



#38
 Acenaphthene-d10
 Concen: 20.00 ng
 RT: 15.21 min Scan# 1236
 Delta R.T. -0.04 min
 Lab File: BF081710.D
 Acq: 21 Sep 2015 12:20

Tgt Ion:164 Resp: 182616
 Ion Ratio Lower Upper
 164 100
 162 102.2 75.2 112.8
 160 47.6 31.5 47.3#





#39

1,2,4,5-Tetrachlorobenzene

Concen: 41.58 ng

RT: 13.19 min Scan# 1059

Delta R.T. -0.03 min

Lab File: BF081710.D

Acq: 21 Sep 2015 12:20


Instrument :

BNA_F

ClientSampleId :

SSTDCCC040

Tgt Ion:	216	Resp:	240139
Ion	Ratio	Lower	Upper
216	100		
214	77.5	63.5	95.3
179	22.4	16.6	24.8
108	21.3	16.3	24.5






Abundance Ion 216.00 (215.70 to 216.70): E

Ion 214.00 (213.70 to 214.70): E

Ion 179.00 (178.70 to 179.70): E

Ion 108.00 (107.70 to 108.70): E



#40

Hexachlorocyclopentadiene

Concen: 38.09 ng


RT: 13.16 min Scan# 1056

Delta R.T. -0.06 min

Lab File: BF081710.D

Acq: 21 Sep 2015 12:20



Tgt Ion:	237	Resp:	107950
Ion	Ratio	Lower	Upper
237	100		
235	61.8	42.0	82.0
272	11.0	0.0	36.1




Abundance Ion 236.80 (236.50 to 237.50): E

Ion 234.90 (234.60 to 235.60): E

Ion 271.75 (271.45 to 272.45): E









#41
 2,4,6-Tribromophenol
 Concen: 86.16 ng
 RT: 17.13 min Scan# 1404
 Delta R.T. -0.03 min
 Lab File: BF081710.D
 Acq: 21 Sep 2015 12:20

Instrument :
 BNA_F
 ClientSampleId :
 SSTDCCC040

Tgt Ion:330 Resp: 140096
 Ion Ratio Lower Upper
 330 100
 332 97.4 76.6 114.8
 141 48.8 29.7 44.5#




Abundance Ion 329.80 (329.50 to 330.50): E
 Ion 331.80 (331.50 to 332.50): E
 Ion 141.00 (140.70 to 141.70): E





#42
 2,4,6-Trichlorophenol
 Concen: 40.36 ng
 RT: 13.55 min Scan# 1090
 Delta R.T. -0.03 min
 Lab File: BF081710.D
 Acq: 21 Sep 2015 12:20

Tgt Ion:196 Resp: 160878
 Ion Ratio Lower Upper
 196 100
 198 94.1 75.7 113.5
 200 29.2 23.9 35.9



Abundance Ion 196.00 (195.70 to 196.70): E
 Ion 198.00 (197.70 to 198.70): E
 Ion 200.00 (199.70 to 200.70): E









#43
 2,4,5-Trichlorophenol
 Concen: 40.02 ng
 RT: 13.66 min Scan# 1100
 Delta R.T. -0.02 min
 Lab File: BF081710.D
 Acq: 21 Sep 2015 12:20

Instrument :
 BNA_F
 ClientSampleId :
 SSTDCCC040

Tgt Ion:196 Resp: 162750
 Ion Ratio Lower Upper
 196 100
 198 95.7 75.4 113.0
 97 62.5 33.3 49.9#
 132 30.2 24.6 37.0




Abundance Ion 196.00 (195.70 to 196.70): E
 Ion 198.00 (197.70 to 198.70): E
 Ion 97.00 (96.70 to 97.70): BF0
 Ion 132.00 (131.70 to 132.70): E






#44
 2-Fluorobiphenyl
 Concen: 83.24 ng
 RT: 13.74 min Scan# 1107
 Delta R.T. -0.03 min
 Lab File: BF081710.D
 Acq: 21 Sep 2015 12:20

Tgt Ion:172 Resp: 1186197
 Ion Ratio Lower Upper
 172 100
 171 35.2 26.1 39.1
 170 24.3 17.4 26.2



Abundance Ion 172.00 (171.70 to 172.70): E
 Ion 171.00 (170.70 to 171.70): E
 Ion 170.00 (169.70 to 170.70): E









#45
 1,1'-Biphenyl
 Concen: 38.93 ng
 RT: 13.93 min Scan# 1124
 Delta R.T. -0.03 min
 Lab File: BF081710.D
 Acq: 21 Sep 2015 12:20




Instrument :
 BNA_F
 ClientSampleId :
 SSTDCCC040


Tgt Ion:154 Resp: 654078
 Ion Ratio Lower Upper
 154 100
 153 40.4 17.1 57.1
 76 11.8 0.0 31.6



#46
 2-Chloronaphthalene
 Concen: 40.20 ng
 RT: 13.92 min Scan# 1123
 Delta R.T. -0.03 min
 Lab File: BF081710.D
 Acq: 21 Sep 2015 12:20

Tgt Ion:162 Resp: 487689
 Ion Ratio Lower Upper
 162 100
 127 49.9 27.6 41.4#
 164 33.9 25.4 38.2





#47

2-Nitroaniline

Concen: 44.69 ng

RT: 14.29 min Scan# 1155

Delta R.T. -0.02 min

Lab File: BF081710.D

Acq: 21 Sep 2015 12:20

Instrument :

BNA_F

ClientSampleId :

SSTDCCC040


Tgt Ion: 65 Resp: 220971

Ion Ratio Lower Upper




65 100

92 59.5 55.4 83.0

138 82.0 115.5 173.3#



Abundance Ion 65.00 (64.70 to 65.70): BF0
Ion 92.00 (91.70 to 92.70): BF0
Ion 138.00 (137.70 to 138.70): E



#48

Acenaphthylene

Concen: 39.21 ng

RT: 14.87 min Scan# 1206

Delta R.T. -0.04 min

Lab File: BF081710.D

Acq: 21 Sep 2015 12:20


Tgt Ion: 152 Resp: 844014

Ion Ratio Lower Upper



152 100

151 20.1 14.6 22.0

153 12.9 10.3 15.5



Abundance Ion 152.00 (151.70 to 152.70): E
Ion 151.00 (150.70 to 151.70): E
Ion 153.00 (152.70 to 153.70): E





#49

Dimethylphthalate

Concen: 41.23 ng

RT: 14.84 min Scan# 1203

Delta R.T. -0.02 min

Lab File: BF081710.D

Acq: 21 Sep 2015 12:20

Instrument :

BNA_F

ClientSampleId :

SSTDCCC040


Tgt Ion:163 Resp: 585003

Ion Ratio Lower Upper

163 100

194 3.0 4.4 6.6#




164 10.0 8.3 12.5



Abundance Ion 163.00 (162.70 to 163.70): E

Ion 194.00 (193.70 to 194.70): E

Ion 164.00 (163.70 to 164.70): E



#50

2,6-Dinitrotoluene

Concen: 36.99 ng

RT: 14.93 min Scan# 1211

Delta R.T. -0.02 min

Lab File: BF081710.D

Acq: 21 Sep 2015 12:20


Tgt Ion:165 Resp: 119121

Ion Ratio Lower Upper

165 100

63 93.6 32.3 48.5#



89 81.1 28.6 43.0#



Abundance Ion 165.00 (164.70 to 165.70): E

Ion 63.00 (62.70 to 63.70): BF0

Ion 89.00 (88.70 to 89.70): BF0





#51

Acenaphthene

Concen: 38.84 ng

RT: 15.31 min Scan# 1244

Delta R.T. -0.03 min

Lab File: BF081710.D

Acq: 21 Sep 2015 12:20

Instrument :

BNA_F

ClientSampleId :

SSTDCCC040


Tgt Ion:154 Resp: 475540

Ion Ratio Lower Upper

154 100

153 113.2 82.1 123.1




152 53.5 37.9 56.9



Abundance Ion 154.00 (153.70 to 154.70): E

Ion 153.00 (152.70 to 153.70): E

Ion 152.00 (151.70 to 152.70): E



#52

3-Nitroaniline

Concen: 41.74 ng

RT: 15.29 min Scan# 1243

Delta R.T. -0.02 min

Lab File: BF081710.D

Acq: 21 Sep 2015 12:20


Tgt Ion:138 Resp: 153022

Ion Ratio Lower Upper

138 100

108 11.1 5.7 8.5#



92 138.3 67.7 101.5#



Abundance Ion 138.00 (137.70 to 138.70): E

Ion 108.00 (107.70 to 108.70): E

Ion 92.00 (91.70 to 92.70): BF0









#53
 2,4-Dinitrophenol
 Concen: 34.80 ng
 RT: 15.58 min Scan# 1268
 Delta R.T. -0.01 min
 Lab File: BF081710.D
 Acq: 21 Sep 2015 12:20




Instrument :
 BNA_F
 ClientSampleId :
 SSTDCCC040

Tgt Ion:184 Resp: 50372
 Ion Ratio Lower Upper
 184 100
 63 99.2 35.4 53.2#
 154 80.4 32.2 48.4#



#54
 Dibenzofuran
 Concen: 39.71 ng
 RT: 15.72 min Scan# 1280
 Delta R.T. -0.04 min
 Lab File: BF081710.D
 Acq: 21 Sep 2015 12:20

Tgt Ion:168 Resp: 710050
 Ion Ratio Lower Upper
 168 100
 139 43.0 24.2 36.2#
 169 13.1 10.6 15.8









#55
4-Nitrophenol
Concen: 32.73 ng
RT: 15.91 min Scan# 1297
Delta R.T. -0.01 min
Lab File: BF081710.D
Acq: 21 Sep 2015 12:20

Instrument :
BNA_F
ClientSampleId :
SSTDCCC040

Tgt Ion:139 Resp: 75372
Ion Ratio Lower Upper
139 100
109 99.3 12.7 52.7#
65 146.4 45.9 85.9#




Abundance Ion 139.00 (138.70 to 139.70): E
Ion 109.00 (108.70 to 109.70): E
Ion 65.00 (64.70 to 65.70): BFO






#56
2,4-Dinitrotoluene
Concen: 42.40 ng
RT: 15.89 min Scan# 1295
Delta R.T. -0.01 min
Lab File: BF081710.D
Acq: 21 Sep 2015 12:20

Tgt Ion:165 Resp: 173863
Ion Ratio Lower Upper
165 100
63 51.5 29.8 44.6#
89 78.7 44.5 66.7#
182 2.0 3.0 4.4#



Abundance Ion 165.00 (164.70 to 165.70): E
Ion 63.00 (62.70 to 63.70): BFO
Ion 89.00 (88.70 to 89.70): BFO
Ion 182.00 (181.70 to 182.70): E









#57
Fluorene
 Concen: 42.89 ng
 RT: 16.53 min Scan# 1351
 Delta R.T. -0.04 min
 Lab File: BF081710.D
 Acq: 21 Sep 2015 12:20




Instrument :
 BNA_F
 ClientSampleId :
 SSTDCCC040


Tgt Ion:166 Resp: 581997
 Ion Ratio Lower Upper
 166 100
 165 96.1 72.6 109.0
 167 13.5 10.6 16.0



#58
2,3,4,6-Tetrachlorophenol
 Concen: 41.68 ng
 RT: 16.09 min Scan# 1313
 Delta R.T. -0.03 min
 Lab File: BF081710.D
 Acq: 21 Sep 2015 12:20

Tgt Ion:232 Resp: 129398
 Ion Ratio Lower Upper
 232 100
 131 55.5 38.5 57.7
 130 2.5 1.8 2.6
 166 34.2 25.0 37.6









#59
Diethylphthalate
Concen: 43.38 ng
RT: 16.52 min Scan# 1350
Delta R.T. -0.03 min
Lab File: BF081710.D
Acq: 21 Sep 2015 12:20




Instrument :
BNA_F
ClientSampleId :
SSTDCCC040


Tgt Ion	Ratio	Lower	Upper
149	100		
177	19.3	17.5	26.3
150	12.3	9.4	14.2



#60
4-Chlorophenyl-phenylether
Concen: 43.28 ng
RT: 16.62 min Scan# 1359
Delta R.T. -0.04 min
Lab File: BF081710.D
Acq: 21 Sep 2015 12:20

Tgt Ion	Ratio	Lower	Upper
204	100		
206	33.3	24.8	37.2
141	68.5	42.2	63.4









#61
4-Nitroaniline
Concen: 37.91 ng
RT: 16.80 min Scan# 1375
Delta R.T. 0.02 min
Lab File: BF081710.D
Acq: 21 Sep 2015 12:20




Instrument :
BNA_F
ClientSampleId :
SSTDCCC040


Tgt Ion:138 Resp: 140427
Ion Ratio Lower Upper
138 100
92 55.4 12.6 52.6#
108 97.6 12.4 52.4#



#62
Azobenzene
Concen: 41.73 ng
RT: 17.00 min Scan# 1392
Delta R.T. -0.03 min
Lab File: BF081710.D
Acq: 21 Sep 2015 12:20

Tgt Ion: 77 Resp: 692436
Ion Ratio Lower Upper
77 100
182 20.3 15.1 55.1
105 20.6 3.4 43.4
51 40.0 12.0 52.0









#63
Phenanthrene-d10
Concen: 20.00 ng
RT: 18.71 min Scan# 1542
Delta R.T. -0.04 min
Lab File: BF081710.D
Acq: 21 Sep 2015 12:20




Instrument :
BNA_F
ClientSampleId :
SSTDCCC040


Tgt Ion:188 Resp: 375335
Ion Ratio Lower Upper
188 100
94 8.3 11.1 16.7#
80 9.3 11.0 16.4#



#64
4,6-Dinitro-2-methylphenol
Concen: 40.66 ng
RT: 16.86 min Scan# 1380
Delta R.T. -0.01 min
Lab File: BF081710.D
Acq: 21 Sep 2015 12:20

Tgt Ion:198 Resp: 85366
Ion Ratio Lower Upper
198 100
51 77.0 39.6 79.6
105 52.7 28.5 68.5





#65

n-Nitrosodiphenylamine

Concen: 38.10 ng

RT: 16.96 min Scan# 1389

Delta R.T. -0.02 min

Lab File: BF081710.D

Acq: 21 Sep 2015 12:20

Instrument :

BNA_F

ClientSampleId :

SSTDCCC040


Tgt Ion:169 Resp: 520294

Ion Ratio Lower Upper

169 100

168 66.3 48.9 73.3




167 36.0 26.2 39.4



Abundance Ion 169.00 (168.70 to 169.70): E

Ion 168.00 (167.70 to 168.70): E

Ion 167.00 (166.70 to 167.70): E



#66

4-Bromophenyl-phenylether

Concen: 41.07 ng

RT: 17.76 min Scan# 1459

Delta R.T. -0.04 min

Lab File: BF081710.D

Acq: 21 Sep 2015 12:20


Tgt Ion:248 Resp: 155850

Ion Ratio Lower Upper

248 100

250 97.3 78.5 117.7



141 96.8 59.0 88.6#




Abundance Ion 248.00 (247.70 to 248.70): E

Ion 250.00 (249.70 to 250.70): E

Ion 141.00 (140.70 to 141.70): E









#67
Hexachlorobenzene
Concen: 40.04 ng
RT: 17.82 min Scan# 1464
Delta R.T. -0.04 min
Lab File: BF081710.D
Acq: 21 Sep 2015 12:20

Instrument :
BNA_F
ClientSampleId :
SSTDCCC040

Tgt Ion: 284 Resp: 158881
Ion Ratio Lower Upper
284 100
142 38.5 29.5 44.3
249 34.0 19.5 29.3#




Abundance Ion 283.80 (283.50 to 284.50): E
Ion 142.00 (141.70 to 142.70): E
Ion 249.00 (248.70 to 249.70): E






#68
Atrazine
Concen: 35.66 ng
RT: 18.37 min Scan# 1512
Delta R.T. -0.03 min
Lab File: BF081710.D
Acq: 21 Sep 2015 12:20

Tgt Ion: 200 Resp: 140933
Ion Ratio Lower Upper
200 100
173 27.4 13.2 53.2
215 42.6 41.8 81.8



Abundance Ion 200.00 (199.70 to 200.70): E
Ion 173.00 (172.70 to 173.70): E
Ion 215.00 (214.70 to 215.70): E









#69
 Pentachlorophenol
 Concen: 33.45 ng
 RT: 18.37 min Scan# 1512
 Delta R.T. -0.03 min
 Lab File: BF081710.D
 Acq: 21 Sep 2015 12:20




Instrument :
 BNA_F
 ClientSampleId :
 SSTCCCC040


Tgt Ion: 266 Resp: 53586
 Ion Ratio Lower Upper
 266 100
 268 63.1 49.8 74.6
 264 66.3 50.2 75.2



#70
 Phenanthrene
 Concen: 39.46 ng
 RT: 18.78 min Scan# 1548
 Delta R.T. -0.04 min
 Lab File: BF081710.D
 Acq: 21 Sep 2015 12:20

Tgt Ion: 178 Resp: 823337
 Ion Ratio Lower Upper
 178 100
 176 19.5 13.8 20.8
 179 15.5 12.2 18.2





#71

Anthracene

Concen: 38.30 ng

RT: 18.91 min Scan# 1559

Delta R.T. -0.04 min

Lab File: BF081710.D

Acq: 21 Sep 2015 12:20

Instrument :

BNA_F

ClientSampleId :

SSTDCCC040


Tgt Ion:178 Resp: 821152

Ion Ratio Lower Upper

178 100

176 19.0 14.1 21.1




179 15.2 12.2 18.2



Abundance Ion 178.00 (177.70 to 178.70): E

Ion 176.00 (175.70 to 176.70): E

Ion 179.00 (178.70 to 179.70): E



#72

Carbazole

Concen: 38.21 ng

RT: 19.39 min Scan# 1601

Delta R.T. -0.03 min

Lab File: BF081710.D

Acq: 21 Sep 2015 12:20


Tgt Ion:167 Resp: 768725

Ion Ratio Lower Upper

167 100

166 21.4 15.7 23.5



139 14.2 9.5 14.3




Abundance Ion 167.00 (166.70 to 167.70): E

Ion 166.00 (165.70 to 166.70): E

Ion 139.00 (138.70 to 139.70): E





#73

Di-n-butylphthalate

Concen: 42.63 ng

RT: 20.43 min Scan# 1692

Delta R.T. -0.04 min

Lab File: BF081710.D

Acq: 21 Sep 2015 12:20

Instrument :

BNA_F

ClientSampleId :

SSTDCCC040





Tgt Ion:149 Resp: 1012588

Ion Ratio Lower Upper

149 100

150 9.2 7.4 11.2

104 5.5 2.4 3.6#



#74

Fluoranthene

Concen: 40.70 ng

RT: 21.40 min Scan# 1777

Delta R.T. -0.03 min

Lab File: BF081710.D

Acq: 21 Sep 2015 12:20




Tgt Ion:202 Resp: 919421


Ion Ratio Lower Upper

202 100

101 10.3 0.0 38.3

203 17.4 0.0 37.3





#75

Chrysene-d12

Concen: 20.00 ng

RT: 23.36 min Scan# 1949

Delta R.T. -0.02 min

Lab File: BF081710.D

Acq: 21 Sep 2015 12:20

Instrument :

BNA_F

ClientSampleId :

SSTDCCC040


Tgt Ion:240 Resp: 308643

Ion Ratio Lower Upper

240 100

120 11.3 16.2 24.2#




236 25.7 18.4 27.6



Abundance Ion 240.00 (239.70 to 240.70): E

Ion 120.00 (119.70 to 120.70): E

Ion 236.00 (235.70 to 236.70): E



#76

Benzidine

Concen: 32.47 ng

RT: 21.72 min Scan# 1805

Delta R.T. -0.02 min

Lab File: BF081710.D

Acq: 21 Sep 2015 12:20


Tgt Ion:184 Resp: 430183

Ion Ratio Lower Upper

184 100

185 14.1 11.8 17.6



183 11.1 7.6 11.4




Abundance Ion 184.00 (183.70 to 184.70): E

Ion 185.00 (184.70 to 185.70): E

Ion 183.00 (182.70 to 183.70): E









#77
 Pyrene
 Concen: 37.93 ng
 RT: 21.75 min Scan# 1808
 Delta R.T. -0.03 min
 Lab File: BF081710.D
 Acq: 21 Sep 2015 12:20

Instrument :
 BNA_F
 ClientSampleId :
 SSTDCCC040

Tgt Ion: 202 Resp: 867170
 Ion Ratio Lower Upper
 202 100
 200 19.9 15.0 22.4
 203 17.2 14.1 21.1




Abundance Ion 202.00 (201.70 to 202.70): E
 Ion 200.00 (199.70 to 200.70): E
 Ion 203.00 (202.70 to 203.70): E






#78
 Terphenyl-d14
 Concen: 80.28 ng
 RT: 22.08 min Scan# 1837
 Delta R.T. -0.03 min
 Lab File: BF081710.D
 Acq: 21 Sep 2015 12:20

Tgt Ion: 244 Resp: 1064419
 Ion Ratio Lower Upper
 244 100
 212 8.0 4.3 6.5#
 122 13.7 17.4 26.2#



Abundance Ion 244.00 (243.70 to 244.70): E
 Ion 212.00 (211.70 to 212.70): E
 Ion 122.00 (121.70 to 122.70): E





#79

Butylbenzylphthalate

Concen: 43.77 ng

RT: 22.78 min Scan# 1898

Delta R.T. -0.02 min

Lab File: BF081710.D

Acq: 21 Sep 2015 12:20

Instrument :

BNA_F

ClientSampleId :

SSTDCCC040


Tgt Ion:149 Resp: 435231

Ion Ratio Lower Upper

149 100

91 68.8 48.6 72.8




206 18.3 14.9 22.3



Abundance Ion 149.00 (148.70 to 149.70): E

Ion 91.00 (90.70 to 91.70): BF0

Ion 206.00 (205.70 to 206.70): E



#80

Benzo(a)anthracene

Concen: 38.90 ng

RT: 23.35 min Scan# 1948

Delta R.T. -0.02 min

Lab File: BF081710.D

Acq: 21 Sep 2015 12:20


Tgt Ion:228 Resp: 764598

Ion Ratio Lower Upper

228 100

226 27.3 20.0 30.0



229 20.2 15.4 23.2




Abundance Ion 228.00 (227.70 to 228.70): E

Ion 226.00 (225.70 to 226.70): E

Ion 229.00 (228.70 to 229.70): E









#81
 3,3'-Dichlorobenzidine
 Concen: 38.55 ng
 RT: 23.36 min Scan# 1949
 Delta R.T. -0.02 min
 Lab File: BF081710.D
 Acq: 21 Sep 2015 12:20




Instrument :
 BNA_F
 ClientSampleId :
 SSTDCCC040


Tgt Ion	Ratio	Lower	Upper
252	100		
254	65.3	51.4	77.2
126	6.1	16.4	24.6



#82
 Chrysene
 Concen: 38.89 ng
 RT: 23.40 min Scan# 1952
 Delta R.T. -0.02 min
 Lab File: BF081710.D
 Acq: 21 Sep 2015 12:20

Tgt Ion	Ratio	Lower	Upper
228	100		
226	30.7	21.0	31.4
229	20.1	15.6	23.4









#83
 Bis(2-ethylhexyl)phthalate
 Concen: 42.61 ng
 RT: 23.47 min Scan# 1958
 Delta R.T. -0.03 min
 Lab File: BF081710.D
 Acq: 21 Sep 2015 12:20




Instrument :
 BNA_F
 ClientSampleId :
 SSTDCCC040


Tgt Ion:149 Resp: 559176
 Ion Ratio Lower Upper
 149 100
 167 25.2 24.2 36.2
 279 2.7 3.8 5.6#



#84
 Di-n-octyl phthalate
 Concen: 40.86 ng
 RT: 24.13 min Scan# 2016
 Delta R.T. -0.02 min
 Lab File: BF081710.D
 Acq: 21 Sep 2015 12:20

Tgt Ion:149 Resp: 882803
 Ion Ratio Lower Upper
 149 100
 167 1.5 1.0 1.6
 43 13.4 10.2 15.2









#85
 Indeno(1,2,3-cd)pyrene
 Concen: 40.55 ng
 RT: 25.84 min Scan# 2166
 Delta R.T. -0.03 min
 Lab File: BF081710.D
 Acq: 21 Sep 2015 12:20




Instrument :
 BNA_F
 ClientSampleId :
 SSTDCCC040


Tgt Ion: 276 Resp: 524169
 Ion Ratio Lower Upper
 276 100
 138 22.4 25.7 38.5#
 277 24.7 15.6 23.4#



#86
 Perylene-d12
 Concen: 20.00 ng
 RT: 24.79 min Scan# 2074
 Delta R.T. -0.02 min
 Lab File: BF081710.D
 Acq: 21 Sep 2015 12:20

Tgt Ion: 264 Resp: 205541
 Ion Ratio Lower Upper
 264 100
 260 24.4 16.7 25.1
 265 22.2 17.2 25.8









#87
Benzo(b)fluoranthene
Concen: 72.35 ng
RT: 24.48 min Scan# 2047
Delta R.T. 0.00 min
Lab File: BF081710.D
Acq: 21 Sep 2015 12:20




Instrument :
BNA_F
ClientSampleId :
SSTDCCC040


Tgt Ion:252 Resp: 1061624
Ion Ratio Lower Upper
252 100
253 22.6 17.5 26.3
125 11.0 17.1 25.7#



#88
Benzo(k)fluoranthene
Concen: 87.19 ng
RT: 24.48 min Scan# 2047
Delta R.T. -0.02 min
Lab File: BF081710.D
Acq: 21 Sep 2015 12:20

Tgt Ion:252 Resp: 1061624
Ion Ratio Lower Upper
252 100
253 22.6 17.0 25.4
125 11.0 16.0 24.0#









#89
Benzo(a)pyrene
Concen: 37.67 ng
RT: 24.75 min Scan# 2070
Delta R.T. -0.02 min
Lab File: BF081710.D
Acq: 21 Sep 2015 12:20




Instrument :
BNA_F
ClientSampleId :
SSTDCCC040


Tgt Ion: 252 Resp: 468411
Ion Ratio Lower Upper
252 100
253 21.8 17.2 25.8
125 11.3 18.0 27.0#



#90
Dibenzo(a,h)anthracene
Concen: 45.92 ng
RT: 25.85 min Scan# 2167
Delta R.T. -0.03 min
Lab File: BF081710.D
Acq: 21 Sep 2015 12:20

Tgt Ion: 278 Resp: 441219
Ion Ratio Lower Upper
278 100
139 13.8 26.3 39.5#
279 23.7 18.2 27.4





#91

Benzo(g,h,i)perylene

Concen: 45.25 ng

RT: 26.14 min Scan# 2192

Delta R.T. -0.03 min

Lab File: BF081710.D


Acq: 21 Sep 2015 12:20

Instrument :

BNA_F

ClientSampleId :

SSTDCCC040



Tgt Ion: 276 Resp: 414947

Ion Ratio Lower Upper

276 100

277 23.6 17.8 26.6

138 19.2 34.6 51.8#

Abundance Ion 276.00 (275.70 to 276.70): E
Ion 277.00 (276.70 to 277.70): E
Ion 138.00 (137.70 to 138.70): E

