

Data Path : Z:\HPCHEM1\BNA F\DATA\BF092117\
 Data File : BF098683.D
 Acq On : 22 Sep 2017 4:51
 Operator : SJ/JU
 Sample : SSTDICV040
 Misc :
 ALS Vial : 9 Sample Multiplier: 1

Instrument :
 BNA_F
 ClientSampleId :
 ICVBF092117

Quant Time: Sep 22 10:46:57 2017
 Quant Method : Z:\HPCHEM1\BNA F\METHODS\8270-BF092117.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Fri Sep 22 10:38:37 2017
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvqRF	CCRF	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	102	0.00
2	1,4-Dioxane	0.622	0.616	1.0	101	0.00
3	Pyridine	1.653	1.701	-2.9	101	0.00
4	n-Nitrosodimethylamine	0.724	0.717	1.0	100	0.00
5 S	2-Fluorophenol	1.317	1.302	1.1	101	0.00
6	Aniline	2.113	2.091	1.0	100	0.00
7 S	Phenol-d6	1.595	1.570	1.6	100	0.00
8	2-Chlorophenol	1.431	1.442	-0.8	103	0.00
9	Benzaldehyde	0.895	0.907	-1.3	100	0.00
10 C	Phenol	1.760	1.717	2.4	101	0.00
11	bis(2-Chloroethyl)ether	1.356	1.351	0.4	102	0.00
12	1,3-Dichlorobenzene	1.511	1.509	0.1	103	0.00
13 C	1,4-Dichlorobenzene	1.530	1.526	0.3	102	0.00
14	1,2-Dichlorobenzene	1.426	1.422	0.3	101	0.00
15	Benzyl Alcohol	1.151	1.152	-0.1	101	0.00
16	2,2'-oxybis(1-Chloropropane	2.161	2.140	1.0	101	0.00
17	2-Methylphenol	1.166	1.155	0.9	101	0.00
18	Hexachloroethane	0.527	0.540	-2.5	104	0.00
19 P	n-Nitroso-di-n-propylamine	0.988	0.977	1.1	103	0.00
20	3+4-Methylphenols	1.503	1.485	1.2	100	0.00
21 I	Naphthalene-d8	1.000	1.000	0.0	101	0.00
22	Acetophenone	0.495	0.480	3.0	103	0.00
23 S	Nitrobenzene-d5	0.295	0.316	-7.1	106	0.00
24	Nitrobenzene	0.341	0.356	-4.4	104	0.00
25	Isophorone	0.643	0.634	1.4	103	0.00
26 C	2-Nitrophenol	0.138	0.158	-14.5	113	0.00
27	2,4-Dimethylphenol	0.316	0.316	0.0	103	0.00
28	bis(2-Chloroethoxy)methane	0.400	0.398	0.5	103	0.00
29 C	2,4-Dichlorophenol	0.272	0.275	-1.1	103	0.00
30	1,2,4-Trichlorobenzene	0.274	0.273	0.4	104	0.00
31	Naphthalene	0.949	0.931	1.9	102	0.00
32	Benzoic acid	0.169	0.234	-38.5#	122	0.00
33	4-Chloroaniline	0.395	0.383	3.0	99	0.00
34 C	Hexachlorobutadiene	0.148	0.146	1.4	103	0.00
35	Caprolactam	0.085	0.087	-2.4	100	0.00
36 C	4-Chloro-3-methylphenol	0.310	0.308	0.6	102	0.00
37	2-Methylnaphthalene	0.610	0.603	1.1	102	0.00
38 I	Acenaphthene-d10	1.000	1.000	0.0	103	0.00
39	1,2,4,5-Tetrachlorobenzene	0.592	0.586	1.0	103	0.00
40 P	Hexachlorocyclopentadiene	0.257	0.278	-8.2	108	0.00
41 S	2,4,6-Tribromophenol	0.183	0.194	-6.0	104	0.00
42 C	2,4,6-Trichlorophenol	0.399	0.411	-3.0	104	0.00
43	2,4,5-Trichlorophenol	0.409	0.416	-1.7	102	0.00
44 S	2-Fluorobiphenyl	1.257	1.219	3.0	104	0.00

Data Path : Z:\HPCHEM1\BNA F\DATA\BF092117\
 Data File : BF098683.D
 Acq On : 22 Sep 2017 4:51
 Operator : SJ/JU
 Sample : SSTDICV040
 Misc :
 ALS Vial : 9 Sample Multiplier: 1

Instrument :
 BNA_F
 ClientSampleId :
 ICVBF092117

Quant Time: Sep 22 10:46:57 2017
 Quant Method : Z:\HPCHEM1\BNA F\METHODS\8270-BF092117.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Fri Sep 22 10:38:37 2017
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvqRF	CCRF	%Dev	Area%	Dev(min)
45	1,1'-Biphenyl	1.718	1.685	1.9	103	0.00
46	2-Chloronaphthalene	1.288	1.279	0.7	104	0.00
47	2-Nitroaniline	0.362	0.383	-5.8	106	0.00
48	Acenaphthylene	1.982	1.955	1.4	102	0.00
49	Dimethylphthalate	1.500	1.487	0.9	104	0.00
50	2,6-Dinitrotoluene	0.296	0.332	-12.2	107	0.00
51 C	Acenaphthene	1.230	1.210	1.6	103	0.00
52	3-Nitroaniline	0.353	0.381	-7.9	105	0.00
53 P	2,4-Dinitrophenol	0.094	0.115	-22.3	128	0.00
54	Dibenzofuran	1.680	1.652	1.7	104	0.00
55 P	4-Nitrophenol	0.307	0.322	-4.9	106	0.00
56	2,4-Dinitrotoluene	0.398	0.428	-7.5	108	0.00
57	Fluorene	1.350	1.304	3.4	102	0.00
58	2,3,4,6-Tetrachlorophenol	0.291	0.307	-5.5	105	0.00
59	Diethylphthalate	1.470	1.456	1.0	103	0.00
60	4-Chlorophenyl-phenylether	0.591	0.574	2.9	102	0.00
61	4-Nitroaniline	0.338	0.373	-10.4	105	0.00
62	Azobenzene	1.375	1.358	1.2	103	0.00
63 I	Phenanthrene-d10	1.000	1.000	0.0	102	0.00
64	4,6-Dinitro-2-methylphenol	0.087	0.101	-16.1	116	0.00
65 c	n-Nitrosodiphenylamine	0.688	0.677	1.6	104	0.00
66	4-Bromophenyl-phenylether	0.202	0.204	-1.0	104	0.00
67	Hexachlorobenzene	0.230	0.231	-0.4	105	0.00
68	Atrazine	0.207	0.205	1.0	103	0.00
69 C	Pentachlorophenol	0.132	0.146	-10.6	105	0.00
70	Phenanthrene	1.085	1.059	2.4	102	0.00
71	Anthracene	1.106	1.080	2.4	104	0.00
72	Carbazole	1.081	1.056	2.3	102	0.00
73	Di-n-butylphthalate	1.256	1.292	-2.9	106	0.00
74 C	Fluoranthene	1.081	1.068	1.2	105	0.00
75 I	Chrysene-d12	1.000	1.000	0.0	107	0.00
76	Benzidine	0.812	0.741	8.7	97	0.00
77	Pyrene	1.485	1.464	1.4	105	0.00
78 S	Terphenyl-d14	0.931	0.895	3.9	104	0.00
79	Butylbenzylphthalate	0.614	0.674	-9.8	110	0.00
80	Benzo(a)anthracene	1.205	1.185	1.7	107	0.00
81	3,3'-Dichlorobenzidine	0.430	0.426	0.9	104	0.00
82	Chrysene	1.163	1.130	2.8	107	0.00
83	Bis(2-ethylhexyl)phthalate	0.850	0.900	-5.9	111	0.00
84 c	Di-n-octyl phthalate	1.276	1.394	-9.2	116	0.00
85	Indeno(1,2,3-cd)pyrene	0.902	0.903	-0.1	114	0.00
86 I	Perylene-d12	1.000	1.000	0.0	115	0.00
87	Benzo(b)fluoranthene	1.160	1.149	0.9	113	0.00

Data Path : Z:\HPCHEM1\BNA F\DATA\BF092117\
Data File : BF098683.D
Acq On : 22 Sep 2017 4:51
Operator : SJ/JU
Sample : SSTDICV040
Misc :
ALS Vial : 9 Sample Multiplier: 1

Instrument :
BNA_F
ClientSampleId :
ICVBF092117

Quant Time: Sep 22 10:46:57 2017
Quant Method : Z:\HPCHEM1\BNA F\METHODS\8270-BF092117.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Fri Sep 22 10:38:37 2017
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
88	Benzo(k)fluoranthene	1.159	1.165	-0.5	115	0.00
89 C	Benzo(a)pyrene	1.053	1.044	0.9	112	0.00
90	Dibenzo(a,h)anthracene	0.891	0.903	-1.3	114	0.00
91	Benzo(a,h,i)perylene	0.896	0.883	1.5	110	0.00

(#) = Out of Range

SPCC's out = 0 CCC's out = 0