

Data Path : Z:\svoasrv\HPCHEM1\BNA_F\Data\BF092122\
 Data File : BF130368.D
 Acq On : 21 Sep 2022 18:47
 Operator : CG\JU
 Sample : N4724-08
 Misc :
 ALS Vial : 13 Sample Multiplier: 1

Instrument :
 BNA_F
 ClientSampleId :
 SB-01-COMP

Quant Time: Sep 22 05:58:39 2022
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_F\Methods\8270-BF091422.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Mon Sep 19 15:06:14 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Dichlorobenzene-d4	6.651	152	109226	20.000	ng	-0.01
21) Naphthalene-d8	7.933	136	435730	20.000	ng	-0.01
39) Acenaphthene-d10	9.686	164	215311	20.000	ng	0.00
64) Phenanthrene-d10	11.163	188	307783	20.000	ng	-0.01
76) Chrysene-d12	13.798	240	207309	20.000	ng	0.00
86) Perylene-d12	15.192	264	217888	20.000	ng	0.00
System Monitoring Compounds						
5) 2-Fluorophenol	5.269	112	900984	143.073	ng	0.01
7) Phenol-d6	6.298	99	1129466	143.342	ng	0.00
23) Nitrobenzene-d5	7.222	82	719497	84.360	ng	-0.01
42) 2,4,6-Tribromophenol	10.474	330	277594	134.553	ng	0.00
45) 2-Fluorobiphenyl	9.016	172	1325097	89.619	ng	0.00
79) Terphenyl-d14	12.757	244	988102	78.954	ng	0.00

Target Compounds Qvalue

 (#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA_F\Data\BF092122\
Data File : BF130368.D
Acq On : 21 Sep 2022 18:47
Operator : CG\JU
Sample : N4724-08
Misc :
ALS Vial : 13 Sample Multiplier: 1

Instrument :
BNA_F
ClientSampleId :
SB-01-COMP

Quant Time: Sep 22 05:58:39 2022
Quant Method : Z:\svoasrv\HPCHEM1\BNA_F\Methods\8270-BF091422.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Mon Sep 19 15:06:14 2022
Response via : Initial Calibration

