

Data Path : Z:\HPCHEM1\BNA F\DATA\BF092215\
 Data File : BF081767.D
 Acq On : 23 Sep 2015 11:55
 Operator : UM/IZ
 Sample : G3708-08
 Misc :
 ALS Vial : 25 Sample Multiplier: 1

Instrument :
 BNA_F
 ClientSampleId :
 MGP201-19-20

Manual Integrations
 APPROVED

MMDadoda
 9/23/2015 4:34:35 PM

Quant Time: Sep 23 14:12:01 2015
 Quant Method : Z:\HPCHEM1\BNA F\METHODS\8270-BF090115.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Tue Sep 22 17:42:58 2015
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Dichlorobenzene-d4	8.14	152	40078	20.00	ng	0.00
21) Naphthalene-d8	11.19	136	146550	20.00	ng	0.15
38) Acenaphthene-d10	15.19	164	65925	20.00	ng	0.02
63) Phenanthrene-d10	18.70	188	116103m	20.00	ng	0.02
75) Chrysene-d12	23.35	240	81983	20.00	ng	0.02
86) Perylene-d12	24.78	264	68352	20.00	ng	0.01
System Monitoring Compounds						
5) 2-Fluorophenol	5.30	112	363164	140.59	ng	0.02
7) Phenol-d6	7.59	99	501821	146.76	ng	0.01
23) Nitrobenzene-d5	9.45	82	333496	109.79	ng	0.00
41) 2,4,6-Tribromophenol	17.11	330	85843	146.24	ng	0.03
44) 2-Fluorobiphenyl	13.71	172	557394	108.35	ng	0.01
78) Terphenyl-d14	22.06	244	366469	104.06	ng	0.01
Target Compounds						
31) Naphthalene	11.29	128	26080434	3336.93	ng	# 92
37) 2-Methylnaphthalene	12.89	142	9067127	1782.72	ng	# 90
45) 1,1'-Biphenyl	13.92	154	1056083	174.12	ng	96
48) Acenaphthylene	14.93	152	5321348	684.82	ng	95
49) Dimethylphthalate	14.80	163	57683	11.26	ng	# 98
51) Acenaphthene	15.29	154	620642	140.40	ng	# 85
54) Dibenzofuran	15.69	168	431622	66.87	ng	# 85
57) Fluorene	16.57	166	2594015	529.57	ng	92
70) Phenanthrene	18.87	178	7355694	1139.59	ng	95
71) Anthracene	18.96	178	2137168m	322.28	ng	
72) Carbazole	19.36	167	86298	13.87	ng	# 94
74) Fluoranthene	21.42	202	1904554	272.57	ng	91
77) Pyrene	21.79	202	3014998	496.47	ng	95
80) Benzo(a)anthracene	23.35	228	887019	169.90	ng	93
82) Chrysene	23.39	228	644765	137.77	ng	92
85) Indeno(1,2,3-cd)pyrene	25.82	276	163148	47.51	ng	# 82
87) Benzo(b)fluoranthene	24.45	252	466993m	95.70	ng	
88) Benzo(k)fluoranthene	24.46	252	190556m	47.06	ng	
89) Benzo(a)pyrene	24.75	252	527847	127.65	ng	# 87
90) Dibenzo(a,h)anthracene	25.82	278	50109	15.68	ng	# 78
91) Benzo(g,h,i)perylene	26.12	276	136582	44.79	ng	# 73

(#) = qualifier out of range (m) = manual integration (+) = signals summed


Data Path : Z:\HPCHEM1\BNA F\DATA\BF092215\
Data File : BF081767.D
Acq On : 23 Sep 2015 11:55
Operator : UM/IZ
Sample : G3708-08
Misc :
ALS Vial : 25 Sample Multiplier: 1


Instrument :
BNA_F
ClientSampleId :
MGP201-19-20

Manual Integrations
APPROVED

MMDadoda
9/23/2015 4:34:35 PM

Quant Time: Sep 23 14:12:01 2015
Quant Method : Z:\HPCHEM1\BNA F\METHODS\8270-BF090115.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Tue Sep 22 17:42:58 2015
Response via : Initial Calibration





#1


1,4-Dichlorobenzene-d4
Concen: 20.00 ng
RT: 8.14 min Scan# 617
Delta R.T. 0.00 min
Lab File: BF081767.D
Acq: 23 Sep 2015 11:55

Instrument :
BNA_F
ClientSampleId :
MGP201-19-20




Tot Ion	Ratio	Lower	Upper
152	100		
150	153.8	124.7	187.1
115	67.2	39.0	58.4#

Manual Integrations
APPROVED

MMDadoda
9/23/2015 4:34:35 PM




Abundance Ion 152.00 (151.70 to 152.70): E
Ion 150.00 (149.70 to 150.70): E
Ion 115.00 (114.70 to 115.70): E





#5


2-Fluorophenol
Concen: 140.59 ng
RT: 5.30 min Scan# 369
Delta R.T. 0.02 min
Lab File: BF081767.D
Acq: 23 Sep 2015 11:55

Tgt Ion	Ratio	Lower	Upper
112	100		
64	79.8	39.7	59.5#
63	44.0	17.4	26.0#



Abundance Ion 112.00 (111.70 to 112.70): E
Ion 64.00 (63.70 to 64.70): BF0
Ion 63.00 (62.70 to 63.70): BF0





#7

Phenol-d6

Concen: 146.76 ng

RT: 7.59 min Scan# 569

Delta R.T. 0.01 min

Lab File: BF081767.D

Acq: 23 Sep 2015 11:55

Instrument :

BNA_F





ClientSampleId :

MGP201-19-20

Tot Ion:	99	Resp:	501821
Ion Ratio	Lower	Upper	
99	100		
42	25.3	13.7	20.5#
71	40.0	14.2	21.2#

Manual Integrations
APPROVED

MMDadoda
9/23/2015 4:34:35 PM



#21

Naphthalene-d8

Concen: 20.00 ng




RT: 11.19 min Scan# 884


Delta R.T. 0.15 min

Lab File: BF081767.D

Acq: 23 Sep 2015 11:55

Tgt Ion:	136	Resp:	146550
Ion Ratio	Lower	Upper	
136	100		
137	11.0	9.0	13.6
54	11.8	8.2	12.2
68	7.5	5.8	8.6









#23
 Nitrobenzene-d5
 Concen: 109.79 ng
 RT: 9.45 min Scan# 732
 Delta R.T. 0.00 min
 Lab File: BF081767.D
 Acq: 23 Sep 2015 11:55

Instrument :
 BNA_F
 ClientSampleId :
 MGP201-19-20

Tot Ion	Ratio	Lower	Upper
82	100		
128	37.9	51.0	76.6#
54	69.3	47.5	71.3




Manual Integrations
 APPROVED


MMDadoda
 9/23/2015 4:34:35 PM



#31
 Naphthalene
 Concen: 3336.93 ng
 RT: 11.29 min Scan# 893
 Delta R.T. 0.21 min
 Lab File: BF081767.D
 Acq: 23 Sep 2015 11:55

Tgt Ion	Ratio	Lower	Upper
128	100		
129	13.0	8.4	12.6#
127	15.6	9.9	14.9#









#37
 2-Methylnaphthalene
 Concen: 1782.72 ng
 RT: 12.89 min Scan# 1033
 Delta R.T. 0.15 min
 Lab File: BF081767.D
 Acq: 23 Sep 2015 11:55

Instrument :
 BNA_F
 ClientSampleId :
 MGP201-19-20

Tot Ion	Ratio	Lower	Upper
142	100		
141	88.0	67.5	101.3
115	37.5	18.2	27.2




Manual Integrations
 APPROVED


MMDadoda
 9/23/2015 4:34:35 PM



#38
 Acenaphthene-d10
 Concen: 20.00 ng
 RT: 15.19 min Scan# 1234
 Delta R.T. 0.02 min
 Lab File: BF081767.D
 Acq: 23 Sep 2015 11:55

Tgt Ion	Ratio	Lower	Upper
164	100		
162	99.8	75.2	112.8
160	46.7	31.5	47.3









#41
 2,4,6-Tribromophenol
 Concen: 146.24 ng
 RT: 17.11 min Scan# 1402
 Delta R.T. 0.03 min
 Lab File: BF081767.D
 Acq: 23 Sep 2015 11:55

Instrument :
 BNA_F
 ClientSampleId :
 MGP201-19-20

Tot Ion	Ratio	Lower	Upper
330	100		
332	96.4	76.6	114.8
141	58.1	29.7	44.5




Manual Integrations
 APPROVED


MMDadoda
 9/23/2015 4:34:35 PM



#44
 2-Fluorobiphenyl
 Concen: 108.35 ng
 RT: 13.71 min Scan# 1104
 Delta R.T. 0.01 min
 Lab File: BF081767.D
 Acq: 23 Sep 2015 11:55

Tgt Ion	Ratio	Lower	Upper
172	100		
171	35.4	26.1	39.1
170	23.9	17.4	26.2





#45

1,1'-Biphenyl

Concen: 174.12 ng

RT: 13.92 min Scan# 1123

Delta R.T. 0.03 min

Lab File: BF081767.D

Acq: 23 Sep 2015 11:55

Instrument :

BNA_F

ClientSampleId :

MGP201-19-20

Tot Ion:154 Resp: 1056083

Ion Ratio Lower Upper

154 100


153 40.2 17.1 57.1

76 12.4 0.0 31.6

Manual Integrations
APPROVED

MMDadoda




9/23/2015 4:34:35 PM



Abundance Ion 154.00 (153.70 to 154.70): E

Ion 153.00 (152.70 to 153.70): E

Ion 76.00 (75.70 to 76.70): BF0



#48

Acenaphthylene

Concen: 684.82 ng

RT: 14.93 min Scan# 1211

Delta R.T. 0.10 min

Lab File: BF081767.D

Acq: 23 Sep 2015 11:55


Tgt Ion:152 Resp: 5321348

Ion Ratio Lower Upper

152 100

151 20.9 14.6 22.0



153 14.2 10.3 15.5




Abundance Ion 152.00 (151.70 to 152.70): E

Ion 151.00 (150.70 to 151.70): E

Ion 153.00 (152.70 to 153.70): E









#49
Dimethylphthalate
Concen: 11.26 ng
RT: 14.80 min Scan# 1200
Delta R.T. 0.02 min
Lab File: BF081767.D
Acq: 23 Sep 2015 11:55

Instrument :
BNA_F
ClientSampleId :
MGP201-19-20

Tot Ion	Ratio	Lower	Upper
163	100		
194	3.5	4.4	6.6#
164	10.4	8.3	12.5




Manual Integrations
APPROVED


MMDadoda
9/23/2015 4:34:35 PM



#51
Acenaphthene
Concen: 140.40 ng
RT: 15.29 min Scan# 1243
Delta R.T. 0.05 min
Lab File: BF081767.D
Acq: 23 Sep 2015 11:55

Tgt Ion	Ratio	Lower	Upper
154	100		
153	116.2	82.1	123.1
152	59.9	37.9	56.9#









#54
Dibenzenofuran
Concen: 66.87 ng
RT: 15.69 min Scan# 1278
Delta R.T. 0.02 min
Lab File: BF081767.D
Acq: 23 Sep 2015 11:55

Instrument :
BNA_F
ClientSampleId :
MGP201-19-20

Tot Ion	Ratio	Lower	Upper
168	100		
139	41.9	24.2	36.2
169	13.1	10.6	15.8




Manual Integrations
APPROVED


MMDadoda
9/23/2015 4:34:35 PM



#57
Fluorene
Concen: 529.57 ng
RT: 16.57 min Scan# 1355
Delta R.T. 0.09 min
Lab File: BF081767.D
Acq: 23 Sep 2015 11:55

Tgt Ion	Ratio	Lower	Upper
166	100		
165	98.6	72.6	109.0
167	15.5	10.6	16.0





#63

Phenanthrene-d10

Concen: 20.00 ng/mL

RT: 18.70 min Scan# 1541

Delta R.T. 0.02 min

Lab File: BF081767.D

Acq: 23 Sep 2015 11:55

Instrument :

BNA_F





ClientSampleId :

MGP201-19-20

Tot Ion:	188	Resp:	116103
Ion Ratio	Lower	Upper	
188	100		
94	8.4	11.1	16.7#
80	8.4	11.0	16.4#

Manual Integrations
APPROVED

MMDadoda
9/23/2015 4:34:35 PM



#70

Phenanthrene

Concen: 1139.59 ng/mL




RT: 18.87 min Scan# 1556


Delta R.T. 0.14 min

Lab File: BF081767.D

Acq: 23 Sep 2015 11:55

Tgt Ion:	178	Resp:	7355694
Ion Ratio	Lower	Upper	
178	100		
176	20.5	13.8	20.8
179	16.6	12.2	18.2





#71

Anthracene

Concen: 322.28 ng/mL

RT: 18.96 min Scan# 1564

Delta R.T. 0.10 min

Lab File: BF081767.D

Acq: 23 Sep 2015 11:55

Instrument :

BNA_F

ClientSampleId :

MGP201-19-20

Tot Ion:178 Resp: 2137168

Ion	Ratio	Lower	Upper
178	100		
176	18.8	14.1	21.1
179	17.0	12.2	18.2

178 100


176 18.8 14.1 21.1

179 17.0 12.2 18.2

Manual Integrations
APPROVED

MMDadoda




9/23/2015 4:34:35 PM



Abundance Ion 178.00 (177.70 to 178.70): E

Ion 176.00 (175.70 to 176.70): E

Ion 179.00 (178.70 to 179.70): E



#72

Carbazole

Concen: 13.87 ng/mL

RT: 19.36 min Scan# 1599

Delta R.T. 0.02 min

Lab File: BF081767.D

Acq: 23 Sep 2015 11:55


Tgt Ion:167 Resp: 86298

Ion	Ratio	Lower	Upper
167	100		
166	21.5	15.7	23.5
139	15.4	9.5	14.3

167 100

166 21.5 15.7 23.5



139 15.4 9.5 14.3#




Abundance Ion 167.00 (166.70 to 167.70): E

Ion 166.00 (165.70 to 166.70): E

Ion 139.00 (138.70 to 139.70): E





#74

Fluoranthene

Concen: 272.57 ng

RT: 21.42 min Scan# 1779

Delta R.T. 0.06 min

Lab File: BF081767.D

Acq: 23 Sep 2015 11:55

Instrument :

BNA_F





ClientSampleId :

MGP201-19-20

Tot Ion:	202	Resp:	1904554
Ion Ratio	Lower	Upper	
202	100		
101	11.4	0.0	38.3
203	18.0	0.0	37.3

Manual Integrations
APPROVED

MMDadoda
9/23/2015 4:34:35 PM



#75

Chrysene-d12

Concen: 20.00 ng




RT: 23.35 min Scan# 1948


Delta R.T. 0.02 min

Lab File: BF081767.D

Acq: 23 Sep 2015 11:55

Tgt Ion:	240	Resp:	81983
Ion Ratio	Lower	Upper	
240	100		
120	13.3	16.2	24.2#
236	25.5	18.4	27.6





#77

Pvrene

Concen: 496.47 ng

RT: 21.79 min Scan# 1811

Delta R.T. 0.07 min

Lab File: BF081767.D

Acq: 23 Sep 2015 11:55

Instrument :

BNA_F

ClientSampleId :

MGP201-19-20

Tot Ion:202 Resp: 3014998

Ion Ratio Lower Upper

202 100


200 20.7 15.0 22.4

203 20.2 14.1 21.1

Manual Integrations
APPROVED

MMDadoda




9/23/2015 4:34:35 PM



Abundance Ion 202.00 (201.70 to 202.70): E

Ion 200.00 (199.70 to 200.70): E

Ion 203.00 (202.70 to 203.70): E



#78

Terphenyl-d14

Concen: 104.06 ng

RT: 22.06 min Scan# 1835

Delta R.T. 0.01 min

Lab File: BF081767.D

Acq: 23 Sep 2015 11:55


Tgt Ion:244 Resp: 366469

Ion Ratio Lower Upper

244 100

212 8.2 4.3 6.5#



122 12.8 17.4 26.2#




Abundance Ion 244.00 (243.70 to 244.70): E

Ion 212.00 (211.70 to 212.70): E

Ion 122.00 (121.70 to 122.70): E









#80
Benzo(a)anthracene
Concen: 169.90 ng
RT: 23.35 min Scan# 1948
Delta R.T. 0.03 min
Lab File: BF081767.D
Acq: 23 Sep 2015 11:55

Instrument :
BNA_F
ClientSampleId :
MGP201-19-20

Tot Ion	Ratio	Lower	Upper
228	100		
226	29.5	20.0	30.0
229	21.2	15.4	23.2




Manual Integrations
APPROVED


MMDadoda
9/23/2015 4:34:35 PM



#82
Chrysene
Concen: 137.77 ng
RT: 23.39 min Scan# 1951
Delta R.T. 0.02 min
Lab File: BF081767.D
Acq: 23 Sep 2015 11:55

Tgt Ion	Ratio	Lower	Upper
228	100		
226	30.9	21.0	31.4
229	22.6	15.6	23.4









#85
 Indeno(1,2,3-cd)pyrene
 Concen: 47.51 ng
 RT: 25.82 min Scan# 2164
 Delta R.T. 0.02 min
 Lab File: BF081767.D
 Acq: 23 Sep 2015 11:55

Instrument :
 BNA_F
 ClientSampleId :
 MGP201-19-20

Tot Ion	Ratio	Lower	Upper
276	100		
138	20.1	25.7	38.5#
277	24.9	15.6	23.4#




Manual Integrations
 APPROVED


MMDadoda
 9/23/2015 4:34:35 PM



#86
 Perylene-d12
 Concen: 20.00 ng
 RT: 24.78 min Scan# 2073
 Delta R.T. 0.01 min
 Lab File: BF081767.D
 Acq: 23 Sep 2015 11:55

Tgt Ion	Ratio	Lower	Upper
264	100		
260	25.6	16.7	25.1#
265	24.6	17.2	25.8





#87

Benzo(b)fluoranthene

Concen: 95.70 ng m

RT: 24.45 min Scan# 2044

Delta R.T. 0.02 min

Lab File: BF081767.D

Acq: 23 Sep 2015 11:55

Instrument :

BNA_F





ClientSampleId :

MGP201-19-20

Tot Ion:	252	Resp:	466993
Ion Ratio	Lower	Upper	
252	100		
253	23.1	17.5	26.3
125	11.7	17.1	25.7

Manual Integrations
APPROVED

MMDadoda
9/23/2015 4:34:35 PM



#88

Benzo(k)fluoranthene

Concen: 47.06 ng m




RT: 24.46 min Scan# 2045


Delta R.T. 0.01 min

Lab File: BF081767.D

Acq: 23 Sep 2015 11:55

Tgt Ion:	252	Resp:	190556
Ion Ratio	Lower	Upper	
252	100		
253	23.5	17.0	25.4
125	8.5	16.0	24.0









#89
Benzo(a)pyrene
Concen: 127.65 ng
RT: 24.75 min Scan# 2070
Delta R.T. 0.02 min
Lab File: BF081767.D
Acq: 23 Sep 2015 11:55

Instrument :
BNA_F
ClientSampleId :
MGP201-19-20

Tot Ion	Ratio	Resp	Lower	Upper
252	100			
253	22.5	17.2	25.8	
125	11.0	18.0	27.0	




Manual Integrations
APPROVED


MMDadoda
9/23/2015 4:34:35 PM



#90
Dibenzo(a,h)anthracene
Concen: 15.68 ng
RT: 25.82 min Scan# 2164
Delta R.T. 0.02 min
Lab File: BF081767.D
Acq: 23 Sep 2015 11:55

Tgt Ion	Ratio	Resp	Lower	Upper
278	100			
139	15.2	26.3	39.5	
279	27.3	18.2	27.4	





#91
 Benzo(a,h,i)perylene
 Concen: 44.79 ng
 RT: 26.12 min Scan# 2190
 Delta R.T. 0.03 min
 Lab File: BF081767.D
 Acq: 23 Sep 2015 11:55

Instrument :
 BNA_F
 ClientSampleId :
 MGP201-19-20

Tot Ion	Ratio	Lower	Upper
276	100		
277	24.8	17.8	26.6
138	18.6	34.6	51.8

Manual Integrations
 APPROVED

MMDadoda
 9/23/2015 4:34:35 PM

