


Data Path : Z:\HPCHEM1\BNA F\Data\BF092217\
 Data File : BF098684.D
 Acq On : 22 Sep 2017 7:51
 Operator : SJ/JU
 Sample : DFTPP
 Misc :
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 BNA_F
 ClientSampleId :
 DFTPP

Integration File: rteint.p

Method : Z:\HPCHEM1\BNA_F\METHODS\8270-BF092117.M
 Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 Last Update : Fri Sep 22 10:38:37 2017



AutoFind: Scans 1632, 1633, 1634; Background Corrected with Scan 1627

Target Mass	Rel. to Mass	Lower Limit%	Upper Limit%	Rel. Abn%	Raw Abn	Result Pass/Fail
51	198	10	80	43.9	42640	PASS
68	69	0.00	2	1.7	642	PASS
69	198	0.00	100	38.9	37786	PASS
70	69	0.00	2	0.3	114	PASS
127	198	10	80	51.6	50093	PASS
197	198	0.00	2	0.8	797	PASS
198	198	100	100	100.0	97037	PASS
199	198	5	9	6.6	6368	PASS
275	198	10	60	24.2	23502	PASS
365	198	1	100	2.7	2660	PASS
441	442	0.01	24	15.3	14308	PASS
442	198	50	100	96.4	93506	PASS
443	442	15	24	18.7	17506	PASS