

Data Path : Z:\HPCHEM1\BNA F\DATA\BF092316\
 Data File : BF090207.D
 Acq On : 23 Sep 2016 19:38
 Operator : UM/SJ
 Sample : H4887-03
 Misc :
 ALS Vial : 14 Sample Multiplier: 1

Instrument :
 BNA_F
 ClientSampleId :
 TRACK-D-GRID-17B

Manual Integrations
 APPROVED

sohil
 9/26/2016 5:59:23 PM

Quant Time: Sep 23 23:38:33 2016
 Quant Method : Z:\HPCHEM1\BNA F\METHODS\8270-BF092016.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Fri Sep 23 23:35:22 2016
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Dichlorobenzene-d4	6.49	152	122602	20.00	ng	0.00
21) Naphthalene-d8	7.78	136	453667	20.00	ng	0.00
38) Acenaphthene-d10	9.52	164	232710	20.00	ng	0.00
63) Phenanthrene-d10	10.98	188	270572	20.00	ng	0.00
75) Chrysene-d12	13.60	240	250426	20.00	ng	0.00
86) Perylene-d12	14.93	264	245505	20.00	ng	0.01
System Monitoring Compounds						
5) 2-Fluorophenol	5.06	112	848797	112.85	ng	0.01
7) Phenol-d6	6.16	99	1203213	129.53	ng	0.01
23) Nitrobenzene-d5	7.06	82	607295	77.60	ng	0.00
41) 2,4,6-Tribromophenol	10.31	330	226497	113.45	ng	0.00
44) 2-Fluorobiphenyl	8.86	172	1040103	87.74	ng	0.00
78) Terphenyl-d14	12.57	244	666925	67.61	ng	0.00
Target Compounds						Qvalue
49) Dimethylphthalate	9.26	163	333535	22.50	ng	98
70) Phenanthrene	11.00	178	44454	3.51	ng	98
74) Fluoranthene	12.18	202	38725	2.85	ng	99
77) Pyrene	12.41	202	69864	4.08	ng	97
80) Benzo(a)anthracene	13.59	228	29855m	2.22	ng	
82) Chrysene	13.62	228	36147m	2.75	ng	

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\HPCHEM1\BNA F\DATA\BF092316\
Data File : BF090207.D
Acq On : 23 Sep 2016 19:38
Operator : UM/SJ
Sample : H4887-03
Misc :
ALS Vial : 14 Sample Multiplier: 1

Instrument :
BNA_F
ClientSampleId :
TRACK-D-GRID-17B

Manual Integrations
APPROVED

sohil
9/26/2016 5:59:23 PM

Quant Time: Sep 23 23:38:33 2016
Quant Method : Z:\HPCHEM1\BNA F\METHODS\8270-BF092016.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Fri Sep 23 23:35:22 2016
Response via : Initial Calibration

