


Data Path : Z:\HPCHEM1\BNA F\DATA\BF092316\
 Data File : BF090194.D
 Acq On : 23 Sep 2016 12:39
 Operator : UM/SJ
 Sample : DFTPP
 Misc :
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 BNA_F
 ClientSampleId :
 DFTPP

Integration File: rteint.p

Method : Z:\HPCHEM1\BNA_F\METHODS\8270-BF092016.M
 Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 Last Update : Mon Sep 26 15:58:38 2016



AutoFind: Scans 847, 848, 849; Background Corrected with Scan 844

Target Mass	Rel. to Mass	Lower Limit%	Upper Limit%	Rel. Abn%	Raw Abn	Result Pass/Fail
51	198	10	80	31.6	10928	PASS
68	69	0.00	2	1.5	171	PASS
69	198	0.00	100	32.3	11176	PASS
70	69	0.00	2	0.0	0	PASS
127	198	10	80	50.5	17450	PASS
197	198	0.00	2	0.5	156	PASS
198	198	100	100	100.0	34565	PASS
199	198	5	9	6.2	2141	PASS
275	198	10	60	24.9	8591	PASS
365	198	1	100	2.7	938	PASS
441	442	0.01	24	15.3	4329	PASS
442	198	50	100	82.0	28343	PASS
443	442	15	24	18.8	5323	PASS