




Data Path : Z:\svoasrv\HPCHEM1\BNA_F\Data\BF092321\
Data File : BF125592.D
Acq On : 23 Sep 2021 15:11
Operator : JU/CG
Sample : DFTPP
Misc :
ALS Vial : 1 Sample Multiplier: 1

Quant Time: Sep 23 16:44:19 2021
Quant Method : Z:\SVOASRV\HPCHEM1\BNA_F\METHODS\8270-BF092121.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Tue Sep 21 19:21:36 2021
Response via : Initial Calibration



TIC: BF125592.D\data.ms

(70) Pentachlorophenol (C)




11.222min (+ 0.000) 32497.77 ng

response 399697

Ion	Exp%	Act%
265.70	100.00	100.00
268.00	61.10	62.15
264.00	69.90	63.45
0.00	0.00	0.00

Data Path : Z:\svoasrv\HPCHEM1\BNA_F\Data\BF092321\
 Data File : BF125592.D
 Acq On : 23 Sep 2021 15:11
 Operator : JU/CG
 Sample : DFTPP
 Misc :
 ALS Vial : 1 Sample Multiplier: 1

Quant Time: Sep 23 16:44:19 2021
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA_F\METHODS\8270-BF092121.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Tue Sep 21 19:21:36 2021
 Response via : Initial Calibration



TIC: BF125592.D\data.ms

(77) Benzidine

12.757min (+ 0.006) 0.00 ng

response 1636968

Ion	Exp%	Act%
184.00	100.00	100.00
185.00	13.50	13.79
183.00	10.10	11.63
0.00	0.00	0.00