

Data Path : Z:\HPCHEM1\BNA F\DATA\BF092415\
 Data File : BF081790.D
 Acq On : 25 Sep 2015 6:08
 Operator : UM/IZ
 Sample : G3708-04MS
 Misc :
 ALS Vial : 20 Sample Multiplier: 1

Instrument :
 BNA_F
 ClientSampleId :
 MGP209BR-29-30MS

Quant Time: Sep 25 07:14:57 2015
 Quant Method : Z:\HPCHEM1\BNA F\METHODS\8270-BF092415.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Fri Sep 25 07:13:40 2015
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Dichlorobenzene-d4	6.95	152	150991	20.00	ng	0.00
21) Naphthalene-d8	8.24	136	620561	20.00	ng	0.00
38) Acenaphthene-d10	10.00	164	326361	20.00	ng	0.00
63) Phenanthrene-d10	11.47	188	611691	20.00	ng	0.00
75) Chrysene-d12	14.12	240	404875	20.00	ng	0.00
86) Perylene-d12	15.58	264	343251	20.00	ng	0.00

System Monitoring Compounds	R.T.	QIon	Response	Conc	Units	Dev(Min)
5) 2-Fluorophenol	5.55	112	1068649	119.73	ng	0.01
7) Phenol-d6	6.58	99	1500989	132.96	ng	0.01
23) Nitrobenzene-d5	7.52	82	859833	91.42	ng	0.00
41) 2,4,6-Tribromophenol	10.79	330	434638	153.40	ng	0.00
44) 2-Fluorobiphenyl	9.31	172	1595556	80.25	ng	0.00
78) Terphenyl-d14	13.06	244	1559080	87.96	ng	0.00

Target Compounds	R.T.	QIon	Response	Conc	Units	Qvalue
2) 1,4-Dioxane	2.69	88	156344	35.39	ng	90
3) Pyridine	3.44	79	443795	35.61	ng	87
4) n-Nitrosodimethylamine	3.41	42	237656	40.41	ng	99
6) Aniline	6.62	93	537452	31.80	ng	91
8) 2-Chlorophenol	6.74	128	432897	44.02	ng	92
9) Benzaldehyde	6.50	77	185201	27.75	ng	# 85
10) Phenol	6.59	94	681679	51.99	ng	75
11) bis(2-Chloroethyl)ether	6.70	93	424098	44.10	ng	87
12) 1,3-Dichlorobenzene	6.89	146	471530	40.52	ng	# 93
13) 1,4-Dichlorobenzene	6.97	146	511008	43.40	ng	# 96
14) 1,2-Dichlorobenzene	6.97	146	511008	46.65	ng	97
15) Benzyl Alcohol	7.10	79	440829	49.51	ng	# 80
16) 2,2'-oxybis(1-Chloropropan	7.23	45	769009	46.75	ng	87
17) 2-Methylphenol	7.20	107	374329	45.07	ng	# 85
18) Hexachloroethane	7.46	117	162030	42.26	ng	# 87
19) n-Nitroso-di-n-propylamine	7.37	70	320910	42.64	ng	# 81
20) 3+4-Methylphenols	7.36	107	496462	47.57	ng	# 71
22) Acetophenone	7.37	105	627040	46.07	ng	# 84
24) Nitrobenzene	7.54	77	537721	51.24	ng	# 76
25) Isophorone	7.78	82	956707	45.92	ng	# 93
26) 2-Nitrophenol	7.85	139	208572	54.72	ng	# 53
27) 2,4-Dimethylphenol	7.89	122	385780	40.82	ng	# 80
28) bis(2-Chloroethoxy)methane	7.99	93	572759	47.06	ng	# 93
29) 2,4-Dichlorophenol	8.09	162	422694	48.05	ng	96
30) 1,2,4-Trichlorobenzene	8.18	180	450450	45.62	ng	99
31) Naphthalene	8.26	128	1303461	45.53	ng	96
32) Benzoic acid	7.94	122	17522	9.52	ng	# 70
33) 4-Chloroaniline	8.31	127	308883	24.58	ng	# 92
34) Hexachlorobutadiene	8.38	225	262088	45.00	ng	99
35) Caprolactam	8.69	113	67377	28.67	ng	# 70
36) 4-Chloro-3-methylphenol	8.79	107	482130	54.69	ng	86
37) 2-Methylnaphthalene	8.95	142	881820	43.22	ng	# 90
39) 1,2,4,5-Tetrachlorobenzene	9.12	216	460323	44.19	ng	99
40) Hexachlorocyclopentadiene	9.11	237	439201	98.38	ng	99


Data Path : Z:\HPCHEM1\BNA F\DATA\BF092415\
 Data File : BF081790.D
 Acq On : 25 Sep 2015 6:08
 Operator : UM/IZ
 Sample : G3708-04MS
 Misc :
 ALS Vial : 20 Sample Multiplier: 1

Instrument :
 BNA_F
 ClientSampleId :
 MGP209BR-29-30MS

Quant Time: Sep 25 07:14:57 2015
 Quant Method : Z:\HPCHEM1\BNA F\METHODS\8270-BF092415.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Fri Sep 25 07:13:40 2015
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
42) 2,4,6-Trichlorophenol	9.22	196	335364	48.86	nq	99
43) 2,4,5-Trichlorophenol	9.27	196	331073	49.03	nq	96
45) 1,1'-Biphenyl	9.42	154	1107868	43.42	nq	93
46) 2-Chloronaphthalene	9.44	162	856934	41.75	nq	# 92
47) 2-Nitroaniline	9.53	65	284578	52.13	nq	# 69
48) Acenaphthylene	9.86	152	1516945	43.67	nq	96
49) Dimethylphthalate	9.73	163	1633652	65.78	nq	# 98
50) 2,6-Dinitrotoluene	9.78	165	252277	52.63	nq	# 70
51) Acenaphthene	10.03	154	920520	45.98	nq	90
52) 3-Nitroaniline	9.95	138	192088	36.95	nq	# 96
53) 2,4-Dinitrophenol	10.06	184	73858	80.30	nq	# 75
54) Dibenzofuran	10.21	168	1345162	45.80	nq	# 92
55) 4-Nitrophenol	10.10	139	354218	98.80	nq	# 46
56) 2,4-Dinitrotoluene	10.18	165	293869	51.21	nq	# 70
57) Fluorene	10.55	166	984960	43.70	nq	92
58) 2,3,4,6-Tetrachlorophenol	10.32	232	270016	52.39	nq	98
59) Diethylphthalate	10.42	149	1167474	48.54	nq	95
60) 4-Chlorophenyl-phenylether	10.54	204	478094	46.21	nq	92
61) 4-Nitroaniline	10.57	138	259562	51.39	nq	# 60
62) Azobenzene	10.70	77	1157116	47.05	nq	94
64) 4,6-Dinitro-2-methylphenol	10.59	198	84624	55.85	nq	# 51
65) n-Nitrosodiphenylamine	10.65	169	900978	43.68	nq	91
66) 4-Bromophenyl-phenylether	11.03	248	325987	47.86	nq	91
67) Hexachlorobenzene	11.09	284	310005	43.62	nq	# 91
68) Atrazine	11.19	200	324795	52.66	nq	84
69) Pentachlorophenol	11.28	266	337193	85.42	nq	98
70) Phenanthrene	11.55	178	1351385	38.64	nq	95
71) Anthracene	11.55	178	1351385	40.10	nq	95
72) Carbazole	11.71	167	1399598	44.74	nq	94
73) Di-n-butylphthalate	12.05	149	1721435	45.21	nq	# 99
74) Fluoranthene	12.70	202	1501954	42.62	nq	86
76) Benzidine	12.81	184	399097	29.05	nq	98
77) Pyrene	12.70	202	1501709	49.20	nq	96
79) Butylbenzylphthalate	13.54	149	699086	57.47	nq	# 88
80) Benzo(a)anthracene	14.10	228	1056095	43.02	nq	95
81) 3,3'-Dichlorobenzidine	14.07	252	271597	34.83	nq	# 96
82) Chrysene	14.10	228	1056095	44.85	nq	96
83) Bis(2-ethylhexyl)phthalate	14.09	149	883951	59.93	nq	# 93
84) Di-n-octyl phthalate	14.71	149	1471800	67.85	nq	95
85) Indeno(1,2,3-cd)pyrene	17.05	276	965627	44.80	nq	# 88
87) Benzo(b)fluoranthene	15.16	252	882070	43.33	nq	# 87
88) Benzo(k)fluoranthene	15.16	252	882070	44.99	nq	# 87
89) Benzo(a)pyrene	15.52	252	828625	45.09	nq	# 87
90) Dibenzo(a,h)anthracene	17.08	278	813349	42.47	nq	# 79
91) Benzo(g,h,i)perylene	17.50	276	753548	38.16	nq	# 76

(#) = qualifier out of range (m) = manual integration (+) = signals summed



#1

1,4-Dichlorobenzene-d4

Concen: 20.00 ng

RT: 6.95 min Scan# 421

Delta R.T. -0.00 min

Lab File: BF081790.D

Acq: 25 Sep 2015 6:08

Instrument :

BNA_F

ClientSampleId :

MGP209BR-29-30MS


Tot Ion:152 Resp: 150991

Ion Ratio Lower Upper

152 100

150 158.9 124.7 187.1




115 62.4 39.0 58.4#



Abundance Ion 152.00 (151.70 to 152.70): E

Ion 150.00 (149.70 to 150.70): E

Ion 115.00 (114.70 to 115.70): E



#2

1,4-Dioxane

Concen: 35.39 ng

RT: 2.69 min Scan# 48

Delta R.T. 0.08 min

Lab File: BF081790.D

Acq: 25 Sep 2015 6:08


Tgt Ion: 88 Resp: 156344

Ion Ratio Lower Upper

88 100

58 85.4 77.7 116.5



43 30.6 26.6 40.0




Abundance Ion 88.00 (87.70 to 88.70): BF0

Ion 58.00 (57.70 to 58.70): BF0

Ion 43.00 (42.70 to 43.70): BF0





#3

Pvridine

Concen: 35.61 ng

RT: 3.44 min Scan# 114

Delta R.T. 0.08 min

Lab File: BF081790.D

Acq: 25 Sep 2015 6:08

Instrument :

BNA_F

ClientSampleId :

MGP209BR-29-30MS


Tot Ion: 79 Resp: 443795

Ion Ratio Lower Upper

79 100

52 79.3 76.8 115.2




51 38.3 32.2 48.4



Abundance Ion 79.00 (78.70 to 79.70): BF0

Ion 52.00 (51.70 to 52.70): BF0

Ion 51.00 (50.70 to 51.70): BF0



#4

n-Nitrosodimethylamine

Concen: 40.41 ng

RT: 3.41 min Scan# 111

Delta R.T. 0.08 min

Lab File: BF081790.D

Acq: 25 Sep 2015 6:08


Tgt Ion: 42 Resp: 237656

Ion Ratio Lower Upper

42 100

74 132.5 105.0 157.6



44 8.1 6.6 10.0




Abundance Ion 42.00 (41.70 to 42.70): BF0

Ion 74.00 (73.70 to 74.70): BF0

Ion 44.00 (43.70 to 44.70): BF0





#5

2-Fluorophenol

Concen: 119.73 ng

RT: 5.55 min Scan# 299

Delta R.T. 0.01 min

Lab File: BF081790.D

Acq: 25 Sep 2015 6:08

Instrument :

BNA_F

ClientSampleId :

MGP209BR-29-30MS


Tot Ion:112 Resp: 1068649

Ion Ratio Lower Upper

112 100

64 66.9 39.7 59.5#




63 35.9 17.4 26.0#



Abundance Ion 112.00 (111.70 to 112.70): E

Ion 64.00 (63.70 to 64.70): BF0

Ion 63.00 (62.70 to 63.70): BF0



#6

Aniline

Concen: 31.80 ng

RT: 6.62 min Scan# 392

Delta R.T. -0.00 min

Lab File: BF081790.D

Acq: 25 Sep 2015 6:08


Tgt Ion: 93 Resp: 537452

Ion Ratio Lower Upper

93 100

66 40.2 27.2 40.8



65 20.3 14.7 22.1




Abundance Ion 93.00 (92.70 to 93.70): BF0

Ion 66.00 (65.70 to 66.70): BF0

Ion 65.00 (64.70 to 65.70): BF0





#7

Phenol-d6

Concen: 132.96 ng

RT: 6.58 min Scan# 389

Delta R.T. 0.01 min

Lab File: BF081790.D

Acq: 25 Sep 2015 6:08

Instrument :

BNA_F

ClientSampleId :

MGP209BR-29-30MS


Tot Ion: 99 Resp: 1500989

Ion Ratio Lower Upper

99 100

42 19.2 13.7 20.5




71 34.1 14.2 21.2#



Abundance Ion 99.00 (98.70 to 99.70): BF0

Ion 42.00 (41.70 to 42.70): BF0

Ion 71.00 (70.70 to 71.70): BF0



#8

2-Chlorophenol

Concen: 44.02 ng

RT: 6.74 min Scan# 403

Delta R.T. 0.01 min

Lab File: BF081790.D

Acq: 25 Sep 2015 6:08


Tgt Ion:128 Resp: 432897

Ion Ratio Lower Upper

128 100

130 34.4 12.2 52.2



64 33.4 20.5 60.5




Abundance Ion 128.00 (127.70 to 128.70): E

Ion 130.00 (129.70 to 130.70): E

Ion 64.00 (63.70 to 64.70): BF0





#9

Benzaldehyde

Concen: 27.75 ng

RT: 6.50 min Scan# 382

Delta R.T. -0.00 min

Lab File: BF081790.D

Acq: 25 Sep 2015 6:08

Instrument :

BNA_F

ClientSampleId :

MGP209BR-29-30MS





Tot Ion: 77 Resp: 185201

Ion Ratio Lower Upper

77 100

105 103.7 91.6 131.6

106 98.3 102.6 142.6#



#10

Phenol

Concen: 51.99 ng

RT: 6.59 min Scan# 390

Delta R.T. 0.01 min

Lab File: BF081790.D

Acq: 25 Sep 2015 6:08




Tgt Ion: 94 Resp: 681679


Ion Ratio Lower Upper

94 100

65 28.0 0.7 40.7

66 36.9 0.8 40.8





#11

bis(2-Chloroethyl)ether

Concen: 44.10 ng

RT: 6.70 min Scan# 399

Delta R.T. 0.01 min

Lab File: BF081790.D

Acq: 25 Sep 2015 6:08

Instrument :

BNA_F

ClientSampleId :

MGP209BR-29-30MS


Tot Ion: 93 Resp: 424098

Ion Ratio Lower Upper

93 100

63 69.2 65.6 105.6




95 34.0 13.6 53.6



Abundance Ion 93.00 (92.70 to 93.70): BF0

Ion 63.00 (62.70 to 63.70): BF0

Ion 95.00 (94.70 to 95.70): BF0



#12

1,3-Dichlorobenzene

Concen: 40.52 ng

RT: 6.89 min Scan# 416

Delta R.T. -0.00 min

Lab File: BF081790.D

Acq: 25 Sep 2015 6:08


Tgt Ion: 146 Resp: 471530

Ion Ratio Lower Upper

146 100

148 64.5 51.0 76.4



75 31.5 15.0 22.6#




Abundance Ion 146.00 (145.70 to 146.70): E

Ion 148.00 (147.70 to 148.70): E

Ion 75.00 (74.70 to 75.70): BF0









#13
 1,4-Dichlorobenzene
 Concen: 43.40 ng
 RT: 6.97 min Scan# 423
 Delta R.T. -0.00 min
 Lab File: BF081790.D
 Acq: 25 Sep 2015 6:08




Instrument :
 BNA_F
 ClientSampleId :
 MGP209BR-29-30MS


Tot Ion:146 Resp: 511008
 Ion Ratio Lower Upper
 146 100
 148 64.7 51.8 77.6
 111 38.7 24.9 37.3#



#14
 1,2-Dichlorobenzene
 Concen: 46.65 ng
 RT: 6.97 min Scan# 423
 Delta R.T. -0.15 min
 Lab File: BF081790.D
 Acq: 25 Sep 2015 6:08

Tgt Ion:146 Resp: 511008
 Ion Ratio Lower Upper
 146 100
 148 64.7 52.7 79.1
 111 38.7 28.8 43.2





#15

Benzyl Alcohol

Concen: 49.51 ng

RT: 7.10 min Scan# 434

Delta R.T. 0.01 min

Lab File: BF081790.D

Acq: 25 Sep 2015 6:08

Instrument :

BNA_F

ClientSampleId :

MGP209BR-29-30MS


Tot Ion: 79 Resp: 440829

Ion Ratio Lower Upper

79 100

108 80.6 88.6 132.8#




77 62.4 47.0 70.4



Abundance Ion 79.00 (78.70 to 79.70): BF0

Ion 108.00 (107.70 to 108.70): BF0

Ion 77.00 (76.70 to 77.70): BF0



#16

2,2'-oxybis(1-chloropropane)

Concen: 46.75 ng

RT: 7.23 min Scan# 446

Delta R.T. -0.00 min

Lab File: BF081790.D

Acq: 25 Sep 2015 6:08


Tgt Ion: 45 Resp: 769009

Ion Ratio Lower Upper

45 100

77 12.8 0.0 28.1



79 9.9 0.0 26.0




Abundance Ion 45.00 (44.70 to 45.70): BF0

Ion 77.00 (76.70 to 77.70): BF0

Ion 79.00 (78.70 to 79.70): BF0









#17
2-Methylphenol
Concen: 45.07 ng
RT: 7.20 min Scan# 443
Delta R.T. -0.00 min
Lab File: BF081790.D
Acq: 25 Sep 2015 6:08




Instrument :
BNA_F
ClientSampleId :
MGP209BR-29-30MS


Tot Ion	Ratio	Lower	Upper
107	100		
108	114.3	96.6	145.0
77	46.7	23.3	34.9#
79	45.4	22.0	33.0#



#18
Hexachloroethane
Concen: 42.26 ng
RT: 7.46 min Scan# 466
Delta R.T. -0.00 min
Lab File: BF081790.D
Acq: 25 Sep 2015 6:08

Tgt Ion	Ratio	Lower	Upper
117	100		
119	95.7	83.8	125.6
201	85.3	55.1	82.7#









#19
n-Nitroso-di-n-propylamine
Concen: 42.64 ng
RT: 7.37 min Scan# 458
Delta R.T. -0.00 min
Lab File: BF081790.D
Acq: 25 Sep 2015 6:08




Instrument :
BNA_F
ClientSampleId :
MGP209BR-29-30MS


Tot Ion	Ion	Ratio	Lower	Upper
70	100			
42	64.7	63.8	95.6	
101	9.1	9.2	13.8	
130	18.4	27.3	40.9	



#20
3+4-Methylphenols
Concen: 47.57 ng
RT: 7.36 min Scan# 457
Delta R.T. 0.01 min
Lab File: BF081790.D
Acq: 25 Sep 2015 6:08

Tgt Ion	Ion	Ratio	Lower	Upper
107	100			
108	92.9	74.6	114.6	
77	92.6	0.7	40.7	
79	30.8	0.0	38.9	









#21
Naphthalene-d8
Concen: 20.00 ng
RT: 8.24 min Scan# 534
Delta R.T. -0.00 min
Lab File: BF081790.D
Acq: 25 Sep 2015 6:08




Instrument :
BNA_F
ClientSampleId :
MGP209BR-29-30MS


Tot Ion:136	Resp:	620561
Ion Ratio	Lower	Upper
136	100	
137	11.4	9.0 13.6
54	9.0	8.2 12.2
68	5.9	5.8 8.6



#22
Acetophenone
Concen: 46.07 ng
RT: 7.37 min Scan# 458
Delta R.T. 0.01 min
Lab File: BF081790.D
Acq: 25 Sep 2015 6:08

Tgt Ion:105	Resp:	627040
Ion Ratio	Lower	Upper
105	100	
71	10.0	4.7 7.1#
51	25.9	22.0 33.0
120	22.1	30.1 45.1#





#23

Nitrobenzene-d5

Concen: 91.42 ng

RT: 7.52 min Scan# 471

Delta R.T. -0.00 min

Lab File: BF081790.D

Acq: 25 Sep 2015 6:08

Instrument :

BNA_F

ClientSampleId :

MGP209BR-29-30MS


Tot Ion: 82 Resp: 859833

Ion Ratio Lower Upper

82 100

128 43.5 51.0 76.6#




54 57.8 47.5 71.3



Abundance Ion 82.00 (81.70 to 82.70): BF0

Ion 128.00 (127.70 to 128.70): BF0

Ion 54.00 (53.70 to 54.70): BF0



#24

Nitrobenzene

Concen: 51.24 ng

RT: 7.54 min Scan# 473

Delta R.T. -0.00 min

Lab File: BF081790.D

Acq: 25 Sep 2015 6:08


Tgt Ion: 77 Resp: 537721

Ion Ratio Lower Upper

77 100

123 50.9 59.8 89.8#



65 13.8 10.2 15.4




Abundance Ion 77.00 (76.70 to 77.70): BF0

Ion 123.00 (122.70 to 123.70): BF0

Ion 65.00 (64.70 to 65.70): BF0





#25

Isophorone

Concen: 45.92 ng

RT: 7.78 min Scan# 494

Delta R.T. -0.00 min

Lab File: BF081790.D

Acq: 25 Sep 2015 6:08

Instrument :

BNA_F

ClientSampleId :

MGP209BR-29-30MS





Tot Ion: 82 Resp: 956707

Ion Ratio Lower Upper

82 100

95 7.2 4.0 6.0#

138 19.2 18.3 27.5



#26

2-Nitrophenol

Concen: 54.72 ng

RT: 7.85 min Scan# 500

Delta R.T. -0.00 min

Lab File: BF081790.D

Acq: 25 Sep 2015 6:08




Tgt Ion: 139 Resp: 208572


Ion Ratio Lower Upper

139 100

109 30.6 11.0 16.4#

65 57.7 24.7 37.1#









#27
 2,4-Dimethylphenol
 Concen: 40.82 ng
 RT: 7.89 min Scan# 503
 Delta R.T. -0.00 min
 Lab File: BF081790.D
 Acq: 25 Sep 2015 6:08




Instrument :
 BNA_F
 ClientSampleId :
 MGP209BR-29-30MS


Tot Ion:122 Resp: 385780
 Ion Ratio Lower Upper
 122 100
 107 115.6 71.8 107.6#
 121 57.7 42.3 63.5



#28
 bis(2-Chloroethoxy)methane
 Concen: 47.06 ng
 RT: 7.99 min Scan# 512
 Delta R.T. -0.00 min
 Lab File: BF081790.D
 Acq: 25 Sep 2015 6:08

Tgt Ion: 93 Resp: 572759
 Ion Ratio Lower Upper
 93 100
 95 33.1 27.5 41.3
 123 10.4 13.7 20.5#









#29
 2,4-Dichlorophenol
 Concen: 48.05 ng
 RT: 8.09 min Scan# 521
 Delta R.T. -0.00 min
 Lab File: BF081790.D
 Acq: 25 Sep 2015 6:08




Instrument :
 BNA_F
 ClientSampleId :
 MGP209BR-29-30MS


Tot Ion:162 Resp: 422694
 Ion Ratio Lower Upper
 162 100
 164 64.1 44.9 84.9
 98 38.8 13.0 53.0



#30
 1,2,4-Trichlorobenzene
 Concen: 45.62 ng
 RT: 8.18 min Scan# 529
 Delta R.T. -0.00 min
 Lab File: BF081790.D
 Acq: 25 Sep 2015 6:08

Tgt Ion:180 Resp: 450450
 Ion Ratio Lower Upper
 180 100
 182 95.8 77.1 115.7
 145 28.3 23.3 34.9









#31
Naphthalene
Concen: 45.53 ng
RT: 8.26 min Scan# 536
Delta R.T. -0.00 min
Lab File: BF081790.D
Acq: 25 Sep 2015 6:08

Instrument :
BNA_F
ClientSampleId :
MGP209BR-29-30MS

Tot Ion:128 Resp: 1303461
Ion Ratio Lower Upper
128 100
129 11.7 8.4 12.6
127 14.0 9.9 14.9




Abundance Ion 128.00 (127.70 to 128.70): E
Ion 129.00 (128.70 to 129.70): E
Ion 127.00 (126.70 to 127.70): E






#32
Benzoic acid
Concen: 9.52 ng
RT: 7.94 min Scan# 508
Delta R.T. -0.03 min
Lab File: BF081790.D
Acq: 25 Sep 2015 6:08

Tgt Ion:122 Resp: 17522
Ion Ratio Lower Upper
122 100
105 117.6 76.1 116.1#
77 92.1 40.5 80.5#



Abundance Ion 122.00 (121.70 to 122.70): E
Ion 105.00 (104.70 to 105.70): E
Ion 77.00 (76.70 to 77.70): BF0









#33
4-Chloroaniline
Concen: 24.58 ng
RT: 8.31 min Scan# 540
Delta R.T. -0.00 min
Lab File: BF081790.D
Acq: 25 Sep 2015 6:08




Instrument :
BNA_F
ClientSampleId :
MGP209BR-29-30MS


Tot Ion:	127	Resp:	308883
Ion	Ratio	Lower	Upper
127	100		
129	33.0	25.8	38.6
65	28.9	17.9	26.9#
92	18.4	10.5	15.7#



#34
Hexachlorobutadiene
Concen: 45.00 ng
RT: 8.38 min Scan# 546
Delta R.T. -0.00 min
Lab File: BF081790.D
Acq: 25 Sep 2015 6:08

Tgt Ion:	225	Resp:	262088
Ion	Ratio	Lower	Upper
225	100		
223	63.1	50.2	75.2
227	63.9	52.2	78.2





#35

Caprolactam

Concen: 28.67 ng

RT: 8.69 min Scan# 573

Delta R.T. -0.00 min

Lab File: BF081790.D

Acq: 25 Sep 2015 6:08

Instrument :

BNA_F

ClientSampleId :

MGP209BR-29-30MS


Tot Ion:113 Resp: 67377

Ion Ratio Lower Upper

113 100

55 215.5 152.4 192.4#




56 157.7 104.6 144.6#



Abundance Ion 113.00 (112.70 to 113.70): E

Ion 55.00 (54.70 to 55.70): BF0

Ion 56.00 (55.70 to 56.70): BF0



#36

4-Chloro-3-methylphenol

Concen: 54.69 ng

RT: 8.79 min Scan# 582

Delta R.T. 0.01 min

Lab File: BF081790.D

Acq: 25 Sep 2015 6:08


Tgt Ion:107 Resp: 482130

Ion Ratio Lower Upper

107 100

144 26.8 25.4 38.0



142 81.8 77.3 115.9




Abundance Ion 107.00 (106.70 to 107.70): E

Ion 144.00 (143.70 to 144.70): E

Ion 142.00 (141.70 to 142.70): E









#37
 2-Methylnaphthalene
 Concen: 43.22 ng
 RT: 8.95 min Scan# 596
 Delta R.T. -0.00 min
 Lab File: BF081790.D
 Acq: 25 Sep 2015 6:08

Instrument :
 BNA_F
 ClientSampleId :
 MGP209BR-29-30MS

Tot Ion	Ratio	Lower	Upper
142	100		
141	89.1	67.5	101.3
115	36.7	18.2	27.2#




Abundance Ion 142.00 (141.70 to 142.70): E
 Ion 141.00 (140.70 to 141.70): E
 Ion 115.00 (114.70 to 115.70): E






#38
 Acenaphthene-d10
 Concen: 20.00 ng
 RT: 10.00 min Scan# 688
 Delta R.T. -0.00 min
 Lab File: BF081790.D
 Acq: 25 Sep 2015 6:08

Tgt Ion	Ratio	Lower	Upper
164	100		
162	105.7	75.2	112.8
160	47.3	31.5	47.3#



Abundance Ion 164.00 (163.70 to 164.70): E
 Ion 162.00 (161.70 to 162.70): E
 Ion 160.00 (159.70 to 160.70): E









#39
 1,2,4,5-Tetrachlorobenzene
 Concen: 44.19 ng
 RT: 9.12 min Scan# 611
 Delta R.T. -0.00 min
 Lab File: BF081790.D
 Acq: 25 Sep 2015 6:08

Instrument :
 BNA_F
 ClientSampleId :
 MGP209BR-29-30MS

Tot Ion:216	Resp:	460323
Ion Ratio	Lower	Upper
216	100	
214	79.8	63.5 95.3
179	21.6	16.6 24.8
108	18.7	16.3 24.5




Abundance Ion 216.00 (215.70 to 216.70): E
 Ion 214.00 (213.70 to 214.70): E
 Ion 179.00 (178.70 to 179.70): E
 Ion 108.00 (107.70 to 108.70): E






#40
 Hexachlorocyclopentadiene
 Concen: 98.38 ng
 RT: 9.11 min Scan# 610
 Delta R.T. 0.01 min
 Lab File: BF081790.D
 Acq: 25 Sep 2015 6:08

Tgt Ion:237	Resp:	439201
Ion Ratio	Lower	Upper
237	100	
235	62.3	42.0 82.0
272	13.8	0.0 36.1



Abundance Ion 236.80 (236.50 to 237.50): E
 Ion 234.90 (234.60 to 235.60): E
 Ion 271.75 (271.45 to 272.45): E









#41
 2,4,6-Tribromophenol
 Concen: 153.40 ng
 RT: 10.79 min Scan# 757
 Delta R.T. -0.00 min
 Lab File: BF081790.D
 Acq: 25 Sep 2015 6:08




Instrument :
 BNA_F
 ClientSampleId :
 MGP209BR-29-30MS


Tot Ion	Ratio	Lower	Upper
330	100		
332	94.5	76.6	114.8
141	36.0	29.7	44.5



#42
 2,4,6-Trichlorophenol
 Concen: 48.86 ng
 RT: 9.22 min Scan# 620
 Delta R.T. -0.00 min
 Lab File: BF081790.D
 Acq: 25 Sep 2015 6:08

Tgt Ion	Ratio	Lower	Upper
196	100		
198	93.4	75.7	113.5
200	29.9	23.9	35.9









#43
 2,4,5-Trichlorophenol
 Concen: 49.03 ng
 RT: 9.27 min Scan# 624
 Delta R.T. 0.01 min
 Lab File: BF081790.D
 Acq: 25 Sep 2015 6:08

Instrument :
 BNA_F
 ClientSampleId :
 MGP209BR-29-30MS

Tot Ion	Ion	Ratio	Lower	Upper
196	100			
198	97.4	75.4	113.0	
97	39.9	33.3	49.9	
132	25.6	24.6	37.0	




Abundance Ion 196.00 (195.70 to 196.70): E
 Ion 198.00 (197.70 to 198.70): E
 Ion 97.00 (96.70 to 97.70): BF0
 Ion 132.00 (131.70 to 132.70): E






#44
 2-Fluorobiphenyl
 Concen: 80.25 ng
 RT: 9.31 min Scan# 628
 Delta R.T. -0.00 min
 Lab File: BF081790.D
 Acq: 25 Sep 2015 6:08

Tgt Ion	Ion	Ratio	Lower	Upper
172	100			
171	37.1	26.1	39.1	
170	25.2	17.4	26.2	



Abundance Ion 172.00 (171.70 to 172.70): E
 Ion 171.00 (170.70 to 171.70): E
 Ion 170.00 (169.70 to 170.70): E





#45

1,1'-Biphenyl

Concen: 43.42 ng

RT: 9.42 min Scan# 637

Delta R.T. -0.00 min

Lab File: BF081790.D

Acq: 25 Sep 2015 6:08

Instrument :

BNA_F

ClientSampleId :

MGP209BR-29-30MS


Tot Ion:154 Resp: 1107868

Ion Ratio Lower Upper

154 100

153 41.9 17.1 57.1




76 12.9 0.0 31.6



Abundance Ion 154.00 (153.70 to 154.70): E

Ion 153.00 (152.70 to 153.70): E

Ion 76.00 (75.70 to 76.70): BF0



#46

2-Chloronaphthalene

Concen: 41.75 ng

RT: 9.44 min Scan# 639

Delta R.T. -0.00 min

Lab File: BF081790.D

Acq: 25 Sep 2015 6:08


Tgt Ion:162 Resp: 856934

Ion Ratio Lower Upper

162 100

127 42.0 27.6 41.4#



164 33.5 25.4 38.2




Abundance Ion 162.00 (161.70 to 162.70): E

Ion 127.00 (126.70 to 127.70): E

Ion 164.00 (163.70 to 164.70): E





#47

2-Nitroaniline

Concen: 52.13 ng

RT: 9.53 min Scan# 647

Delta R.T. -0.00 min

Lab File: BF081790.D

Acq: 25 Sep 2015 6:08

Instrument :

BNA_F

ClientSampleId :

MGP209BR-29-30MS


Tot Ion: 65 Resp: 284578

Ion Ratio Lower Upper




65 100

92 62.9 55.4 83.0

138 92.9 115.5 173.3#



Abundance Ion 65.00 (64.70 to 65.70): BF0
Ion 92.00 (91.70 to 92.70): BF0
Ion 138.00 (137.70 to 138.70): E



#48

Acenaphthylene

Concen: 43.67 ng

RT: 9.86 min Scan# 676

Delta R.T. -0.00 min

Lab File: BF081790.D

Acq: 25 Sep 2015 6:08


Tgt Ion: 152 Resp: 1516945

Ion Ratio Lower Upper



152 100


151 21.1 14.6 22.0

153 13.6 10.3 15.5



Abundance Ion 152.00 (151.70 to 152.70): E
Ion 151.00 (150.70 to 151.70): E
Ion 153.00 (152.70 to 153.70): E









#49
Dimethylphthalate
Concen: 65.78 ng
RT: 9.73 min Scan# 664
Delta R.T. 0.01 min
Lab File: BF081790.D
Acq: 25 Sep 2015 6:08

Instrument :
BNA_F
ClientSampleId :
MGP209BR-29-30MS

Tot Ion:163 Resp: 1633652
Ion Ratio Lower Upper
163 100
194 3.9 4.4 6.6#
164 10.7 8.3 12.5




Abundance Ion 163.00 (162.70 to 163.70): E
Ion 194.00 (193.70 to 194.70): E
Ion 164.00 (163.70 to 164.70): E






#50
2,6-Dinitrotoluene
Concen: 52.63 ng
RT: 9.78 min Scan# 669
Delta R.T. -0.00 min
Lab File: BF081790.D
Acq: 25 Sep 2015 6:08

Tgt Ion:165 Resp: 252277
Ion Ratio Lower Upper
165 100
63 57.0 32.3 48.5#
89 55.7 28.6 43.0#



Abundance Ion 165.00 (164.70 to 165.70): E
Ion 63.00 (62.70 to 63.70): BF0
Ion 89.00 (88.70 to 89.70): BF0





#51

Acenaphthene

Concen: 45.98 ng

RT: 10.03 min Scan# 691

Delta R.T. -0.00 min

Lab File: BF081790.D

Acq: 25 Sep 2015 6:08

Instrument :

BNA_F

ClientSampleId :

MGP209BR-29-30MS


Tot Ion:154 Resp: 920520

Ion Ratio Lower Upper

154 100

153 112.8 82.1 123.1




152 55.0 37.9 56.9



Abundance Ion 154.00 (153.70 to 154.70): E

Ion 153.00 (152.70 to 153.70): E

Ion 152.00 (151.70 to 152.70): E



#52

3-Nitroaniline

Concen: 36.95 ng

RT: 9.95 min Scan# 684

Delta R.T. 0.01 min

Lab File: BF081790.D

Acq: 25 Sep 2015 6:08


Tgt Ion:138 Resp: 192088

Ion Ratio Lower Upper

138 100

108 8.7 5.7 8.5#



92 81.4 67.7 101.5




Abundance Ion 138.00 (137.70 to 138.70): E

Ion 108.00 (107.70 to 108.70): E

Ion 92.00 (91.70 to 92.70): BF0









#53
2,4-Dinitrophenol
Concen: 80.30 ng
RT: 10.06 min Scan# 693
Delta R.T. 0.01 min
Lab File: BF081790.D
Acq: 25 Sep 2015 6:08

Instrument :
BNA_F
ClientSampleId :
MGP209BR-29-30MS

Tot Ion:184 Resp: 73858
Ion Ratio Lower Upper
184 100
63 26.6 35.4 53.2#
154 54.4 32.2 48.4#




Abundance Ion 184.00 (183.70 to 184.70): E
Ion 63.00 (62.70 to 63.70): BF0
Ion 154.00 (153.70 to 154.70): E






#54
Dibenzofuran
Concen: 45.80 ng
RT: 10.21 min Scan# 706
Delta R.T. 0.01 min
Lab File: BF081790.D
Acq: 25 Sep 2015 6:08

Tgt Ion:168 Resp: 1345162
Ion Ratio Lower Upper
168 100
139 36.4 24.2 36.2#
169 13.8 10.6 15.8



Abundance Ion 168.00 (167.70 to 168.70): E
Ion 139.00 (138.70 to 139.70): E
Ion 169.00 (168.70 to 169.70): E









#55
4-Nitrophenol
Concen: 98.80 ng
RT: 10.10 min Scan# 697
Delta R.T. 0.01 min
Lab File: BF081790.D
Acq: 25 Sep 2015 6:08




Instrument :
BNA_F
ClientSampleId :
MGP209BR-29-30MS


Tot Ion:139 Resp: 354218
Ion Ratio Lower Upper
139 100
109 71.6 12.7 52.7#
65 102.6 45.9 85.9#



#56
2,4-Dinitrotoluene
Concen: 51.21 ng
RT: 10.18 min Scan# 704
Delta R.T. -0.00 min
Lab File: BF081790.D
Acq: 25 Sep 2015 6:08

Tgt Ion:165 Resp: 293869
Ion Ratio Lower Upper
165 100
63 53.8 29.8 44.6#
89 79.4 44.5 66.7#
182 2.1 3.0 4.4#









#57
Fluorene
 Concen: 43.70 ng
 RT: 10.55 min Scan# 736
 Delta R.T. 0.01 min
 Lab File: BF081790.D
 Acq: 25 Sep 2015 6:08




Instrument :
 BNA_F
 ClientSampleId :
 MGP209BR-29-30MS


Tot Ion:166 Resp: 984960
 Ion Ratio Lower Upper
 166 100
 165 99.6 72.6 109.0
 167 13.3 10.6 16.0



#58
2,3,4,6-Tetrachlorophenol
 Concen: 52.39 ng
 RT: 10.32 min Scan# 716
 Delta R.T. 0.01 min
 Lab File: BF081790.D
 Acq: 25 Sep 2015 6:08

Tgt Ion:232 Resp: 270016
 Ion Ratio Lower Upper
 232 100
 131 45.5 38.5 57.7
 130 2.1 1.8 2.6
 166 31.4 25.0 37.6









#59
Diethylphthalate
Concen: 48.54 ng
RT: 10.42 min Scan# 725
Delta R.T. 0.01 min
Lab File: BF081790.D
Acq: 25 Sep 2015 6:08

Instrument :
BNA_F
ClientSampleId :
MGP209BR-29-30MS

Tot Ion:149 Resp: 1167474
Ion Ratio Lower Upper
149 100
177 24.7 17.5 26.3
150 12.7 9.4 14.2




Abundance Ion 149.00 (148.70 to 149.70): E
Ion 177.00 (176.70 to 177.70): E
Ion 150.00 (149.70 to 150.70): E






#60
4-Chlorophenyl-phenylether
Concen: 46.21 ng
RT: 10.54 min Scan# 735
Delta R.T. -0.00 min
Lab File: BF081790.D
Acq: 25 Sep 2015 6:08

Tgt Ion:204 Resp: 478094
Ion Ratio Lower Upper
204 100
206 34.6 24.8 37.2
141 58.9 42.2 63.4



Abundance Ion 204.00 (203.70 to 204.70): E
Ion 206.00 (205.70 to 206.70): E
Ion 141.00 (140.70 to 141.70): E





#61

4-Nitroaniline

Concen: 51.39 ng

RT: 10.57 min Scan# 738

Delta R.T. 0.01 min

Lab File: BF081790.D

Acq: 25 Sep 2015 6:08

Instrument :

BNA_F

ClientSampleId :

MGP209BR-29-30MS


Tot Ion:138 Resp: 259562

Ion Ratio Lower Upper

138 100

92 43.2 12.6 52.6




108 66.2 12.4 52.4#



Abundance Ion 138.00 (137.70 to 138.70): E

Ion 92.00 (91.70 to 92.70): BF0

Ion 108.00 (107.70 to 108.70): E



#62

Azobenzene

Concen: 47.05 ng

RT: 10.70 min Scan# 749

Delta R.T. -0.00 min

Lab File: BF081790.D

Acq: 25 Sep 2015 6:08

Tgt Ion: 77 Resp: 1157116


Ion Ratio Lower Upper

77 100

182 27.3 15.1 55.1

105 24.1 3.4 43.4

51 31.8 12.0 52.0






Abundance Ion 77.00 (76.70 to 77.70): BF0

Ion 182.00 (181.70 to 182.70): E

Ion 105.00 (104.70 to 105.70): E

Ion 51.00 (50.70 to 51.70): BF0





#63

Phenanthrene-d10

Concen: 20.00 ng

RT: 11.47 min Scan# 817

Delta R.T. -0.00 min

Lab File: BF081790.D

Acq: 25 Sep 2015 6:08

Instrument :

BNA_F

ClientSampleId :

MGP209BR-29-30MS


Tot Ion:188 Resp: 611691

Ion Ratio Lower Upper

188 100

94 12.3 11.1 16.7

80 13.7 11.0 16.4





Abundance Ion 188.00 (187.70 to 188.70): E

Ion 94.00 (93.70 to 94.70): BF0

Ion 80.00 (79.70 to 80.70): BF0

11.47

Time-->



#64

4,6-Dinitro-2-methylphenol

Concen: 55.85 ng

RT: 10.59 min Scan# 740

Delta R.T. 0.01 min

Lab File: BF081790.D

Acq: 25 Sep 2015 6:08


Tgt Ion:198 Resp: 84624

Ion Ratio Lower Upper

198 100

51 17.2 39.6 79.6#

105 21.7 28.5 68.5#




Abundance Ion 198.00 (197.70 to 198.70): E


Ion 51.00 (50.70 to 51.70): BF0

Ion 105.00 (104.70 to 105.70): E

10.59

Time-->









#65
 n-Nitrosodiphenylamine
 Concen: 43.68 ng
 RT: 10.65 min Scan# 745
 Delta R.T. -0.00 min
 Lab File: BF081790.D
 Acq: 25 Sep 2015 6:08




Instrument :
 BNA_F
 ClientSampleId :
 MGP209BR-29-30MS


Tot Ion	Ratio	Lower	Upper
169	100		
168	68.4	48.9	73.3
167	38.0	26.2	39.4



#66
 4-Bromophenyl-phenylether
 Concen: 47.86 ng
 RT: 11.03 min Scan# 778
 Delta R.T. -0.00 min
 Lab File: BF081790.D
 Acq: 25 Sep 2015 6:08

Tgt Ion	Ratio	Lower	Upper
248	100		
250	95.2	78.5	117.7
141	59.3	59.0	88.6









#67
Hexachlorobenzene
Concen: 43.62 ng
RT: 11.09 min Scan# 783
Delta R.T. -0.00 min
Lab File: BF081790.D
Acq: 25 Sep 2015 6:08




Instrument :
BNA_F
ClientSampleId :
MGP209BR-29-30MS


Tot Ion	Ratio	Lower	Upper
284	100		
142	39.0	29.5	44.3
249	33.4	19.5	29.3



#68
Atrazine
Concen: 52.66 ng
RT: 11.19 min Scan# 792
Delta R.T. 0.01 min
Lab File: BF081790.D
Acq: 25 Sep 2015 6:08

Tgt Ion	Ratio	Lower	Upper
200	100		
173	22.8	13.2	53.2
215	50.5	41.8	81.8





#69

Pentachlorophenol

Concen: 85.42 ng

RT: 11.28 min Scan# 800

Delta R.T. -0.00 min

Lab File: BF081790.D

Acq: 25 Sep 2015 6:08

Instrument :

BNA_F

ClientSampleId :

MGP209BR-29-30MS


Tot Ion:266 Resp: 337193

Ion Ratio Lower Upper

266 100

268 63.6 49.8 74.6




264 65.0 50.2 75.2



Abundance Ion 265.70 (265.40 to 266.40): E

Ion 268.00 (267.70 to 268.70): E

Ion 264.00 (263.70 to 264.70): E



#70

Phenanthrene

Concen: 38.64 ng

RT: 11.55 min Scan# 824

Delta R.T. 0.05 min

Lab File: BF081790.D

Acq: 25 Sep 2015 6:08


Tgt Ion:178 Resp: 1351385

Ion Ratio Lower Upper

178 100

176 20.2 13.8 20.8



179 16.5 12.2 18.2




Abundance Ion 178.00 (177.70 to 178.70): E

Ion 176.00 (175.70 to 176.70): E

Ion 179.00 (178.70 to 179.70): E









#71
 Anthracene
 Concen: 40.10 ng
 RT: 11.55 min Scan# 824
 Delta R.T. -0.00 min
 Lab File: BF081790.D
 Acq: 25 Sep 2015 6:08

Instrument :
 BNA_F
 ClientSampleId :
 MGP209BR-29-30MS

Tot Ion:178 Resp: 1351385
 Ion Ratio Lower Upper
 178 100
 176 20.2 14.1 21.1
 179 16.5 12.2 18.2




Abundance Ion 178.00 (177.70 to 178.70): E
 Ion 176.00 (175.70 to 176.70): E
 Ion 179.00 (178.70 to 179.70): E






#72
 Carbazole
 Concen: 44.74 ng
 RT: 11.71 min Scan# 838
 Delta R.T. 0.01 min
 Lab File: BF081790.D
 Acq: 25 Sep 2015 6:08

Tgt Ion:167 Resp: 1399598
 Ion Ratio Lower Upper
 167 100
 166 23.2 15.7 23.5
 139 13.3 9.5 14.3



Abundance Ion 167.00 (166.70 to 167.70): E
 Ion 166.00 (165.70 to 166.70): E
 Ion 139.00 (138.70 to 139.70): E





#73

Di-n-butylphthalate

Concen: 45.21 ng

RT: 12.05 min Scan# 867

Delta R.T. 0.01 min

Lab File: BF081790.D

Acq: 25 Sep 2015 6:08

Instrument :

BNA_F

ClientSampleId :

MGP209BR-29-30MS


Tot Ion:149 Resp: 1721435

Ion Ratio Lower Upper

149 100

150 9.2 7.4 11.2




104 3.6 2.4 3.6#



Abundance Ion 149.00 (148.70 to 149.70): E

Ion 150.00 (149.70 to 150.70): E

Ion 104.00 (103.70 to 104.70): E



#74

Fluoranthene

Concen: 42.62 ng

RT: 12.70 min Scan# 924

Delta R.T. 0.01 min

Lab File: BF081790.D

Acq: 25 Sep 2015 6:08


Tgt Ion:202 Resp: 1501954

Ion Ratio Lower Upper

202 100

101 6.9 0.0 38.3



203 18.0 0.0 37.3




Abundance Ion 202.00 (201.70 to 202.70): E

Ion 101.00 (100.70 to 101.70): E

Ion 203.00 (202.70 to 203.70): E





#75

Chrysene-d12

Concen: 20.00 ng

RT: 14.12 min Scan# 1048

Delta R.T. -0.00 min

Lab File: BF081790.D

Acq: 25 Sep 2015 6:08

Instrument :

BNA_F

ClientSampleId :

MGP209BR-29-30MS


Tot Ion:240 Resp: 404875

Ion Ratio Lower Upper

240 100

120 14.9 16.2 24.2#




236 26.7 18.4 27.6



Abundance Ion 240.00 (239.70 to 240.70): E

Ion 120.00 (119.70 to 120.70): E

Ion 236.00 (235.70 to 236.70): E



#76

Benzidine

Concen: 29.05 ng

RT: 12.81 min Scan# 934

Delta R.T. -0.00 min

Lab File: BF081790.D

Acq: 25 Sep 2015 6:08


Tgt Ion:184 Resp: 399097

Ion Ratio Lower Upper

184 100

185 15.1 11.8 17.6



183 11.2 7.6 11.4




Abundance Ion 184.00 (183.70 to 184.70): E

Ion 185.00 (184.70 to 185.70): E

Ion 183.00 (182.70 to 183.70): E









#77
Pvrene
Concen: 49.20 ng
RT: 12.70 min Scan# 924
Delta R.T. -0.22 min
Lab File: BF081790.D
Acq: 25 Sep 2015 6:08

Instrument :
BNA_F
ClientSampleId :
MGP209BR-29-30MS

Tot Ion:202 Resp: 1501709
Ion Ratio Lower Upper
202 100
200 22.0 15.0 22.4
203 18.0 14.1 21.1




Abundance Ion 202.00 (201.70 to 202.70): E
Ion 200.00 (199.70 to 200.70): E
Ion 203.00 (202.70 to 203.70): E






#78
Terphenyl-d14
Concen: 87.96 ng
RT: 13.06 min Scan# 956
Delta R.T. -0.00 min
Lab File: BF081790.D
Acq: 25 Sep 2015 6:08

Tgt Ion:244 Resp: 1559080
Ion Ratio Lower Upper
244 100
212 8.3 4.3 6.5#
122 12.5 17.4 26.2#



Abundance Ion 244.00 (243.70 to 244.70): E
Ion 212.00 (211.70 to 212.70): E
Ion 122.00 (121.70 to 122.70): E





#79

Butylbenzylphthalate

Concen: 57.47 ng

RT: 13.54 min Scan# 998

Delta R.T. 0.01 min

Lab File: BF081790.D

Acq: 25 Sep 2015 6:08

Instrument :

BNA_F

ClientSampleId :

MGP209BR-29-30MS


Tot Ion:149 Resp: 699086

Ion Ratio Lower Upper

149 100

91 52.2 48.6 72.8




206 24.7 14.9 22.3#



Abundance Ion 149.00 (148.70 to 149.70): E

Ion 91.00 (90.70 to 91.70): BF0

Ion 206.00 (205.70 to 206.70): E



#80

Benzo(a)anthracene

Concen: 43.02 ng

RT: 14.10 min Scan# 1047

Delta R.T. -0.00 min

Lab File: BF081790.D

Acq: 25 Sep 2015 6:08


Tgt Ion:228 Resp: 1056095

Ion Ratio Lower Upper

228 100

226 28.5 20.0 30.0



229 20.7 15.4 23.2




Abundance Ion 228.00 (227.70 to 228.70): E

Ion 226.00 (225.70 to 226.70): E

Ion 229.00 (228.70 to 229.70): E





#81

3,3'-Dichlorobenzidine

Concen: 34.83 ng

RT: 14.07 min Scan# 1044

Delta R.T. -0.00 min

Lab File: BF081790.D

Acq: 25 Sep 2015 6:08

Instrument :

BNA_F

ClientSampleId :

MGP209BR-29-30MS


Tot Ion:252 Resp: 271597

Ion Ratio Lower Upper

252 100

254 64.2 51.4 77.2




126 12.4 16.4 24.6#



Abundance Ion 252.00 (251.70 to 252.70): E

Ion 254.00 (253.70 to 254.70): E

Ion 126.00 (125.70 to 126.70): E



#82

Chrysene

Concen: 44.85 ng

RT: 14.10 min Scan# 1047

Delta R.T. -0.05 min

Lab File: BF081790.D

Acq: 25 Sep 2015 6:08


Tgt Ion:228 Resp: 1056095

Ion Ratio Lower Upper

228 100

226 28.5 21.0 31.4



229 20.7 15.6 23.4




Abundance Ion 228.00 (227.70 to 228.70): E

Ion 226.00 (225.70 to 226.70): E

Ion 229.00 (228.70 to 229.70): E





#83

Bis(2-ethylhexyl)phthalate

Concen: 59.93 ng

RT: 14.09 min Scan# 1046

Delta R.T. -0.00 min

Lab File: BF081790.D

Acq: 25 Sep 2015 6:08

Instrument :

BNA_F

ClientSampleId :

MGP209BR-29-30MS


Tot Ion:149 Resp: 883951

Ion Ratio Lower Upper

149 100

167 26.1 24.2 36.2




279 2.6 3.8 5.6#



Abundance Ion 149.00 (148.70 to 149.70): E

Ion 167.00 (166.70 to 167.70): E

Ion 279.00 (278.70 to 279.70): E



#84

Di-n-octyl phthalate

Concen: 67.85 ng

RT: 14.71 min Scan# 1100

Delta R.T. -0.00 min

Lab File: BF081790.D

Acq: 25 Sep 2015 6:08


Tgt Ion:149 Resp: 1471800

Ion Ratio Lower Upper

149 100

167 1.6 1.0 1.6



43 10.6 10.2 15.2




Abundance Ion 149.00 (148.70 to 149.70): E

Ion 167.00 (166.70 to 167.70): E

Ion 43.00 (42.70 to 43.70): BF0





#85

Indeno(1,2,3-cd)pyrene

Concen: 44.80 ng

RT: 17.05 min Scan# 1305

Delta R.T. 0.01 min

Lab File: BF081790.D

Acq: 25 Sep 2015 6:08

Instrument :

BNA_F

ClientSampleId :

MGP209BR-29-30MS


Tot Ion:276 Resp: 965627

Ion Ratio Lower Upper

276 100

138 25.6 25.7 38.5#




277 24.8 15.6 23.4#



Abundance Ion 276.00 (275.70 to 276.70): E

Ion 138.00 (137.70 to 138.70): E

Ion 277.00 (276.70 to 277.70): E



#86

Perylene-d12

Concen: 20.00 ng

RT: 15.58 min Scan# 1176

Delta R.T. -0.00 min

Lab File: BF081790.D

Acq: 25 Sep 2015 6:08


Tgt Ion:264 Resp: 343251

Ion Ratio Lower Upper

264 100

260 23.9 16.7 25.1



265 22.0 17.2 25.8




Abundance Ion 264.00 (263.70 to 264.70): E

Ion 260.00 (259.70 to 260.70): E

Ion 265.00 (264.70 to 265.70): E









#87
Benzo(b)fluoranthene
Concen: 43.33 ng
RT: 15.16 min Scan# 1139
Delta R.T. -0.00 min
Lab File: BF081790.D
Acq: 25 Sep 2015 6:08




Instrument :
BNA_F
ClientSampleId :
MGP209BR-29-30MS


Tot Ion:252 Resp: 882070
Ion Ratio Lower Upper
252 100
253 23.0 17.5 26.3
125 9.9 17.1 25.7#



#88
Benzo(k)fluoranthene
Concen: 44.99 ng
RT: 15.16 min Scan# 1139
Delta R.T. -0.02 min
Lab File: BF081790.D
Acq: 25 Sep 2015 6:08

Tgt Ion:252 Resp: 882070
Ion Ratio Lower Upper
252 100
253 23.0 17.0 25.4
125 9.9 16.0 24.0#









#89
Benzo(a)pyrene
Concen: 45.09 ng
RT: 15.52 min Scan# 1171
Delta R.T. -0.00 min
Lab File: BF081790.D
Acq: 25 Sep 2015 6:08

Instrument :
BNA_F
ClientSampleId :
MGP209BR-29-30MS

Tot Ion:252 Resp: 828625
Ion Ratio Lower Upper
252 100
253 22.4 17.2 25.8
125 11.4 18.0 27.0#




Abundance Ion 252.00 (251.70 to 252.70): E
Ion 253.00 (252.70 to 253.70): E
Ion 125.00 (124.70 to 125.70): E








#90
Dibenzo(a,h)anthracene
Concen: 42.47 ng
RT: 17.08 min Scan# 1307
Delta R.T. 0.01 min
Lab File: BF081790.D
Acq: 25 Sep 2015 6:08

Tgt Ion:278 Resp: 813349
Ion Ratio Lower Upper
278 100
139 14.2 26.3 39.5#
279 24.1 18.2 27.4



Abundance Ion 278.00 (277.70 to 278.70): E
Ion 139.00 (138.70 to 139.70): E
Ion 279.00 (278.70 to 279.70): E





#91

Benzo(a,h,i)perylene

Concen: 38.16 ng

RT: 17.50 min Scan# 1344

Delta R.T. 0.01 min

Lab File: BF081790.D

Acq: 25 Sep 2015 6:08

Instrument :

BNA_F

ClientSampleId :

MGP209BR-29-30MS

Tot Ion:276 Resp: 753548

Ion Ratio Lower Upper

276 100

277 23.8 17.8 26.6

138 20.5 34.6 51.8#

