

Data Path : Z:\HPCHEM1\BNA_F\DATA\BF092415\
 Data File : BF081783.D
 Acq On : 25 Sep 2015 1:03
 Operator : UM/IZ
 Sample : G3708-08DL2 100X
 Misc :
 ALS Vial : 16 Sample Multiplier: 1

Instrument :
 BNA_F
 ClientSampleId :
 MGP201-19-20DL2

Quant Time: Sep 25 07:07:51 2015
 Quant Method : Z:\HPCHEM1\BNA_F\METHODS\8270-BF092415.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Fri Sep 25 06:04:07 2015
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Dichlorobenzene-d4	6.95	152	207609	20.00	ng	0.00
21) Naphthalene-d8	8.24	136	863630	20.00	ng	0.00
38) Acenaphthene-d10	9.99	164	400553	20.00	ng	-0.01
63) Phenanthrene-d10	11.47	188	739244	20.00	ng	0.00
75) Chrysene-d12	14.12	240	596317	20.00	ng	0.00
86) Perylene-d12	15.58	264	490628	20.00	ng	0.00
System Monitoring Compounds						
5) 2-Fluorophenol	5.53	112	11845	0.97	ng	-0.01
7) Phenol-d6	6.55	99	15200	0.98	ng	-0.02
23) Nitrobenzene-d5	7.51	82	7145	0.55	ng	-0.01
41) 2,4,6-Tribromophenol	10.78	330	1374	0.40	ng	-0.01
44) 2-Fluorobiphenyl	9.30	172	19296	0.79	ng	-0.01
78) Terphenyl-d14	13.05	244	18563	0.71	ng	-0.01
Target Compounds						Qvalue
31) Naphthalene	8.26	128	1144658	28.73	ng	97
37) 2-Methylnaphthalene	8.95	142	364677	12.84	ng	# 90
48) Acenaphthylene	9.85	152	224310	5.26	ng	97
57) Fluorene	10.54	166	126968	4.59	ng	91
70) Phenanthrene	11.50	178	367969	8.71	ng	96
71) Anthracene	11.54	178	98043	2.41	ng	99
74) Fluoranthene	12.69	202	108651	2.55	ng	88
77) Pyrene	12.91	202	155434	3.46	ng	97

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\HPCHEM1\BNA_F\DATA\BF092415\
Data File : BF081783.D
Acq On : 25 Sep 2015 1:03
Operator : UM/IZ
Sample : G3708-08DL2 100X
Misc :
ALS Vial : 16 Sample Multiplier: 1

Instrument :
BNA_F
ClientSampleId :
MGP201-19-20DL2

Quant Time: Sep 25 07:07:51 2015
Quant Method : Z:\HPCHEM1\BNA_F\METHODS\8270-BF092415.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Fri Sep 25 06:04:07 2015
Response via : Initial Calibration

