

Data Path : Z:\HPCHEM1\BNA_F\DATA\BF092415\
 Data File : BF081795.D
 Acq On : 25 Sep 2015 8:39
 Operator : UM/IZ
 Sample : G3753-04
 Misc :
 ALS Vial : 25 Sample Multiplier: 1

Instrument :
 BNA_F
 ClientSampleId :
 SP-2

Manual Integrations
 APPROVED

MMDadoda
 9/25/2015 3:36:55 PM

Quant Time: Sep 25 13:28:21 2015
 Quant Method : Z:\HPCHEM1\BNA_F\METHODS\8270-BF092415.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Fri Sep 25 06:02:49 2015
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Dichlorobenzene-d4	6.95	152	146203	20.00	ng	0.00
21) Naphthalene-d8	8.24	136	591114	20.00	ng	0.00
38) Acenaphthene-d10	9.99	164	254026	20.00	ng	-0.01
63) Phenanthrene-d10	11.47	188	413238	20.00	ng	0.00
75) Chrysene-d12	14.12	240	322570	20.00	ng	0.00
86) Perylene-d12	15.59	264	309950	20.00	ng	0.01
System Monitoring Compounds						
5) 2-Fluorophenol	5.55	112	936490	108.36	ng	0.01
7) Phenol-d6	6.57	99	1113420	101.86	ng	0.00
23) Nitrobenzene-d5	7.52	82	820824	91.62	ng	0.00
41) 2,4,6-Tribromophenol	10.78	330	263727	119.58	ng	-0.01
44) 2-Fluorobiphenyl	9.31	172	1262565	81.59	ng	0.00
78) Terphenyl-d14	13.06	244	985888	69.82	ng	0.00
Target Compounds						Qvalue
49) Dimethylphthalate	9.70	163	375781	19.44	ng	# 97
70) Phenanthrene	11.50	178	187361	7.93	ng	98
71) Anthracene	11.55	178	47769	2.10	ng	98
74) Fluoranthene	12.69	202	380873	16.00	ng	93
77) Pyrene	12.91	202	358157	14.73	ng	96
80) Benzo(a)anthracene	14.10	228	241545	12.35	ng	# 91
82) Chrysene	14.14	228	171067	9.12	ng	92
83) Bis(2-ethylhexyl)phthalate	14.09	149	33407	2.84	ng	94
85) Indeno(1,2,3-cd)pyrene	17.05	276	106498	6.20	ng	# 82
87) Benzo(b)fluoranthene	15.16	252	240905m	13.11	ng	
88) Benzo(k)fluoranthene	15.17	252	113833m	6.43	ng	
89) Benzo(a)pyrene	15.52	252	184382	11.11	ng	# 88
91) Benzo(g,h,i)perylene	17.50	276	100977	5.66	ng	# 78

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\HPCHEM1\BNA_F\DATA\BF092415\
Data File : BF081795.D
Acq On : 25 Sep 2015 8:39
Operator : UM/IZ
Sample : G3753-04
Misc :
ALS Vial : 25 Sample Multiplier: 1

Instrument :
BNA_F
ClientSampleId :
SP-2

Manual Integrations
APPROVED

MMDadoda
9/25/2015 3:36:55 PM

Quant Time: Sep 25 13:28:21 2015
Quant Method : Z:\HPCHEM1\BNA_F\METHODS\8270-BF092415.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Fri Sep 25 06:02:49 2015
Response via : Initial Calibration

